Evaluation of the Immunoquick+4 malaria rapid diagnostic test in a non-endemic setting

D. P. J. Dijk, P. Gillet, E. Vlieghe, L. Cnops, M. Esbroeck, J. Jacobs

► To cite this version:

D. P. J. Dijk, P. Gillet, E. Vlieghe, L. Cnops, M. Esbroeck, et al.. Evaluation of the Immunoquick+4 malaria rapid diagnostic test in a non-endemic setting. *European Journal of Clinical Microbiology and Infectious Diseases*, 2010, 29 (5), pp.577-583. 10.1007/s10096-010-0898-y . hal-00577008

HAL Id: hal-00577008

<https://hal.science/hal-00577008>

Submitted on 16 Mar 2011

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Evaluation of the Immunoquick+4 malaria rapid diagnostic test in a non-endemic setting

D. P. J. van Dijk · P. Gillet · E. Vlieghe · L. Cnops ·
M. Van Esbroeck · J. Jacobs

Received: 22 September 2009 / Accepted: 13 February 2010 / Published online: 16 March 2010
© Springer-Verlag 2010

Abstract The aim of this retrospective study was to evaluate the Immunoquick+4 (BioSynex, Strasbourg, France), a three-band malaria rapid diagnostic test (MRDT) targeting histidine-rich protein-2 (HRP-2) and pan *Plasmodium*-specific parasite lactate dehydrogenase, in a non-endemic reference setting. Stored whole-blood samples ($n=613$) from international travellers suspected of malaria were used, with microscopy corrected by polymerase chain reaction (PCR) as the reference method. Samples infected by *P. falciparum* ($n=323$), *P. vivax* ($n=97$), *P. ovale* ($n=73$) and *P. malariae* ($n=25$) were selected, as well as 95 malaria-negative samples. The overall sensitivities

of the Immunoquick+4 for the diagnosis of *P. falciparum*, *P. vivax*, *P. malariae* and *P. ovale* were 88.9, 75.3, 56.0 and 19.2%, respectively. Sensitivity was significantly related to parasite density for *P. falciparum* (93.6% versus 71.4% at parasite densities $>100/\mu\text{l}$ and $\leq 100/\mu\text{l}$, respectively) and *P. vivax* (86.8% versus 48.3% at parasite densities $>500/\mu\text{l}$ and $\leq 500/\mu\text{l}$, respectively). The Immunoquick+4 showed good reproducibility and reliability for both test results and line intensities. The Immunoquick+4 performed well for the detection of *P. falciparum* and *P. vivax*.

Background

For the diagnosis of malaria, microscopy is considered as the reference method, but expert microscopy may be lacking in both endemic and non-endemic settings. In resource-poor endemic settings, there may be problems related to equipment, expertise and workload, whereas in non-endemic settings in industrialised countries, there may be a lack of routine among the laboratory staff, resulting in low expertise [1]. In these circumstances, the use of malaria rapid diagnostic tests (MRDTs) can be valuable in the diagnosis of malaria [1, 2]. MRDTs detect antigens specific to one or more of the *Plasmodium* species. The initially developed tests were two-band tests, including a control line and a *P. falciparum*-specific line targeting histidine-rich protein-2 (HRP-2) or *P. falciparum*-specific parasite lactate dehydrogenase (Pf-pLDH). Later developed so-called three-band MRDTs also detect antigens common to the four *Plasmodium* species, such as aldolase or the pan *Plasmodium*-specific pLDH (pan-pLDH). The Immunoquick malaria +4 (BioSynex, Strasbourg, France), further referred to as Immunoquick+4, is a three-band MRDT targeting HRP-2 and pan-pLDH. We evaluated the

D. P. J. van Dijk · J. Jacobs
Medical Microbiology, Faculty of Health,
Medicine and Life Sciences (FHML), Maastricht University,
Maastricht, The Netherlands

D. P. J. van Dijk
e-mail: D.vanDijk@student.maastrichtuniversity.nl

J. Jacobs
e-mail: jjacobs@itg.be

P. Gillet (✉) · E. Vlieghe · L. Cnops · M. Van Esbroeck ·
J. Jacobs
Department of Clinical Sciences,
Unit of Tropical Laboratory Medicine,
Institute of Tropical Medicine (ITM),
Nationaalestraat 155,
2000 Antwerp, Belgium
e-mail: pgillet@itg.be

E. Vlieghe
e-mail: evlieghe@itg.be

L. Cnops
e-mail: lcnops@itg.be

M. Van Esbroeck
e-mail: mvesbroeck@itg.be

Immunoquick+4 for use on returned international travellers in a non-endemic reference centre.

Methods

Study design

In this retrospective study, the Immunoquick+4 was evaluated against a collection of stored samples obtained from international travellers. Tests were carried out in the reference laboratory of the Institute of Tropical Medicine (ITM), Antwerp, Belgium.

Patients and samples

Samples were selected from a collection of EDTA blood samples stored at -70°C . They were obtained from patients attending the travel clinic at the ITM. Included patients were international travellers or immigrants returning from visits to their native countries. In addition, samples sent by Belgian laboratories to the ITM in the scope of the national reference function were included. All patients had symptoms suspect for malaria (e.g. fever, headache, nausea). All positive samples collected between December 1995 to August 2008 were selected, including the four malaria species with varying parasite densities. In addition, samples from symptomatic patients without malaria parasites (negative samples) attending the ITM between December 1995 to 2008 were included. There were no exclusion criteria except a blood sample lower than 100 μl .

Reference method

Microscopy, corrected by polymerase chain reaction (PCR), was used as the reference method. Standard microscopy including the determination of parasite density was performed on thick blood films as described previously [3]. PCR analysis was performed on all samples, with a species-specific real-time PCR adapted from Rougemont et al. [4], as described previously [3].

Test platform

The Immunoquick+4 is a lateral-flow immunochromatographic MRDT in a dipstick format. In brief, 20 μl of whole blood is deposited on the test strip, which is subsequently transferred to a plastic tube containing six drops of buffer. After 15 to 30 min, the results can be read. Three lines are present; a control line which indicates whether the test is valid, a HRP-2 line and a pan-pLDH line. According to the manufacturer's instructions, a combination of a HRP-2 line

and a pan-pLDH line indicates an infection with *P. falciparum* or a mixed infection with *P. falciparum* and one or more of the other *Plasmodium* species. A single HRP-2 line indicates an infection with *P. falciparum* and a single pan-pLDH line indicates an infection with one or more of the non-*falciparum* species. For the evaluation, test kits of two different lot numbers were used, M4D060508 and M4D091708, with expiration dates 01/2010 and 04/2010, respectively.

Test procedure

Tests were performed according to the instructions of the manufacturer, except that a scoring system was used to assess the intensity of the test lines. As recommended by the manufacturer in the instructions for application of the test on venous whole blood, we used a transfer pipette of 20 μl (Finnpipette, Helsinki, Finland) to apply blood to the test strip. In case the control line did not appear, the result was interpreted as invalid and the test was repeated. In order to score test line intensities, we used a scoring system of five categories: none (no line visible), faint (barely visible line), weak (paler than the control line), medium (equal to the control line) or strong (stronger than the control line) [3]. Readings were performed with daylight assisted by a standard halogen lamp. To homogenise the results, all samples were read after 20 to 30 min, as recommended by the manufacturer for low positive samples. Readings were carried out by three subsequent observers, of whom the one who performed the test procedure invariably was the first. Observers were blinded to the results of microscopy, PCR and to each others' readings. The results of the readings considered were based on consensus agreement, which means that a positive result was defined as a result read positive by at least two out of three different observers. When there was no consensus, the result of the first reader was considered. Inter-reader reliabilities were assessed for the test results, expressed as positive and negative readings, as well as for the intensity readings. To assess reproducibility, we tested a panel of 16 samples with different parasite densities (including four *P. falciparum* samples, four *P. vivax* samples, four *P. ovale* samples and four *P. malariae* samples) on five successive occasions.

Statistical analysis

Samples infected with *P. falciparum* and the non-*falciparum* species were considered separately. For the detection of *P. falciparum*, any test that generated a HRP-2 line was considered positive. Any test that did not generate a HRP-2 line was considered negative. For the detection of the non-*falciparum* species, any test that generated a

unique pan-pLDH line was considered positive. Any test that did not generate a test line or showed a HRP-2 line was considered negative. Samples that generated a wrong test line (e.g. a *P. falciparum* sample generating only a pan-pLDH line) were considered as species mismatch and assigned to the false-positives or false-negatives, depending on the situation. Samples with pure gametocytaemia were included among the positive *P. falciparum* samples. Sensitivity and specificity were calculated for both *P. falciparum* and the non-*falciparum* species with 95% confidence intervals (CIs) and differences were tested for significance using the Pearson Chi-square test or, in case of small sample numbers, Fisher's exact probability test. A *p*-value of <0.05 was considered as significant. Reliabilities for positive and negative readings and line intensities were calculated as percentage agreements for all three readers and kappa values for each pair of readers. Associations between line intensity readings and parasite densities were assessed for strength of association with Cramer's *V* for categorical variables, using interpretative criteria published previously [3].

Duration of storage

Long-term storage of samples could affect antigen stability. To minimise the possibility that the duration of storage affects the test performance, we compared the sensitivities of samples obtained between 1995 to 2000 with those obtained between 2001 to 2008.

Ease of use

Three experienced laboratory technicians scored the ease of use of the Immunoquick+4 test and the clarity of the manufacturer's instructions with a standardised list [3].

Ethical review

The study was reviewed and approved by the Institutional Review Board of the ITM and by the Ethical Committee of Antwerp University, Belgium.

Results

Sample collection

A total of 613 samples were selected, of which 83 samples were sent by Belgian laboratories to the ITM for second opinion and confirmation. According to microscopy and after correction by PCR analysis, 323 of these samples were positive for *P. falciparum*, 97 for *P. vivax*, 73 for *P. ovale* and 25 for *P. malariae*. The results of microscopy were corrected in 11 out of the 518 (2.1%) positive samples.

These corrections were uniquely related to *P. vivax*–*P. ovale* mismatches: in the final collection, 4/97 (4.1%) *P. vivax* and 7/73 (9.5%) *P. ovale* samples had originally been identified as *P. ovale* and *P. vivax*, respectively. In addition, 95 microscopic and PCR-negative samples of symptomatic travellers were included in the panel. The majority (270/323) of *P. falciparum* infection was acquired in Africa. The samples were obtained from 613 patients, with a male:female ratio of 2.07:1 and median age of 36.5 years (range 1–84 years). Only a minority (nine patients, 1.5%) were children less than five years old.

Invalid test results

A single sample gave an invalid result at initial testing. After 30 min, there was no control line visible. Upon repetition, the test performed well.

Sensitivity and specificity

Table 1 lists the test results of all samples and Table 2 lists the sensitivities matched with species identification and parasite densities. For the detection of *P. falciparum*, the sensitivity was 88.9%; it increased with parasite density, with values at parasite densities >100/μl being significantly higher as compared to those at parasite densities ≤100/μl (93.6% versus 71.4%, respectively, *p*<0.001). For the detection of the non-*falciparum* species, the overall sensitivities for *P. vivax*, *P. ovale* and *P. malariae* were 75.3, 19.2 and 56.0%, respectively. The difference in sensitivity between parasite densities above and below 500/μl was statistically significant in the case of *P. vivax* (86.8% versus 48.3%, *p*<0.001). Specificities for *P. falciparum* and the non-*falciparum* species were 97.9% (95.3%–99.2%) and 98.3% (96.4%–99.3%), respectively. Eight out of the 518 (1.5%) *Plasmodium*-positive samples resulted in a species mismatch (Table 1). Four *P. falciparum* samples reacted only with the pan-pLDH line; they had parasite densities of 180, 210, 1,123 and 147,000/μl, respectively, and the travel destinations were Democratic Republic of the Congo (*n*=2), Nigeria (*n*=1) and Ghana (*n*=1), respectively. One *P. vivax* sample (India; 4,150/μl), one *P. ovale* sample (Mali; 864/μl) and one *P. malariae* sample (Nigeria; 2,200/μl) reacted with both HRP-2 and pan-pLDH lines. Finally, there was one *P. malariae* sample (Togo, 712/μl) that reacted only with the HRP-2 line.

Line intensities

The line intensity readings for both HRP-2 and pan-pLDH lines were faint or weak in almost half of the true-positive readings, 136/287 (47.4%) and 131/301 (43.5%), respectively. Line intensity readings for HRP-2 and especially pan-pLDH

Table 1 Test results of all samples ($n = 613$)

	Samples	HRP-2 line-positive		HRP-2 line-negative	
		pan-pLDH line-positive	pan-pLDH line-negative	pan-pLDH line-positive	pan-pLDH line-negative
	<i>P. falciparum</i> ($n=323$)	194	93	4 ^a	32
	<i>P. vivax</i> ($n=97$)	1 ^a	-	73	23
	<i>P. ovale</i> ($n=73$)	1 ^a	-	14	58
	<i>P. malariae</i> ($n=25$)	1 ^a	1 ^a	14	9
	Negative ($n=95$)	-	2	3	90

HRP-2 = histidine-rich protein-2; pan-pLDH = pan *Plasmodium*-specific parasite lactate dehydrogenase

^a Species mismatch

were related to parasite densities (HRP-2: $V=0.296$, $p<0.001$; pan-pLDH: $V=0.620$, $p<0.001$). Of interest is that, in the case of positive *P. falciparum* samples, the unique presence of a HRP-2 line almost exclusively (98.9%, 92/93 samples) pointed to a parasite density below 1,000/μl (Table 3). Conversely, the co-presence of both HRP-2 and pan-pLDH lines predicted parasite densities above 1,000/μl in 131/194 (67.5%) samples.

Inter-reader reliability

For the HRP-2 line, the inter-reader reliability for positive and negative test results was excellent, with 97.2% agreement among the three readers and kappa values between the three pairs of readers of 0.97, 0.96 and 0.95, respectively. For the pan-pLDH line, there was 91.2% overall agreement and the kappa values were 0.91, 0.88 and 0.86. For the HRP-2 line intensity readings, the overall agreement among the three readers was 80.6%, with kappa values of 0.81, 0.80 and 0.79 for the different pairs of readers. For the pan-pLDH line intensity readings, there

was 76.7% overall agreement and kappa values of 0.79, 0.78 and 0.74.

Reproducibility

The test results (positive or negative for relevant species detection) and HRP-2 line intensity readings were reproducible. Consistent test results upon five times repetition were obtained for 10 out of 16 samples. Of the remaining six samples, five had consistent test results upon four times repetition and one sample upon three times repetition. HRP-2 line intensity readings were consistent upon five times repetition for 13 out of 16 samples. Two out of the three non-consistent samples had identical test results upon four times repetition and one sample upon three times repetition. The pan-pLDH intensity readings were consistent upon five times repetition for five out of 16 samples. Eight out of the 11 non-consistent samples had identical results upon four times repetition and three samples upon three times repetition. All but one inconsistent result for line intensity readings (13 out of 14) of the HRP-2 and pan-pLDH samples had

Table 2 Sensitivities of the Immunoquick+4 for the detection of all *Plasmodium* species related to parasite densities

Species	No.	Correctly identified by Immunoquick+4	Sensitivity in % (95% CI)
All <i>P. falciparum</i> samples	323	287	88.9 (84.8–92.0)
Pure gametocytaemia	17	13	76.5 (50.0–92.2)
Parasite density 0–100/μl	56	40	71.4 (57.6–82.3)
Parasite density 101–1,000/μl	116	102	87.9 (80.3–93.0)
Parasite density >1,000/μl	134	132	98.5 (94.2–99.7)
Parasite density >100/μl	250	234	93.6 (89.6–96.2)
All <i>P. vivax</i> samples	97	73	75.3 (65.3–83.2)
Parasite density 0–500/μl	29	14	48.3 (30.0–67.1)
Parasite density >500/μl	68	59	86.8 (75.9–93.4)
All <i>P. ovale</i> samples	73	14	19.2 (11.2–30.4)
Parasite density ≤500/μl	36	6	16.7 (7.0–33.5)
Parasite density >500/μl	37	8	21.6 (10.4–38.7)
All <i>P. malariae</i> samples	25	14	56.0 (35.3–75.0)
Parasite density ≤500/μl	9	4	44.4 (15.3–77.3)
Parasite density >500/μl	16	10	62.5 (35.9–83.7)

CI = confidence interval

Table 3 Presence of HRP-2 and pan-pLDH lines for the Immunoquick+4 in *Plasmodium falciparum* samples related to parasite density

Parasite density	Test line(s) visible			
	None	Only HRP-2	HRP-2 and pan-pLDH	Total
Pure gametocytaemia	4	5	8	17
0–100/μl	16	36	4	56
101–1,000/μl	14	51	51	116
>1,000/μl	2	1	131	134
Total	32	93	194	323

HRP-2 = histidine-rich protein-2; pan-pLDH = pan *Plasmodium*-specific parasite lactate dehydrogenase

discordances within one category of difference in line intensity. One pan-pLDH result differed by three categories in line intensity (medium instead of negative).

Duration of storage

There was no significant difference between the test sensitivities of samples that had been stored for long versus shorter periods, for any of the four *Plasmodium* species. The sensitivities of samples obtained between 1995 to 2000 and samples obtained between 2001 to 2008 were, respectively, 86.9 and 89.7% for *P. falciparum*, 73.2 and 76.8% for *P. vivax*, 16.1 and 21.4% for *P. ovale*, and 66.7 and 52.4% for *P. malariae*. None of these differences were statistically significant.

Ease of use

The Immunoquick+4 was evaluated as easy to use and the instructions were scored as clear and simple to follow by all three technicians. However, because of its design (a strip soaked in a test tube), it was considered to be less practical as compared to one-step MRDTs in cassette format. Further, the clearance was not optimal, causing the test strip to remain red-coloured after completion of the test procedure.

Discussion

In this retrospective study, the performance of the Immunoquick+4 was evaluated on a large panel of stored samples obtained from international travellers. The sensitivity for the detection of *P. falciparum* was 88.9%, reaching 98.5% at parasite densities above 1,000/μl. The overall sensitivities for *P. vivax*, *P. ovale* and *P. malariae* were 75.3, 19.2 and 56%, respectively. The specificity was above 97.9%. Species mismatches occurred in 1.5% of the samples and were limited to non-*falciparum* species that were identified as *P. falciparum*. The test results were reliable and reproducible.

The present study has some limitations. For instance, we did not challenge the test with samples that are known to

give false test results (such as rheumatoid factor) [5]. Further, tests were performed in a non-endemic reference centre, which tends—for reasons of resources and trained staff—to give better results compared to field settings [6, 7]. Finally, the storage time of the present samples could affect antigen stability [6]. However, samples were not exposed to repeated thawing and freezing, and no differences in test performances were observed between samples that had been stored for long versus short periods. In addition, a prospective study revealed similar results between fresh and stored samples with regard to HRP-2 detection [8].

The Immunoquick+4 has recently been evaluated as part of the World Health Organisation (WHO) and the Foundation for Innovative New Diagnostics (FIND) Malaria RDT Evaluation Programme [9]. This evaluation showed for *P. falciparum* and *P. vivax* sensitivities of, respectively, 93.7 and 30.0% at parasite densities of 200/μl, and sensitivities of 98.7 and 100% at parasite densities of 2,000–5,000/μl. In addition, this study reported species mismatches in 2 (1.6%) out of 120 *P. vivax* samples (which showed an additional HRP-2 test line) and a single (0.6%) false-positive reaction among 168 malaria-negative samples. The sensitivities reported by the WHO/FIND study are higher than those obtained in the present study, but the nature of samples and the parasite densities evaluated by both studies were different: for instance, the WHO/FIND study also used culture-derived samples and used samples that were diluted to parasite densities of 200/μl and 2,000–5,000/μl. Unlike the present study, the WHO/FIND study did not include *P. ovale* and *P. malariae* samples and it assessed species mismatches only between *P. falciparum* and *P. vivax* [9].

The sensitivities of the Immunoquick+4 for the detection of *P. falciparum* and *P. vivax* are in line with those for other MRDTs in non-endemic settings. Reported sensitivities for *P. falciparum* range from 87.5 to 99.0%, with one exception of 76.2% [3, 10–16]. However, the Immunoquick+4 did not reach the 95% sensitivity for parasite densities >100/μl recommended by the WHO [17]. For *P. vivax*, MRDTs targeting pan-pLDH show sensitivities from 33.5% and 62.0–95.0% [3, 12, 15, 18–20], compared to 46.0–93.0% [20] for those MRDTs targeting aldolase. The higher

sensitivities at increasing parasite densities are a well-known phenomenon, with breakpoints around 100/μl (*P. falciparum*) and 500/μl (*P. vivax*), respectively [3, 10, 18, 20, 21]. There are only a few studies that allow the comparison of test characteristics for the detection of *P. ovale* and *P. malariae* separately. The sensitivities for the detection of *P. ovale* and *P. malariae* of the Immunoquick+4 tended to be lower compared to other MRDTs, with combined sensitivities of 36.0–95.0% for MRDTs targeting pan-pLDH [18, 20] and 7.0–80.0% for MRDTs targeting aldolase [20]. Previously, we demonstrated for a three-band HRP-2/pan-pLDH MRDT in a study design similar to the present one sensitivities for *P. ovale* and *P. malariae* of 76.3 and 45.2%, respectively [3].

In line with other publications [3, 10, 22, 23], line intensities were in relation with the parasite densities, but with considerable overlaps. The finding that a unique HRP-2 line was indicative for a low parasite density of *P. falciparum* ≤1,000/μl is in line with a previous observation for another MRDT based on HRP2/pan-pLDH detection [3].

Although the Immunoquick+4 was reliable and reproducible, we observed incomplete clearance of the test strip. In addition, nearly half of the test lines showed weak or faint intensities. These findings are of concern, as they can result in false-negative interpretations. Indeed, disregarding faint line intensities is one of the most common mistakes among end-users in both endemic and non-endemic settings, even when provided with adapted job aids [24–28].

The Immunoquick+4 was scored as a less practical test compared to MRDTs in cassette format. In particular, the manipulation of the strip was felt to be more difficult compared to the manipulation of a cassette. However, although not evaluated as part of the present study, we assume that the recommended transfer of the blood by applying the soak pad of the strip to the blood on the finger is easier and less error-prone than other application systems, such as loops and straws [25, 29]. On the other hand, the use of the absorption pad of the dipstick as a sampling system for capillary blood may be less accurate. Further studies at the end-user level are required to assess these assumptions.

In conclusion, the Immunoquick+4 performed well for the detection of *P. falciparum* and *P. vivax*, but moderately and poorly for the detection of *P. malariae* and *P. ovale*, respectively. The Immunoquick+4 can be a valuable adjunct to microscopy, especially in those settings where *P. falciparum* and *P. vivax* are predominant. Possible test improvements include an increase in sensitivity, especially for non-*falciparum* species detection, a better clearance of the test strip and an increase in test line intensities.

Acknowledgements We would like to thank the staff of the Central Laboratory of Clinical Biology for their technical support.

Authors' contributions DvD, PG and JJ designed the study protocol. MvE and EV organised prospective sample collection. DvD and PG carried out the test evaluations, LC performed PCR analysis. DvD, PG and JJ analysed and interpreted the results and drafted the manuscript. DvD performed the statistical analysis. All authors contributed to the discussion of the results and the redaction of the manuscript; they all approved the final manuscript.

References

- Moody A (2002) Rapid diagnostic tests for malaria parasites. Clin Microbiol Rev 15:66–78
- Hänscheid T (2003) Current strategies to avoid misdiagnosis of malaria. Clin Microbiol Infect 9:497–504
- Van der Palen M, Gillet P, Bottieau E, Cnops L, Van Esbroeck M, Jacobs J (2009) Test characteristics of two rapid antigen detection tests (SD FK50 and SD FK60) for the diagnosis of malaria in returned travellers. Malar J 8:90
- Rougemont M, Van Saanen M, Sahli R, Hinrikson HP, Bille J, Jaton K (2004) Detection of four *Plasmodium* species in blood from humans by 18S rRNA gene subunit-based and species-specific real-time PCR assays. J Clin Microbiol 42:5636–5643
- Iqbal J, Sher A, Rab A (2000) *Plasmodium falciparum* histidine-rich protein 2-based immunocapture diagnostic assay for malaria: cross-reactivity with rheumatoid factors. J Clin Microbiol 38:1184–1186
- Bell D, Peeling RW (2006) Evaluation of rapid diagnostic tests: malaria. Nat Rev Microbiol 4:S34–S38
- Murray CK, Gasser RA Jr, Magill AJ, Miller RS (2008) Update on rapid diagnostic testing for malaria. Clin Microbiol Rev 21:97–110
- Mayxay M, Pukrittayakamee S, Chotivanich K, Looareesuwan S, White NJ (2001) Persistence of *Plasmodium falciparum* HRP-2 in successfully treated acute falciparum malaria. Trans R Soc Trop Med Hyg 95:179–182
- World Health Organization (2009) Malaria rapid diagnostic test performance: executive summary. Results of WHO product testing of malaria RDTs: Round 1 (2008). Available online at: <http://apps.who.int/tdr/publications/tdr-research-publications/rdt-performance/pdf/executive-summary-malaria-RDTs.pdf>. Accessed 13 June 2009
- Farcas GA, Zhong KJ, Lovegrove FE, Graham CM, Kain KC (2003) Evaluation of the Binax NOW® ICT test versus polymerase chain reaction and microscopy for the detection of malaria in returned travelers. Am J Trop Med Hyg 69:589–592
- Gatti S, Gramegna M, Bisoffi Z, Raglio A, Gulletta M, Klersy C, Bruno A, Maserati R, Madama S, Scaglia M (2007) A comparison of three diagnostic techniques for malaria: a rapid diagnostic test (NOW Malaria), PCR and microscopy. Ann Trop Med Parasitol 101:195–204
- Gillet P, Bosselaers K, Cnops L, Bottieau E, Van Esbroeck M, Jacobs J (2009) Evaluation of the SD FK70 malaria Ag *Plasmodium vivax* rapid diagnostic test in a non-endemic setting. Malar J 8:129
- Grobusch MP, Hänscheid T, Göbels K, Slevogt H, Zoller T, Rögler G, Teichmann D (2003) Comparison of three antigen detection tests for diagnosis and follow-up of falciparum malaria in travellers returning to Berlin, Germany. Parasitol Res 89:354–357
- Jelinek T, Schulte C, Behrens R, Grobusch MP, Coulaud JP, Bisoffi Z, Matteelli A, Clerinx J, Corachán M, Puente S, Gjørup I, Harms G, Kollaritsch H, Kotlowski A, Björkmann A, Delmont JP, Knobloch J, Nielsen LN, Cuadros J, Hatz C, Beran J, Schmid ML, Schulze M, Lopez-Velez R, Fleischer K, Kapaun A, McWhinney P, Kern P, Atougia J, Fry G, da Cunha S, Boeckx G (2002)

- Imported *Falciparum* malaria in Europe: sentinel surveillance data from the European network on surveillance of imported infectious diseases. *Clin Infect Dis* 34:572–576
15. Wiese L, Bruun B, Baek L, Friis-Møller A, Gahrn-Hansen B, Hansen J, Heltberg O, Højbjerg T, Hornstrup MK, Kvinesdal B, Gomme G, Kurtzhals JA (2006) Bedside diagnosis of imported malaria using the Binax Now malaria antigen detection test. *Scand J Infect Dis* 38:1063–1068
 16. Wongsrichanalai C, Barcus MJ, Muth S, Sutamihardja A, Wernsdorfer WH (2007) A review of malaria diagnostic tools: microscopy and rapid diagnostic test (RDT). *Am J Trop Med Hyg* 77:119–127
 17. World Health Organization. Regional Office for the Western Pacific (2003) Malaria rapid diagnosis: making it work. Meeting report 20–23 January 2003. Manila, the Philippines. Available online at: http://www.searo.who.int/LinkFiles/Malaria_MalariaRDT.pdf. Accessed 19 June 2009
 18. De Monbrison F, Gérome P, Chaulet JF, Wallon M, Picot S, Peyron F (2004) Comparative diagnostic performance of two commercial rapid tests for malaria in a non-endemic area. *Eur J Clin Microbiol Infect Dis* 23:784–786
 19. Grobusch MP, Hänscheid T, Göbels K, Slevogt H, Zoller T, Rögler G, Teichmann D (2003) Sensitivity of *P. vivax* rapid antigen detection tests and possible implications for self-diagnostic use. *Travel Med Infect Dis* 1:119–122
 20. Marx A, Pewsner D, Egger M, Nüesch R, Bucher HC, Genton B, Hatz C, Jüni P (2005) Meta-analysis: accuracy of rapid tests for malaria in travelers returning from endemic areas. *Ann Intern Med* 142:836–846
 21. Richardson DC, Ciach M, Zhong KJ, Crandall I, Kain KC (2002) Evaluation of the Makromed dipstick assay versus PCR for diagnosis of *Plasmodium falciparum* malaria in returned travelers. *J Clin Microbiol* 40:4528–4530
 22. Mason DP, Kawamoto F, Lin K, Laoboonchai A, Wongsrichanalai C (2002) A comparison of two rapid field immunochromatographic tests to expert microscopy in the diagnosis of malaria. *Acta Trop* 82:51–59
 23. Pieroni P, Mills CD, Ohrt C, Harrington MA, Kain KC (1998) Comparison of the ParaSight™-F test and the ICT Malaria Pf™ test with the polymerase chain reaction for the diagnosis of *Plasmodium falciparum* malaria in travellers. *Trans R Soc Trop Med Hyg* 92:166–169
 24. Craig MH, Bredenkamp BL, Williams CH, Rossouw EJ, Kelly VJ, Kleinschmidt I, Martineau A, Henry GF (2002) Field and laboratory comparative evaluation of ten rapid malaria diagnostic tests. *Trans R Soc Trop Med Hyg* 96:258–265
 25. Harvey SA, Jennings L, Chinyama M, Masaninga F, Mulholland K, Bell DR (2008) Improving community health worker use of malaria rapid diagnostic tests in Zambia: package instructions, job aid and job aid-plus-training. *Malar J* 7:160
 26. Mayxay M, Newton PN, Yeung S, Pongvongsa T, Phompida S, Phetsouvanh R, White NJ (2004) Short communication: an assessment of the use of malaria rapid tests by village health volunteers in rural Laos. *Trop Med Int Health* 9:325–329
 27. McMorrow ML, Masanja MI, Abdulla SM, Kahigwa E, Kachur SP (2008) Challenges in routine implementation and quality control of rapid diagnostic tests for malaria—Rufiji District, Tanzania. *Am J Trop Med Hyg* 79:385–390
 28. Rennie W, Phetsouvanh R, Lupisan S, Vanisaveth V, Hongvanthong B, Phompida S, Alday P, Fulache M, Lumagui R, Jorgensen P, Bell D, Harvey S (2007) Minimising human error in malaria rapid diagnosis: clarity of written instructions and health worker performance. *Trans R Soc Trop Med Hyg* 101:9–18
 29. Luchavez J, Lintag ME, Coll-Black M, Baik F, Bell D (2007) An assessment of various blood collection and transfer methods used for malaria rapid diagnostic tests. *Malar J* 6:149