Determination of acrylamide in coffee and coffee products by GC-MS using an improved SPE clean-up

Journal:	<i>Food Additives and Contaminants</i>
Manuscript ID:	TFAC-2006-141.R1
Manuscript Type:	Original Research Paper
Date Submitted by the Author:	22-Jun-2006
Complete List of Authors:	Dias Soares, Cristina; FFUP, Food Science Cunha, Sara; FFUP, Food Science Fernandes, José; FFUP, Food Science
Methods/Techniques:	Chromatography - GC/MS, Clean-up, Extraction, GC/MS
Additives/Contaminants:	Acrylamide, Process contaminants – acrylamide
Food Types:	Coffee, Processed foods
Note: The following files were submitted by the author for peer review, but cannot be converted to PDF. You must view these files (e.g. movies) online.	
Table I.doc	

SCHOLARONE™
Manuscripts

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Determination of acrylamide in coffee and coffee products by GC-MS using an improved SPE clean-up

Cristina Soares, Sara Cunha, and José Fernandes

FFUP Food Science, Rua Aníbal Cunha, 164 Porto 4050-047 (Portugal)

Abstract

An improved GC-MS method to determine acrylamide (AA) in coffee and coffee products was developed. The method was based in two main purification steps, the first one with ethanol and Carrez solutions in order to precipitate polysaccharides and proteins, respectively, and the second with a layered solid phase extraction (SPE) column which proved to be efficient to eliminate the main chromatographic interferences. The method is applicable to a wide range of coffee products. Twenty six samples of different coffee products were analysed. The levels of AA were in the range of 11.4 – 36.2 µg L⁻¹ for “espresso coffee” and 200.8 – 229.4 µg L⁻¹ for coffee blends with cereals. The results indicated that the presence of cereals significantly increased the levels of AA.

Keywords: Coffee, SPE clean up, acrylamide, GC-MS

Introduction

1
2
3
4
5 25 The occurrence of acrylamide (AA) in foodstuffs particularly in certain baked and fried
6
7 26 products such as potato chips and French fries was first reported by Tareke *et al.* (2002).
8
9 27 AA is now known to be formed by the Maillard reaction during industrial food
10
11 28 processing, retail, catering and home food preparation ((Mottram *et al.*, 2002; Stadler *et*
12
13 29 *al.*, 2002; Becalski *et al.*, 2003).. The foodstuffs that contribute most to AA exposure
14
15
16 30 vary depending upon the population's nutritional habits and the way the food is
17
18 31 prepared and processed (Dybing *et al.*, 2005). Generally, the most important sources of
19
20 32 AA appear to be potato products (potato chips, French fries, and potato snacks), cereals
21
22 33 (breakfast cereals, roasted cereals), baked goods (bread, cookies, biscuits), and brewed
23
24 34 coffee.
25
26
27
28
29

30
31 36 Recent reports showed that coffee is among the highest contributors to the AA intake in
32
33 37 some countries in Europe. In Sweden the main dietary intake of AA from coffee was
34
35 38 estimated to be 12 $\mu\text{g}/\text{day}$, representing ~39% of the total dietary intake of 31 $\mu\text{g}/\text{day}$
36
37 39 (Svensson *et al.*, 2003). These results were based on ready-to-drink coffee samples that
38
39 40 contained around 25 $\mu\text{g kg}^{-1}$ AA. Dybing and Sanner (2003) reported that a mean intake
41
42 41 of AA from coffee was estimated to contribute to 28% of the 38 $\mu\text{g}/\text{day}$ for males and
43
44 42 29% of the 29 $\mu\text{g}/\text{day}$ for females in Norway. Their results were based on a study
45
46 43 (Norwegian Food Control Authority, 2002) that reported concentrations of AA in filter
47
48 44 coffee of 25 $\mu\text{g kg}^{-1}$ and 10 $\mu\text{g kg}^{-1}$ in instant coffee. A Swiss total diet study of AA in
49
50 45 the diet of 14 men and 13 women showed that coffee contributed 36% of the total intake
51
52 46 (Brunner *et al.*, 2002). Granby and Fagt (2004) found in their study 2-16 $\mu\text{g L}^{-1}$ of AA
53
54 47 in brewed coffee and 10 $\mu\text{g L}^{-1}$ AA in instant coffee. They concluded that the AA intake
55
56 48 from coffee consumption in Denmark was 10 $\mu\text{g}/\text{day}$ for males and 9 $\mu\text{g}/\text{day}$ for
57
58
59
60

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

females. Assuming that the total exposure of Danish consumers is comparable to those of Sweden and Norway, the mean exposure to AA from coffee contributes to 20% of the total AA exposure. In Netherlands, Boon *et al.* (2005) accounted an AA concentration in coffee in the range 4 – 45 $\mu\text{g kg}^{-1}$ that contributes with 13% of the total AA intake in adults. The US Food and Drug Administration (FDA) reported levels of AA in brewed coffee ranging from 6 to 16 $\mu\text{g L}^{-1}$ (Andrzejewski *et al.*, 2004).

In Portugal, coffee is highly consumed as “espresso” which has a peculiar brewing technique: a small amount of hot water ($\pm 30\text{ mL}$) is percolated in a very short time at high pressure through a layer of ground roasted coffee ($\pm 6\text{-}7\text{ g}$), the coffee cake, to produce efficiently a very concentrated brew ($\approx 200\text{ g}$ of ground coffee per litre) comparing with the 20 – 60 g of ground coffee per litre related in the above mentioned countries (Nunes & Coimbra, 1987).

Several methods to determine AA in food have been developed and they are based on either GC-MS with derivatisation with bromine (Andrawes *et al.*, 1987; Nemoto *et al.*, 2002; Ono *et al.*, 2003; Pittet *et al.*, 2004) and without derivatisation (Jezussek & Schieberle, 2003; Biedermann *et al.*, 2002; Amrein *et al.*, 2005) or LC-MS with derivatisation with mercaptobenzoic acid (Jezussek & Schieberle, 2003) and without derivatisation (Roach *et al.*, 2003; Becalski *et al.*, 2003; Ahn *et al.*, 2002; Rosén *et al.*, 2002; Şenyuva & Gökmen, 2005 & 2006; Delatour *et al.*, 2004; Andrzejewski *et al.*, 2004; Yusà *et al.*, 2006). Independently of the chromatographic technique adopted, the success of the methods is very dependent of the extraction and clean-up steps. The most used procedures usually fail when dealing with coffee products, due to their inability to

73 avoid the presence of interferences that co-elute with the analyte thus preventing its
74 correct quantification (Delatour *et al.*, 2004).

75
76 The methods developed so far to determine AA in coffee and coffee products, are based
77 mainly in LC-MS/MS with variations in the clean-up conditions employed. Roach *et al.*
78 (2003) used a combination of Oasis HLB and Bond Elut-Accucat SPE cartridges before
79 the LC analysis but found coffee to be a troublesome matrix. The method still presents
80 some problems regarding contaminations, as it requires frequently reconditioning the
81 HPLC column (Andrzejewski *et al.*, 2004). Granby and Fagt (2004) performed a single-
82 step clean-up on 300 mg Isolute Multimode SPE columns of five hundreds microlitres
83 of sample to make the clean-up more efficient and eliminate almost completely the
84 colour from the coffee. However, they referred that the loading of the SPE cartridge
85 with the sample exceeding 0.5 mL resulted in a suppression of the MS response.
86 Şenyuva and Gökmen (2005) extracted AA from ground coffee using methanol instead
87 of water and the clean-up was made with Oasis HLB SPE cartridges prior to LC
88 analysis. Delatour *et al.* (2004) used 500 mg of Isolute Multimode SPE cartridges to
89 clean-up 2 mL of samples but realized that a single step was not effective and developed
90 a combination of liquid-liquid and solid phase extraction methods. Ono *et al.* (2003)
91 determined acrylamide in several food items including roasted coffee by GC-MS.
92 However no reference was made about the efficiency of their method to coffee analysis.
93 The set of methods developed so far demonstrate a lack in the methodologies employing
94 GC-MS with the objective to specifically determine AA in coffee samples.

95

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

The objective of this work was to develop a method to determine AA in several coffee products using GC-MS as analytical technique. The method was successfully applied to determine acrylamide in ready-to-drink products and ground coffee, that was analysed as “espresso”.

Experimental

Sampling and coffee preparation

Eighteen samples of roasted coffee beans, five samples of instant coffee, two of them decaffeinated, two samples of coffee blends with cereals and one soluble “cappuccino” were collected from several supermarkets and stored at room temperature. The most common brands were selected. The roasted coffee beans were ground in a coffee grinder incorporated in a Philips coffee machine, model Espresso Professional selected for a fine grain size typical for espresso coffee preparation. Each espresso was prepared in duplicate according to the typical conditions used in Portugal for an “espresso” preparation, i.e.: 7 g of ground coffee extracted with hot water during 30 ± 5 s (about 30 to 35 mL of water) at an inlet pressure of 9 ± 2 bar. Two grams of instant coffee, coffee blends and cappuccino were dissolved in 10 mL of deionised water.

Chemicals

Acrylamide (AA) 99% purity grade was acquired from Aldrich (Steinheim, Germany). Acrylamide 1,2,3- C_{13} ($^{13}C_3$ -AA) 99% in methanol was purchased from Cambridge Isotope Laboratories (Andover, MA, USA). The solid-phase extraction (SPE) columns, Isolute Multimode, 3 g custom made, were from International Sorbent Technology (Hengoed, Mid Glamorgan, UK). The preparative C_{18} sorbent 125 Å, 55-105 µm was

from Waters (Milford, USA). The filters were polyethylene single fritted reservoir 25 mL, filter pore size 10 μm (Symta, Madrid, Spain). The n-hexane and ethyl-acetate were of pesticide residue analysis grade and methanol and acetonitrile were of ultrapure grade, all from Fluka. The potassium bromide for IR spectroscopy grade and the bromine analytical grade were from Merck (Darmstadt, Germany). Sodium chloride analytical grade was from J.T. Baker (Deventer, Holland). Potassium hexacyanoferrate (II) trihydrate, zinc sulphate heptahydrate, acetic acid (glacial) and absolute ethanol were analytical grade obtained from Panreac (Barcelone, Spain). Hydrobromic acid 48% and sodium tiosulphate volumetric solution 1 mol L⁻¹ were from Riedel-de Hen (Seelze, Germany).

Standards and reagents

A stock solution of AA (2 g L⁻¹) was prepared by dissolving the compound in acetonitrile, and appropriately diluted to prepare a working standard solution at 4 mg L⁻¹. A stock solution of internal standard, ¹³C₃-AA, was prepared by the same way with the difference that the working standard solution was made at 40 mg L⁻¹. All stock and working solutions were stored at 4C. Saturated bromine water solution and Carrez reagents were prepared according to Pittet *et al.* (2004).

Sample preparation

To 10 mL of “espresso” coffee or 2 g of other coffee product dissolved in 10 mL of water, 15 mL of absolute ethanol were added and the solution shaken for ~ 1 min. 0.5 μg of IS was added and the solution was shaken again and then kept at -20C during 15 minutes. It was centrifuged at 15000 g (5 minutes, 4C) and the resulting solution was

transferred to a centrifugal flask. Afterwards, it was acidified with acetic acid until pH 4-5 (0.1-0.2 mL). 1 mL of Carrez I and 1 mL of Carrez II solutions were added and then the flask shaken and kept at 4 °C for 15 minutes. The solution was centrifuged at 15000 g (15 minutes, 4°C) and filtrated to a 40 mL vial through the polyethylene filters. The volume of the solution was reduced to 2-3 mL in a rotary evaporator at 55°C and transferred to a 20 mL vial. The rotary evaporator vessel was washed with 2 mL of water that was added to the solution in the vial. If necessary a centrifugation at 3000 g was performed to deposit suspended solids in the solution prior to the SPE clean-up.

SPE clean-up

The SPE columns were prepared by adding a 1g layer of C18 sorbent to the Isolute Multimode 3 g column. This layer was covered with a polyethylene frit (pore size 10 µm). The columns were conditioned first with 20 mL of methanol and then with 20 mL of water. The sample solution (~5 mL) was loaded on the column and the first 2 mL of the eluate were discarded. The remaining solution (~3 mL) was collected in a 40 mL vial and the column was washed with 10 mL of water that was collected to the same vial.

For instant coffee it was necessary to perform a previous SPE cleaning with a 3g C18 columns home made prior to the Isolute cleaning. The column was conditioned under the same conditions as the Isolute column and the sample was loaded and eluted with just 1 mL of water. This solution was collected and loaded in the prepared Isolute columns.

168 *Bromination*

169 To the collected extract 1 g of calcinated KBr was added. The solution was then
170 acidified with HBr until pH 1-3 (100-150 μL) and 2 mL of saturated bromine solution
171 were added. The solution was allowed to stand in an ice bath and kept from the light at
172 least during 1 h. The excess bromine in the solution was decomposed by the addition of
173 1 mol L^{-1} sodium tiosulphate until the yellow colour disappeared (50-150 μL). The
174 extract was saturated with 4 g of NaCl and extracted twice with 10 mL and 5 mL
175 portions of ethyl acetate/n-hexane 4:1 (v/v). The volume was reduced to 3 mL under a
176 stream of nitrogen and a small quantity of anhydrous Na_2SO_4 was added. Finally, the
177 solution was centrifuged at 3000 rpm during 3 min, the upper layer was transferred to
178 another vial and evaporated to 0.5 mL under a gentle stream of nitrogen. The solutions
179 were then injected in the gas-chromatograph.

181 *Calibration standard*

182 Aliquots of the working standard AA solution (equivalent to 0 to 3 μg of AA
183 corresponding to 0 to 300 $\mu\text{g L}^{-1}$ in the samples) were taken into 50 mL centrifugal
184 tubes and 0.5 μg of internal standard was added to each tube and made up to 10 mL
185 with water. These solutions were treated according to the overall procedure (clean-up
186 included) described for the samples.

188 *Equipment*

189 GC-MS analysis were performed in a gas-chromatograph, model HP GC-6890, split-
190 splitless injector, coupled to a Mass Selective Detector model Agilent MSD-5973N
191 (Agilent, Palo Alto, CA, USA). The analytical separation was performed in a capillary

column DB 1301 (30 m x 0.25 μm , 0.25 mm i.d.) from J&W Scientific (Folsom, CA, USA). The centrifugations were made in an ultra-centrifuge from Eppendorf, model 5810 R (Hamburg, Germany) at 15000 g and in a Heraeus Sepatek, model Labofuge Ae (Osterode, Germany) at 3000 g. The SPE clean-up was made in a Visiprep Solid Phase Extraction Manifold from Supelco (Taufkirchen, Germany) with capacity for 12 columns. Evaporation of the solvents was performed in a Büchi Rotavapor model RE 111 and 461 water bath (Flawil, Switzerland). The evaporation under a stream of nitrogen was carried out on a Pierce, model Reacti-therm 18790 (Rockford, IL, USA) with capacity for 9 vials.

201

202 *GC-MS Operating conditions*

203 *Gas-chromatography.* Carrier gas: helium (constant flow at 1 mL min⁻¹). Sample injection volume: 1 μL (splitless, pulsed pressure 32 psi, 60 sec). Injector temperature: 280 °C. Oven temperature: 85°C (1 min), 15°C min⁻¹ to 280°C, hold 10 min (24 min), transfer line, 240 °C.

207 *Mass-spectrometry.* Electron energy, 70 eV (EI mode). Mode of acquisition: selected ion monitoring (SIM), m/z 106, 108, 150 and 152 for 2,3-dibromopropionamide (2,3-DBPA) and m/z 110, 153, 155 for 2,3-¹³C₃- dibromopropionamide (2,3-¹³C₃-DBPA). The ions m/z 150 for 2,3-DBPA, m/z 155 for 2,3-DBPA(¹³C₃) were used for quantification and the others for confirmation. AA was determined with the internal standard using the ratio of peak area of 2,3-DBPA to 2,3-¹³C₃-DBPA. The identity of the peak was confirmed by retention time and by comparing the relative abundance ratios of the confirmatory ions with those of standard solution.

215

216 Results and discussion

217 Method development

218 Aiming for screening of the presence of AA in food products largely consumed in
219 Portugal, we developed a GC-MS methodology (not published) based on the several
220 papers published in this field (Tareke *et al.*, 2002; Nemoto *et al.*, 2002; Ono *et al.*, 2003;
221 Pittet *et al.*, 2004; Biedermann *et al.*, 2002; Delatour *et al.*, 2004). Briefly, AA was
222 extracted from the samples with water at 65 °C, Carrez solutions were added in order to
223 precipitate proteins, the upper layer partially evaporated and finally the compound was
224 brominated, extracted twice by ethyl acetate:n-hexane 4:1 (v/v) and then analysed by
225 GC-MS in SIM mode.

226
227 As reported by other authors (Delatour *et al.*, 2004; Andrzejewski *et al.*, 2004) that used
228 similar methods, some drawback occurred when difficult matrices were analysed,
229 namely coffee products. Taking this into account some modifications were introduced
230 in our previous developed method.

231
232 The addition of absolute ethanol to coffee samples intended to control the great quantity
233 of foam that was constantly being formed during almost all steps of the sample
234 preparation and resulting in severe losses of the analyte and contamination of the used
235 instruments. The problem was particularly serious during the evaporation step on the
236 rotary evaporator because the foam was always being projected. The adopted solution
237 was an adaptation from the work of Nunes and Coimbra (1998), where the precipitation
238 of polymeric polysaccharides, responsible for the “foam stability” in espresso coffees,
239 was achieved by adding 55 and 75 % ethanol solution at 4°C during one hour. After

1
2
3
4
5 240 trials with 5, 10, 15, 20, 25 mL of ethanol, 15 mL were found to give the best results.
6
7 241 The precipitation time was decreased to 15 minutes by changing the precipitation
8
9 242 temperature from 4°C to -20°C.
10
11 243
12
13
14 244 The SPE clean-up was a challenge in the sense that it was difficult to achieve the correct
15
16 245 ratio between the two used sorbents. The use of a single purification step with C₁₈
17
18 246 revealed to be useful to eliminate the coffee colour but it was not effective to remove
19
20
21 247 other compounds that overlapped the 2,3-DBPA peak no matter what quantity of
22
23 248 sorbent was used. When using Isolute Multimode cartridges (0.5 or 1 g) it was verified
24
25 249 that the sorbent layer was completely saturated with very small quantities of coffee
26
27
28 250 samples (0.5-1.0 ml). Because it was difficult to achieve such small sample volumes
29
30 251 during the evaporation step, the rejection of an important portion of the extract was
31
32 252 necessary, with the inherent loss of sensitivity of the overall method. The solution found
33
34 253 was to increase the quantity of sorbent material instead of decreasing the sample
35
36 254 volume. The use of higher quantities of Isolute Multimode (3g) resulted in an increase
37
38 255 of the analytical signal but it was not effective to eliminate the coloured compounds and
39
40 256 suppression of the MS response was observed. The addition of both sorbents in a single
41
42 257 step was the solution found to surpass both problems. It was found that a ratio of 1:3 of
43
44 258 C₁₈: Isolute was ideal to eliminate the most relevant contaminants present in “espresso”
45
46 259 coffee.
47
48
49
50
51 260
52
53
54 261 Soluble coffees, revealed to be a troublesome matrix due to its high percentage of
55
56 262 concentrated polysaccharides. The addition of ethanol was not enough to eliminate the
57
58 263 majority of the polysaccharides that increased the formation of foam and the 1:3 ratio
59
60

between the sorbents was incapable of eliminate all the interferences observed in the analytical signal. Besides, these samples provoked contamination of the injector, what demanded constant changing of the liner sometimes just after few injections. The addition of a bigger top layer of C₁₈ was impracticable due to the dimensions of the used cartridges. The use of a previous clean-up step with a C₁₈ sorbent, in which the majority of the coloured compounds were eliminated, followed by a second clean-up step with the “C₁₈ layered” Isolute column was the best solution found. Notwithstanding the slight decrease observed in the analytical signal, the use of a “clean” extract provided a much better quality of the chromatographic results.

The analytical separation of acrylamide was performed in a DB 1301 capillary column. In Figure 1 is presented a chromatogram for an “espresso” coffee sample with 20,8 µg L⁻¹ AA and 0.5 µg ¹³C₃-AA corresponding 50 µg L⁻¹ in the sample.

[Insert Figure 1 about here]

Method assessment

Linearity. The linearity of the method was tested several times using standard (calibrating) solutions performed in water and treated using the same method developed for the samples. Six standards were simultaneously prepared and treated in parallel with the sample. The range of concentrations varied according to the expected levels of AA in coffee samples. Typical ranges were 0-300 µg L⁻¹ AA in the sample. The quantity of IS was 0.5 µg corresponding to 50 µg L⁻¹ in the sample for all samples and standards.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Calibration curves were constructed by plotting the AA/IS area ratio against the concentration of AA in the standard. The correlation coefficients were usually higher than 0.997.

Precision. To study the precision of the method, an "espresso" coffee (~30 mL) was divided in three aliquots of 10 mL each, which were submitted to the overall developed method and injected twice. The RSD obtained was 6.3 %.

Recovery. For the recovery test, an instant coffee surrogate solution (containing 200.8 $\mu\text{g L}^{-1}$ of AA) was prepared with 8 g of powder dissolved in 40 ml of water and divided into four aliquots of 10 ml each. AA was added at each aliquot at the levels from 0 to 2 μg corresponding to 0-200 $\mu\text{g L}^{-1}$. After addition of IS the samples were analysed by the described overall method. The AA recovery from the spiked solutions varied between 97.4 to 108.4 %.

Limit of detection and limit of quantification

No exhaustive studies for determining the limit of detection were made. However, standards containing 10 $\mu\text{g L}^{-1}$ of AA, were used to construct the calibration curves. Under conditions of ideal performance of the MS detector, it was possible to integrate the corresponding peaks of AA and for the IS. Replicate analysis (extraction procedure included) of these standards showed that the values obtained for the ratio area of AA/IS, presented a R.S.D. less than 15%. This fact was confirmed when samples containing very low levels of AA were analysed. For example, a sample of Espresso coffee containing 20.5 $\mu\text{g L}^{-1}$ of AA was analysed in triplicate ant the RSD obtained was 13.6

312 %. Below this level of concentration the values obtained registered an R.S.D. above
313 15%.

314

315 *Method application*

316 The method developed was applied to 26 coffee samples (roasted beans, instant and
317 coffee blends). Samples were randomly selected from supermarkets, and therefore may
318 not be representative of coffee supply. Nevertheless they represent a general guide to
319 AA concentration in a selected segment of coffee supply. The results obtained are
320 presented in Table I. Accordingly AA levels ranged from 11.4 to 36.2 $\mu\text{g L}^{-1}$ in ground
321 roasted coffee analysed as “espresso”, 47.4 to 95.2 $\mu\text{g L}^{-1}$ for instant coffee and 200.8 to
322 229.4 $\mu\text{g L}^{-1}$ for coffee blends with cereals. Soluble “cappuccino” contained 6.4 $\mu\text{g L}^{-1}$.
323 Considering that 6 g of ground coffee extracted with ~30 mL water are used to prepare
324 an “espresso”, the concentration of AA per cup is therefore 0.32 to 1.46 $\mu\text{g}/30\text{ mL}$. To
325 prepare a cup of instant coffee the usual measure is 2 g per individual portion making
326 0.47 to 0.95 μg per 30 mL. Considering that coffee blends with cereals were prepared as
327 soluble coffee (2 g/ 30 mL cup) the concentration of AA per cup is therefore 2.01 to
328 2.09 μg per 30 mL cup. “Cappuccino” as analysed is a soluble mixture of coffee, milk
329 powder, cocoa and sugar and each individual dose correspond to 14 g, resulting in an
330 AA concentration of 0.45 μg per 30 mL cup. These results highlight the fact that the
331 addition of cereal products to coffee increased the amount of AA. This in agreement
332 with studies (Yusà *et al.*, 2006) reporting that cereal products form substantial amounts
333 of AA due to their chemical constitution and processing.

334

335 [Insert Table I about here]

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

336

337

338

339

340

341

342

343

344

345

346

347

348

349

350

351

352

353

354

355

356

357

358

359

Conclusion

To date, LC/MS has been the most widely used analytical approach to assay AA in coffee. GC/MS methods usually suffer from the presence of co-extractives which rapidly contaminate the chromatographic system, degrading the respective performance and avoiding the correct quantification of AA. This paper describes a GC/MS method based on an improved sample preparation procedure that enables the correct quantification of AA in coffee extracts without the problems generally associated with those kind of samples. The high levels of sensitivity, and reproducibility achieved recommend it when dealing with this type of food matrices. The method was applied to several coffee products and the quantity of AA was found to be dependent on the coffee processing and preparation, namely the addition of cereals.

Acknowledgements

The authors are thankful to FCT for financial support in the framework of the project POCI/AGR/61543/2004.

References

Ahn JS, Castle L, Clarke DB, Lloyd AS, Philo MR, Speck DR. 2002. Verification of the findings of acrylamide in heated foods. Food Additives and Contaminants 19:1116-1124.

Amrein TM, Lukac H, Andres L, Perren R, Escher F, Amado R. 2005. Acrylamide in Roasted Almonds and Hazelnuts. Journal of Agricultural and Food Chemistry 53:7819-7825.

- 1
2
3
4
5 360 Andrawes F, Greenhouse S, Draney D. 1987. Chemistry of acrylamide bromination for
6
7 361 trace analysis by Gas Chromatography and Gas Chromatography-Mass
8
9 362 Spectrometry. *Journal of Chromatography* 399:269-275.
10
11 363 Andrzejewski D, Roach JA, Gay ML, Musser SM. 2004. Analysis of Coffee for the
12
13 364 presence of Acrylamide by LC-MS/MS. *Journal of Agricultural of Food Chemistry*
14
15 365 52:1996-2002.
16
17
18 366 Becalski A, Lau BP, Lewis D, Seaman SW. 2003. Acrylamide in Foods: Occurrence,
19
20 367 Sources, and Modeling. *Journal of Agricultural and Food Chemistry* 51:802-808.
21
22 368 Biedermann M, Biedermann-Brem S, Noti A, Grob K, Egli P, Mändli H. 2002. Two
23
24 369 GC-MS Methods for the Analysis of Acrylamide in Foodstuffs. *Mitteilungen aus*
25
26 370 *Lebensmitteluntersuchung und Hygiene* 93:638-652.
27
28
29 371 Boon PE, de Mul A, van der Voet H, van Donkersgoed G, Brette M, van Klaveren JD.
30
31 372 2005. Calculations of dietary exposure to acrylamide. *Mutation Research* 580:143-
32
33 373 155.
34
35
36 374 Brunner K, Dudler V, Reinhard H, Rhyn P, Rupp H, Sager F, Streule M, Zimmermann
37
38 375 H. and. Zoller O. 2002. Assessment of acrylamide intake by duplicate diet study.
39
40 376 Swiss Federal Office of Public Health, Bern. Available from:
41
42 377 [http://www.bfr.bund.de/cm/208/assessment_of_acrylamide_intake_by_duplicate_di](http://www.bfr.bund.de/cm/208/assessment_of_acrylamide_intake_by_duplicate_diet_study.pdf)
43
44 378 [et_study.pdf](http://www.bfr.bund.de/cm/208/assessment_of_acrylamide_intake_by_duplicate_diet_study.pdf). Accessed 2006 March 15.
45
46
47 379 Delatour T, Perisset A, Goldmann T, Riediker S, Stadler RH. 2004. Improved Sample
48
49 380 Preparation to Determine Acrylamide in Difficult Matrixes Such as Chocolate
50
51 381 Powder, Cocoa, and Coffee by Liquid Chromatography Tandem Mass
52
53 382 Spectrometry. *Journal of Agricultural and Food Chemistry* 52:4625-4631.
54
55
56
57
58
59
60

- 383 Dybing E, Farmer PB, Andersen M, Fennell TR, Lalljie SPD, Müller DJG, Olin S,
384 Petersen BJ, Schlatter J, Scholz G, Scimeca JA, Slimani N, Törnqvist M, Tuijelaars
385 S, Verger P. 2005. Human exposure and internal dose assessments of acrylamide in
386 food. *Food and Chemical Toxicology* 43:365-410.
- 387 Dybing E, Sanner T. 2003. Risk assessment of acrylamide in foods. *Toxicological*
388 *Sciences* 75:7-15.
- 389 Granby K, Fagt S. 2004. Analysis of acrylamide in coffee and dietary exposure to
390 acrylamide from coffee. *Analytica Chimica Acta* 520:177-182.
- 391 Jezussek M, Schieberle P. 2003. A New LC/MS-Method for the Quantification of
392 Acrylamide Based on a Stable Isolute Dilution Assay and Derivatization with 2-
393 Mercaptobenzoic Acid. Comparison with Two GC/MS Methods. *Journal of*
394 *Agricultural Food Chemistry* 51:7866-7871.
- 395 Mottram DS, Wedzicha BL, Dodson AT. 2002. Food chemistry: Acrylamide is formed
396 in the Maillard reaction. *Nature* 419:448-449.
- 397 Nemoto S, Takatsuki S, Sasaki K, Maitani T. 2002. Determination of Acrylamide in
398 Foods by GC/MS Using ¹³C-Labeled Acrylamide as an Internal Standard. *Journal of*
399 *the Food Hygienic Society of Japan* 43:371-376.
- 400 Norwegian Food Control Authority. 2002. Assesement of cancer risk due to
401 acrylamide intake from coffee consumption. Oslo 13 December 2002. Available at
402 <http://www.snt.no>.
- 403 Nunes FM, Coimbra MA, Duarte AC, Delgadillo I. 1997. Foamability, foam stability,
404 and chemical composition of espresso coffee as affected by the degree of roast.
405 *Journal of Agricultural and Food Chemistry* 45:3238-3243.

- 1
2
3
4
5 406 Nunes FM, Coimbra MA. 1998. Influence of polysaccharide in foam stability of
6
7 407 espresso coffee. *Carbohydrate Polymers* 37:283-285.
8
9 408 Ono H, Chuda Y, Ohnishi-Kameyama M, Yada H, Ishizaka M, Kobayashi H, Yoshida
10
11 409 M. 2003. Analysis of acrylamide by LC-MS/MS and GC-MS in processed Japanese
12
13 410 foods. *Food Additives and Contaminants* 20:215-220.
14
15
16 411 Pittet A, Périsset A, Oberson J-M. 2004. Trace level determination of acrylamide in
17
18 412 cereal-based foods by gas chromatography-mass spectrometry. *Journal of*
19
20 413 *Chromatography A* 1035:123-130.
21
22
23 414 Rosén J, Hellenäs K-E. 2002. Analysis of acrylamide in cooked foods by liquid
24
25 415 chromatography tandem mass spectrometry. *Analyst* 127:880-882.
26
27
28 416 Şenyuva HZ ,Gökmen V. 2005. Study of acrylamide in coffee using an improved liquid
29
30 417 chromatography mass spectrometry method: Investigation of colour changes and
31
32 418 acrylamide formation in coffee during roasting. *Food Additives and Contaminants*
33
34 419 22:214-220.
35
36
37 420 Şenyuva HZ ,Gökmen V. 2006. Interference-free determination of acrylamide in potato
38
39 421 and cereal-based foods by a laboratory validated liquid chromatography-mass
40
41 422 spectrometry method. *Food Chemistry* 97:539-545.
42
43
44 423 Stadler RS, Blank I, Varga N, Robert F, Hau J, Guy PA, Robert M-C, Riediker S. 2002.
45
46 424 Acrylamide from Maillard reaction products. *Nature* 419: 449-450.
47
48
49 425 Svensson K., Abramsson L, Becker W, Glynn A, Hellenas KE, Lind Y, Rosen J. 2003.
50
51 426 Dietary intake of acrylamide in Sweden. *Food and Chemical Toxicology* 41:1581-
52
53 427 1586.
54
55
56
57
58
59
60

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

428 Tareke E, Rydberg P, Karlsson P, Eriksson S, Tornqvist M. 2002. Analysis of
429 Acrylamide, a Carcinogen Formed in Heated Foodstuffs. *Journal of Agricultural and*
430 *Food Chemistry* 50:4998-5006.

431 Yusà V, Quintás G, Pardo O, Martí P, Pastor A. 2006. Determination of acrylamide in
432 foods by pressurized fluid extraction and liquid chromatography-tandem mass
433 spectrometry used for a survey of Spanish cereal-based foods. *Food Additives and*
434 *Contaminants* 23:237-244.

435 Zhang Y, Zhang G, Zhang Y. 2005. Occurrence and analytical methods of acrylamide
436 in heat-treated foods: Review and recent developments. *Journal of Chromatography*
437 *A* 1075:1-21.

Figure 1. Chromatogram of a ground coffee sample analysed as “espresso” coffee with $20.8 \mu\text{g L}^{-1}$ AA and addition of $0.5 \mu\text{g}$ of $^{13}\text{C}_3$ -AA corresponding to a concentration of $50 \mu\text{g L}^{-1}$ AA. m/z 150 is the quantification ion for AA and ion m/z 155 is the quantification ion for $^{13}\text{C}_3$ -AA.

Table I. Acrylamide levels measured in 26 coffee products. The minimum and maximum values are presented and also the mean level of AA in the different coffee products analysed.

Samples	n	AA found $\mu\text{g L}^{-1}$			$\mu\text{g AA} / 30 \text{ mL cup}$		
		Min	Mean	Max	Min	Mean	Max
"Espresso"	18	11.4	21	36.2	0.32	0.62	1.46
Soluble coffee	5	47.4	72.4	95.2	0.47	0.72	0.95
Coffee blends with cereals	2	200.8	215.2	229.4	2.01	2.15	2.29
Cappuccino	1	-	6.4	-	-	0.45	-