New characterization methods for monitoring small resonant frequency variation: Experimental tests in the case of hydrogen detection with uncoated silicon microcantilever-based sensors

Mohand-Tayeb Boudjiet, Johan Bertrand, Claude Pellet, Isabelle Dufour

► To cite this version:

Mohand-Tayeb Boudjiet, Johan Bertrand, Claude Pellet, Isabelle Dufour. New characterization methods for monitoring small resonant frequency variation: Experimental tests in the case of hydrogen detection with uncoated silicon microcantilever-based sensors. *Sensors and Actuators B: Chemical*, 2014, 199, pp.269-276. 10.1016/j.snb.2014.03.098 . hal-00964871

HAL Id: hal-00964871

<https://hal.science/hal-00964871>

Submitted on 13 Nov 2014

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

New characterization methods for monitoring small resonant frequency variation: Experimental tests in the case of hydrogen detection with uncoated silicon microcantilever-based sensors.

M.T. Boudjiet¹, J. Bertrand², C. Pellet¹, I. Dufour¹

¹ Univ. Bordeaux, IMS, UMR 5218, F-33400 Talence, France

² Andra, F-92298 Châtenay-Malabry, France

Abstract

The uncoated silicon microcantilever (USMC) operated in the dynamic mode is a new concept in the field of microcantilever-based chemical sensors. Due to the absence of a sensitive layer, this kind of microsensor can only be used for specific applications where it is known that only one chemical species may be varying in concentration, such as monitoring hydrogen release in radioactive waste disposal facilities. Usually, the relative variation of the USMC resonant frequency expected for low concentrations ($\leq 2\%$) of hydrogen in nitrogen is below 50ppm. As a result, the measurement of both the resonant frequency, f_r , and the quality factor, Q , by classical methods, based on the gain spectrum (resonant peak and -3dB bandwidth), is not sufficiently accurate. In this paper, new measurement methods for monitoring f_r and Q variations are proposed: (1) variation of gain and phase at fixed frequencies and (2) polynomial approximations of gain and phase spectra. The performance study of these characterization methods shows that monitoring f_r by using phase linearization yields the best signal-to-noise ratio (e.g., 100 at 0.6% of H₂ in N₂), with 0.02% as a limit of detection for hydrogen.

Keywords

Resonant frequency, quality factor, uncoated silicon microcantilever, chemical detection, signal-to-noise ratio.

1 INTRODUCTION

1.1 Hydrogen release in radioactive waste disposal facility

After having concluded a feasibility study on deep geological disposal for high-level and Intermediate level long-lived radioactive waste in 2005, the French national radioactive waste management agency (Andra) was charged by the Planning Act n°2006-739 to study the design and the creation of an industrial center for geological disposal called Cigéo which must be reversible for at least a century-long period. Within the framework of this geological repository project, the observation and surveillance must contribute to acquire the knowledge required to run the disposal and its reversible management. Hydrogen release is expected in the radioactive waste disposal facility. It originates from (i) radioactive waste release (ii) and anoxic corrosion of metallic materials. In fact, some radioactive wastes (containing α , β , γ radioactivity) resulting from the reprocessing of irradiated fuels are embedded in bitumen matrix. The self-irradiation of the bituminized waste leads mainly to the production of radiolytic hydrogen (75–95 vol. % of produced radiolytic gas). In the deep geological disposal environments steel components would corrode to more oxidized corrosion products and hydrogen gas. The kinetic of Anoxic corrosion of Fe components is not well known, it is actually considered an average rate about 10 $\mu\text{m}/\text{year}$. This assumption still awaits experimental verification in the Underground Research Laboratory sited at Bure in the Meuse district and aims at studying the feasibility of the reversible geological disposal of high-level and long-lived intermediate-level radioactive waste in the Callovo-Oxfordian clay formation (east of France). Despite the fact that hydrogen releases are expected to be small (in the order of 430 mmol/hour for each intermediate level nuclear waste), when ventilation stops with cell closure, concentrations would slowly and regularly increase.

The first calculation gives 4 % hydrogen content in the atmosphere of the radioactive waste cell in less than one year by the end of the oxic corrosion period.

Monitoring of repository structures contributes to security, safety and reversible management of the repository [1]. It is important as it relates to the guidance of the disposal process and to the corresponding decision-making process. In the upstream repository-design phases, the hydrogen monitoring system was planned to work under normal operating conditions and to withstand radiation exposure in case of an accidental event. There exist different kinds of hydrogen sensors, including those that are commercially available and others still under development [2]. In the present paper, the preliminary results concerning the development of microcantilever-based sensors without sensitive coating (Uncoated Silicon MicroCantilever, USMC) for hydrogen detection are presented.

1.2 Resonant frequency shifts of USMC-based sensors

In recent years, interest in microcantilever-based chemical sensing systems has risen due in part to their projected high sensitivity [3-4]. Because of the large ratio of surface area to volume, the microcantilevers are extremely sensitive to surface processes and can be used as micromechanical sensors [5] for high-resolution sensing and probing [6]. For chemical detection, the USMC-based sensors operated in the dynamic mode are used to measure mass density changes of one scalable analyte in gas mixture [7]. The operating principle of this unconventional use of a microcantilever is based on the fact that the resonant frequency of the microresonator depends on both the density and the viscosity of the surrounding medium. In the case of the out-of-plane mode of a wide microcantilever, it has been shown that the predominant effect causing the change in the microcantilever resonant frequency is due to the gas mass density variation [7]. In fact, when the surrounding fluid mass density increases (decreases), the equivalent effective mass of the microcantilever increases (decreases), thereby causing the resonant frequency to decrease (increase). The absence of a sensitive coating leads to a more reliable and reversible behavior because there is no significant absorption and desorption phenomena. However, the major drawbacks of such sensors is that there is no intrinsic selectivity; thus, these sensors can only be used for specific applications, such as when monitoring environments where it is known that only one gas concentration can vary [8].

The conventional method for monitoring the resonant frequency shift consists of searching for the resonant peak on the amplitude spectrum and tracking the change in the corresponding frequency. The relative variation of the resonant frequency in the case of low concentration of hydrogen (H_2) in nitrogen (N_2) is very small. For instance, using an USMC (length=5mm, width=1mm and thickness=170 μ m, resonant frequency approximately 10 kHz), a concentration of 0.2% of H_2 in N_2 corresponds to a resonant frequency change of approximately 0.02 mHz. Therefore, the conventional method does not prove to be sufficiently accurate to measure such a small shift in resonant frequency. This is mainly due to the difficulty of identifying precisely the resonant peak location due to the measurement noise.

The quality factor is a key parameter in resonating sensors because it determines the stability of the resonator in terms of resonant frequency. It is defined as the ratio of the vibrational energy stored in the system to the total energy lost per cycle. This parameter is also commonly used to characterize resonating sensors [9, 10] that are typically used for viscosity measurements [11, 12]. In high quality factor systems, and when the resonant peak is sufficiently high to contain the -3dB bandwidth, the -3dB measurement method is often used to calculate the quality factor. However, when the magnitude of the resonant peak is less than 3dB, this method cannot be applied.

Other methods for monitoring the shift in resonant frequency consist of making measurements at a fixed frequency instead of using a frequency sweep. By fixing the measurement frequency at the value corresponding to the maximum slope of the gain [13], impedance [14] or phase spectra, changes in the resonant frequency can be detected simply by monitoring variations of response magnitude.

In order to reduce the impact of the noise, polynomial approximations for the gain or phase spectrum can be used. These polynomial approximations can also be used to analytically extract various response characteristics such as natural frequency, resonant frequency and quality factor.

In this paper, the different characterization methods using gain and phase spectra have been tested on measurement data obtained with a USMC device with integrated actuation and read-out systems. To generate measurement data, the sensor has been subjected to various concentrations of hydrogen in nitrogen. The study is focused on the performance of the different characterization

methods in terms of signal-to-noise ratio (SNR) and the limit of detection (LOD) corresponding to a signal-to-noise ratio of 3. Following a description of the different methods to measure small shifts in resonant characteristics in section 2, the experiments using an uncoated silicon microcantilever-based sensor (USMC) for hydrogen detection are presented in section 3. Finally, a comparison of the results of the different methods applied to these measurements are presented and discussed in section 4.

2 CHARACTERIZATION METHODS

In this section, two conventional methods (sections 2.1 and 2.2) and four novel characterization methods (sections 2.3-2.6) are presented.

2.1 Variation of resonant frequency: Δf_r

The measurement of the resonant frequency shift, Δf_r , can be simply performed by looking for the resonant peak in the gain spectrum, then associating it with the corresponding frequency as shown in *Fig. 1.a*. However, for small changes of the resonant frequency (USMC case) this technique is not very efficient because of the noise on the gain spectrum inducing significant measurement uncertainty. Nevertheless, eliminating the noise on the gain using appropriate fitting methods (section 2.5) can considerably improve the efficiency of this characterization method.

2.2 Variation of quality factor: ΔQ

The conventional measurement method for extracting the quality factor from experimental measurements is the $-3dB$ bandwidth method, as shown in *Fig. 1.b*. This technique can only be applied to high quality factor systems and it consists of calculating the ratio between the resonant frequency and the $-3dB$ bandwidth (Δf_{-3dB}), measured on the gain spectrum:

$$Q = \frac{f_r}{\Delta f_{-3dB}} \quad (1)$$

Just as was the case for the Δf_r method of the previous section, measuring changes in the quality factor by the $-3dB$ bandwidth method is not effective in the case of small resonant frequency changes and small quality factor changes, because this method requires not only the measurement of Δf_r , but requires also the measurement of $\Delta(\Delta f_{-3dB})$, i.e., the change in the bandwidth.

2.3 Variation of gain at fixed frequency: ΔG_{FF}

After setting the measurement (actuation) frequency, f_{mes} , on the point approximately representing the highest slope in the gain spectrum, the characterization of the resonant frequency variation is possible, thanks to the gain variation obtained at f_{mes} as shown in *Fig. 1.c*.

Moreover, to increase the measurement sensitivity and eliminate the noise on the gain, a linear fit can be applied to the portion of the gain spectrum that includes the maximum slope (inflection point). In practice, it is not efficient to use a general equation to determine the inflection point, because, the shape of the spectrum depends strongly on the coupling between the actuation and sensing circuits (crosstalk). However, a visual inspection of the gain spectrum is sufficient to determine approximately the inflection point, which allows determining the appropriate measurement frequency, f_{mes} , thus improving the efficiency of this characterization method.

Another approach can also be used to eliminate the noise present on the gain spectrum around the resonant peak. In fact, around the resonant frequency, f_r , the gain spectrum can be approximated by a parabolic curve by using the least-squares method. This approach allows not only the optimization of the ΔG_{FF} and Δf_r methods for small-span measurements (span $\leq 4Hz$ in our case), but also the determination of other characteristic parameters as presented in section 2.5.

2.4 Variation of phase at fixed frequency: ΔP_{FF}

The linear appearance and high slope around the resonant frequency shown in the phase spectrum (*Fig. 1.d*) is a major asset for the measurement of small resonant frequency variation, as was seen in the ΔG_{FF} method presented previously. The ΔP_{FF} method is

based on the same principle as the ΔG_{FF} method; however, in this case the value of f_{mes} is fixed at the inflection point of the phase spectrum (corresponding to a phase of $-\pi/2$) as shown in Fig.1.d.

The fit of the phase spectrum around the resonant frequency, f_r , by a straight line, using least-squares method, allows one to not only eliminate noise on the phase and, thus, optimize the performance of this method, but also to deduce other characteristic parameters as discussed in the section 2.6.

2.5 Parabolic approximation of the gain spectrum near the resonance peak

As mentioned in the Δf_r method, to reduce the impact of noise-induced measurement uncertainty on the gain spectrum, a parabolic approximation can be performed around the resonance peak. Using the least-squares method, the a_G , b_G and c_G coefficients in the following fitting equation can be calculated:

$$|H| \simeq a_G + b_G f + c_G f^2 \quad (2)$$

The resonant frequency, f_r , is the maximum of equation (2); thus,

$$f_r = \frac{-b_G}{2c_G} \quad (3)$$

The gain transfer function of the microcantilever operating in the linear bending domain is similar to that of a second-order system and is given by

$$|H| = \frac{H_0}{\sqrt{\left(1 - \frac{f^2}{f_0^2}\right)^2 + \left(2\xi \frac{f}{f_0}\right)^2}} \quad (4)$$

where H_0 is the static gain, f_0 is the natural frequency and ξ the damping ratio.

To determine an analytical parabolic expression of the gain around the resonance peak in terms of system parameters f_0 and ξ , a second-order Taylor series expansion is performed on expression (4), yielding

$$|H| \simeq \frac{H_0}{2\xi} \left[3 - \frac{1}{2\xi^2} + \left(-3 + \frac{1}{\xi^2}\right) \frac{f}{f_0} + \left(1 - \frac{1}{2\xi^2}\right) \frac{f^2}{f_0^2} \right] \quad (5)$$

Matching coefficients on equations (2) and (5), we obtain

$$\begin{cases} \frac{a_G c_G}{b_G^2} = \frac{\left(3 - \frac{1}{2\xi^2}\right) \left(1 - \frac{1}{2\xi^2}\right)}{\left(-3 + \frac{1}{\xi^2}\right)^2} & (a) \\ H_0 = \frac{2a_G \xi}{3 - \frac{1}{2\xi^2}} & (b) \\ \frac{a_G}{b_G} = f_0 \frac{3 - \frac{1}{2\xi^2}}{-3 + \frac{1}{\xi^2}} & (c) \end{cases} \quad (6)$$

Equation 6.a may be rewritten to give an explicit expression for the quality factor in terms of the fitting coefficients:

$$Q = \frac{1}{2\xi} = \sqrt{\frac{(3A-1) - \frac{1}{2}\sqrt{1-3A}}{4A-1}} \quad (7)$$

with $A = \frac{a_G c_G}{b_G^2}$.

Remarks:

• The parabolic approximation of the gain is only valid on the part of the gain spectrum very close to the resonance (frequency span $< \Delta f_{3dB}$).

• The second root, $\frac{1}{2\xi} = \sqrt{\frac{(3A-1) + \frac{1}{2}\sqrt{1-3A}}{4A-1}}$, of equation (6a) gives a false Q -factor value, so only the root given by equation (7) is considered.

2.6 Linear approximation of the phase spectrum near the resonance peak

In the linear bending regime the phase transfer function of the microcantilever can be written as

$$\varphi(f) = -\text{Arctan} \left(\frac{2\xi \frac{f}{f_0}}{1 - \frac{f^2}{f_0^2}} \right) \quad (8)$$

A linear fit of the phase spectrum around f_0 using the least-squares method allows for the numerical determination of the coefficients a_p and b_p in the fitting equation below:

$$\varphi(f) = a_p + b_p f \quad (9)$$

Furthermore, an analytical expression of the phase in the vicinity of the eigenfrequency f_0 may be used to determine f_0 and ξ of the microcantilever. To obtain this expression, a first-order Taylor series expansion is performed on the phase transfer function around f_0 (i.e., where the phase is to $-\pi/2$). The result of this calculation yields

$$\varphi(f) = -\frac{\pi}{2} + \frac{1}{\xi} \left(1 - \frac{f}{f_0} \right) \quad (10)$$

from which one may obtain the following:

$$\begin{cases} Q = \frac{1}{2\xi} = -\frac{1}{2} \left(\frac{\pi}{2} + a_p \right) & (a) \\ f_0 = \frac{1}{\xi b_p} = \frac{-\frac{\pi}{2} - a_p}{b_p} & (b) \end{cases} \quad (11)$$

Remarks:

• The linear approximation of the phase is only valid in the part of the phase spectrum that is very close to resonance (span $< \Delta f_{3dB}$).

- Once f_0 and ξ are calculated, it is easy to deduce the resonant frequency: $f_r = f_0 \sqrt{1 - 2\xi^2}$

Fig. 1 : Principle of the different characterization methods: (a) Δf_r : measurement of the resonant frequency variation corresponding to the maximum of the gain. (b) ΔQ : measurement of the quality factor variation using the -3dB bandwidth method, (c) ΔG_{FF} : measurement of the gain variation at a fixed frequency, f_{mes} , corresponding approximately to the maximum slope of the gain, (d) ΔP_{FF} : measurement of the phase variation at a fixed frequency, f_{mes} , corresponding approximately to the maximum slope of the phase.

3 EXPERIMENTS

In order to apply the different characterization methods presented in the section 2, measurements of hydrogen in nitrogen have been conducted using: (1) a USMC with integrated electromagnetic actuation and piezoresistive sensing (*Fig.2*), and (2) a gas line for mixing gas with controlled flow and concentrations (*Fig.3*).

3.1 Microcantilever design

The microcantilever has been fabricated by ESIEE group (Paris, France, www.esiee.fr) using classical silicon technologies [15].

To generate vibrations, an AC current is passed through the conductive wire placed at the cantilever periphery. In the presence of a magnetic field collinear to the longitudinal axis of the beam, an AC Lorentz force is created at the microcantilever free-end and induces out-of-plane vibrations. In order to detect these vibrations, semiconductor strain gauges (boron-doped piezoresistors) are fabricated during the process. They are arranged in a half Wheatstone bridge configuration: a first gauge is located where the strains are maximum (at the clamped-end of the beam) and the other one is on the rigid substrate (*Fig.2*). The USMC characteristics in air at 23°C are: length=5mm, width=1mm, thickness=170μm, $f_r \approx 9662\text{Hz}$ and $Q \approx 1350$.

Fig. 2 : Uncoated silicon microcantilever-based sensor (USMC) design: a conductive wire is placed at the free-end of the cantilever for the electromagnetic actuation by Lorentz force; piezoresistors are placed on the clamped-end of the cantilever and on the substrat for cantilever vibration measurement.

3.2 Experimental setup

The microcantilever is placed in a hermetic gas chamber (total volume: 500 μl) under a controlled gas flow. The gas line (Fig.3) contains mass-flowmeters allowing control of flow (100–1000 ml/min) and concentration of each gas. To perform chemical detection, a gas stream containing a $\text{H}_2\text{-N}_2$ mixture flows into the gas chamber in which the USMC has been placed. The tested succession of the different concentrations of H_2 in N_2 is reported in different detection curves.

Fig. 3 : Experimental setup. Gas line constituted of hydrogen (4% $\text{H}_2\text{-N}_2$) and nitrogen (100% N_2) gas bottles and mass-flowmeters controlled by a computer. The uncoated silicon microcantilever-based sensor is put in a hermetic gas chamber connected to the $\text{H}_2\text{-N}_2$ mixture gas flow (usually 100ml/min but can be increased up to 1000ml/min).

In order to study the behavior of the USMC in the presence of different concentrations of H_2 in N_2 , a spectrum analyzer (HP4194A) is used to acquire gain and phase spectra with the following configuration: *INTEG.TIME=Short*, *AVRG=8bits*. The feature of this device is that it has only 400 measurement points. Therefore, for small f_r shifts ($\leq 1\text{Hz}$), small spans must be applied to increase the measurement sensitivity. For example, to detect f_r variations of at least 20mHz, a measurement span of no more than 8Hz (8Hz/400 points) can be used. In our case a span of 4Hz is used for the detection of different H_2 in N_2 concentrations. Each H_2 in N_2 concentration lasts 3 minutes with a constant flow of 100 ml/min.

After following the experimental protocol detailed previously, the gain and phase spectra obtained for the concentrations of H_2 in N_2 are presented in Fig.4. The gain and phase translations to the right observed in Fig.4.a and Fig.4.b, respectively, are consistent with the decrease in the gas mass density when the H_2 concentration increases. This type of resonant frequency shift in the USMC due to density changes in the surrounding gas has been discussed in an earlier work [7].

Fig. 4 : Evolution of spectra of the USMC (5mm x 1mm x 170μm) near the resonance as H₂ in N₂ concentration is changed from 0% to 2% with a gas flow of 100ml/min. (a) Gain spectrum. (b) Phase spectrum. Small modifications of both the gain and phase spectra are observed due to the increase of the resonant frequency with the hydrogen concentration.

4 RESULTS AND DISCUSSION

In this section, the different characterization methods presented in this paper are tested. In order to compare performances of these methods in terms of signal-to-noise ratio, SNR, the relative variation of f_r , f_0 , Q , G_{FF} and P_{FF} are presented.

The results of Fig.5 demonstrate the validity of polynomial approximations of gain and phase spectra around resonance for a frequency span of 4Hz. These approximations are used to fit spectra, enabling one to increase the SNR (section 4.1) and to deduce f_r , f_0 and Q (section 4.2) using equations (3), (7) and (11). The measurement frequencies (f_{mes}), which are different for the ΔG_{FF} method and ΔP_{FF} method, are also indicated in the gain (Fig.5.a) and phase (Fig.5.b) spectra; they are respectively placed at the maximum slope of the gain ($f_{mes}=9564\text{Hz}$) for the ΔG_{FF} method and at the natural frequency ($f_{mes}=f_0=9562\text{Hz}$) for the ΔP_{FF} method.

Fig. 5 : Agreement between least-squares polynomial approximations and raw spectra. (a) Gain spectrum. (b) Phase spectrum. The measurement frequencies are shown for both the gain measurement (highest frequency of the measurements which corresponds to the maximum slope of the gain) and the phase measurement (frequency corresponding to a phase equal to -90° which corresponds to the maximum slope of the phase).

4.1 Performance of the Δf_r , ΔG_{FF} and ΔP_{FF} methods

The detection results of the Δf_r , ΔG_{FF} and ΔP_{FF} characterization methods, obtained for concentrations of H₂ in N₂ ranging from 2% to 0.2% are presented in Fig. 6.

As expected, monitoring f_r by the conventional method without fitting the gain spectrum is not efficient because of the noise on the gain (Fig.6.a). However, fitting the gain with a parabolic approximation and using relation (3) to monitor f_r drastically improves the SNR (blue line in Fig.6.a).

The ΔG_{FF} (Fig.6.b) and ΔP_{FF} (Fig.6.c) methods show good sensitivity; they are able to detect a concentration of 0.2% of H_2 in N_2 even without fitting. Moreover, the polynomial fits have greatly improved the SNR of both methods (blue lines). Table 1 presents the performance of each method, calculated at 0.6% of H_2 in N_2 . The noise is estimated by calculating the standard deviation (σ_x). From Table 1, it can be seen that the ΔP_{FF} method has the best SNR (S/N=100), the best relative variation ($\Delta x/x = 0.86\%$) and the best limit of detection (hydrogen concentration corresponding to a SNR equal to 3, = 0.018% of H_2 in N_2).

The performance of the ΔG_{FF} method can be optimized by fixing f_{mes} at the point of maximum slope of the gain spectrum. Thus, another experiment is carried in the same conditions (the same concentrations of H_2 in N_2 and the performances are calculated at 0.6% of H_2 in N_2), but using a frequency span of 10Hz in order to identify the maximum slopes of both the gain and phase spectra. In these conditions, the parabolic fitting cannot be applied to the whole spectrum (Fig.7, red line with 'x' marks); it can only be applied closely around resonance (Fig.7, green line with triangle marks). However, for large span measurements only the tangent to the curve at the inflection point (Fig.7, black line with circle marks) is used to improve the SNR. The performance of each method is reported in Table 2. The comparison between the ΔG_{FF} and ΔP_{FF} methods in their optimal conditions shows that the ΔP_{FF} method yields a SNR of 32 in comparison with 25 for the ΔG_{FF} method, thereby confirming that ΔP_{FF} is the best characterization method that we have examined to date. This is due to the fact that the phase spectrum presents a higher slope and a lower noise than the gain spectrum.

Remark:

Because of the large frequency span (10Hz) chosen in this measurement, the signal-to-noise ratios (Table 2) are relatively small compared to the values found in Table 1. This is due to the configuration of the gain/phase analyzer (HP5194A) that is maintained at the same level (INTEG.TIME=Short, AVRGE=8bits) as for the measurements performed with the short span (4Hz). This results in a higher noise level associated with the larger frequency span.

Fig. 6 : Detection curves of the USMC (5mm x 1mm x 170 μ m) as H_2 in N_2 concentration is changed from 2% to 0.2% with a gas flow of 100ml/min. The duration of each concentration step is 3min and between each concentration step pure nitrogen is injected in the gas chamber during 3min. The same measurements are used to test the three different methods: (a) Δf_r method. (b) ΔG_{FF} method. (c) ΔP_{FF} method.

Fig. 7 : Comparison between the real gain spectrum and the various fits adopted for the USMC (5mm x 1mm x 170µm) with a 10Hz span around the resonant frequency.

Table 1 : Performance of ΔP_{FF} and Δf_r methods ($\Delta x/x$: relative variation. σ_x : noise estimation (standard deviation of the measured relative variation). LOD: limit of detection). Frequency span =4 Hz.

Characterization methods		$\left(\frac{\Delta x}{x}\right)_{0.6\% H_2} (\%)$	$\sigma_x (\%)$	$\left(\frac{S}{N}\right)_{0.6\% H_2}$	LOD
Without fitting	Δf_r	-----	0.82×10^{-3}	-----	-----
	ΔG_{FF}	0.72	82×10^{-3}	8.8	0.2%
	ΔP_{FF}	0.88	54×10^{-3}	16	0.11%
With fitting	Δf_r	0.72×10^{-3}	15×10^{-6}	48	0.037%
	ΔG_{FF}	0.60	16×10^{-3}	38	0.047%
	ΔP_{FF}	0.86	8.5×10^{-3}	100	0.018%

Table 2 : Performance of ΔP_{FF} and ΔG_{FF} methods ($\Delta x/x$: relative variation. σ_x : noise estimation (standard deviation of the measured relative variation). LOD: limit of detection). Frequency span =10 Hz.

Characterization methods		$\left(\frac{\Delta x}{x}\right)_{0.6\% H_2} (\%)$	$\sigma_x (\%)$	$\left(\frac{S}{N}\right)_{0.6\% H_2}$
Without fitting	ΔG_{FF}	0.71	140×10^{-3}	5.1
	ΔP_{FF}	0.77	94×10^{-3}	8.2
	ΔG_{FF}	0.52	20×10^{-3}	26
With fitting	ΔP_{FF}	0.72	22×10^{-3}	33

4.2 Performance of polynomial approximation methods

In the previous part, polynomial approximations using the least-squares method were applied to fit the gain and phase spectra near resonance in order to decrease the impact of the noise. Using the polynomial coefficients of the fitting methods, equations (3), (7), (6.c) and (11) can be used to deduce f_r , Q and f_0 . In this experiment, we compare only the performance of each fitting method in terms of the determination of f_0 and Q . The results of this processing and the corresponding performance values are presented in *Fig. 8* and *Table 3*, respectively, for a parabolic approximation of the gain and a linear approximation of the phase.

It can be observed that variations of Q and f_0 in the presence of different concentrations of H_2 in N_2 are detected using the different polynomial methods. All performance metrics are reported in *Table 3*, where it can be seen that Q measurements (green lines with stars in *Fig. 8*) present a poor SNR (2.5 and 1.8 for gain and phase fits, respectively) and cannot detect H_2 concentrations in N_2 below 1%. Also, the actual Q value is around 1350, while the measured Q values based on the polynomial approximations are approximately 900-950. This estimation error is mostly caused by the span measurement (4 Hz) which is relatively large in comparison with the range of validity of these polynomial methods. However, the exact Q measurement is not necessary; only the Q variation is important in the present application, i.e., to sufficiently characterize USMC response to a change in H_2 concentration in N_2 . On the other hand, f_0 measurements utilizing both the gain and phase polynomial methods (blue lines in *Fig. 8*) yield a good SNR (46 and 82 for gain and phase spectra, respectively) and a good f_0 estimation (at 0% of H_2 in N_2 , temperature $\approx 23^\circ C$ and $f_0 \approx 9562$ Hz). The comparison between the f_0 determinations shows that the linear approximation of the phase near resonance results in the best SNR and consequently the best LOD.

Fig. 8 : Detection curves of the USMC (5mm x 1mm x 170 μ m) as H₂ in N₂ concentration is changed from 2% to 0.2% with a gas flow of 100ml/min. The duration of each concentration step is 3min and between each concentration step pure nitrogen is injected in the gas chamber during 3min. It is the same data measurements as the one of Fig 6 and the methods presented in sections 2.5 and 2.6 are tested. (a) The parabolic approximation of the gain spectrum near the resonance peak is used and both the eigenfrequency f_0 and the quality factor Q are plotted using Eqs. 6c and 7; (b) the linear approximation of the phase spectrum near the resonance peak is used and both the eigenfrequency f_0 and the quality factor Q are plotted using Eqs. 11a and 11b.

Table 3 : Performance of polynomial gain and phase approximations ($\Delta x/x$: relative variation. σ_x : noise estimation (standard deviation of the measured relative variation). LOD: limit of detection). Frequency span= 4Hz.

		$\left(\frac{\Delta x}{x}\right)_{0.6\% H_2}$ (%)	σ_x (%)	$\left(\frac{S}{N}\right)_{0.6\% H_2}$	LOD
Characterization methods					
Parabolic approximation: gain fitting	ΔQ	330×10^{-3}	130×10^{-3}	2.5	0.72%
	Δf_0	0.73×10^{-3}	16×10^{-6}	46	0.04%
Linear approximation : phase fitting	ΔQ	74×10^{-3}	41×10^{-3}	1.8	1%
	Δf_0	0.74×10^{-3}	9×10^{-6}	82	0.022%

4.3 Limit of detection

The analysis of the performance tables (Tables 1-3) shows that phase linearization around resonance optimizes the measurement sensitivity of the resonant frequency variation using the ΔP_{FF} method (SNR=100). Therefore, it would be interesting to assess the corresponding limit of detection (LOD).

In theory, the LOD of the ΔP_{FF} method is approximately 0.02% of H_2 in N_2 . To verify this value, another experiment is carried out using a concentration of 0.025% of H_2 in N_2 . Because of the accuracy limit of the flowmeters, the gas flow is increased in this experiment to 200ml/min in order to achieve this low concentration of H_2 in N_2 . The measurement result using the ΔP_{FF} method is presented in Fig. 9, where it can be seen that a concentration of 200 ppm (0.025%) has been successfully detected.

Fig. 9 : Detection curve of the USMC (5mm x 1mm x 170µm) using the ΔP_{FF} method as H_2 in N_2 concentration is changed from 0% to 0.025% and to 0% with a gas flow of 200ml/min..

5 CONCLUSION

In summary, the phase linearization using the least-squares method and the linear model around resonance to measure the phase variation at a fixed frequency (Table 1) or the natural frequency variation (Table 3) present the best SNRs among all characterization methods studied in this work. This is due to the fact that the phase spectrum results in lower noise and a higher slope compared to the gain spectrum. Using the ΔP_{FF} method, 0.025% of H_2 in N_2 has been detected using an USMC with the characteristics mentioned in section 3.1.

Unlike the conventional resonant frequency variation method (Δf_r), the ΔP_{FF} method allows accurate measurement of small resonant frequency variation of the USMC, thanks to its high phase slope around resonance (high Q). In addition, the linearized phase model enables one to extract an accurate value of natural frequency variation (Δf_0) and to monitor accurately both small and large (if the frequency span is large enough) resonant frequency variations (but with less sensitivity than the ΔP_{FF} method for small resonant frequency variation).

Another advantage of the ΔP_{FF} method is the simplicity involved in making an integrated measurement with only one circuit that performs the phase comparison between the actuation signal at a fixed frequency ($f \approx f_r$) and the read-out signal. The major drawback of this method is that it can only be used for small resonant frequency variation (due to the smallness of phase linearity domain).

Acknowledgement

This work is supported by Andra. The authors are grateful for interesting discussions with Stephen Heinrich on the paper content and for his suggestions for improving the paper.

References

- [1] S. Buschaert, S. Lesoille-Delepine, J. Bertrand, S. Mayer, P. Landais, Developing the tools for geologic repository monitoring - Andra's monitoring R&D program, *Proc. of the WM2012 conference, Phoenix, 2012*.
- [2] T. Hübert, L. Boon-Brett, G. Black, U. Banach, Hydrogen Sensors- A review, *Sensors and Actuators B*, 157 (2011), 329-352.
- [3] K.M. Goeders, J.S. Colton, L. A. Bottomley, Microcantilevers: sensing chemical interactions via mechanical motion, *Chemical Reviews*, 108 (2008), 522-542.
- [4] Q. Zhu, Microcantilever sensors in biological and chemical detections, *Sensors & Transducers Journal*, 125 (2011), 1-21.
- [5] A. Boisen, S. Dohn, S.S. Keller, S. Schmid, M. Tenje, Cantilever-like micromechanical sensors, *Reports on Progress in Physics*, 74 (2011), 036101.
- [6] D. W. Lee, X. Li, Integrated microcantilevers for high-resolution sensing and probing, *Measurement Science and Technology* 23 (2012), 029501.
- [7] S. Tétin, B. Caillard, F. Ménil, H. Debéda, C. Lucat, C. Pellet, I. Dufour, Modeling and performance of uncoated microcantilever-based chemical sensors, *Sensors and Actuators B*, 143 (2010), 555-560.
- [8] I. Dufour, F. Josse, S. Heinrich, C. Lucat, C. Ayela, F. Ménil, O. Brand, Unconventional uses of microcantilevers as chemical sensors in gas and liquid media, *Sensors and Actuators B*, 170 (2012), 115-121.
- [9] S. Schmid, P. Senn, C. Hierold, Electrostatically actuated nonconductive polymer microresonators in gaseous and aqueous environment, *Sensors and Actuators A*, 145-146 (2008), 442-448.
- [10] J.E.Y. Lee, Y. Xu, Direct inference of parameters for piezoresistive micromechanical resonators embedded in feedthrough, *Sensors and Actuators A*, 186 (2012), 257-263.
- [11] A. Abdallah, M. Heinisch, B. Jakoby, Measurement error estimation and quality factor improvement of an electrodynamic-acoustic resonator sensor for viscosity measurement, *Sensors and Actuators A*, 199 (2013), 318-324.
- [12] A.M. Schilowitz, D.G. Yablona, E. Lansey, F.R. Zypman, Measuring hydrocarbon viscosity with oscillating microcantilevers, *Measurement*, 41 (2008), 1169-1175.
- [13] T.R. Albrecht, P. Grütter, D. Horne, D. Rugar, Frequency modulation detection using high-Q cantilevers for enhanced force microscope sensitivity, *Journal of Applied Physics*, 69 (1991), 668-673.
- [14] S. Xu, R. Mutharasan, A novel method for monitoring mass-change response of piezoelectric-excited millimeter-sized cantilever (PEMC) sensors, *Sensors and Actuators B*, 143 (2009), 144-151.
- [15] L. Fadel, F. Lochon, I. Dufour, O. Français, Chemical sensing: millimeter size resonant microcantilever performance, *Journal of Micromechanics and Microengineering*, 14 (2004), S23-S30.

Biographies

Mohand-Tayeb Boudjiet received in 2012 the Master degree in embedded electronics systems from University of Bordeaux in France. He is currently a Ph.D student at IMS Laboratory and he is working on microcantilevers for density measurement and chemical detections.

Johan Bertrand received in 2005 the M.S. degree solid and inorganic chemistry from Rennes 1 University, France. In 2009, he got a Ph.D degree in physic and chemistry from the University of Tübingen and École des Mines de Saint-Étienne. Since 2009, he works for the National Radioactive Waste Management Agency as a Research Engineer working an overall monitoring strategy and sensors development for the deep geological disposal for high-level and long-lived radioactive waste.

Claude Pellet was a researcher at the "Institut d'Electronique Fondamentale" from the university Paris XI-Orsay, where he has studied the deposition of thin film by ion beam sputtering from 1982 to 1993. He joined university of Bordeaux as full professor in 1993. His work focuses on micro -technology, micro-system development (humidity sensor), assembly technology and micro-systems reliability. He is currently the director of the IMS Laboratory (400 people).

Isabelle Dufour graduated from Ecole Normale Supérieure de Cachan in 1990 and received the Ph.D. and H.D.R. degrees in engineering science from the University of Paris-Sud, Orsay, France, in 1993 and 2000, respectively. She was a CNRS researchfellow from 1994 to 2007, first in Cachan working on the modelling of electrostatic actuators (micromotors, micropumps) and then after 2000 in Bordeaux working on microcantilever-based chemical sensors. She is currently Professor of electrical

engineering at the University of Bordeaux and her research interests are in the areas of microcantilever-based sensors for chemical detection, rheological measurements, material characterization and energy harvesting.