



HAL
open science

Dynamic channel allocation for interference mitigation in relay-assisted wireless body networks

Mohamad Jaafar Ali, Hassine Moun gla, Ahmed Mehaoua, Yong Xu

► **To cite this version:**

Mohamad Jaafar Ali, Hassine Moun gla, Ahmed Mehaoua, Yong Xu. Dynamic channel allocation for interference mitigation in relay-assisted wireless body networks. 2015 2nd International Symposium on Future Information and Communication Technologies for Ubiquitous HealthCare (Ubi-HealthTech'15), May 2015, Beijing, China. pp.1 - 6, 10.1109/Ubi-HealthTech.2015.7203331 . hal-01399595

HAL Id: hal-01399595

<https://hal.archives-ouvertes.fr/hal-01399595>

Submitted on 21 Nov 2016

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Dynamic Channel Allocation for Interference Mitigation in Relay-assisted Wireless Body Networks

Mohamad Jaafar Ali*, Hassine MOUNGLA*, Ahmed Mehaoua*, Yong XU†

*LIPADE, University of Paris Descartes, Sorbonne Paris Cité, Paris, France

†School of Computer Science and Engineering, South China University of Technology, China,

Email: {mohamad.ali; hassine.moungla; ahmed.mehaoua}@parisdescartes.fr

Abstract—We focus on interference mitigation and energy conservation within a single wireless body area network (WBAN). We adopt two-hop communication scheme supported by the the IEEE 802.15.6 standard 2012 [14]. In this paper, we propose a dynamic channel allocation scheme, namely DCAIM to mitigate node-level interference amongst the coexisting regions of a WBAN. At the time, the sensors are in the radius communication of a relay, they form a relay region (RG) coordinated by that relay using time division multiple access (TDMA). In the proposed scheme, each RG creates a table consisting of interfering sensors which it broadcasts to its neighboring sensors. This broadcast allows each pair of RGs to create an interference set (IS). Thus, the members of IS are assigned orthogonal sub-channels whereas other sensors that do not belong to IS can transmit using the same time slots. Experimental results show that our proposal mitigates node-level interference and improves node and WBAN energy savings. These results are then compared to the results of other schemes. As a result, our scheme outperforms in all cases. Node-level signal to interference and noise ratio (SINR) improved by 11dB whilst, the energy consumption decreased significantly. We further present a probabilistic method and analytically show the outage probability can be effectively reduced to the minimal.

I. INTRODUCTION

A WBAN is a wireless network of wearable sensors that may be embedded inside or attached on the human body. The sensors are used in various applications such as health monitoring, ubiquitous healthcare, sports and entertainment. These networks mainly monitor vital signs as glucose percentage in blood, heart beats, respiration, body temperature and/or can record electrocardiography (ECG) [2], [10], [13].

WBANs are more likely to coexist with each other due to their dynamic and mobile nature. Interferences can happen due to the absence of coordination amongst them. Also, concurrent transmissions of a particular WBAN can interfere with each others as well as introduce interferences at some other nodes (intra-WBAN interference). Firstly, radio interference is of the paramount importance since it can quickly degrade a WBAN's performance. Secondly, the stringent factor is energy and so this requires to always keep as low power consumption as possible. However, the latest version of the IEEE 802.15.6 standard proposed in February 2012 for WBANs supports two-hop communication and states that up to 10 co-located WBANs can function properly within a transmission range of 3 meters [12], [13].

Due to the constrained nature of WBANs in terms of energy, size and cost, advanced antenna techniques cannot be used for interference mitigation as well as power control mechanisms used in cellular networks are not applicable to WBANs. Thus, novel methods and schemes are required for interference mitigation and consequently for energy conservation in both intra- and inter-WBANs.

The rest of the paper is organized as follows. Section II presents the recent works addressed problems related to interference mitigation and energy conservation for WBANs. Section III describes the network model and presents a virtual relaying scheme for WBANs. Section IV shows the proposed dynamic channel allocation for interference mitigation (DCAIM) scheme and also presents a probabilistic approach to reduce outage probability. Section V shows and explains the experimental results. The conclusions and future works are drawn in section VI.

II. RELATED WORK

Recent studies show multi-hop scheme has a lower power consumption in comparison to one-hop scheme. Using relays reduces the whole WBAN interference and its power consumption [7], [8], [9]. On the other hand, other works prove that TDMA scheme is an attractive solution to avoid interference within an intra-WBAN. These schemes require time synchronization which is infeasible when large number of WBANs coexist [1], [3], [13]. Some efforts focus on co-channel interference analysis of non-overlapping WBANs. They develop a model for efficient network planning and resource management by using geometrical probability approach [4]. Other research works assume one desired and multiple coexisting interferers WBANs. The coordinator of the desired WBAN calculates SINR periodically and afterwards commands its nodes to select an appropriate interference mitigation scheme [1].

Other studies as in [6] analyze performance parameters of a reference WBAN in the presence of another interfering network. Moreover, these parameters have been improved by adoption of an optimized time hopping code assignment strategy. Works in [5] consider a WBAN where the coordinator periodically polls its sensors. The nodes however calculate SINR then select the modulation scheme according to the

experienced channel quality. Authors of [7] propose a single-relay cooperative scheme. Where a set of relays computes individually the required transmission power to participate in the cooperative communication. Eventually, the best relay is selected in a distributed fashion.

Furthermore, authors of [6] explore the problem of interference among multiple coexisting WBANs. The proposed scheme enables two or three WBANs to agree on a common TDMA schedule. This scheme reduces the whole interference amongst them. The work in [3] investigates the problem of co-existing WBANs. It adopts a TDMA scheme within a WBAN and a carrier sensing mechanism to deal with inter-WBAN interference. The coordinator checks the channel whether free before polling the sensor to avoid inter-WBANs interference.

Authors of [8] consider a sensor network with slot based transmission scheme. This work provides an analytical framework to optimize the size of relay-zone around each source node. Authors of [11] propose a resource allocation scheme for interference mitigation among multiple coexisting WBANs. Each pair of interfering WBANs form an interference region. However, the nodes belonging to this region are later allocated orthogonal sub-channels. Whilst, other nodes that do not belong to this region transmit using the same time slot.

Most of the recent works show problems related to interference mitigation and power conservation in inter-WBANs environment. They do not address problems related to **node-level** interference and energy consumption within an intra-WBAN. However, in this work, we are interested to address these problems. Our proposed DCAIM scheme exploits the benefits of multi-hop communication schemes for interference mitigation and energy conservation within a single WBAN. Thus, it allows the high interfering nodes of a single WBAN to transmit through using orthogonal channels whilst, others nodes also with acceptable low levels of interferences to transmit using the same channel.

III. NETWORK MODEL

A. Virtual Network Model

To address these interference and energy concerns, we propose in this work a reliable topology design for WBANs as shown in Fig. 1. that takes into account the mobility of the patient while guaranteeing a reliable data delivery. We consider a tree-based multi-hop WBAN topology where in addition to sink and sensor nodes, we introduce a set of relay nodes distributed along the body to serve as transport network for sensory and control data. We call this relaying structure virtual backbone network where each node has some fixed neighbors to which it has constant Euclidian distance independently of body gestures. Accordingly, this backbone structure has relatively stable and reliable links in the presence of postural mobility related to body movement.

B. Network Assumptions and Description

More specifically, we consider a single WBAN formed of a set of sources N_s , set of relays N_r and coordinator C. We adopt two-hop communication (source - relay - C) as an energy




Figure 1. A WBAN-based relaying virtual backbone

efficient scheme. The WBAN is divided into multiple relay regions (RGs) as shown in Fig. 2. Where each region is formed of a set of sources S_s and a set of relays R_r . Any region is coordinated by a relay. This requires that the sources are in the radius communication of a relay. Whilst, the relay transmits sources' data to a common destination C. Thus, in our model we assume the followings:

- The underlying sources of each RG use TDMA to communicate with the relay
- Any TDMA schedule of any specific RG may overlap with other RGs' schedules
- The relays compete for the shared channel using slotted CSMA/CA to communicate with the sink
- Within each RG, all the sources are in the radius communication of at least one relay
- Within each RG, the sources are distributed based on the application requirements in use

C. Channel Model

The radio wave propagation in a WBAN is noticeably different compared to the other environments since it takes place inside or close to the human body. WBAN channel modeling is challenging due to the complex shape and composition of human body. In WBANs, the small-scale fading is dominant compared to the large-scale fading due to the short distance between the transmitter and the receiver. It is claimed that the lognormal distribution is the most fitting to describe the fading channel of WBANs since the human body is the dominating shadowing factor. Furthermore, the lognormal distribution is the best choice for modeling the small-scale fading in WBANs because there are only a small number of multi-path components from the diffraction around the body [13].

The path loss (PL) can be described in the Eq. 1 below:

$$PL(d) = PL(d_0) + 10 \times \log_{10} \left(\frac{d}{d_0} \right)^\alpha + X_\sigma \quad (1)$$

Where

- $PL(d)$: is the path loss at distance d from the transmitter

- $PL(d_0)$: is the path loss at reference distance d_0 from the transmitter
- α : is the path loss exponent
- X_{σ} : is a log normal distributed random variable

Algorithm 1 Dynamic Channel Assignment for Interference Mitigation (DCAIM)

```

Step 1: Orthogonal Transmission
for i ← 1 to Nr do
  for j ← 1 to Ss do
    Sensor (i, j) is transmitting
    RGk estimates the received signal power, δi,j
    from sensor (i, j)
  end
end
Step 2: Determine the inter-interference set
for i ← 1 to Nr do
  δmin,i = min{δi,j}
  for k ← 1 to Nr, k ≠ i do
    for j2 ← 1 to Ss do
      if δk,j2 > δmin,i - δThr then
        Add (k, j2) to set ILi
      end
    end
  end
end
Step 3: Broadcast for i ← 1 to Nr do
  Broadcast ILi
end
Step 4: Determine the interference set
for i ← 1 to Nr do
  ISi = ILi ∪ {(i,j) | (i,j) ∈ ILk, k ≠ i}
end
Step 5: Channel assignment
for i ← 1 to Nr do
  Leave the time slots for nodes in ISi unchanged
  Equally assign the remaining channels to nodes that do
  not belong to ISi
end

```

D. Dynamic Channel Assignment Proposed Scheme

Inter-RGs interference can happen due to the concurrent transmissions of sources of different RGs. Based on our assumptions, a source transmission of a specific RG may introduce higher interference level on the nodes of other nearby RGs whilst other sensors of that RG may have a very low interference on the surrounding RGs. A simple approach to avoid inter-RGs interference is to assign orthogonal channel to those nodes which experience higher interference levels.

We consider a total of N_r RGs are co-located within a WBAN, each RG consists of S_s source nodes. The shared channel among the sources of any RG is divided evenly into S_s time-slots. On the first side, we assume that k^{th} source sensor of RG_i transmits to its relay using its time slot $T_{i,k}$. On the other side, other relays try to compute the interference level of that source based on the power received from that signal.

1) *Interference Lists*: Let $\delta_{i,j,k}$ denote the received power from k^{th} source of RG_j at RG_i . After the orthogonal transmissions of sources of all RGs are complete in the first round. This enables each relay to compute and maintain a table consisting of the received powers from each source of all RGs. Also, each relay finds the minimum received power from

sensors within its RG_i which denoted by $\delta_{min,i} = \min\{\delta_{i,j,k}\}$. It then compares it to the received power from sources of other RGs. We present the interference list of RG_i as IL_i , where $IL_i = \{(j,k) | \delta_{i,j,k} > \delta_{min,i} - \delta_{Thr}\}$. The relay R_i of RG_i includes a sensor from other RGs in its IL_i if its received power is larger than $\delta_{min,i} - \delta_{Thr}$.




Figure 2. An example of distribution of nodes amongst three RGs and their Interference Region

2) *Interference Sets*: Based on interference lists broadcasts amongst relays, each relay can verify which of its sources introduce severe interference on other RGs. And also, which sources of other RGs impose severe interference on its RG. Afterwards, each relay creates an interference set formed of all sources that have a severe interference level. We denote interference set of RG_i by $IS_i = IL_i \cup \{(i,k) | (i,k) \in IL_j, j \neq i\}$

E. Channel Assignment Mechanism - Illustrative Example

Three coexisting relay regions RG1, RG2 and RG3 of a WBAN are shown in Fig. 2. We present an illustrative example to clarify the channel assignment mechanism in our proposed Algo. 1 for typical scenario of a WBAN. Each circumference of a circle denotes the coverage range of each RG and the interference region amongst them is represented by the intersection of their coverage ranges. The interference list of each RG_i is as follows:

- $IL_1 = \{(2,D), (3,d)\}$
- $IL_2 = \{(1,2), (1,4), (3,d)\}$

RG 1	d	1	3	1	3	1	2	4	3	1	3	D
RG 2	A	B	A	d	B	A	2	4	A	B	C	D
RG 3	a	b	c	d	a	b	c	d	a	b	c	d
	← RG 3				← RG 1				← RG 2			

Figure 3. Proposed channel assignment for nodes in Fig. 2

Parameters List	
transmission power	-10 dBm
sensitivity	-84.7 dBm
noise floor	-102 dBm
interference distance d	10 cm
data rate	250 kbps
base frequency	2.4 GHz
path loss exponent	4.22
sub-channels	8 channels

Table I
SIMULATION PARAMETERS

- $IL_3 = \{(2,C)\}$

Generally, $IL_{ID1} = \{(ID2, ID3), \dots\}$ where ID2 is the identifier of RG interfering with the specified RG_{ID1} and ID3 is the identifier of the node of the RG_{ID2} which is interfering with RG of interest RG_{ID1} . Each relay broadcasts its Interference list (IL) and based on other RGs broadcasts, it creates its interference set (IS). Doing so, each RG forms an interference set as follows:

- $IS_1 = \{(2,D), (3,d), (1,2)\}$
- $IS_2 = \{(1,2), (1,4), (3,d), (2,C), (2,D)\}$
- $IS_3 = \{(2,C), (3,d)\}$

However, each relay of RG assigns orthogonal channels for each of its nodes that are in the interference set of other co-located RGs. Whilst, other nodes are allowed to use all the free time slots until they are all occupied. Therefore, each node will be allocated and assigned more slots which increases the overall spatial reuse. As an example, Fig. 3 shows that node 1 of RG_1 have been assigned four slots. That means four times more spatial reuse than the case where each node is assigned only one slot. It is clear to see in Fig. 3 that the nodes of interfering sets (ISs) are not transmitting concurrently. This ensures the fact that the interference level is kept as low as possible. The following pseudo code shown in Algo. 1 clarifies the different steps of the the DCAIM proposed scheme.

IV. DCAIM PROPOSED SCHEME ANALYSIS - OUTAGE PROBABILITY

We denote the probability that the total interference at time instant i is larger than δ_{Thr} at RG_0 by P_{out}^i . Then, we calculate this probability by the following formula:

$$P_{out}^i = \left(\sum_{j=1}^{N-1} \delta_j > \delta_{Thr} \right) \quad (2)$$

We present a probabilistic approach which we prove analytically it lowers the outage probability. As mentioned above, any sensor whose received SINR at RG_0 is higher than a threshold is added to the interference region of RG_0 . We denote by δ_j the received SINR from a sensor in RG_j at RG_0 . An orthogonal channel is assigned to that sensor with certain probability which equals $\frac{\delta_i}{\delta_{Thr}}$. Thus, at time instant

i , we can calculate the average interference level using the proposed probabilistic approach as follows:

$$\delta_i = \sum_{j=1}^{N-1} \delta_j \left(1 - \frac{\delta_j}{\delta_{Thr}} \right) \quad (3)$$

Based on the probabilistic approach, any sensor with probability $\frac{\delta_i}{\delta_{Thr}}$ is assigned with an orthogonal channel.

Lemma 1: We denote by $P_{Probilistic}$ and $P_{Original}$ the outage probability of probabilistic approach and the outage probability of the original scheme respectively. Then, $P_{Probilistic} < P_{Original}$. If we let the average reuse factor R_{prob} and R_0 for the proposed scheme with and without probabilistic channel assignment respectively, then $R_{prob} < R_0$.

Proof: Based on outage probability definition, we have:

$$P_{Probilistic} = p \left(\sum_{i=1}^{N-1} \delta_i \left(1 - \frac{\delta_i}{\delta_{Thr}} \right) > \delta_{Thr} \right) \quad (4)$$

$$= p \left(\sum_{i=1}^{N-1} \delta_i > \delta_{Thr} + \sum_{i=1}^{N-1} \frac{\delta_i^2}{\delta_{Thr}} \right) \quad (5)$$

$$< p \left(\sum_{j=1}^{N-1} \delta_j > \delta_{Thr} \right) = P_{Original} \quad (6)$$

Where the last line of $P_{Probilistic}$ is based on the fact that the CDF is an increasing function of its argument. We define $P_{probabilistic, I, i}$ as the probabilistic approach deployment probability that a sensor node of RG_i is in the interference region of RG_0 . Then:

$$P_{prob, I, i} = P(\delta_i > \delta_{Thr}) + P(\delta_i < \delta_{Thr}) \frac{\delta_i}{\delta_{Thr}}, \quad (7)$$

which is greater than $P_i = P(\delta_i > \delta_{Thr})$. To complete the proof, let $P_{I, i}$ as the probability that a sensor node of RG_i exists in the interference region of RG_0 . Then:

$$R_0 = N - \left(\sum_{i=1}^{N-1} P_{I, i} \right) \quad (8)$$

To complete the proof, the average reuse factor of probabilistic approach is calculated as the following:

$$R_{prob} = N - \sum_{i=1}^{N-1} P_{prob, I, i} \quad (9)$$

$$= N - \sum_{i=1}^{N-1} P_I - \sum_{i=1}^{N-1} P(\delta_i < \delta_{Thr}) E \left(\frac{\delta_i}{\delta_{Thr}} \right) \quad (10)$$

$$< N - \sum_{i=1}^{N-1} P_I = R_0 \quad (11)$$

Where P_I denotes the probability that the received power is above the threshold value δ_{Thr}

V. EXPERIMENTAL RESULTS

In our simulation scenario, we consider a WBAN network consisting of three relay regions ($N_r = 3$). Each RG consisting of set $R_r = 2$ relay nodes and set $S_s = 4$ source sensors. The WBAN is modeled for on-body communication and is located within an area of $1 \times 2 \text{ m}^2$. We set the same transmission power for all the nodes in the network except the coordinator. For simplicity, the simulation parameters are listed in Table I above.

A. WBAN Energy Consumption

As shown in Fig. 4, we compare the sum of energy consumption of each WBAN versus time for all WBANs of the same simulation setup. The first WBAN implements our proposed DCAIM scheme, the second implements CSMA/CA access mechanism with opportunistic relaying (OR) scheme and the third WBAN allows its sources to transmit their data directly to common coordinator using the traditional Zigbee MAC scheme (one-hop). However, as can be clearly seen in this figure, firstly, the energy consumption of DCAIM scheme is always less than the energy consumption of the other schemes. Secondly, a lower energy consumption is obtained for both two-hop schemes than single-hop scheme. It can be evidently seen from the same figure that the energy consumption of our proposed DCAIM scheme increases slightly versus time, whilst in the other both schemes increases sharply, specifically, in Zigbee MAC scheme. This is because each high interfering node is assigned orthogonal channel to avoid interference. Hence, the collisions and retransmissions are reduced. Doing so makes less interference introduced on other nodes and consequently conserves their energy. However, this justifies that our DCAIM proposed scheme outperforms other schemes and extends longer a WBAN's energy lifetime.




Figure 4. Sum of energy consumption of all WBANs versus time for all different schemes

B. Nodes Signal to Interference and Noise Ratio (SINR)

As shown in Fig. 5 and based on the same simulation setup, we compare the node's SINR results of a WBAN that

implements our proposed DCAIM scheme with another that implements OR scheme. Both schemes adopt two-hop communication. However, it can be clearly seen from this figure that SINR results of our proposed scheme outperforms and always larger than the SINR values of OR scheme at each node. In other words, the interference is mitigated at each node level by improving its SINR. These nodes which were supposed to introduce interference on others nodes, after our proposed scheme, they are assigned orthogonal sub-channels. Doing so, lets them avoiding interference and consequently increases their SINR. Also, our proposed scheme increases the SINR by 11dB on average. This justifies that our proposed scheme mitigates node-level interference and hence this decreases the energy consumption of each node and WBAN as whole as well as extends its lifetime.




Figure 5. SINR versus node index of each WBAN for DCAIM and OR schemes

VI. CONCLUSION

A distributed and dynamic channel allocation scheme for interference mitigation (DCAIM) is presented in this paper to mitigate node-level interference amongst coexisting regions of a WBAN. The proposed scheme states that the sensor nodes that introduce significant interferences on other RGs are assigned orthogonal channels whereas other nodes can use the same time slots allocated before. The simulation results show that our proposed DCAIM scheme can reduce the interference at both node and WBAN levels and hence extend WBAN lifetime. We further compare the simulation results of our proposal with other scheme using OR and Zigbee single-hop scheme, the results show that ours outperforms in terms of interference mitigation and power savings in all cases. Additionally, we study and evaluate our proposed approach by theoretical analysis and show the outage probability can be effectively reduced to the minimal which allows to better and efficient use of the limited resources of WBANs. However, the study of our proposed scheme for a dynamic WBAN topology is left as future research issue.

REFERENCES

- [1] Yang, Wen-Bin and Sayrafian-Pour, Kamran, Interference Mitigation Using Adaptive Schemes in Body Area Networks, *International Journal of Wireless Information Networks*, pages 193-200, 2012
- [2] Latré, Benoît and Braem, Bart and Moerman, Ingrid and Blondia, Chris and Demeester, Piet, A Survey on Wireless Body Area Networks, *journal Wirel. Netw.*, pages 1-18, 2011
- [3] Chen, G. Chen, W. and Shen, S., 2L-MAC: A MAC Protocol with Two-Layer Interference Mitigation in Wireless Body Area Networks for Medical Applications, *IEEE International Conference on Communications (ICC)*, pages 3523-3528, 2014
- [4] Xuan Wang and Lin Cai, *IEEE Global Telecommunications Conference (GLOBECOM)*, Interference Analysis of Co-Existing Wireless Body Area Networks, 2011
- [5] Martelli, F. and Verdone, R. and Buratti, C., Link Adaptation in IEEE 802.15.4-based Wireless Body Area Networks, *Personal, Indoor and Mobile Radio Communications Workshops (PIMRC Workshops)*, 2010 *IEEE 21st International Symposium*, pages 117-121,2010
- [6] Mahapatro, J. and Misra, S. and Manjunatha, M. and Islam, N., Interference mitigation between WBAN equipped patients, *Ninth International Conference on Wireless and Optical Communications Networks (WOCN)*, pages 1-5,2012
- [7] Dong, Jie and Smith, David, Opportunistic relaying in wireless body area networks: Coexistence performance, *IEEE International Conference on Communications (ICC)*, pages 5613-5618, 2013
- [8] Vakil, S. and Ben Liang, Relaying in Wireless Sensor Networks with Interference Mitigation, *IEEE Global*
- [9] Braem, B. and Latre, B. and Moerman, I. and Blondia, C. and Reusens, E. and Joseph, W. and Martens, L. and Demeester, P., The Need for Cooperation and Relaying in Short-Range High Path Loss Sensor Networks, *International Conference on Sensor Technologies and Applications, SensorComm*, 2007
- [10] Javaid, Nadeem and Khan, NA and Shakir, M and Khan, MA and Bouk, Safdar Hussain and Khan, ZA , Ubiquitous healthcare in wireless body area networks-a survey, *journal arXiv preprint arXiv:1303.2062*, 2013
- [11] Movassaghi, Samaneh and Abolhasan, Mehran and Smith, David, Smart spectrum allocation for interference mitigation in Wireless Body Area Networks, *IEEE International Conference on Communications (ICC)*, pages 5688-5693, 2014
- [12] Javaid, Nadeem and Khan, NA and Shakir, M and Khan, MA and Bouk, Safdar Hussain and Khan, ZA, Ubiquitous healthcare in wireless body area networks-a survey, *arXiv preprint arXiv:1303.2062*, 2013
- [13] Movassaghi, Samaneh and Abolhasan, Mehran and Lipman, Justin and Smith, David and Jamalipour, Abbas, *Wireless Body Area Networks: A Survey*, *IEEE Communications Surveys Tutorials*, pages 1658-1686, 2014
- [14] *IEEE Standard for Local and metropolitan area networks - Part 15.6: Wireless Body Area Networks*, pages 1-271,2012