



**HAL**  
open science

## Brain-shift compensation using intraoperative ultrasound and constraint-based biomechanical simulation

Fanny Morin, Hadrien Courtecuisse, Ingerid Reinertsen, Florian Le Lann, Olivier Palombi, Yohan Payan, Matthieu Chabanas

### ► To cite this version:

Fanny Morin, Hadrien Courtecuisse, Ingerid Reinertsen, Florian Le Lann, Olivier Palombi, et al.. Brain-shift compensation using intraoperative ultrasound and constraint-based biomechanical simulation. *Medical Image Analysis*, 2017, 40, pp.133 - 153. 10.1016/j.media.2017.06.003 . hal-01560157

**HAL Id: hal-01560157**

**<https://hal.science/hal-01560157>**

Submitted on 11 Jul 2017

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# Brain-shift compensation using intraoperative ultrasound and constraint-based biomechanical simulation

Fanny Morin<sup>a,b</sup>, Hadrien Courtecuisse<sup>b,c</sup>, Ingerid Reinertsen<sup>d,e</sup>, Florian Le Lann<sup>f</sup>, Olivier Palombi<sup>f</sup>, Yohan Payan<sup>a</sup>, Matthieu Chabanas<sup>a,\*</sup>

<sup>a</sup>*Univ. Grenoble Alpes, CNRS, Grenoble INP, TIMC-IMAG, F38000 Grenoble France*

<sup>b</sup>*University of Strasbourg, CNRS, AVR-ICube, F67000 Strasbourg France*

<sup>c</sup>*MIMESIS, INRIA Nancy, France*

<sup>d</sup>*SINTEF, Department of Medical Technology, Trondheim, Norway*

<sup>e</sup>*Norwegian National Advisory Unit for Ultrasound and Image-guided Therapy, St. Olav University Hospital, Trondheim, Norway*

<sup>f</sup>*Grenoble Alpes University Hospital, Department of Neurosurgery, F38000 Grenoble France*

---

## Abstract

*Purpose.* During brain tumor surgery, planning and guidance are based on preoperative images which do not account for brain-shift. However, this deformation is a major source of error in image-guided neurosurgery and affects the accuracy of the procedure. In this paper, we present a constraint-based biomechanical simulation method to compensate for craniotomy-induced brain-shift that integrates the deformations of the blood vessels and cortical surface, using a single intraoperative ultrasound acquisition.

*Methods.* Prior to surgery, a patient-specific biomechanical model is built from preoperative images, accounting for the vascular tree in the tumor region and brain soft tissues. Intraoperatively, a navigated ultrasound acquisition is performed directly in contact with the organ. Doppler and B-mode images are recorded simultaneously, enabling the extraction of the blood vessels and probe footprint respectively. A constraint-based simulation is then executed to register the pre- and intraoperative vascular trees as well as the cortical surface with

---

\*Corresponding author

*Email address:* [Matthieu.Chabanas@univ-grenoble-alpes.fr](mailto:Matthieu.Chabanas@univ-grenoble-alpes.fr) (Matthieu Chabanas)

the probe footprint. Finally, preoperative images are updated to provide the surgeon with images corresponding to the current brain shape for navigation.

*Results.* The robustness of our method is first assessed using sparse and noisy synthetic data. In addition, quantitative results for five clinical cases are provided, first using landmarks set on blood vessels, then based on anatomical structures delineated in medical images. The average distances between paired vessels landmarks ranged from 3.51 to 7.32 (in mm) before compensation. With our method, on average 67% of the brain-shift is corrected (range [1.26; 2.33]) against 57% using one of the closest existing works (range [1.71; 2.84]). Finally, our method is proven to be fully compatible with a surgical workflow in terms of execution times and user interactions.

*Conclusion.* In this paper, a new constraint-based biomechanical simulation method is proposed to compensate for craniotomy-induced brain-shift. While being efficient to correct this deformation, the method is fully integrable in a clinical process.

*Keywords:* Brain-shift, Intraoperative ultrasound, Constraint-based biomechanical simulation, Elastic registration, Boundary conditions, Lagrangian Multipliers.

---

## 1. Introduction

During brain tumor surgery, accurate localization of the brain tumor is essential to both ensure its total resection and reduce the morbidity of surrounding healthy tissues. Images of the patient’s brain are thus acquired prior to surgery  
5 and used by neuronavigation systems to assist the surgeon. However, due to the intraoperative deformation of soft tissues, called ”brain-shift”, these images no longer correspond to brain morphology of the patient throughout the entire procedure.

The impacting factors on the amount and direction of brain-shift are numerous:  
10 patient positioning during surgery, craniotomy size and dura opening,

loss of Cerebrospinal Fluid (CSF), actions of the surgeon, swelling due to drugs, anesthetics or edema, *etc.* Even if the brain-shift mainly occurs in the direction of gravity, the magnitude is hard to predict and brain deformations are thus estimated by the surgeon. Magnitude and direction of brain-shift have been  
15 the subject of several studies reporting cortical displacements up to 20 mm and subsurface movements up to 7 mm (Hill et al., 1998; Roberts et al., 1998; Nimsky et al., 2000; Nabavi et al., 2001). A recent overview of the causes and measurements has been proposed by Gerard et al. (2017).

Brain-shift compensation methods propose to update preoperative images  
20 according to the intraoperative deformations of the organ. All of them rely on the acquisition of data during the procedure, providing information about the current tissue deformations. While Magnetic Resonance Imaging (MRI) is the reference preoperative exam, various intraoperative imaging systems are used in the literature such as MRI (Clatz et al., 2005; Vigneron et al., 2012), laser  
25 range scanners (LRS) (Sun et al., 2014; Miga et al., 2015), stereo cameras (Sun et al., 2005a) or ultrasound (US) imaging (Reinertsen et al., 2014; Rivaz & Collins, 2015). However, the data obtained cannot be directly used for surgical navigation: US images are of poor quality compared to preoperative MRI, and surface data alone are not clinically relevant. The goal of brain-shift compensation methods is then to register images acquired before surgery with these  
30 intraoperative data. Full brain deformations can thus be recovered based on sparse intraoperative information.

Even if this topic has been widely studied in the literature, very few methods are usable in a clinical practice. Methods aiming to be used in a surgical pro-  
35 cess have to satisfy three essential criteria. First, acquisition of intraoperative data must be easily integrated in the standard surgical workflow, with limited additional devices and no significant increase of the operative time and cost. Then, the compensation method must be relevant in terms of execution times and as automatic as possible. This criteria should be valid for every step of  
40 the proposed method, including image data pre-processing (*e.g.* patient-specific biomechanical model preparation) and intraoperative computations. Finally,

accuracy and robustness must obviously be validated on actual surgical cases.

In this paper, we present a method that intends to fulfill these requirements. First, we chose to rely on the acquisition of US images during surgery. Compared  
45 to other intraoperative imaging systems, US is portable, compatible with other surgical equipment and already available in most operating rooms. In addition, US provides subsurface information and not only the cortical surface. While soft tissues can be viewed using B-mode US, the Doppler mode enables specific imaging of the blood vessels. Next, a constraint-based biomechanical simulation  
50 method is proposed to compute the MR/US registration. Displacements are imposed over a finite element (FE) brain model to account for the deformations of the vascular tree and the exposed cortical surface.

The contributions of this work are summarized below:

1. We propose to account for the deformation of the blood vessels and ex-  
55 posed cortical surface using a single intraoperative US acquisition (Doppler and B-mode images recorded simultaneously).
2. The whole compensation method is fully detailed, including the data ex-  
traction from MR and US images and constraint-based biomechanical sim-  
ulation. The choices regarding the construction of the brain model, the  
60 constraints definition and their filtering are described. Our work combines these specific technical contributions with existing ones within a new complete craniotomy-induced brain-shift compensation system.
3. The performance of our method is evaluated over synthetic and patient  
data. For five clinical cases, a quantitative analysis is proposed first using  
65 landmarks set on blood vessels, then based on anatomical structures (such as *sulci* or ventricles) delineated in intraoperative B-mode images.
4. Improvements, in terms of accuracy and robustness, compared to one of  
the closest existing method (Reinertsen et al., 2014) are presented.
5. Finally, our method is fully compatible with a surgical process in terms of  
70 user interactions and execution times.

This article is organized as follows. First, detailed related works are pre-

sented in Section 2. Then, the method is described in the four following sections. Experiments on synthetic and patient data are detailed in Section 7, and the results are presented in Section 8. Finally, the method and results are discussed  
75 in Section 9.

## 2. Related works

After a description of the existing brain-shift compensation methods, this section focuses on the biomechanical simulation of brain deformations.

### 2.1. Brain-shift compensation methods for brain tumor surgery

80 The main methods to compensate for brain-shift occurring during tumor surgery can be classified as image-based and biomechanical simulation-based techniques. These approaches are presented in the following subsections. Each of them is organized according to the kind of intraoperative data.

#### 2.1.1. Image-based registration methods

85 Nimsy et al. (2001) presented a rigid registration between pre- and intra-operative MR images, using fiducial markers placed around the craniotomy. For their part, a non-linear registration method based on mutual information was introduced by Hastreiter et al. (2004). Regardless of the accuracy, the use of such intraoperative MR devices is cumbersome. The acquisition procedure is  
90 complex and increases the operating time considerably: the patient must be transferred to the scanner and specific surgical tools are required due to the magnetic field generated. Moreover, these devices are expensive and require large dedicated operating rooms. For these reasons, this intraoperative modality is thus rarely used in clinical routine.

95 Pereira et al. (2016) proposed to use pre- and intraoperative Cone-Beam Computed Tomography (CBCT) enhanced with intravenous injection of contrast agent. The volumetric deformation vector field was computed combining first a rigid registration of the skull then an elastic transformation based on

vessels, ventricles and image intensities. However, CBCT images suffer from  
100 poor soft tissue contrast and are thus rarely used for brain tumor surgery.

**A geometric 3D brain model was proposed by Nakajima et al. (1997). This 3D model was built on the preoperative MRI and integrated brain anatomical structures and blood vessels. Intraoperatively, the surface blood vessels were tracked using video cameras and registered with the preoperative ones, deforming the geometric model. No subsurface features were thus considered to compute the brain deformations. This compensation method was validated using data from 17 clinical cases.**

105

Finally, several methods rely on the acquisition of intraoperative US images.  
110 Ultrasound scanners are portable systems, already available in most operating theaters, and cost less than 10% of a classic intraoperative MR device. In addition, intraoperative US acquisitions do not require important changes in the operating procedure and are compatible with other surgical equipment.

On the one hand, deformations of soft brain tissues can be visualized using  
115 B-mode US imaging. These images can be registered with preoperative MRI using mutual information (Ji et al., 2008), cross correlation (Rivaz & Collins, 2015) or linear correlation of linear combination (Fuerst et al., 2014). Finally, Mercier et al. (2012) generated pseudo US images from preoperative MRI, then registered them with the real intraoperative US images. These methods were  
120 validated respectively over 6, 13, 14 and 15 patients.

On the other hand, Doppler US imaging provides flow visualization. Vessels  
around the craniotomy region can thus be viewed using this US modality. Reinertsen et al. (2007a) proposed a modified Iterative Closest Point (ICP) algorithm to rigidly register the blood vessels extracted from preoperative MR Angiography (MRA) to the ones extracted from intraoperative Doppler US acquisitions.  
125 A non-linear transformation using thin-plate spline was also presented. However, a clinical validation using data from 5 patients (Reinertsen et al., 2007b) showed that no major improvements are provided by the elastic registration technique over the rigid one. Finally, this rigid registration approach was vali-

130 dated in a recent clinical study including 7 patients (Reinertsen et al., 2014).

### 2.1.2. Biomechanical simulation-based registration methods

Key surfaces, such as the cortical surface and the surface of the lateral ventricles, were tracked in pre- and intraoperative MR exams by Ferrant et al. (2001) and Ferrant et al. (2002). Displacements were then defined using a matching algorithm and imposed on a biomechanical model. An improvement of this matching algorithm was proposed by Clatz et al. (2005). The computed displacement field was based on matching block features distributed on the whole image volume instead of using only a few key surfaces. Using this registration method, an augmented reality visualization of critical and functional brain regions was presented by Archip et al. (2007). MRI of various modalities (T1, fMRI and DTI) were acquired before surgery. All images were then aligned with the intraoperative T1 volume providing the surgeon with augmented information. More recently, Vigneron et al. (2012) introduced extended FEM to handle discontinuities between consecutive MR acquisitions. Finally, all these methods were respectively evaluated over 1, 9, 11 and 2 clinical cases. However, as stated above, MR acquisitions are rarely performed in clinical practice during surgery. Relying on such intraoperative acquisitions is thus a serious limitation regarding the usability of the methods in the operating room.

Several groups proposed to intraoperatively track the exposed cortical surface. First, Sun et al. (2005a) and De Lorenzo et al. (2012) introduced models where the displacements of the brain surface were directly driven by the motions acquired with stereo-cameras. For a better integration to the surgical process, Sun et al. (2005b) suggested to attach these cameras to the stereoscopic operating microscope. Their system was evaluated on 10 patients.

Next, similar methods were presented by Mostayed et al. (2013) and Galapati et al. (2014). Sparse point sets of the exposed part of the brain were

160 used to deform a brain model. These point sets were assimilated to the ones that could be acquired during surgery using the tracking pointer tool available in classic neuronavigation systems. These methods were respectively validated over 13 and 33 retrospective clinical cases.

Then, Sun et al. (2014) computed before surgery an atlas of brain deformation solutions calculated using a distribution of boundary and deformation-inducing forces conditions (e.g. sag, tissue contraction, tissue swelling). Intra-operatively, a LRS was used to record surface deformation data. The optimum brain shift solution was then determined using an inverse problem approach which linearly combines solutions from the atlas. This method has been recently validated by Miga et al. (2015) using a clinical study over 16 patients.

A limitation of all these methods is the assumption that the entire 3D deformation can be extrapolated by the biomechanical model from the exposed brain surface only. However, according to Wittek et al. (2007), the registration accuracy improves when data are collected from both the exposed and unexposed surface of the brain. Moreover, no subsurface features (e.g. the ones visible with MR or US images) can be accounted for during registration.

Mohammadi et al. (2015) thus proposed to combine the tracking using stereo cameras with Doppler US imaging. The exposed cortical surface as well as the vascular tree close to the tumor were registered. To do so, a constraint-based biomechanical simulation was used. However, this requires to bring two imaging systems (*i.e.* the stereo cameras and US scanner) in the operating room and therefore complicates the clinical process. Finally, Bucki et al. (2012) presented a similar model-based method. Only the blood vessels were registered, consequently a single Doppler US acquisition is necessary. In addition, results on one surgical case were provided.

## 2.2. Brain biomechanical simulations

A recent survey focusing on brain biomechanical simulation has been proposed by Morin et al. (2017). In this section, works dealing with the determination of the mechanical properties of brain tissues are first introduced, pointing

190 out the highly non-linear behavior of the organ. The use of such complex behavior laws is then nuanced in a second paragraph, according to the kind of loadings applied on the brain during simulations. Finally, the main biomechanical models published in the literature are presented.

### 2.2.1. Mechanical properties

195 The mechanical properties of the brain (with its constitutive law and corresponding parameters) have been widely studied in the literature. These properties can be determined using rheological studies on animal or human brains. Experimental measurements, obtained by compression, tension or suction of the tissues, are correlated with numerical simulation using a biomechanical model.  
200 Even if a non-linear behavior of the organ is highlighted by the studies cited in the next paragraph, the equivalent Young's modulus  $E$  at small deformations and the Poisson's ratio  $\nu$  are given for comparison.

Among studies realized on porcine brains, Paulsen et al. (1999) introduced a biphasic poroelastic model where the organ was considered as a sponge-like  
205 material. It was modeled using a porous solid tissue ( $E = 2100$  Pa and  $\nu = 0.45$ ) with interstitial fluid. This hypothesis was nuanced by Tavner et al. (2016) who demonstrated that there was no bulk fluid flow through the brain *in-vivo*. The authors argued that computing such a biphasic model was unnecessary and a single phase viscoelastic model should provide enough accuracy. Based  
210 on swine brains observations, Miller & Chinzei (2002) proposed an Ogden-like hyper-viscoelastic model with a Prony-series relaxation modulus. The very low stiffness of porcine brain tissues was also pointed out by Kaster et al. (2011), reporting Young's modulus for the white and grey matter respectively equal to  $1787 \pm 186$  Pa and  $1195 \pm 157$  Pa. Finally, the only rheological experiment on  
215 an *in vivo* human brain was presented by Schiavone et al. (2009). The authors used a light aspiration device and proposed a modified 2-term Mooney-Rivlin law, with an equivalent  $E = 1440$  Pa and  $\nu = 0.45$ .

Brain stiffness can also be determined *in vivo* using MR Elastography (MRE). This imaging technique estimates the shear modulus  $\mu$  from harmonic shear

220 waves velocity. Since this parameter is linked to the Young's modulus and Poisson's ratio by  $E = 2\mu(1 + \nu)$ , the value of  $E$  corresponding to the measured  $\mu$  can be computed for indication (here with a  $\nu = 0.45$ ). With the measurements reported by Kruse et al. (1999); Uffmann et al. (2004); McCracken et al. (2005); Kruse et al. (2008), the estimated Young's moduli were significantly  
225 higher (between 30 and 45 kPa for the white matter and 15 and 40 kPa for the grey matter) than the ones found with rheological experiments. However, the estimated  $\mu$  decreased with the shear wave frequency (Chatelin et al., 2012). For numerical simulation, a *static* Young's modulus (corresponding to a null excitation frequency, to avoid any viscosity effect) is required. Then, the Young's  
230 moduli computed above probably overestimate this static value.

### 2.2.2. Boundary conditions and loadings

While getting an accurate description of the tissues is essential, boundary conditions (*e.g.* contacts with the dura mater) and loadings are at least as important for the accuracy of comprehensive simulations. These loadings can  
235 be applied through forces or displacements.

On the one hand, Valencia et al. (2012) applied a pressure (*i.e.* a force field) on the surface of a brain model, while running the simulation with various elastic and hyper-elastic constitutive laws. Relevant differences are shown for resulting displacements and especially stresses at equilibrium. Moreover, Morin  
240 et al. (2015) pointed out the importance of accounting for the internal pre-stress of the organ. Indeed, the study exhibited significant differences for the resulting displacements when applying the same free force over a pre-stressed and an unpre-stressed brain model. In case of force loading, the biomechanical model is thus sensitive to the constitutive law and parameters, including internal  
245 pre-stress.

On the other hand, Wittek et al. (2009) and Miller & Lu (2013) investigated the impact of tissues mechanical properties when loadings are imposed through displacements. The authors showed that the solution in displacement is weakly sensitive to the chosen constitutive laws and parameters. In that sit-

250 uation, boundary conditions and loadings are then much more important than  
mechanical properties.

### 2.2.3. Biomechanical models

Many biomechanical models of the brain have been proposed in the literature  
in order to compensate for brain-shift, to simulate the growth of a pathological  
255 region (tumor, edema, ...) or within surgical simulators to train medical stu-  
dents. Most of these models are run using linear elasticity (**Ferrant et al.,  
2001, 2002; Castellano-Smith et al., 2003; Clatz et al., 2005; Archip  
et al., 2007; Wittek et al., 2009; Vigneron et al., 2012; De Lorenzo  
et al., 2012; Bucki et al., 2012; Yousefi et al., 2013; Mohammadi et al.,  
260 2015**). This choice is explained by the authors according to the use of imposed  
displacements (see previous paragraph) or to the small deformations observed.  
The corotational formulation (Müller et al., 2002) was used by Bilger et al.  
(2011), Dequidt et al. (2015) and Sase et al. (2015) to increase the range of  
authorized displacements (removing the artifacts of linear elasticity in large  
265 rotation) while running real-time simulations. A few brain models simulate  
hyper-elastic laws. Prastawa et al. (2009) proposed to use the constitutive law  
introduced by Miller & Chinzei (2002) whereas the models presented by Sun  
et al. (2005a) and Miga et al. (2015) are based on the one proposed by Paulsen  
et al. (1999). Finally, Miller & Lu (2013), Mostayed et al. (2013) and Garlapati  
270 et al. (2014) proposed fully non-linear models using Neo-Hookean constitutive  
law solved with the Total Lagrangian Explicit Dynamics algorithm (Miller et al.,  
2007). An efficient implementation of this algorithm on a Graphic Processing  
Unit (GPU) was presented by Joldes et al. (2010).

In all these works, brain models are run using small Young's modulus, or  
275 equivalent hyper-elastic parameter, values ( $E \in [694; 8000]$  Pa) and considering  
quasi-incompressible tissues ( $\nu \in [0.4; 0.495]$ ). Tissue are most of the time mod-  
eled as homogeneous, although some authors have integrated local properties.  
For example, Castellano-Smith et al. (2003) set different stiffness parameters  
to the white and grey matter, and Miller & Lu (2013) and Mostayed et al.

280 (2013) fixed a Young's modulus higher for the tumor than for healthy tissues. In addition, very soft and compressible properties are assigned to the ventricles ( $E = 10$  Pa and  $\nu = 0.05$ , against  $E = 694$  Pa and  $\nu = 0.45$  for brain tissues) by Clatz et al. (2005) and Bucki et al. (2012), in order to simulate the CSF loss during brain-shift deformation. This CSF is thus modeled as a liquid  
285 applying pressure on the brain surface by Clatz et al. (2003) and Bilger et al. (2011).

For the constraints, Dumpuri et al. (2006) proposed an automatic method to generate the boundary conditions. While the upper part of the brain, close to the craniotomy, is stress free, the lower cerebellum part is fixed. Sliding  
290 constraints are applied on the remaining brain surface, similarly to the ones imposed between the brain and the skull by Prastawa et al. (2009). Similar constraints are also used by Chen et al. (2011) to simulate interactions between the brain and the dural septa (*falx cerebri* and *tentorium cerebelli*). A very complex model was also introduced by Hu et al. (2007) where the dura mater  
295 (including the *falx cerebri* and the *tentorium cerebelli*) was simulated using isotropic elastic shell elements with high stiffness.

Finally, most of the biomechanical models mentioned in the above paragraphs rely on finite element (FE) meshes. Their generation includes soft tissues segmentation and then meshing steps that can be time-consuming. In order to  
300 avoid these limitations, brain biomechanical simulations based on meshless techniques (Miller et al., 2012; Zhang et al., 2013) were proposed in the literature. Even if a slight segmentation step is still needed, no meshing process is thus required.

### 2.3. Conclusion

305 In terms on intraoperative data our method rely on ultrasound images, to us the best compromise between provided data (surface and subsurface information, about tissues and vessels) and practical use in the operating room. For the MR/US registration, we propose a biomechanical simulation-based approach to extrapolate sparse deformation data to the entire brain based on a physical and




Figure 1: Brain-shift compensation process

310 not only geometrical background. Our modeling choices are based on the above  
 review of the literature.

### 3. Method overview

First, the overall process of our method is described. Second, the modeling choices made to perform the registration are detailed.

#### 315 3.1. Overall process

The overall process of our method is presented in Figure 1. Preoperatively, an anatomic patient-specific brain model is built from the images acquired prior

to surgery. Brain soft tissues and blood vessels around the tumor are extracted from MRI and MRA respectively. Boundary conditions (mainly contacts with the dura mater, *falx cerebri* and *tentorium cerebelli*) are then determined and transferred to the generated brain model. As preoperative images are often acquired one day before surgery, several hours are therefore available to build the brain model. However, the generation of the patient-specific model should be as automatic as possible.

Intraoperatively, a series of navigated 2D US slices is acquired directly in contact with the brain after opening the skull and dura mater (Reinertsen et al., 2014). Power Doppler as well as B-mode information are recorded simultaneously. A single US acquisition is thus necessary to obtain the whole intraoperative data. Two 3D volumes are reconstructed from the Doppler and B-mode slices, enabling the extraction of the blood vessels and probe footprint respectively. A constraint-based biomechanical simulation is then used to perform the registration of the pre- and intraoperative vascular trees and of the probe footprint with the surface of the brain model. Finally, the preoperative MRI is updated using the displacement field computed over the brain model. In order to comply with a clinical workflow, the designed extraction algorithms have to be robust, fast and must limit user interactions.

### 3.2. Brain modeling

Figure 2 illustrates the modeling choices and the use of each component during the constraint-based biomechanical simulation.

The brain is represented using several models, each one being optimized for a different task (upper panel in Figure 2). First, a *deformable model* describes the internal mechanical behavior of the soft tissues. Its geometry is represented by a tetrahedral FE mesh. Next, two *collision models*, a brain surface mesh and a skeleton of the blood vessels, are exclusively employed to express the *Lagrangian Multipliers* used to impose displacements (*i.e.* constraints) on the brain. During FE resolution, *constraint forces* are then computed so that the imposed displacement are satisfied at the end of the simulation step, whatever




Figure 2: Different components and data types of our brain model. Their interaction and use are highlighted. Preoperative and intraoperative data are respectively shown within a blue and orange box.

the biomechanical properties of the model. The deformable and collision models are coupled using a *barycentric mapping*.

350 The reference data (lower panel in Figure 2) are used as a base to express the constraints forces. These data are not deformed and remain fixed during all the simulation. First, the dura surface mesh is employed to express the boundary conditions between the surface of the organ and the dura mater. Then, the registration data (*i.e.* the probe footprint point set and the vessels  
355 skeleton) are extracted from intraoperative US images. They are used to drive the deformation of the biomechanical model to compensate for brain-shift.

The technical background that is necessary to understand our constraint-based biomechanical simulation is first detailed in Section 4. The extraction and generation methods of the meshes, skeletons and point sets are then described in Section 5, highlighting the brain modeling choices. A last section  
360 presents the constraint-based biomechanical simulation, focusing on the main technical contribution of this paper: the definition of the Lagrangian Multipliers constraints and their filtering.

#### 4. Simulation background

365 After introducing the corotational formulation and barycentric mapping, this section details the static governing equations used for the simulation of the brain and their numerical resolution. While these notions are not original to this paper, they are essential to understand our constraint-based simulation method.

##### 370 4.1. Corotational formulation

First introduced in the field of computer graphics (Müller et al., 2002), this formulation is frequently used for medical simulations due to its low computational cost. For such a formulation, the elastic forces applied on the nodes of a tetrahedral element of the brain FE mesh can be written as:

$$\mathbf{f}_e = \mathbf{R}_e \mathbf{K}_e^l (\mathbf{R}_e^T \bar{\mathbf{u}}_e - \mathbf{u}_e) \quad (1)$$

where  $\mathbf{u}_e$  and  $\bar{\mathbf{u}}_e$  are respectively the initial and deformed positions of the four nodes of the tetrahedron and  $\mathbf{f}_e$ , the elastic forces applied on these nodes. While  $\mathbf{K}_e^l$  corresponds to the local linear stiffness matrix of the element, parametrized by the Young's modulus  $E$  and the Poisson's ratio  $\nu$ ,  $\mathbf{R}_e$  is a bloc diagonal matrix defined as:

$$\mathbf{R}_e = \begin{bmatrix} \mathbf{R} & & & 0 \\ & \mathbf{R} & & \\ & & \mathbf{R} & \\ 0 & & & \mathbf{R} \end{bmatrix} \quad (2)$$

with  $\mathbf{R}$  being the  $3 \times 3$  rotation matrix of the tetrahedral element. The rotation of each element, computed based on the *QR decomposition* proposed by Nesme et al. (2005), is evaluated independently accounting for finite rotations. Swelling artifacts showed by classic linear elasticity are thus removed.

The local stiffness matrix  $\mathbf{K}_e$  can be written with the synthetic formulation (Felippa & Haugen, 2005):

$$\mathbf{K}_e = \mathbf{R}_e \int_{V_e} (\mathbf{C}_e \mathbf{D}_e \mathbf{C}_e \partial V_e) \mathbf{R}_e^T \quad (3)$$

375 where  $\mathbf{C}_e$  and  $\mathbf{D}_e$  are the strain-displacement matrix and the stress-strain matrix respectively. While these matrices are constant during the simulation,  $\mathbf{R}_e$  needs to be updated at each iteration.

#### 4.2. Barycentric mapping

As described above, several meshes and skeleton are used to describe the  
 380 brain model (see Figure 2). Soft tissue modeling is solved over the deformable model whereas the constraint forces are computed and applied on the vertices of the collision models. During each resolution step, the computed constraint forces (*i.e.* Lagrangian Multipliers) have thus to be transferred from the collision models to the FE model. Conversely, once the deformation is computed, the FE  
 385 nodes displacements must be propagated to the collision vertices. As a result, a barycentric mapping is used to link the different representations of the brain.

As proposed by Faure et al. (2012), each vertex of the collision models is associated with its closest FE tetrahedron at the beginning of the simulation. Since the collision and FE models are deformed at the same time, this association remains constant over the simulation. A vertex position can thus be defined by a linear combination of the nodes positions of its associated tetrahedron using barycentric coefficients. All these linear combinations are gathered by the function  $\mathcal{J}$ , used to map the positions  $\mathbf{u}_{FE}$  of the deformable model to the positions  $\mathbf{u}_{col}$  of the collision models:

$$\mathbf{u}_{col} = \mathcal{J}(\mathbf{u}_{FE}) \quad (4)$$

Using the *Jacobian matrix* of this function, defined as  $\mathbf{J} = \partial\mathbf{u}_{col}/\partial\mathbf{u}_{FE}$ , equation (4) can be written as:

$$\mathbf{u}_{col} = \mathbf{J}\mathbf{u}_{FE} \quad (5)$$

This *Jacobian matrix* is thus used to deduce  $\mathbf{u}_{col}$  from the positions  $\mathbf{u}_{FE}$ .

For the constraint forces  $\boldsymbol{\lambda}_{col}$  applied to the collision models, the mapping computes and accumulates the equivalent forces on the FE nodes. The transferred constraint forces  $\boldsymbol{\lambda}_{FE}$  are then weighted with the same barycentric coefficients using:

$$\boldsymbol{\lambda}_{FE} = \mathbf{J}^T \boldsymbol{\lambda}_{col} \quad (6)$$

In the following sections, the notation  $\mathbf{u}$  and  $\boldsymbol{\lambda}$  will refer to the nodes positions of the FE brain model ( $\mathbf{u} = \mathbf{u}_{FE}$ ) and constraint forces (*i.e.* Lagrangian Multipliers) applied on this model ( $\boldsymbol{\lambda} = \boldsymbol{\lambda}_{FE} = \mathbf{J}^T \boldsymbol{\lambda}_{col}$ ) respectively.

### 4.3. Governing equation

The governing equation is given by the static formulation:

$$\mathcal{F}(\mathbf{u}) + \mathcal{H}(\mathbf{u})\boldsymbol{\lambda} = \mathbf{f}_{ext} = 0 \quad (7)$$

where  $\mathcal{F}(\mathbf{u})$  and  $\mathcal{H}(\mathbf{u})$  are both non-linear functions providing the internal forces and gathering the constraint directions for the positions  $\mathbf{u}$  respectively. In our case, no external forces  $\mathbf{f}_{ext}$  (*e.g.* gravity or CSF pressure) are applied to the model.

Instead of directly solving this non-linear problem, each simulation step  $i$  consists in solving a linearized problem according to the two unknown  $\mathbf{u}$  and  $\boldsymbol{\lambda}$ . Note that computing the solution of this linearized problem is equivalent to performing the first iteration of the Newton-Raphson algorithm for the static problem (7). After several simulation steps, the positions at equilibrium (i.e.  $\Delta \mathbf{u}^i = \mathbf{u}^{i+1} - \mathbf{u}^i = 0$ ) correspond to the solution positions of this formulation.

The linearized equation is given by:

$$\mathcal{F}(\mathbf{u}^i) + \mathcal{H}(\mathbf{u}^i)\boldsymbol{\lambda}^i + \left( \frac{\partial \mathcal{F}}{\partial \mathbf{u}} \Big|_{\mathbf{u}^i} + \frac{\partial \mathcal{H}}{\partial \mathbf{u}} \Big|_{\mathbf{u}^i} \boldsymbol{\lambda}^i \right) (\mathbf{u}^{i+1} - \mathbf{u}^i) + \mathcal{H}(\mathbf{u}^i)(\boldsymbol{\lambda}^{i+1} - \boldsymbol{\lambda}^i) = 0 \quad (8)$$

At the beginning of each iteration step, no constraint forces are applied to the FE model which yields  $\boldsymbol{\lambda}^i = 0$  and then:

$$\mathcal{F}(\mathbf{u}^i) + \frac{\partial \mathcal{F}}{\partial \mathbf{u}} \Big|_{\mathbf{u}^i} (\mathbf{u}^{i+1} - \mathbf{u}^i) + \mathcal{H}(\mathbf{u}^i)\boldsymbol{\lambda}^{i+1} = 0 \quad (9)$$

On the one hand, the global stiffness matrix of the model is defined as:

$$\mathbf{K}^i = \frac{\partial \mathcal{F}}{\partial \mathbf{u}} \Big|_{\mathbf{u}^i} = \sum (\mathbf{G}_e \mathbf{K}_e^i \mathbf{G}_e^T) \quad (10)$$

where  $\mathbf{G}_e$  is the *globalization matrix* transferring the local stiffness matrix  $\mathbf{K}_e^i$  of an element  $e$  to the global stiffness matrix  $\mathbf{K}^i$ . Since the rotation matrices of the corotational formulation are not constant (see Section 4.1),  $\mathbf{K}^i$  must be recomputed at each simulation step.

On the other hand, the formalism of the Lagrangian Multipliers is used to express the constraints. A collision detection is first performed at the beginning of each iteration.  $\mathcal{H}(\mathbf{u}^i)$  then remains constant during all the step allowing to define the *constraint matrix*  $\mathbf{H}^i$  as:

$$\mathbf{H}^{iT} = \mathcal{H}(\mathbf{u}^i) \quad (11)$$

The number of lines in  $\mathbf{H}^i$  is equal to three times the total number of constraints in the simulation. For each constraint, this matrix gathers the weighted normals. The core of the constraint-based biomechanical simulation proposed in this paper is then to define the contents of the matrix  $\mathbf{H}^i$  (collision detection and weighed normal computation).

Once  $\mathbf{H}^i$  is defined, a constraint law can be written as:

$$\boldsymbol{\delta} = \tilde{\mathbf{H}}^i \mathbf{p}^i - \mathbf{H}^i \mathbf{u} \quad (12)$$

with  $\boldsymbol{\delta}$  being the violation of constraints,  $\mathbf{p}$  the targeted positions of the constraint and  $\tilde{\mathbf{H}}^i$  the matrix of the unweighted constraint normals. The linearization of this constraint law in  $\mathbf{u}^i$  gives:

$$\boldsymbol{\delta}^{i+1} = \tilde{\mathbf{H}}^i \mathbf{p}^i - \mathbf{H}^i \mathbf{u}^i - \mathbf{H}^i (\mathbf{u}^{i+1} - \mathbf{u}^i) = \mathbf{c}^i - \mathbf{H}^i (\mathbf{u}^{i+1} - \mathbf{u}^i) \quad (13)$$

where  $\mathbf{c}^i = \tilde{\mathbf{H}}^i \mathbf{p}^i - \mathbf{H}^i \mathbf{u}^i$  remains constant during each iteration.

The final linearized equation of (7) is obtained by injecting equations (10) and (11) in equation (9):

$$\mathbf{K}^i \boldsymbol{\Delta} \mathbf{u}^i + \mathbf{H}^{iT} \boldsymbol{\lambda} = -\mathcal{F}(\mathbf{u}^i) \quad \text{with} \quad \boldsymbol{\Delta} \mathbf{u}^i = \mathbf{u}^{i+1} - \mathbf{u}^i \quad (14)$$

where the unknowns are  $\boldsymbol{\Delta} \mathbf{u}^i$  and  $\boldsymbol{\lambda}$ .

For the sake of simplicity, superscript  $i$  is omitted in the following paragraphs, which describe the resolution for a single step only.

#### 415 4.4. Numerical resolution

Rewriting equations (14) and (13) provides the following Karush-Kuhn-Tucker (KKT) system:

$$\begin{cases} \mathbf{K} \mathbf{x} + \mathbf{H}^T \boldsymbol{\lambda} = \mathbf{b} & (15) \\ \mathbf{c} - \mathbf{H} \mathbf{x} = \boldsymbol{\delta} & (16) \end{cases}$$

where  $\mathbf{x} = \boldsymbol{\Delta} \mathbf{u}$  and  $\mathbf{b} = -\mathcal{F}(\mathbf{u})$ . At each iteration, the KKT problem is solved in five steps with the Schür complement method.

##### 4.4.1. Free motion

We first compute the free deformation  $\mathbf{x}_{free} = \mathbf{K}^{-1} \mathbf{b}$ , that does not depend on  $\boldsymbol{\lambda}$ , using a preconditioned Conjugate Gradient (Courtecuisse et al., 2011). The fixed Dirichlet conditions imposed on the base of the brainstem ensure that  $\mathbf{K}$  is invertible. Injecting  $\mathbf{x}_{free}$  in equation (15) gives:

$$\mathbf{x} = \mathbf{x}_{free} - \mathbf{K}^{-1} \mathbf{H}^T \boldsymbol{\lambda} \quad (17)$$

#### 4.4.2. Collision detection

420 A proximity-based detection is performed for each constraint, allowing to define the matrix  $\mathbf{H}$  described above. The free violation of constraint  $\boldsymbol{\delta}_{free}$  is defined as  $\boldsymbol{\delta}_{free} = \mathbf{c} - \mathbf{H}\mathbf{x}_{free}$ , with  $\mathbf{x}_{free}$  being the node positions of the brain model if no constraint force  $\boldsymbol{\lambda}$  is applied.

#### 4.4.3. Compliance computation

Injecting (17) in (16) gives:

$$\boldsymbol{\delta}_{free} + \mathbf{H}\mathbf{K}^{-1}\mathbf{H}^T\boldsymbol{\lambda} = \boldsymbol{\delta} \quad (18)$$

425 allowing to define the *compliance matrix*  $\mathbf{W} = \mathbf{H}\mathbf{K}^{-1}\mathbf{H}^T$  which traduces the mechanical coupling of the constraints. Its computation is the most time consuming step in the iteration since it involves the inversion of the large matrix  $\mathbf{K}$ . In order to get a clinically time relevant simulation,  $\mathbf{K}^{-1}$  is approximated by  $\mathbf{K}^{0-1}$ , with  $\mathbf{K}^0$  being the stiffness matrix at the first iteration.  $\mathbf{K}^{0-1}$  is  
 430 computed preoperatively and stored as a dense matrix. As the mechanical coupling (due to small brain stiffness parameters, see Section 2.2.1) and the brain deformations are low, this approximation is consistent in our simulation case.

#### 4.4.4. Constraint solving

Equation (18) thus provides the Non Linear Complementarity Problem:

$$\mathbf{W}\boldsymbol{\lambda} = \boldsymbol{\delta} - \boldsymbol{\delta}_{free} \quad (19)$$

where both  $\boldsymbol{\lambda}$  and  $\boldsymbol{\delta}$  are unknown. The problem is solved using a modified  
 435 Gauss-Seidel algorithm as detailed by Duriez et al. (2006).

#### 4.4.5. Corrective motion

Once the constraint forces  $\boldsymbol{\lambda}$  are known, they are replaced in equation (15) giving:

$$\mathbf{x} = \mathbf{K}^{-1}(\mathbf{b} - \mathbf{H}^T\boldsymbol{\lambda}) \quad (20)$$

As for the *Free Motion* step, this equation is solved using a preconditioned Conjugate Gradient providing the final positions  $\mathbf{u}^{i+1}$  that fulfill the constraints.

## 5. Brain modeling and data extraction from medical images

440 This section is divided into two subsections describing how pre- then intra-operative medical images are processed to build the brain models respectively.

### 5.1. Anatomic patient-specific brain model generated from preoperative images

#### 5.1.1. Soft tissues segmentation

Soft tissues are segmented from preoperative MRI (in our case a T2-FLAIR  
445 sequence). The whole brain is first extracted using the fast and automatic method proposed by Smith (2002). A region growing algorithm is then used to segment the tumor. Finally, the hemispheres are separated and the brainstem, cerebellum and *falx cerebri* are segmented by an operator.

#### 5.1.2. Soft tissues meshing

450 For the FE brain mesh, only the hemisphere affected by the tumor is accounted for. Soft tissues, excluding the cerebellum, are uniformly meshed with tetrahedrons. The surface of the previous FE mesh is used to represent the brain surface collision mesh. It is then decimated in order to reduce the number of constraints during the simulation and thus consequently reduce computation  
455 time. Finally, the same decimated surface mesh is used to describe the dura mesh.

#### 5.1.3. Determination of boundary conditions

Boundary conditions are used to simulate the interactions between the brain and its surrounding anatomical structures. On the one hand, sliding contacts  
460 between the brain surface collision model and dura mater, including the *falx cerebri* and *tentorium cerebelli*, are accounted for and illustrated by Figure 3. On the other hand, FE nodes located at the base of the brainstem are fixed during the simulation.

For all these boundary conditions, indexes of the constrained nodes are au-  
465 tomatically selected. To do so, the corresponding brain model meshes are compared to segmentation volume boundaries of each anatomic element. The mesh vertices close to this surface are then selected.




Figure 3: Boundary conditions applied on the brain surface collision mesh. The segmentation volumes of the dura mater, *falx cerebri* and *tentorium cerebelli* are respectively shown in red, blue and green. Sliding contacts are applied on the mesh vertices close to these volumes.

#### 5.1.4. Segmentation of the vascular tree

Blood vessels are segmented from the preoperative MRA exam using the  
 470 Maximum of Intensity Projection (MIP) image analysis method adapted from  
 the one proposed by Vermandel et al. (2007). After segmenting targeted vascular  
 structures, the binarized MIP image is used as a mask to recover 3D positions  
 of vessels from MRA 2D slices.

In order to get more contrast in the region of interest, the MRA is cropped  
 475 to tumor region before computing the MIP image. Two peaks are then visible in  
 the histogram of the MIP image, corresponding to the background (black) and  
 the soft tissues (dark grays) respectively. The blood vessels, highlighted with  
 a contrast product, are thus represented by the higher intensities. The MIP  
 image is therefore segmented with a threshold set just after the second peak of  
 480 the histogram. The segmented MIP image is shown in Figure 4(a).

However, artificial noise can appear due to the 3D reconstruction of the seg-  
 mented volume. In order to increase the robustness of our vessels extraction  
 method, the whole segmentation process is performed for the x, y and z di-




Figure 4: Blood vessels extraction from MRA

rections of the 3D space, thus obtaining three segmentation volumes. Finally,  
 485 the segmented vascular tree (see Figure 4(b)) is computed as the intersection of  
 these volumes.

#### 5.1.5. Skeletonization of the vascular tree

The blood vessels skeletonization process is based on the modified Dijkstra  
 algorithm detailed by Wan et al. (2002). First, the Euclidean distance to the  
 490 closest vessel boundary (DBF-distance) is computed for each voxel inside the  
 vessels. Images are then converted into a connected graph, using the inverse  
 value of the DBF-distance. Finally, vascular branches are extracted using a  
 modified Dijkstra algorithm (Figure 4(c)).

As the regularity of the extracted skeleton strongly depends on the quality  
 495 and voxel size of the input segmented image, each branch is smoothed and  
 re-sampled using B-splines. Finally, points and their connectivity (edges) are  
 exported.

## 5.2. Data extraction from intraoperative US images

### 5.2.1. Blood vessels skeleton extraction

500 The vascular tree skeleton is extracted from the Doppler US images. Since these images provide high contrast information on the blood vessels, their segmentation is simply achieved using a threshold set by an operator (Reinertsen et al., 2007b). Finally, the skeletonization step is realized with the algorithm described in Section 5.1.5.

### 505 5.2.2. Probe footprint point set extraction

The probe footprint point set is extracted from the B-mode US images. As the B-mode volume is performed (image acquisition and 3D volume reconstruction) at the same time as the Doppler one, its quality is quite low compared to a dedicated B-mode acquisition. However, as the probe footprint is easily identifiable (localization in the US image, pixels intensities and plane geometry), 510 such image quality is sufficient to automatically extract it.

Since the probe footprint is localized on the edge of the B-mode US volume, the boundaries of this volume are first extracted. In order to only keep the highest intensities, this surface is then segmented using a threshold set by an operator. While the probe footprint is extracted, some noise, spread over the 515 surface of the B-mode volume, also remains. The obtained binarized volume is therefore sampled and turned into a graph where close points are connected. Finally, the largest connected component is extracted.

### 5.2.3. Data decimation

520 As mentioned above, constraints are directly expressed based on the intraoperative vessels skeleton and probe footprint point set (see Figure 2). The number of constraints will then be linked to the number of points composing these elements. While a high number of constraints has a strong impact on the computation times, it can also lead to an over-constrained simulation. They are 525 therefore decimated in order to reduce the number of constraints. With respect

to the resolution of the FE and surface collision meshes, the blood vessels skeleton and the probe footprint point set are sampled at 1 point per 2.5 mm and 1 point per 10 mm respectively.

#### 5.2.4. Deactivation of boundary conditions at the craniotomy location

530 Vertices on which boundary conditions are applied are automatically determined preoperatively (see Section 5.1.3 for details). However, the position and size of the craniotomy are unknown before surgery. When the dura mater is opened, contacts between the brain and this membrane at the craniotomy location have therefore to be deactivated (*i.e.* corresponding vertices have to be  
535 removed from the list of the constrained indexes).

Intraoperatively, the craniotomy region is assimilated to the area where US slices are acquired, and then, to the probe footprint location in US images. The extracted point set is thus rigidly registered with the brain surface using the ICP algorithm (Besl & McKay, 1992). Finally, all the boundary conditions close to  
540 the probe footprint are deactivated.

## 6. Constraint-based biomechanical simulation

The constraint-based biomechanical simulation used to compensate for brain-shift is described in this section. First, the choices about the soft tissues modeling are briefly justified. Next, this section focuses on the main technical contribution of our registration method: the definition of the constraints applied on  
545 the collision models and their filtering. Finally, the update of the preoperative images with the simulated brain deformations is succinctly described.

### 6.1. Soft tissue modeling

As detailed in Section 2.2.1, brain mechanical behavior is complex and  
550 should be described using hyper-elastic laws. In our context, displacements are imposed on the brain model to simulate the contacts with the dura mater, register the blood vessels and constrain the cortical surface. The solution in displacements should thus be weakly sensitive to the biomechanical properties

(see Section 2.2.2 for details). A linear elastic law is therefore accounted for.  
 555 In addition, the corotational formulation has the strong advantage to avoid the swelling artifacts observed with pure linear elastic models while requiring low computational efforts (see Section 4.1 for details). This formulation thus appeared to us as being a good trade-off in our context. For the parametrization,  $E$  and  $\nu$  are respectively set to 1.5 kPa and 0.45, following Schiavone et al.  
 560 (2009). A higher stiffness equal to 10 kPa, close to the value chosen by Miller & Lu (2013), is used for the tumor region.

### 6.2. Constraints

On the one hand, constraints are used to solve the contacts between the brain and dura mater, including the *falx cerebri* and *tentorium cerebelli*, and  
 565 between the brain and probe footprint. On the other hand, displacements are imposed through forces applied on the blood vessels to register of the pre and intraoperative vascular trees. For all these constraints, the same formalism of Lagrangian Multipliers is used (see Section 4.3): constraint forces  $\boldsymbol{\lambda}$  are computed in order to cancel the violation of constraints  $\boldsymbol{\delta}$ . This section thus  
 570 describes how the matrix  $\mathbf{H}$  is filled, especially detailing the collision detection and weighted normals computation phases for each constraint type. Even if some parts of this process could be improved, for example by detecting collisions in a more efficient way (Joldes et al., 2009), this is beyond the scope of this paper.

As mentioned above (see Figure 2), collision detection and response are  
 575 computed over the collision models. In the next paragraphs, the constraint forces  $\boldsymbol{\lambda}$  thus correspond to  $\boldsymbol{\lambda}_{col}$  defined in Section 4.2. Vertices positions of the collision meshes, before and after the constraint forces are applied, are noted  $\mathbf{u}$  and  $\bar{\mathbf{u}}$  respectively and drawn in red and green in the figures. They are associated with fixed target positions chosen among the reference data meshes  
 580 (see Figure 2) called  $\mathbf{p}$  and drawn in blue. Finally, the violation of constraints after the resolution is noted  $\bar{\boldsymbol{\delta}}$ .

### 6.2.1. Bilateral dura constraint $\Omega$

This constraint is used to simulate interactions between the brain and the folds of the dura mater. It is applied on the vertices of the brain surface mesh in contact with the *falx cerebri* and *tentorium cerebelli*. The collision detection is processed between these vertices and the dura surface mesh. Each vertex, with position  $\mathbf{u}_j$ , is associated with a target position  $\mathbf{p}_j$ , computed as the closest projection of  $\mathbf{u}_j$  over the Bezier surface computed over the dura mesh. Since the dura mesh is fixed, Bezier coefficients are evaluated once for each triangle on the beginning of the simulation. This is used to obtain smooth contacts with the membrane, with constraints being continuous all along the surface.

The violation of constraint  $\delta_j$  is computed as:

$$\delta_j = \mathbf{n}_j \cdot (\mathbf{p}_j - \mathbf{u}_j) \quad (21)$$

with  $\cdot$  being the Euclidean dot product and  $\mathbf{n}_j$  the normalized normal to the Bezier surface at position  $\mathbf{p}_j$ . During the *constraint solving* step (described in Section 4.4.4), the corresponding constraint force  $\lambda_j$ , applied in the direction of  $\mathbf{n}_j$ , is computed so that no violation remains ( $\bar{\delta}_j = 0$ ) in that direction.

After the resolution, the vertex  $j$ , with position  $\bar{\mathbf{u}}_j$ , is located on the tangential plane given by  $\mathbf{n}_j$  (see Figure 5(a) for details). During the simulation, the brain model then remains in contact with the *falx cerebri* and the *tentorium cerebelli* but is authorized to slide along these surfaces.

### 6.2.2. Unilateral dura constraint $\Psi$

This constraint is applied to the brain surface collision mesh to simulate interactions between the organ and the dura mater. It is similar to  $\Omega$  except that it satisfies the Signorini conditions  $\lambda \perp \delta$  (Duriez et al., 2006). If a vertex, with position  $\mathbf{u}_j$ , is located inside the dura mater ( $\delta_j \geq 0$ ), no constraint force is applied and  $\lambda_j = 0$ . Otherwise, a negative constraint force ( $\lambda_j < 0$ ) is used to cancel the penetration ( $\bar{\delta}_j = 0$ ) and to bring back the vertex on the tangential plane given by  $\mathbf{n}_j$ .

At the end of the resolution, the node  $j$ , with position  $\bar{\mathbf{u}}_j$ , is located on or




Figure 5: For both figures, the blue broken line and curve are respectively representing the dura mesh and the Bezier surface defined over the triangles of the dura mesh. For  $\mathbf{u}_j$  (in red), two cases are shown. When  $\mathbf{u}_j$  is located outside of the dura mesh, the violation is negative ( $\delta_j < 0$ ). Conversely,  $\mathbf{u}_j$  inside the dura mater corresponds to a positive violation ( $\delta_j \geq 0$ ). For both cases, this point is projected on  $\mathbf{p}_j$  (in blue). After the resolution, its position  $\bar{\mathbf{u}}_j$  (in green) is located on the tangential plane for  $\Omega$  and under this plane for  $\Psi$ .

below this plane (see Figure 5(b)). The brain model is thus prohibited to exit  
 610 the cavity formed by the dura mater but it is however authorized to sag inside  
 it.

### 6.2.3. Unilateral probe constraint $\beta$

$\beta$  is used to impose interactions between the brain surface mesh and the  
 probe footprint point set extracted from intraoperative B-mode US images.  
 615 Conversely to the dura mesh used for  $\Omega$  and  $\Psi$ , the probe footprint is repre-  
 sented using a point set without any connectivity (edges, triangles, *etc.*). Pro-  
 jections in the collision detection phase are thus inverted. Each footprint point,  
 with position  $\mathbf{p}_j$ , is associated with  $\mathbf{q}_j$ , its nearest projection over the triangles  
 of the brain surface collision mesh.  $\mathbf{q}_j$  can then be described using barycen-  
 620 tric coefficients computed over the vertices positions  $\mathbf{u}_{j_a}$ ,  $\mathbf{u}_{j_b}$  and  $\mathbf{u}_{j_c}$  of its  
 corresponding triangle  $j_a j_b j_c$ .

The violation of constraint is computed as:

$$\delta_j = \mathbf{n}_j \cdot (\mathbf{p}_j - \mathbf{q}_j) \quad (22)$$

with  $\mathbf{n}_j$  being the normalized normal of the triangle  $j_a j_b j_c$ . As for  $\Psi$ , the Signorini conditions are satisfied for  $\beta$ . The resulting constraint force  $\lambda_j$  is applied along  $\mathbf{n}_j$  and distributed on the three vertices  $j_a$ ,  $j_b$  and  $j_c$ , according  
625 to the barycentric coefficients of  $\mathbf{q}_j$ .

After the resolution, these vertices have moved to positions  $\bar{\mathbf{u}}_{j_a}$ ,  $\bar{\mathbf{u}}_{j_b}$  and  $\bar{\mathbf{u}}_{j_c}$ , such as  $\mathbf{p}_j$  is located on or above the plane that they form. The brain surface mesh is then situated under the probe footprint point set, but the two are not necessarily in contact. Indeed, during the US acquisition, the probe might not  
630 be perfectly in contact with the exposed cortical surface and a gap can appear in the images between the soft tissues and the probe footprint.

#### 6.2.4. Bilateral vessels constraint $\Phi$

This constraint is used to register the skeleton of the preoperative vascular tree  $\zeta_{MRA}$  with the one extracted from the intraoperative Doppler US images  
635  $\zeta_{US}$ . As  $\zeta_{US}$  is more sparse than  $\zeta_{MRA}$ , this skeleton is projected towards  $\zeta_{MRA}$  for the registration. Each point of  $\zeta_{US}$ , with position  $\mathbf{p}_j$ , is associated with  $\mathbf{q}_j$ , being the projection of  $\mathbf{p}_j$  onto its nearest segment of  $\zeta_{MRA}$ .  $\mathbf{q}_j$  can then be defined from the two extremities of this segment, with positions  $\mathbf{u}_{j_a}$  and  $\mathbf{u}_{j_b}$ , using barycentric coefficients.

The violation of constraint is computed such as:

$$\delta_j = \begin{cases} \delta_{j_1} \\ \delta_{j_2} \end{cases} = \begin{cases} \mathbf{n}_{j_1} \cdot (\mathbf{p}_j - \mathbf{q}_j) \\ \mathbf{n}_{j_2} \cdot (\mathbf{p}_j - \mathbf{q}_j) \end{cases} \quad (23)$$

640 with  $\mathbf{n}_{j_1}$  and  $\mathbf{n}_{j_2}$  being two orthogonal and normalized normals to the segment  $j_a j_b$ . The constraint forces  $\lambda_{j_1}$  and  $\lambda_{j_2}$  are then computed such that all penetrations are canceled ( $\bar{\delta}_{j_1} = \bar{\delta}_{j_2} = 0$ ).  $\lambda_{j_1}$  and  $\lambda_{j_2}$  are respectively applied along  $\mathbf{n}_{j_1}$  and  $\mathbf{n}_{j_2}$  and distributed on the vertices  $j_a$  and  $j_b$  according to the barycentric coordinates of  $\mathbf{q}_j$ .

645 At the end of the resolution,  $\zeta_{MRA}$  has deformed so that  $\mathbf{p}_j$  is located on the line engendered by the vertices  $j_a$  and  $j_b$ , with positions  $\bar{\mathbf{u}}_{j_a}$  and  $\bar{\mathbf{u}}_{j_b}$  (see Figure 6).




Figure 6: Constraint resolution for  $\Phi$ . The red and green broken lines corresponds to the position of  $\zeta_{MRA}$  before and after the resolution respectively. The point from  $\zeta_{US}$  (in blue), with position  $\mathbf{p}_j$  is associated with its closest projection  $\mathbf{q}_j$  (in orange) among the  $\zeta_{MRA}$  segments. At the end of the resolution, forces are applied on  $\mathbf{u}_{j_a}$  and  $\mathbf{u}_{j_b}$  such as the two violations  $\delta_{j_1}$  and  $\delta_{j_2}$  are canceled and  $\mathbf{p}_j$  is located on the line formed by  $\bar{\mathbf{u}}_{j_a}$  and  $\bar{\mathbf{u}}_{j_b}$ .

### 6.2.5. Outliers filtering for $\Phi$

The quality of the vessels registration strongly depends on the quality of the pairings (association of  $\mathbf{p}_j$  with  $\mathbf{q}_j$ ) performed during the collision detection of  $\Phi$ . However, incoherent or antagonistic pairings (called outliers) can appear due to the noise (image artefact, missing data, *etc.*) contained in skeletons. These outliers have then to be ignored in order to make our elastic registration algorithm robust.

For this, an outliers filtering step is first added after the collision detection phase of  $\Phi$ . For all pairings, the distance between a vertex of  $\zeta_{US}$  and its corresponding segment in  $\zeta_{MRA}$  is defined as  $\mathbf{d}_j = \|\mathbf{p}_j - \mathbf{q}_j\|$ . When several vertices of  $\zeta_{US}$  are projected onto the same segment of  $\zeta_{MRA}$ , only the nearest one is kept. In addition, a pairing is considered as an outlier, and then ignored, if it does not satisfy the following constraint:

$$\mathbf{d}_j \in [\tilde{\mathbf{d}} - \mathbf{d}_t; \tilde{\mathbf{d}} + \mathbf{d}_t] \quad (24)$$

where  $\tilde{\mathbf{d}}$  is the median distance computed over all pairings and  $\mathbf{d}_t$ , a threshold. For example, if a vessel appears in  $\zeta_{US}$  and not in  $\zeta_{MRA}$ , incoherent pairings might be created with another distant vessel. Such incoherent pairings will then be ignored using the outliers filtering step described above.

Furthermore, during the resolution phase of  $\Phi$ , the constraint forces  $\lambda_{j_1}$  and  $\lambda_{j_2}$  are computed such as:

$$(\lambda_{j_1}, \lambda_{j_2}) \in [-\mathbf{f}_{max}; \mathbf{f}_{max}]^2 \quad (25)$$

with  $\mathbf{f}_{max}$  a parameter defining the maximal force applied on the vessels. Since the constraint forces are bounded, all constraints might not be satisfied at the end of the resolution ( $(\bar{\delta}_{j_1}, \bar{\delta}_{j_2}) \neq (0, 0)$ ). When vessels are close or skeletons are noisy, a segment of  $\zeta_{MRA}$  may be associated with a vertex of  $\zeta_{US}$  in a different direction than its neighbors. In such a case, huge constraint forces will be needed to satisfy these pairings, which might destabilized the simulation. Bounding the constraint forces thus enables the simulation convergence even if some outliers pairings remain. In addition, each constraint response ( $\lambda_{j_1}$  and

$\lambda_{j_2}$ ) being individually bounded, the total registration force applied on the brain model depends on the number of constraints. However, since the US skeleton is decimated, the local constraint number is controlled, guaranteeing that no huge  
670 force is applied locally.

Finally, the registration constraint parameters  $\mathbf{d}_t$  and  $\mathbf{f}_{max}$ , defined in the above paragraphs, are set to 1.5 mm and 0.001 N respectively. These *ad hoc* values were optimized over the first clinical case and remain the same for all other simulations.

### 675 6.3. Update of preoperative images


After running the constraint-based biomechanical simulation, the brain model has deformed, registering the blood vessels and constraining the cortical surface (see Figures 7(a) and 7(b)). Preoperative MR images have thus to be updated to account for this simulated brain-shift deformation. A displacement  
680 field is first computed on the entire MR volume based on the deformations of the FE mesh, combined with barycentric coefficients. Preoperative images are then warped with the previous displacement field to create navigable intraoperative MR images, that now correspond to the current patient anatomy. As highlighted by (Li et al., 2014), warping a set of images is a complex problem. In this paper,  
685 this process is based on linear interpolations to determine the voxels' grey levels, as implemented in the VTK library (Schroeder et al., 2006). Preoperative and updated MR images are shown in Figures 7(c) and 7(d) respectively.

## 7. Experiments


The experiments performed on our method are introduced in this section.  
690 First, details about the implementation are provided. The generation of synthetic data is then described before presenting the five surgical cases.

### 7.1. Implementation


For each clinical case, brain tissues were segmented from T2-FLAIR MR images. The BET tool available in the FSL software (Smith, 2002) was used for




(a) Brain model before compensation



(b) Brain model after compensation



(c) Preoperative MRI



(d) Updated preoperative MRI

Figure 7: Brain-shift compensation using our constraint-based method and update of preoperative images. For the brain models, blood vessels extracted from the preoperative MRA and intraoperative Doppler US images are respectively displayed in blue and orange. The probe footprint point set coming from the B-mode US acquisitions is shown in green.

695 the whole brain while the various anatomical parts were extracted using ITK-  
SNAP (Yushkevich et al., 2006). The FE meshes were then created with CGAL  
(CGAL, 2016). All others steps, including the definition of the boundary con-  
ditions, the segmentation and skeletonization of the vessels and the extraction  
of the probe footprint, were implemented as plug-ins within the visualization  
700 framework Paraview (Ayachit, 2015). Finally, the biomechanical model was  
developed using the simulation framework Sofa (Faure et al., 2012).

As detailed above, several software frameworks were used to develop our  
method, limiting a practical use in a clinical environment. However, the main  
objective of this paper is to provide a proof of concept of a new method that  
705 could be used, in term of user interactions and execution times, in a surgical  
workflow. In order to actually bring our system in a clinical environment, we  
aim to develop plug-ins to connect each step of our method to the navigation  
system CustusX (Askeland et al., 2015), therefore providing a unique software  
tool. Moreover, CustusX has already been used in clinical studies, as the one  
710 presented by Reinertsen et al. (2014).

## 7.2. Creation of synthetic data

The objective of the next paragraphs is to create synthetic datasets to study  
the impact of manually set US segmentation thresholds on the *ad-hoc* registra-  
tion constraint parameters (see Section 6.2.5). Instead of extracting the vascular  
715 tree skeleton and probe footprint from US acquisitions, fake intraoperative data  
are generated using a simulated brain deformation. In real clinical conditions,  
the quality of these input data relies on both the US images quality and the  
manually chosen segmentation thresholds. To simulate their variability, various  
controlled noise levels are added to the generated skeleton and probe footprint  
720 in order to evaluate the robustness of the registration process. As the same  
biomechanical model is used to generate and compensate for the deformation  
(i.e. same meshes, constitutive laws and parameters, boundary conditions, *etc.*),  
the registration results should only depend on the input data (US skeleton and  
probe footprint) and constraint parameters.

Table 1: Mean distances $\pm$ std between vessels and tumor landmarks after the generation of synthetic deformation. Maximum distances are given in parenthesis. All distances are provided in mm.

Landmarks type	Number of landmarks	Distance
Vessels landmarks	7	$4.27\pm 1.18$ (6.49)
Tumor landmarks	33	$5.94\pm 1.04$ (7.77)




Figure 8: Generation of a synthetic skeleton (orange) compared to the initial skeleton (blue)

725 *7.2.1. Generation of a brain-shift deformation*

A biomechanical model, built from the preoperative MR images of clinical case 1 (see Section 7.3), is used to generate the deformation. Arbitrary forces are applied on the brain surface, close to the craniotomy region, and around the tumor in order to simulate a brain-shift deformation (see Figure 8(a)). This deformation is then obtained using meaningless applied forces, simplified behavior models and boundary conditions and does not account for registration errors between preoperative MRI and intraoperative US images. Therefore, it cannot correspond to a physically plausible deformation obtained under real clinical conditions. However, the resulting displacements (see Table 1) can be compared to the ones observed in surgical cases (see Table 5). While it is not sufficient to assess the accuracy of the complete brain-shift compensation method, this deformation is nevertheless used to evaluate the robustness of its *ad-hoc* parameters. The deformed blood vessels skeleton  $\zeta_{generated}$  and brain surface mesh are then extracted. Finally, the brain surface mesh is cropped to the craniotomy region in order to represent the probe footprint that could be extracted from B-mode US images.

The deformation is quantified using 40 landmarks distributed on the vascular tree and on the tumor surface, summarized in Table 1. Seven landmarks are set on the blood vessels and located on the bifurcations of the skeleton, similarly to the ones used in Section 7.3.1 to validate our method on clinical cases. The other landmarks are uniformly placed on the tumor surface in order to monitor the impact of the registration method on the surgical target.

*7.2.2. Addition of noise*

For real clinical cases, only noisy skeletons, partially corresponding to the preoperative MR one, are extracted from intraoperative US images. This noise has two main sources: the data themselves and their segmentation. As we treat images from very different modalities and qualities, a vessel can be visible in MRA and not in US acquisitions (*e.g.* vessels in the direction of the US wave propagation) and vice versa (*e.g.* US acquisition noise and/or 3D reconstruction




Figure 9: Addition of noise in synthetic skeletons

artifacts). In addition, as the vessels often appear thicker in US images than in MRA, the position of the extracted skeleton can be affected. In order to obtain a synthetic skeleton  $\zeta_{synthetic}$  similar to the US extracted ones, some noise is added to  $\zeta_{generated}$ .

*3D coordinates perturbation.* A perturbation is added to the 3D coordinates of every node of  $\zeta_{generated}$ . This perturbation is independently and randomly chosen for each direction of the 3D space in the interval  $[-0.5; 0.5]$  (in mm). Using such an interval, each point is then displaced within a cube of 1 mm side (thus with a diagonal of 1.73 mm), correlated to the vessels thickness in US images (measured between 1.0 and 2.0 mm). The resulting perturbed skeleton  $\zeta_{perturbed}$  can be seen in Figure 8(b).

*Nodes removal.* Iteratively, a node is randomly selected and removed from  $\zeta_{perturbed}$ . However, in order to remove pieces of branch and not only isolated points, its first to fifth level neighbors are also deleted, as shown in Figure 9(a).

The sparsity of the output skeleton is controlled by removing a given percentage  
 770 of nodes from  $\zeta_{perturbed}$ . For example, 50% of points were deleted in the skeleton  
 presented in Figure 8(b) to obtain the sparse skeleton  $\zeta_{sparse}$  of Figure 8(c).

*Vessels addition.* Extra vessels are added to  $\zeta_{sparse}$ . A new vessel is defined  
 by an origin  $\mathbf{o}$ , with coordinates randomly chosen in the 3D space, a random  
 number of points  $N$  and an orientation vector  $\mathbf{v}$ . In order to maximize the  
 probability of erroneous pairings during the registration process, the distance  
 between  $\mathbf{o}$  and its closest point in  $\zeta_{generated}$  is included between 2.0 and 5.0  
 mm. In addition,  $\mathbf{v}$  is initialized parallel to the closest vessel with a random  
 perturbation added at each iteration. Finally, the vessel is built point per point  
 from its origin  $\mathbf{o}$ , until  $N$  is reached:

$$\mathbf{p}_i = \mathbf{p}_{i-1} + \mathbf{v}_{i-1}, \quad \forall i \in \{1, \dots, N-1\}, \quad \mathbf{p}_0 = \mathbf{o} \quad \text{and} \quad \mathbf{v}_0 = \mathbf{v} \quad (26)$$

with  $\mathbf{p}_i$  and  $\mathbf{v}_i$  respectively the built position and normalized orientation vector  
 at iteration  $i$  (see Figure 9(b) for details). The total number of added points  
 (combining all extra vessels) is set as a percentage of the initial number of  
 775 points of  $\zeta_{sparse}$ . The final synthetic skeleton  $\zeta_{synthetic}$  shown in Figure 8(d)  
 was obtained by adding 50% of points to the sparse one presented in Figure 8(c).

Even with a segmentation threshold set manually, the acquisition and ex-  
 traction noises observed for the probe footprint are far less important than  
 for the US skeleton. However, some reconstruction artifacts can result on a  
 780 misalignment of few millimeters between two consecutive US slices and then  
 irregularities on the point set extracted as the probe footprint. Therefore, only  
 the 3D perturbation detailed above is added to the coordinates of each point  
 of the generated probe footprint. Correlated to the size of the reconstruction  
 artifacts, the perturbation interval is also set to  $[-0.5; 0.5]$  (in mm).

785 Finally, the synthetic skeleton and probe footprint are prepared for the sim-  
 ulation. As described in Section 5.2.3, these data are decimated and contacts  
 with the dura mater at the craniotomy location are deactivated.

Table 2: Clinical cases, all of them operated for a low-grade tumor. For the validation, the segmented anatomical structure and the number of landmarks identified on blood vessels are given for each case.


Patient	Location	Vessels landmarks	B mode structures
1	Left frontal lobe	12	Prefrontal sulcus
2	Insula and left temporal lobe	14	Sylvian fissure
3	Insula and right temporal lobe	16	Right ventricle
4	Right insula	17	Superior temporal sulcus
5	Left prefrontal lobe	-	Cortical sulcus

### 7.3. Patient data


In this study, five patients suffering from a low-grade tumor are included. They underwent brain tumor surgery at Saint Olav University Hospital (Trondheim, Norway). Data were collected by the SINTEF Medical Technology (Trondheim, Norway) with Sonowand Invite (Sonowand AS, Trondheim, Norway). The study was approved by the local ethics committee, and patients gave informed consent prior to the procedure.

The protocol is similar to the one followed by Reinertsen et al. (2014). For each case, MR scans (T2-FLAIR and Angiography) were acquired before surgery. Intraoperatively, US acquisitions were performed for navigation as well as to monitor the soft tissues and vessels deformations. The clinical cases are presented in Table 2.

For the validation of our compensation method, the warped preoperative MRI must be compared with actual intraoperative information. While objective methods exist in the literature (Garlapati et al., 2015), they require data of the same modality (intraoperative MRI in that case). In our context, US images are the only intraoperative information available. Hence, a more subjective and potentially less-repeatable validation process had to be used. Two quantitative analysis are thus proposed, on vessel landmarks then on soft tissues structures.



(a) Paired landmarks identified by the second operator on the pre-operative (blue) and intra-operative (orange) blood vessels. Even if the landmarks are directly identified on the medical images, the vessels are shown in this figure as skeletons for clarity.



(b) Sets of points representing the superior temporal sulcus segmented in the pre-operative MRI (blue) and intra-operative B-mode US (orange)

Figure 10: Example of validation data for clinical case 4

### 7.3.1. Identification of vessels landmarks

A few landmarks were manually set on two vessels datasets: the one extracted from MR images and the one segmented from intra-operative Power  
810 Doppler US images. Depending on the data, 5 to 9 landmarks are identified for each patient by two operators. Salient features are looked for, like bifurcations or high curvature points (see Figure 10(a)). For patient 5, no landmarks could be set due to the lack of reliable vessel structures (very sparse signal of poor quality) in the region of interest. The registration accuracy is then measured  
815 by computing distances between paired vessels landmarks before and after compensation.

Although this process is common in the literature and followed by Reinertsen et al. (2014), two limitations could be pointed out. First, a bias exists when measuring errors on vessels while these same vessels are used for the registra-  
820 tion although they are not treated similarly. Next, positioning landmarks on vessels with high accuracy is quite difficult, especially on US data. In addition, this manual process is subjective and probably non repeatable. To account for these drawbacks, landmarks positioning was realized twice for each patient by two different users. Average and maximum distances between paired vessels  
825 are displayed in Table 3. Even if obtained values are in the same ranges, some inter-operator differences are also shown. These differences are obviously explained by the difficulty and subjectivity of the process. However, since it was achieved blindly (*i.e.* an operator does not know the positioning of the other operator’s landmarks), landmarks were not set on the same salient features by  
830 the two operators, thus implying variations in the obtained distances. Finally, our method is evaluating based on the reunification of these two landmarks sets for each clinical case. Such data thus provide higher variability in the positioning of landmarks (location and repartition) and account for the subjectivity of the process.

Table 3: Mean distances $\pm$ sd between paired vessels landmarks blindly identified by two operator then reunified in a unique dataset (synthesis). Maximum distances are given in parenthesis. All distances are provided in mm.

Patient	Operator 1		Operator 2		Synthesis	
	number	distances	number	distances	number	distances
1	7	3.97 $\pm$ 1.30 (5.86)	5	4.49 $\pm$ 1.02 (5.79)	12	4.19 $\pm$ 1.22 (5.86)
2	5	3.56 $\pm$ 0.27 (3.95)	9	3.48 $\pm$ 0.86 (4.77)	14	3.51 $\pm$ 0.71 (4.77)
3	7	7.67 $\pm$ 1.19 (9.04)	9	7.06 $\pm$ 1.23 (8.97)	16	7.32 $\pm$ 1.25 (9.04)
4	8	6.27 $\pm$ 0.67 (7.02)	9	6.41 $\pm$ 0.63 (7.45)	17	6.35 $\pm$ 0.65 (7.45)
5	-	-	-	-	-	-

835 *7.3.2. Delineation of anatomic structures*

So far, US B-mode images were only used to segment the probe footprint (see Section 5.2.2) later used as a boundary condition. Tissue information for US B-mode is never used in the registration process. Therefore, it is totally independent from the vessel features driving the model and can be used without  
840 any bias to evaluate the procedure.

Anatomical structures were manually delineated by a clinician in both the MR and US B-mode images. The most salient and easily identifiable structures were chosen to obtain reliable data. Depending of the case, it was either a *sulcus* or the ventricle border (see Table 2). These contours are at different  
845 depth, around or within the tumor, but are not correlated with the presence of vessels. For each patient, two clouds of points representing the same structure were thus created (see Figure 10(b)), the denser coming from MR images.

The Hausdorff distance is a classic metric to determine the distance between two point sets (Taha & Hanbury, 2015). However, it is generally very sensitive  
850 to outliers. In this paper, we compared structures come from images of very different modalities. A *sulcus* can thus be only partially visible in B-mode US images while being completely delineated within MR slices. As a result, huge outliers can appear that even robust versions of Hausdorff distance (Fedorov

et al., 2008) are not able to treat. That’s why a more global metric, translating  
855 the average registration score, is used to quantify the brain-shift compensation.  
The mean closest-point Euclidean distance is then computed before and after  
registration, similarly to Clements et al. (2016). However this measurement  
is a minimization of the actual error: the closest neighbor of a given point is  
not necessarily its adequate match in the other surface. Hence, if two points  
860 would be associated to the same correspondent, only the closest pair is kept.  
The total number of associated points gives an indication of the pairing quality  
(100% meaning each US point has a single pair in the MR contour, which would  
corresponds to a quasi-perfect superposition).

## 8. Results

865 In this section, quantitative and qualitative results are provided to evaluate  
our method. While the first paragraphs focus on its compatibility with a surgical  
process, its efficiency to compensate for brain-shift is shown in the next sections,  
first over synthetic data and finally over patient data.

### 8.1. Integration to the surgical pipeline

870 In order to be compatible with a surgical process, a method should be as  
automatic as possible, limiting the user interactions, and clinically relevant in  
terms of execution times.

#### 8.1.1. User interactions

875 Limiting the user interactions, especially during the intervention, is a pre-  
dominant criterion in order to be compatible with a surgical process. Our  
method was thus developed so that most of the steps, and mainly the intraop-  
erative ones, are automatically executed. However, it is not fully automatic and  
some steps still require interactions with an operator.


880 Our method does not need a detailed soft tissues segmentation of the pre-  
operative MRI: only the envelopes of the brain, cerebellum, tumor and the  
*falx cerebri* are needed. For the clinical cases presented in Section 7.3, the

Table 4: Summary of the parameters, their types, uses, values and how they are set


Type	Section	Parameter/Name	Value	Configuration
Data extraction	5.2.1	Vessels segmentation threshold	-	Fixed by an operator
	5.2.2	Footprint segmentation threshold	-	
	5.2.3	Vessels decimation threshold Footprint decimation threshold	2.5 mm 10 mm	Set regarding meshes resolution
Biomechanical	6.1	Young’s modulus $E$	1.5 kPa	Taken from the literature
		Tumor Young’s modulus	10 kPa	
		Poisson’s ratio $\nu$	0.45	
Registration constraint	6.2.5	Outliers median threshold $\mathbf{d}_t$ Maximal registration force $\mathbf{f}_{max}$	1.5 mm 0.001 N	Optimized over a first patient

segmentation of some brain anatomical structures was realized by an operator (see Section 5.1.1 for the details), which is not a satisfactory solution in a surgical process. However, this problem is fundamental in medical simulation and has  
885 been widely addressed in the literature. For example, automatic atlas-based segmentation algorithms experimented on brain MRI were recently presented at the MICCAI’2016 conference and workshops (Shakeri et al., 2016; Benkarim et al., 2016; Arthofer et al., 2016). The obtained brain segmentations are then much more detailed and accurate than the ones needed for our method. Finally,  
890 although this is not the topic of this paper, some solutions exist in the literature in order to automatically extract the brain tissues and anatomical structures.

Moreover, our method still requires several parameters, summarized in Table 4. Even if most of them are fixed for all simulation cases, the segmentation thresholds for the vessels and the probe footprint extracted from the intra-  
895 operative US images have to be manually set by an operator for each patient. These thresholds depend on the quality of the input US image but also of the US scanner settings selected by the surgeon (gain, *etc.*). Depending on the operator, the chosen thresholds may thus be different, leading to data with various levels of noise. However, since these data (the intra-operative vessels and probe



(a) Pre-operative execution times. For building the brain model, the segmentation time is not accounted for.



(b) Intra-operative execution times, after US acquisition and 3D volume reconstruction.

Figure 11: Execution times of each step of our method

900 footprint) are used as references for the registration algorithm, such variability could have a strong impact on our method and its parameters, especially the parameters set for the registration constraint. The robustness of our method to these parameters is evaluated over synthetic data in Section 8.2.

### 8.1.2. Execution times

905 Proposing a method that fits with the clinical workflow in terms of execution times is another essential criterion. The pre- and intraoperative execution times of each step are respectively presented in Figure 11(a) and 11(b).

Since MR brain images are usually acquired one day prior to surgery, several hours are available to process the preoperative data. Using our method, the  
 910 brain FE and surface meshes can be built simultaneously as the blood vessels extraction from MRA. As shown in Figure 11(a), all these steps are executed in less than 16 minutes and are thus completely compatible with a clinical workflow. However, we assumed that the semi-automatic soft tissues segmentation, that currently needs approximately one hour and occurs before the meshing  
 915 step, could be processed with a method from the literature (discussed in Sec-




Figure 12: Validation on synthetic data. For each configuration, the median mean landmarks distance over three synthetic skeletons is displayed.

tion 8.1.1). Only our additional computing time is therefore evaluated in Figure 11(a). As a result, if a time relevant solution exists in the literature for segmenting soft brain tissues, our method will also be time relevant.

Intraoperative execution times are presented in Figure 11(b). Reinertsen et al. (2014) already showed that performing US acquisition and 3D volume reconstruction during brain surgery is time relevant and thus, only our additional executions times are evaluated. Since the blood vessels and probe footprint are extracted from the US images in parallel, less than 2 minutes are necessary to simulate the brain deformations and obtain the updated preoperative MRI. In conclusion, this additional intraoperative time is acceptable considering the fact that a typical brain tumor resection procedure lasts several hours.

## 8.2. Validation on synthetic data

The efficiency and robustness of our method, and more specifically of its *ad-hoc* registration parameters (see Table 4), was first assessed over synthetic data sets. US skeletons and probe footprint were thus created using the algorithms described in Section 7.2. Deformed skeletons with various levels of sparsity (between 0 and 96% points removed) and noise (between 0 and 100% points added) were generated. For each configuration, three skeletons were cre-

ated. Our constraint-based simulation method was then run for each one of  
935 these skeletons, computing the mean distances over the vessels and tumor land-  
marks. The median of these mean distances is displayed in Figure 12, for each  
configuration.

A clear correlation between the results of the vascular tree and tumor sur-  
face registrations is shown by Figures 12(a) and 12(b). When the registration  
940 accuracy improves over the vessels, it is also increased for the surgical target.  
In addition, the method seems more robust to the addition of noise than to  
the removal of vessels, especially for the tumor registration. Indeed, if a vessel  
is associated with a wrong neighboring vessel, the distances computed on the  
vessels landmarks might be highly destabilized. Conversely, deformations of the  
945 tumor are interpolated by the biomechanical model from the ones induced to  
the vascular tree and by the probe footprint. Then, a slight error on the associ-  
ation of the vessels does not significantly disturb the registration results of the  
tumor.

Finally, as shown in Figure 12, our constraint-based method is efficient to  
950 compensate for the generated brain-shift even for configurations where 75%  
vessels were removed, whatever the percentage of points added. Beyond this  
percentage, the method has difficulties to compensate for the deformations but  
does not diverge: whatever the configuration, the initial mean distances (see  
Table 1) are never increased. A large range of input data can thus be treated  
955 with our method, limiting the impact of the inter-operator differences for the  
settings of the segmentation thresholds (see Table 4). In addition, if the regis-  
tration constraint parameters were optimized over the first surgical case, they  
are robust enough to obtain good registration results for various input skeletons.

### 8.3. Validation on patient data

960 After studying the method on synthetic data, the next validation step was  
carried out retrospectively on actual patient images. Two quantitative analy-  
sis are proposed on vessel landmarks and then on soft tissue structures. The  
validation protocols are detailed in Section 7.3 and only the results are pre-

Table 5: Mean distances $\pm$ sd between paired vessels landmarks before compensation, after rigid registration proposed by Reinertsen et al. (2014) and following our constraint-based registration method. Maximum distances are given in parenthesis. All distances are provided in mm.

Patient	Before compensation	Rigid registration	Constraint-based registration
1	4.19 $\pm$ 1.22 (5.86)	2.29 $\pm$ 1.36 (5.25)	1.79 $\pm$ 0.88 (3.66)
2	3.51 $\pm$ 0.71 (4.77)	1.71 $\pm$ 0.77 (3.88)	1.26 $\pm$ 0.18 (2.70)
3	7.32 $\pm$ 1.25 (9.04)	2.84 $\pm$ 1.27 (5.61)	2.33 $\pm$ 0.22 (4.85)
4	6.35 $\pm$ 0.65 (7.45)	1.91 $\pm$ 1.06 (4.37)	1.40 $\pm$ 0.44 (3.13)
5	-	-	-

sented in the following paragraphs. The compensation results obtained with  
965 our constraint-based method are compared to the ones obtained with the rigid  
registration algorithm proposed by Reinertsen et al. (2014) and available in the  
CustusX navigation tool (Askeland et al., 2015). Finally, a qualitative analysis  
presents our results with respect to clinical expectations.

### 8.3.1. Quantitative evaluation on vessels landmarks

970 Distances between paired vessels landmarks are given in Table 5, before  
compensation, after rigid registration proposed by Reinertsen et al. (2014) and  
following our registration method. Figure 13 also shows the repartition of these  
distances. In every case, the mean error with our process is reduced in compari-  
son to the existing CustusX method. Moreover, the standard deviation is lower  
975 and the maximum error is importantly lower. This means our results are more  
accurate not only globally but for every vessel.

As stated in Section 7.3.1, such evaluation process, based on vessels land-  
marks, has several drawbacks. However, it is the one followed by Reinertsen  
et al. (2014), and still shows a quantitative improvement with our method.  
980 Especially, accounting for all patient, on average 67% of the deformation is  
corrected with our constraint-based method, against 57% with the rigid one.




Figure 13: Dispersion of the distances between the paired landmarks defined on the blood vessels, for patients 1 to 4. While the mean distances  $\pm$ std and minimal and maximal distances are shown by the boxes, the distances between each associated landmarks are displayed using red dots.

Table 6: Mean closest-point Euclidean distances $\pm$ sd between the anatomical structures (represented as point sets) described in Table 2. Maximum closest-point distance is given in parenthesis and the number of associated points is shown in square brackets. These values are provided before compensation, after rigid registration proposed by Reinertsen et al. (2014) and following our constraint-based registration method.

Patient	Before compensation	Rigid registration	Constraint-based registration
1	1.91 $\pm$ 1.04 (6.18) [63%]	1.53 $\pm$ 0.92 (4.59) [72%]	1.03 $\pm$ 0.76 (5.17) [90%]
2	1.48 $\pm$ 1.07 (6.18) [66%]	1.28 $\pm$ 0.82 (4.66) [79%]	1.18 $\pm$ 0.73 (4.16) [77%]
3	3.86 $\pm$ 2.64 (10.93) [42%]	1.85 $\pm$ 1.56 (6.89) [75%]	3.21 $\pm$ 2.72 (11.34) [55%]
4	2.48 $\pm$ 1.45 (5.64) [59%]	1.27 $\pm$ 0.82 (4.36) [82%]	0.98 $\pm$ 0.64 (3.90) [86%]
5	3.44 $\pm$ 1.76 (7.23) [42%]	3.04 $\pm$ 3.00 (12.34) [12%]	2.73 $\pm$ 1.39 (6.18) [48%]

### 8.3.2. Quantitative evaluation using anatomical structures

While the measured distances before and after registration with CustusX and our constraint-based method are given in Table 6, the dispersion of these  
985 distances is shown in Figure 14. As can be seen, the registration accuracy is clearly improved for patients 1, 2 and 4. Not only the mean, standard-deviation and maximum errors are reduced, but the number of associated points is increased. That means the error measurement is more reliable, less minimized, with our method.

990 For patient 3, the error is only slightly reduced, far less than with the CustusX rigid method. Even if the vessel quantification was improved (see Table 5) for this patient, the simulated state is thus not satisfactory. In fact, the probe footprint measured in the US images (supposed to be in contact with the brain) is out of the cranial cavity in the MR images/model. Although tissue could  
995 locally and slightly sag out through the craniotomy, this is probably due to a large error of localization during the acquisition process. This error cannot be corrected with our compensation method since the brain model is constrained by the dura surface (see Section 6.2.2). If this constraint was released, results on both the vessels and tissue would be better with our method. This case mostly  
1000 stresses the importance of an accurate pre- to intraoperative registration before




Figure 14: Dispersion of the closest-points distances between the associated anatomical structures delineated in MR and B-mode US images, for patients 1 to 5. While the mean distances  $\pm$  std and minimal and maximal distances are shown by the boxes, the distances between each matched point are displayed using red dots.

acquiring the US images.

Concerning case 5, vessel data are so sparse in the region of interest that it was not possible to identify paired landmarks in the US and MR images. However, both registration methods were still run normally. As can be seen in  
1005 Table 6, the rigid method actually fails. Even if the mean error seems acceptable, the maximum distance and number of associated points show that the algorithm diverged. On the other hand, our method still yields a coherent result even if the improvement over the initial position is low. In conclusion, not only our constraint-based method is more accurate but it is also more robust to low-  
1010 quality sparse data.

### 8.3.3. Qualitative results

Beyond quantitative measurements, results must also be qualitatively evaluated with respect to surgeons expectations, especially the updated MR images that would eventually be used during surgical navigation. In the current Cus-  
1015 tusX navigation process (Askeland et al. (2015)), the initial MR images are displayed next to the intraoperative US images. In addition to these views, Figure 15 also shows the MR images updated with the computed deformation field for patients 1, 2 and 4. For each clinical case, a 3D view of the brain and tumor is presented, highlighting the 3D localization of the surgical target within  
1020 the organ.

Two points are localized in the medical images, first on the exposed cortical surface then in subsurface structures. The deep border of the tumor is pointed for clinical cases 1 and 4 whereas a *sulcus* curvature point is shown for patient 2. In each case, the updated MR images clearly fit the actual US data which  
1025 proves the quality of our method.

## 9. Discussion

After comparing our compensation results with existing works, limits and perspectives of our method are discussed.




Figure 15: Brain-shift correction for clinical cases 1, 2 and 4. For each patient, a 3D view of its brain and tumor is shown next to medical images. Preoperative MRI (left column), intraoperative b-mode US images (middle column) and updated MRI with our constraint-based compensation method (right column) are displayed within the navigation software CustusX (Askeland et al., 2015). A pointer first shows the borders of the exposed cortical surface (top row) then the bottom of the tumor for cases 1 and 4 and a *sulcus* curvature point for patient 2 (bottom row).

### 9.1. Comparison with existing methods

1030 As stated in the related works, brain-shift compensation is a widely studied topic. However, comparing the accuracy of existing methods remains difficult due to the different input data (US acquisitions, laser range scanner, *etc.*) and validation protocols proposed in the literature. Consequently, very few authors present a direct comparison between their method and existing ones.

1035 In this paper, improvements are shown over the image-based technique proposed by Reinertsen et al. (2014), comparing compensation results on the same surgical cases. While our method cannot be directly compared with other existing techniques, this point can still be discussed. To do so, we have selected recent clinical studies also proposing methods that could be easily embedded  
1040 into a surgical workflow. Obviously, comparisons have to be carefully interpreted, since results are not presented over the same surgical cases and methods involve different validation protocols.

First, Rivaz & Collins (2015) introduced an image-based method to register preoperative MRI with intraoperative US images. The registration accuracy is  
1045 evaluated over 13 patients, based on landmarks set on soft tissues. Their input data (US images) and validation protocol are thus very similar to ours. Results on pre- and post-resection US image registration are presented. However, only the pre-resection ones are accounted for in this paragraph, in order to compare brain-shift compensation results at the same time of the surgery (*i.e.* just  
1050 after the opening the dura mater). Initial shifts ranging from 1.5 to 9.4 mm are reported. These values are in the same order of magnitude as the ones measured on our first four patients, using landmarks set on blood vessels. Rivaz and Collins report results after compensation that range from 1.4 to 4.2 mm, which corresponds to an average correction of 36% of the shift. For very small  
1055 initial shifts ( $\leq 3.5$  mm), computing a percentage of corrected deformation may not be meaningful. Without accounting for these cases, the correction rate increases to 45%. In comparison, this average correction appears globally lower than the 67% obtained on our cases.

Next, Miga et al. (2015) presented a model-based method using LRS ac-

1060 quisitions of the cortical surface. Compensation results are evaluated based on  
shift vectors computed over bitmap images of the craniotomy. Input data and  
validation process are thus very different from ours, complicating the compar-  
ison. In addition, a wider range of measured shift is reported including very  
large displacements (from 2.5 to 21.3 mm). Two reasons explaining such differ-  
1065 ence could be pointed out. First, the amount of brain-shift is measured after  
tumor resection, thus accounting for these deformations. Next, it is evaluated  
on the cortical surface, often showing higher displacements than for deep struc-  
tures (in our case, the blood vessels). The mean corrected brain-shift is ranging  
from 0.7 to 4.0 mm, for an average compensation of 69% of the deformations.  
1070 Considering only the patients exhibiting an initial shift in the same range of  
magnitude observed for our cases ( $\leq 10$  mm), this percentage decreases to 65%.  
The registration accuracy is thus very similar the one reported in this paper  
(67%).

Finally, Mostayed et al. (2013) and Garlapati et al. (2014) proposed a model-  
1075 based method relying on the registration of few points on the exposed cortical  
surface. In real clinical conditions, these points could be acquired by the surgeon  
using a tracking pointer tool available in most classic neuronavigation systems.  
No extra imaging system is thus required which is a strong advantage for the  
practicality within the operating room. In these papers, the 3D points are how-  
1080 ever extracted from intraoperative MR images. It is thus very complicated to  
compare our results to theirs since intraoperative data were not acquired in  
real surgical conditions. Indeed, the method does not account for the poten-  
tial calibration errors of the tracking tool nor for the manipulation errors of  
the surgeon that could substantially affect the compensation results. In addi-  
1085 tion, compensation results are computed based on objective methods aiming  
to compare the intraoperative MRI with the generated warped MRI. No global  
registration scores are thus provided, making a numerical comparison with our  
results impossible. Nevertheless, similarly to our method these techniques show  
a registration improvement when using model-based instead of pure image-based  
1090 approaches.

As stated above, these comparisons should be cautiously interpreted. While we cannot conclude on an accuracy improvement, they at least show the relevancy and reliability of our registration results, considering the initial and compensated measured brain-shifts. Obviously, the necessity to validate on extra surgical cases is stressed out, including initial shifts with magnitudes of higher variability. Finally, our method is only evaluated at the opening of the dura mater and its ability to compensate for deformations induced by tumor ablation still has to be proven.

### 9.2. Limits and perspectives

The proposed method has limitations and several aspect could be improved.

Among the parameters gathered in Table 4, the biomechanical ones are set following the literature. However, high interpersonal differences are highlighted by Sack et al. (2009). Even if these parameters may have a low importance (Wittek et al., 2009), it would be interesting to study their impact on the final registration. Especially, the effect of an heterogeneous elasticity for the tumor has to be evaluated. In addition, these parameters could be estimated intraoperatively using US elastography. While computing true static Young’s modulus using shear waves is complex, the ratio between the tumor and surrounding soft tissues elasticity could at least be estimated. Finally, only some of the intrinsic characteristics of the organ are simulated by our brain model (morphology, soft tissues elasticity, contacts with neighbor anatomical structures, *etc.*). External phenomenons could then be accounted for to improve the method accuracy, such as the gravity-induced stress (Morin et al., 2015).

Like other intraoperative compensation methods, we are sensitive to localization errors (mostly due to probe calibration and image-to-patient registration). Essentially, this was highlighted by our results on the third clinical case. Even if the vessels are well registered, the compensation of the tissue deformations is not satisfying due to the registration errors between the preoperative MRI and intraoperative US images (see Tables 5 and 6). In order to detect and account for these localization errors, a rigid transformation could first be applied on the

brain model to register the cortical surface with the extracted probe footprint. Once corrected, the non-rigid tissue deformations could thus be compensated using our constraint-based method

A last limit is the dependency of our method to the presence of vessels  
1125 around the surgical target. When very few vessels are located close to the tumor for patient 5, our method has difficulties to compensate for brain-shift. Similar results were observed on synthetic data: the method fails to correct the deformation for very sparse input US skeletons (up to 75% points removed). To overcome this limit, salient corresponding features could be extracted in  
1130 the soft tissues respectively from the preoperative MR and intraoperative US images. The same constraint-based formalism could be run for the registration of these matching features. Furthermore, our process requires preoperative MRA or another sequence enhanced with contrast agent, to extract the vascular tree. As this may not be a standard protocol for brain tumor surgery, relying on soft  
1135 tissue features only could enable to spare this additional exam.

## 10. Conclusion

In this paper, a new constraint-based method to compensate for the craniotomy-induced brain-shift observed during tumor ablation procedure was proposed. A solution easily integrable in the operating room, in terms of intraoperative  
1140 acquisitions (US images), execution times and user interactions is presented. While quantitative improvements over one of the closest methods in the literature (Reinertsen et al., 2014) are shown on five clinical cases, the robustness of the method was also proven using synthetic data. This method addresses some of the current limitations, towards optimal solutions in image-guided neurosurgery. Next stage of this work will be to extend the proposed method in  
1145 order to compensate for resection-induced brain-shift deformations.

## Acknowledgments

This work was partly supported by the French ANR within the Investissements d’Avenir program under references ANR-11-LABX-0004 (Labex CAMI) and ANR-11-INBS-0006 (Infrastructure d’avenir en Biologie Santé) and by a France-Norway partnership (PHC Aurora 2015/Research Council of Norway). The authors would like to thank Clément Guyomard and Vincent Genty, interns at TIMC-IMAG Laboratory, Grenoble.

## Ethic committee

Anonymous and retrospective images are used for this clinical study. It was approved by the local ethics committee and formal consent was obtained from all individual participants.

## References

- Archip, N., Clatz, O., Whalen, S., Kacher, D., Fedorov, A., Kot, A., Christodoulos, N., Jolesz, F. A., Golby, A. J., Black, P. M., & Warfield, S. K. (2007). Non-rigid alignment of pre-operative MRI, fMRI, and DT-MRI with intra-operative MRI for enhanced visualization and navigation in image-guided neurosurgery. *NeuroImage*, *35*, 609–624.
- Arthofer, C., Morgan, P. S., & Pitiot, A. (2016). Hierarchical Multi-Atlas Segmentation Using Label-Specific Embeddings, Target-Specific Templates and Patch Refinement. *International Workshop on Patch-based Techniques in Medical Imaging*, (pp. 84–91).
- Askeland, C., Solberg, O. V., Beate, J. B. L., Reinertsen, I., Tangen, G. A., Hofstad, E. F., Iversen, D. H., Vapenstad, C., Selbekk, T., Lango, T., Hernes, T. A., Leira, H. O., Unsgard, G., & Lindseth, F. (2015). CustusX: an open-source research platform for image-guided therapy. *International journal of computer assisted radiology and surgery*, *11*, 505–519.

- Ayachit, U. (2015). *The ParaView Guide: A Parallel Visualization Application*.  
Kitware.
- 1175 Benkarim, O. M., Piella, G., Gonzalez Ballester, M. A., & Sanroma, G.  
(2016). Enhanced Probabilistic Label Fusion by Estimating Label Confidences  
Through Discriminative Learning. *Medical Image Computing and Computer-  
Assisted Intervention - MICCAI 2016*, 2, 505–512.
- Besl, P. J., & McKay, N. D. (1992). A Method for Registration of 3-D Shapes.  
1180 *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 14, 239–  
256.
- Bilger, A., Dequidt, J., Duriez, C., & Cotin, S. (2011). Biomechanical simulation  
of electrode migration for deep brain stimulation. *Medical Image Computing  
and Computer-Assisted Intervention - MICCAI 2011*, 14, 339–346.
- 1185 Bucki, M., Palombi, O., Bailet, M., & Payan, Y. (2012). Doppler Ultrasound  
Driven Biomechanical Model of the Brain for Intraoperative Brain-Shift Com-  
pensation: A Proof of Concept in Clinical Conditions. In *Soft Tissue Biome-  
chanical Modeling for Computer Assisted Surgery* (pp. 135–165). Springer.
- Castellano-Smith, A. D., Crum, W. R., Hill, D. L. G., Thacker, N. A., & Bromi-  
1190 ley, P. A. (2003). Biomechanical simulation of atrophy in MR images. *Medical  
Imaging*, (pp. 481–490).
- CGAL (2016). *CGAL User and Reference Manual*. (4th ed.). CGAL Editorial  
Board.
- Chatelin, S., Vappou, J., Roth, S., Raul, J. S., & Willinger, R. (2012). Towards  
1195 child versus adult brain mechanical properties. *Journal of the mechanical  
behavior of biomedical materials*, 6, 166–173.
- Chen, I., Coffey, A. M., Ding, S., Dumpuri, P., Dawant, B. M., Thompson, R. C.,  
& Miga, M. I. (2011). Intraoperative Brain Shift Compensation: Accounting  
for Dural Septa. *IEEE Transactions on Biomedical Engineering*, 58, 499 –  
1200 508.

- Clatz, O., Delingette, H., Bardinet, E., Dormont, D., & Ayache, N. (2003). Patient-specific biomechanical model of the brain: application to Parkinson's disease procedure. *Surgery Simulation and Soft Tissue Modeling*, (pp. 321–331).
- 1205 Clatz, O., Delingette, H., Talos, I.-F., Golby, A. J., Kikinis, R., Jolesz, F. A., Ayache, N., & Warfield, S. K. (2005). Robust Nonrigid Registration to Capture Brain Shift From Intraoperative MRI. *IEEE Transactions on Medical Imaging*, *24*, 1417–1427.
- Clements, L. W., Collins, J. A., Weis, J. A., Simpson, A. L., Adams, L. B.,  
1210 Jarnagin, W. R., & Miga, M. I. (2016). Evaluation of model-based deformation correction in image-guided liver surgery via tracked intraoperative ultrasound. *Journal of Medical Imaging*, *3*, 015003.
- Courtecuisse, H., Allard, J., Duriez, C., & Cotin, S. (2011). Preconditioner-Based Contact Response and Application to Cataract Surgery. *International  
1215 Conference on Medical Image Computing and Computer-Assisted Intervention*, (pp. 315–322).
- De Lorenzo, C., Papademetris, X., Staib, L. H., Vives, K. P., Spencer, D. D., & Duncan, J. S. (2012). Volumetric Intraoperative Brain Deformation Compensation: Model Development and Phantom Validation. *IEEE Transactions  
1220 on Medical Imaging*, *31*, 1607–1619.
- Dequidt, J., Coevoet, E., Thines, L., & Duriez, C. (2015). Vascular neurosurgery simulation with bimanual haptic feedback. In *12th Workshop on Virtual Reality Interaction and Physical Simulation*.
- Dumpuri, P., Thompson, R. C., Sinha, T. K., & Miga, M. I. (2006). Automated  
1225 Brain Shift Correction Using A Pre-computed Deformation Atlas. *Medical Imaging*, (pp. 61411F–61411F–8).
- Duriez, C., Dubois, F., Kheddar, A., & Andriot, C. (2006). Realistic Haptic

Rendering of Interacting Deformable Objects in Virtual Environments. *IEEE transactions on visualization and computer graphics*, 12, 36–47.

1230 Faure, F., Duriez, C., Delingette, H., Allard, J., Gilles, B., Marchesseau, S., Talbot, H., Courtecuisse, H., Bousquet, G., Peterlik, I., & Cotin, S. (2012). SOFA: A Multi-Model Framework for Interactive Physical Simulation. In *Soft Tissue Biomechanical Modeling for Computer Assisted Surgery* (pp. 283–321). Springer Berlin Heidelberg.

1235 Fedorov, A., Billet, E., Prastawa, M., Gerig, G., Radmanesh, A., Warfield, S. K., Kikinis, R., & Chrisochoides, N. (2008). Evaluation of Brain MRI Alignment with the Robust Hausdorff Distance Measures. Las Vegas, NV, USA.

Felippa, C. A., & Haugen, B. (2005). Unified Formulation of Small-Strain Corotational Finite Elements: I. Theory. *Computer Methods in Applied Mechanics and Engineering*, 194, 2285–2335.

1240 Ferrant, M., Nabavi, A., Macq, B., Black, P. M., Jolesz, F. A., Kikinis, R., & Warfield, S. K. (2002). Serial registration of intraoperative MR images of the brain. *Medical Image Analysis*, 6, 337–359.

Ferrant, M., Nabavi, A., Macq, B., Jolesz, F. A., Kikinis, R., & Warfield, S. K. (2001). Registration of 3-D Intraoperative MR Images of the Brain Using a Finite-Element Biomechanical Model. *IEEE Transactions on Medical Imaging*, 20, 1384–1397.

Fuerst, B., Wein, W., Müller, M., & Navab, N. (2014). Automatic ultrasound–MRI registration for neurosurgery using the 2d and 3d LC2 Metric. *Medical Image Analysis*, (pp. 1312–1319).

1250 Garlapati, R. R., Mostayed, A., Joldes, G. R., Wittek, A., Doyle, B., & Miller, K. (2015). Towards measuring neuroimage misalignment. *Computers in Biology and Medicine*, 64, 12–23.

1255 Garlapati, R. R., Roy, A., Joldes, G. R., Wittek, A., Mostayed, A., Doyle, B., Warfield, S. K., Kikinis, R., Knuckey, N., Bunt, S., & Miller, K. (2014). More

accurate neuronavigation data provided by biomechanical modeling instead of rigid registration. *Journal of neurosurgery*, 120, 1477–1483.

Gerard, I. J., Kersten-Oertel, M., Petrecca, K., Sirhan, D., Hall, J. A., & Collins, D. L. (2017). Brain shift in neuronavigation of brain tumors: A review. *Medical Image Analysis*, 35, 403–420.

Hastreiter, P., Rezk-Salama, C., Soza, G., Bauer, M., Greiner, G., Fahlbush, R., Ganslandt, O., & Nimsky, C. (2004). Strategies for brain-shift evaluation. *Medical Image Analysis*, 8, 447–464.

Hill, D. L. G., Maurer Jr, C. R., Maciunas, R. J., Barwise, J. A., Fitzpatrick, M. J., & Wang, M. Y. (1998). Measurement of intraoperative brain surface deformation under a craniotomy. *Neurosurgery*, 43, 514–526.

Hu, J., Jin, X., Lee, J. B., Zhang, L., Chaudary, V., Guthikonda, M., Yang, K. H., & King, A. I. (2007). Intraoperative brain shift prediction using a 3d inhomogeneous patient-specific finite element model. *Journal of neurosurgery*, 106, 164–169.

Ji, S., Wu, Z., Hartov, A., Roberts, D. W., & Paulsen, K. D. (2008). Mutual-information-based image to patient re-registration using intraoperative ultrasound in image-guided neurosurgery. *Medical physics*, 35.

Joldes, G. R., Wittek, A., & Miller, K. (2009). Suite of finite element algorithms for accurate computation of soft tissue deformation for surgical simulation. *Medical Image Analysis*, 13, 912–919.

Joldes, G. R., Wittek, A., & Miller, K. (2010). Real-time nonlinear finite element computations on GPU – Application to neurosurgical simulation. *Computer Methods in Applied Mechanics and Engineering*, 199, 3305–3314.

Kaster, T., Sack, I., & Samani, A. (2011). Measurement of the hyperelastic properties of ex vivo brain tissue slices. *Journal of Biomechanics*, 44, 1158–1163.

- 1285 Kruse, S. A., Dresner, M. A., Rossman, P. J., Felmlee, J. P., Jack, C. R., & Ehman, R. L. (1999). Palpation of the Brain Using Magnetic Resonance Elastography. *International Society for Magnetic Resonance in Medicine*, (p. 258).
- Kruse, S. A., Rose, G. H., Glaser, K. J., Manduca, A., Felmlee, J. P., Jack Jr., C. R., & Ehman, R. L. (2008). Magnetic resonance elastography of the brain. *NeuroImage*, *39*, 231–237.
- 1290 Li, M., Wittek, A., & Miller, K. (2014). Efficient Inverse Isoparametric Mapping Algorithm for Whole-Body Computed Tomography Registration Using Deformations Predicted by Nonlinear Finite Element Modeling. *Journal of Biomechanical Engineering*, *136*, 084503.
- 1295 McCracken, P. J., Manduca, A., Felmlee, J. P., & Ehman, R. L. (2005). Mechanical Transient-Based Magnetic Resonance Elastography. *Magnetic Resonance in Medicine*, *53*, 628–639.
- Mercier, L., Fonov, V., Haegelen, C., Del Maestro, R. F., Petrecca, K., & Collins, D. L. (2012). Comparing two approaches to rigid registration of three-dimensional ultrasound and magnetic resonance images for neurosurgery. *International journal of computer assisted radiology and surgery*, *7*, 125–136.
- 1300 Miga, M. I., Sun, K., Chen, I., Clements, L. W., Pheiffer, T. S., Simpson, A. L., & Thompson, R. C. (2015). Clinical evaluation of a model-updated image-guidance approach to brain shift compensation: experience in 16 cases. *International journal of computer assisted radiology and surgery*, *11*, 1467–1474.
- 1305 Miller, K., & Chinzei, K. (2002). Mechanical properties of brain tissue in tension. *Journal of Biomechanics*, *35*, 483–490.
- 1310 Miller, K., Horton, A., Joldes, G. R., & Wittek, A. (2012). Beyond Finite Elements: A Comprehensive, Patient-Specific Neurosurgical Simulation Utilizing a Meshless Method. *Journal of Biomechanics*, *45*, 2698–2701.

- Miller, K., Joldes, G. R., Lance, D., & Wittek, A. (2007). Total Lagrangian explicit dynamics finite element algorithm for computing soft tissue deformation. *COMMUNICATIONS IN NUMERICAL METHODS IN ENGINEERING*, *23*, 121–134.
- 1315 Miller, K., & Lu, J. (2013). On the prospect of patient-specific biomechanics without patient-specific properties of tissues. *Journal of the mechanical behavior of biomedical materials*, *27*, 154–166.
- Mohammadi, A., Ahmadian, A., Darbandi Azar, A., Darban Sheykh, A., Amiri, F., & Alirezaie, J. (2015). Estimation of intraoperative brain shift by combination of stereovision and doppler ultrasound: phantom and animal model  
1320 study. *International journal of computer assisted radiology and surgery*, *10*, 1753–1764.
- Morin, F., Chabanas, M., Courtecuisse, H., & Payan, Y. (2017). Biomechanical modeling of brain soft tissues for medical applications. In *Biomechanics of  
1325 Living Organs: Hyperelastic Laws for Finite Element Modeling*.
- Morin, F., Courtecuisse, H., Chabanas, M., & Payan, Y. (2015). Rest shape computation for highly deformable model of brain. *Computer Methods in Biomechanics and Biomedical Engineering*, *18*, 2006–2007.
- Mostayed, A., Garlapati, R. R., Joldes, G. R., Wittek, A., Roy, A., Kikinis,  
1330 R., Warfield, S. K., & Miller, K. (2013). Biomechanical Model as a Registration Tool for Image-Guided Neurosurgery: Evaluation Against BSpline Registration. *Annals of Biomechanical Engineering*, *41*, 2409–2425.
- Müller, M., Dorsey, J., McMillan, L., Jagnow, R., & Cutler, B. (2002). Stable real-time deformations. *Proceedings of ACM SIGGRAPH Symposium on  
1335 Computer Animation (SCA)*, (pp. 49–54).
- Nabavi, A., Black, P. M., Gering, D. T., Westin, C.-F., Mehta, V., Pergolizzi, R. S., Ferrant, M., Warfield, S. K., Hata, N., Scharz, R. B., Wells, W. M.,

- Kikinis, R., & Jolesz, F. A. (2001). Serial Intraoperative Magnetic Resonance Imaging of Brain Shift. *Neurosurgery*, *48*, 787–798.
- 1340 Nakajima, S., Atsumi, H., Kikinis, R., Moriarty, T. M., Metcalf, D. C., Jolesz, F. A., & Black, P. M. (1997). Use of cortical surface vessel registration for image-guided neurosurgery. *Neurosurgery*, *40*, 1201–1210.
- Nesme, M., Payan, Y., & Faure, F. (2005). Efficient, physically plausible finite elements. *Eurographics*, .
- 1345 Nimsky, C., Ganslandt, O., Cerny, S., Hastreiter, P., Greiner, G., & Fahlbush, R. (2000). Quantification of, Visualization of, and Compensation for Brain Shift Using Intraoperative Magnetic Resonance Imaging. *Neurosurgery*, *47*, 1070–1080.
- Nimsky, C., Ganslandt, O., Hastreiter, P., & Fahlbush, R. (2001). Intraoperative  
1350 Compensation for Brain Shift. *Surgical Neurology*, *56*, 357–364.
- Paulsen, K. D., Miga, M. I., Kennedy, F. E., Hoopes, J., Hartov, A., & Roberts, D. W. (1999). A Computational Model for Tracking Subsurface Tissue Deformation During Stereotactic Neurosurgery. *IEEE Transactions on Biomedical Engineering*, *46*, 213–225.
- 1355 Pereira, V. M., Smith-Ockeloen, I., Brina, O., Babic, D., Breeuwer, M., Schaller, K., Lovblad, K.-O., & Ruijters, D. (2016). Volumetric Measurements of Brain Shift Using Intraoperative Cone-Beam Computed Tomography: Preliminary Study. *Operative Neurosurgery*, *12*, 4–13.
- Prastawa, M., Bullit, E., & Gerig, G. (2009). Simulation of brain tumors in MR  
1360 images for evaluation of segmentation efficacy. *Medical Image Analysis*, *13*, 297–311.
- Reinertsen, I., Descoteaux, M., Siddiqi, K., & Collins, D. L. (2007a). Validation of vessel-based registration for correction of brain-shift. *Medical Image Analysis*, *11*, 374–388.

- 1365 Reinertsen, I., Lindseth, F., Askeland, C., Iversen, D. H., & Unsgard, G. (2014).  
Intra-operative correction of brain-shift. *Acta Neurochirurgica*, *156*, 1301–  
1310.
- Reinertsen, I., Lindseth, F., Unsgard, G., & Collins, D. L. (2007b). Clinical  
validation of vessel-based registration for correction of brain-shift. *Medical*  
1370 *Image Analysis*, *11*, 673–684.
- Rivaz, H., & Collins, D. L. (2015). Deformable registration of preoperative  
MR, pre-resection ultrasound, and post-resection ultrasound images of neu-  
rosurgery. *International journal of computer assisted radiology and surgery*,  
*10*, 1017–1028.
- 1375 Roberts, D. W., Hartov, A., Kennedy, F. E., Miga, M. I., & Paulsen, K. D.  
(1998). Intraoperative Brain Shift and Deformation: A Quantitative Analysis  
of Cortical Displacement in 28 Cases. *Neurosurgery*, *43*, 749–758.
- Sack, I., Beierbach, B., Wuerfel, J., Dieter, K., Hamhaber, U., Papazoglou, S.,  
Martus, P., & Braun, J. (2009). The impact of aging and gender on brain  
1380 viscoelasticity. *NeuroImage*, *46*, 652–657.
- Sase, K., Fukuhara, A., Tsujita, T., & Konno, A. (2015). GPU-accelerated  
surgery simulation for opening a brain fissure. *Robomech Journal*, *2*, 1–16.
- Schiavone, P., Chassat, F., Boudou, T., Promayon, E., Valdivia, F., & Payan, Y.  
(2009). In vivo measurement of human brain elasticity using a light aspiration  
1385 device. *Medical Image Analysis*, *13*, 673–678.
- Schroeder, W., Martin, K., & Lorensen, B. (2006). *The Visualization Toolkit*  
(4th ed.). Kitware.
- Shakeri, M., Ferrante, E., Tsogkas, S., Lippe, S., Kadoury, S., Kokkinos, I., &  
Paragios, N. (2016). Prior-based Coregistration and Cosegmentation. *Medical*  
1390 *Image Computing and Computer-Assisted Intervention - MICCAI 2016*, *2*,  
529–537.

- Smith, S. M. (2002). Fast Robust Automated Brain Extraction. *Human Brain Mapping, 17*, 143–155.
- Sun, H., Lunn, K. E., Farid, H., Wu, Z., Roberts, D. W., Hartov, A., & Paulsen, K. D. (2005a). Stereopsis-Guided Brain Shift Compensation. *IEEE Transactions on Medical Imaging, 24*, 1039–1052.
- Sun, H., Roberts, D. W., Farid, H., Wu, Z., Hartov, A., & Paulsen, K. D. (2005b). Cortical Surface Tracking Using a Stereoscopic Operating Microscope. *Neurosurgery, 56*, 86–97.
- 1400 Sun, K., Pfeiffer, T. S., Simpson, A. L., Weis, J. A., Thompson, R. C., & Miga, M. I. (2014). Near Real-Time Computer Assisted Surgery for Brain Shift Correction Using Biomechanical Models. *IEEE Journal of Translational Engineering in Health and Medicine, 2*, 1–13.
- Taha, A. A., & Hanbury, A. (2015). Metrics for evaluating 3d medical image segmentation: analysis, selection, and tool. *BMC Medical Imaging, 15*, 29.
- 1405 Tavnier, A., Roy, T. D., Hor, K., Majimbi, M., Joldes, G. R., Wittek, A., Bunt, S., & Miller, K. (2016). On the appropriateness of modelling brain parenchyma as a biphasic continuum. *Journal of the mechanical behavior of biomedical materials, 61*, 511–518.
- 1410 Uffmann, K., Maderwald, S., de Greiff, A., & Ladd, M. E. (2004). Determination of Gray and White Matter Elasticity with MR Elastography. *Proc. Intl. Soc. Mag. Reson. Med, 11*.
- Valencia, A., Benjamin, B., & Ortega, J. H. (2012). Modeling of Brain Shift Phenomenon for Different Craniotomies and Solid Models. *Journal of Applied Mathematics, 2012*.
- 1415 Vermandel, M., Betrouni, N., Taschner, C., Vasseur, C., & Rousseau, J. (2007). From MIP to MRA segmentation using fuzzy set theory. *Computerized Medical Imaging and Graphics, 31*, 128–140.

- Vigneron, L. M., Noels, L., Warfield, S. K., Verly, J. G., & Robe, P. A. (2012).  
1420 Serial FEM/XFEM-Based Update of Preoperative Brain Images Using Intra-  
operative MRI. *International Journal of Biomedical Imaging*, 2012, 872783.
- Wan, M., Liang, Z., Ke, Q., Hong, L., Bitter, I., & Kaufman, A. (2002). Auto-  
matic Centerline Extraction for Virtual Colonoscopy. *IEEE Transactions on  
Medical Imaging*, 21, 1450–1460.
- 1425 Wittek, A., Hawkins, T., & Miller, K. (2009). On the unimportance of con-  
stitutive models in computing brain deformation for image-guided surgery.  
*Biomechanics and modeling in mechanobiology*, 8, 77–84.
- Wittek, A., Miller, K., Kikinis, R., & Warfield, S. K. (2007). Patient-specific  
model of brain deformation: Application to medical image registration. *Jour-  
1430 nal of Biomechanics*, 40, 919–929.
- Yousefi, H., Ahmadian, A., Khodadad, D., Saberi, H., & Daneshmehr, A. (2013).  
An optimised linear mechanical model for estimating brain shift caused by  
meningioma tumours. *International Journal of Biomedical Science and En-  
gineering*, 1, 1–9.
- 1435 Yushkevich, P. A., Piven, J., Cody Hazlett, H., Gimpel Smith, R., Ho, S., Gee,  
J. C., & Gerig, G. (2006). User-guided 3d active contour segmentation of  
anatomical structures: Significantly improved efficiency and reliability. *Neu-  
roImage*, 31, 1116–1128.
- Zhang, J. Y., Joldes, G. R., Wittek, A., & Miller, K. (2013). Patient-specific  
1440 computational biomechanics of the brain without segmentation and meshing.  
*International journal for numerical methods in biomedical engineering*, 29,  
293–308.