Development of an analytical methodology for obtaining quantitative mass concentrations from LAAP-ToF-MS measurements

R Gemayel, B Temime-Roussel, N Hayeck, A Gandolfo, S Hellebust, S Gligorovski, Henri Wortham

► To cite this version:

R Gemayel, B Temime-Roussel, N Hayeck, A Gandolfo, S Hellebust, et al.. Development of an analytical methodology for obtaining quantitative mass concentrations from LAAP-ToF-MS measurements. *Talanta*, 2017, 174, pp.715-724. 10.1016/j.talanta.2017.06.050 . hal-01635700

HAL Id: hal-01635700

<https://hal.science/hal-01635700>

Submitted on 25 Apr 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

**Development of an analytical methodology for obtaining quantitative mass
concentrations from LAAP-ToF-MS measurements**

Rachel Gemayel^{1*}, Brice Temime-Roussel¹, Nathalie Hayeck^{1[‡]}, Adrien Gandolfo¹, Stig
Hellebust^{1#}, Sasho Gligorovski^{*1,2}, Henri Wortham¹.¹ Aix Marseille Univ, CNRS, LCE,
Marseille, France.

²State Key Laboratory of Organic Geochemistry, Guangzhou Institute of Geochemistry,
Chinese Academy of Sciences, Guangzhou 510 640, China

Submitted to

Talanta

Correspondence to:

*e-mail: rachel.gemayel@etu.univ-amu.fr

*e-mail: gligorovski@gig.ac.cn

Now at: Dept. Chemistry, University College Cork, College Road, Cork, Ireland

[‡]Now at: Univ Lyon, Université Claude Bernard Lyon 1, CNRS, IRCELYON, F-69626, Villeurbanne,
France

1 *Abstract*

2 Laser ablation aerosol particle-time of flight mass spectrometer (LAAP-ToF-MS) measures the
3 size number of particles, and chemical composition of individual particles in real-time. LAAP-
4 ToF-MS measurements of chemical composition are difficult to quantify, mostly because the
5 instrument sensitivities to various chemical species in the multicomponent atmospheric aerosol
6 particles are unknown. In this study, we investigate a field-based approach for quantitative
7 measurements of ammonium, nitrate, sulfate, OC, and EC, in size-segregated atmospheric
8 aerosols, by LAAP-ToF-MS using concurrent measurements from high-resolution time-of-
9 flight aerosol mass spectrometer (HR-ToF-AMS), and multi-angle absorption photometer
10 (MAAP). An optical particle counter (OPC) and a high-resolution nanoparticle sizer (scanning
11 mobility particle sizer, or SMPS), were used to measure the particle size distributions of the
12 particles in order to correct the number concentrations. The intercomparison reveals that the
13 degree of agreement of the mass concentrations of each compound measured with LAAP-ToF-
14 MS and HR-ToF-AMS/MAAP increases in the following order $\text{NH}_4^+ < \text{SO}_4^{2-} < \text{NO}_3^- < \text{EC} < \text{OC}$
15 $< \text{Cl}^-$ with r^2 values in the range of 0.4-0.95 and linear regression slopes ranging between 0.62
16 and 1.2. The factors that affect the mass concentrations measured by LAAP-ToF-MS are also
17 discussed in details. Yet, the matrix effect remains one of the strongest limiting factor to achieve
18 an absolute quantification of the aerosol chemical composition.
19 In the future we suggest the development of a methodology based on the calculation of the
20 response factors generated by different types of particles, which could possibly resolve certain
21 difficulties associated with the matrix effect.

22 **Keywords:** single aerosol; laser ablation, LAAP-ToF-MS; matrix effect; quantification.

1

2 *Introduction*

3 Atmospheric aerosol particles largely influence the air quality in urban and semi urban
4 environments. It is known that high particle mass concentrations can induce hazardous
5 respiratory health effects [1]. Adverse health effects are associated with particles having
6 aerodynamic diameters smaller than 2.5 μm (PM 2.5). Namely, the size range of particles
7 inferior of 2.5 μm is well correlated with the human mortality rates [2]. A recently performed
8 global atmospheric chemistry model [3] aimed to calculate the concentrations of pollutants,
9 reported that 3.3 million people die prematurely every year as a result of air pollution, mostly
10 associated with PM 2.5. These authors [3] predicted that by 2050 the number of particles-
11 induced deaths worldwide could rise to 6.6 million per year.

12 Also, aerosol particles have a large impact on Earth's climate by scattering and absorbing
13 radiation and by serving as nuclei for cloud formation [4].

14 Comprehensive knowledge about the physical and chemical properties of aerosols is crucial for
15 properly evaluating the effects of aerosols on human health, air quality and climate changes [5].

16 Therefore, long-term monitoring of atmospheric particles and quantification of chemically
17 resolved composition in ambient particles is of paramount importance for developing legislative
18 tools and creating clean air action plans. Recently, Gemayel et al. (2016) [6] presented the
19 ability of a recently launched commercial single particle mass spectrometer, the laser ablation
20 aerosol particle-time of flight mass spectrometer (LAAP-ToF-MS; Aeromegt GmbH) to detect
21 and analyze atmospheric particles. However, the quantification of the chemical composition by
22 laser desorption ionization (LDI) technique appears to be a challenging task due to
23 inhomogeneous laser beam profile [7] (the effects of the shot-to-shot single particle ion signal
24 fluctuations), size-dependent particle transmission efficiency and matrix effect on detection and
25 ionization [8]. There is an important number of studies [9][10][11][12][13][14] which
26 demonstrates that quantification of the chemical composition by LDI techniques is feasible by

1 mathematical and comparative methods. The comparative approach consist to converts the ion
2 intensity produced by LDI into mass concentration. The conversion is based on comparison to
3 another techniques for particle analysis [15][10][8] [16] [17][18][19].

4 It has been questioned as to whether laser desorption ionization (LDI) analysis of aerosols can
5 be used to quantitatively assess the amounts of specific chemical species such as organic
6 compounds (OC) and elemental carbon (EC) in ambient particles. . Healy et al. (2013) and
7 Jeong et al. (2011) developed the idea of quantifying aerosol chemical composition based on
8 the specific ion intensities response factors using concurrent quantitative measurements
9 including an Organic/Elemental carbon analyser and a high resolution time-of-flight aerosol
10 mass spectrometer (HR-ToF-AMS) [18,19].

11 For example, aerosol time of flight mass spectrometer (ATOFMS) has been used in a number
12 of atmospheric measurement campaigns and provides complementary information on aerosol
13 chemistry and particle variability in the environment [20,21]. Single particle mass
14 spectrometers such as ATOFMS provide information on the number concentrations of
15 chemically distinct particle types, showing how these concentrations change over time [22,23].
16 The same methodology [10,15,24] was also adopted in this study with an additional step based
17 on the calculation of a response factor of the LAAP-ToF-MS for different size ranges in order
18 to quantify several chemical compounds.

19 In this study we compared the performances of LAAP-ToF-MS against another well-
20 established commercial mass spectrometer, high resolution-time of flight-aerosol mass
21 spectrometer (HR-ToF-AMS) (Aerodyne research Inc.,) with the aim of quantitative
22 determination of chemical compounds within the atmospheric aerosol particles. In addition, a
23 multi-angle absorption photometer (MAAP, Thermo Sci., model 5012) was also employed to
24 analyze the elemental carbon. An optical particle counter (OPC) and a high-resolution
25 nanoparticle sizer (scanning mobility particle sizer, or SMPS), were measuring the particle size

distributions of the aerosol particles for the purpose of correcting the number concentrations as measured by the LAAP-ToF-MS. The emerging outcomes of these comprehensive measurements together with the developed methodology to quantify the chemical composition of the atmospheric particles are described below.

Methodology

Sampling sites and instruments

The measurements were performed from 19 January until 29 January 2015 at the campus of Aix Marseille University situated in the city center of Marseille (43.30° N, 5.38° E).

The site is influenced by many aerosol sources since the sampling point was located about 300 m from the railway station, 500 m from the highway and 1.3 km from the harbor. For these measurements a single particle laser ablation mass spectrometer, LAAP-ToF-MS [6], was applied for on-line and continuous monitoring of single aerosol particles. The working principle of LAAP-ToF-MS is detailed by Gemayel et al. (2016) [6]. Briefly, LAAP-ToF-MS is equipped with aerodynamic lenses which focus particles in the range between 80 nm and 700 nm into a narrow beam. The particle beam passes through a nozzle where the particles are accelerated depending on their size (vacuum aerodynamic diameter, d_{va}). Then the particle beam travels through two light scattering detection stages (laser diodes, wavelength: 405 nm). The vacuum aerodynamic diameter is calculated from the time of flight of the particle between the two detection stages [6].

The second detection stage triggers an excimer laser ($\lambda = 193$ nm; irradiance: approx. 10^9 W cm⁻²) (EX 5, GAM Laser, Inc) to pulse and ablate the particles. The Gaussian profile of the laser beam, hot spots in the laser beam and shot-to-shot variation in laser pulse characteristics [7] [8]), are three parameters that strongly affect the ionization and detection of the particles. Therefore, two mathematical quantities are introduced to quantify the detection efficiency, i) the scattering efficiency which corresponds to the ratio between the number of particles detected by the laser diodes and the number of

particles entering the instrument and ii) the hit rate which is defined as a ratio between the number of particles ionized by the ionization laser and the number of particles detected by the laser diode ([6][25][26]). The generated ions are analyzed in a bipolar time of flight mass spectrometer with the resolution $m/\Delta m > 600$.

In this study, LAAP-ToF-MS was used to measure the size-resolved chemical composition of single particles in the size range between 150 nm and 1 μ m. The dual ion mass spectrum was treated subsequently in MATLAB (R2013b). A HR-ToF-AMS[27,28] was used for real-time measurements of size resolved aerosol chemical composition. A Multi-Angle Absorption Photometer (MAAP, Thermo Sci., model 5012) [29] was used to measure elemental carbon, an Optical Particle Counter (OPC 1.109, Grimm) and a high resolution nanoparticle sizer Scanning Mobility Particle Sizer (TSI, SMPS 3936) were used to measure the particle size distributions to correct the number concentrations measured by LAAP-ToF-MS. All instruments were connected to the same sampling system consisting of a TSP (Total suspended particulates) inlet located approximately 3 m above the ground level. The sampled air was dried using an aerosol diffusion dryer system maintaining the relative humidity in the line below 30%. The sample air flow was then split and diverted to each of the instruments using a custom built flow splitter. In order to allow a direct comparison between the HR-ToF-AMS the LAAP-ToF-MS and the MAAP, a PM1 cyclone was installed at the inlet of the MAAP.

Comparison of the analyzed data

In a very recent study, Gemayel et al. (2016) reported the low detection efficiency of 2.5 % for the laser scattering diodes with respect to the particles having a diameter of 450 nm and even lower detection efficiency for smaller particles. Therefore, it is essential to scale the LAAP-ToF-MS data with those obtained by OPC and SMPS. Then, the next step would be the comparison of the data obtained by LAAP-ToF-MS with the data obtained by HR-ToF-AMS and MAAP. Hence, it is crucial that HR-ToF-AMS and

MAAP are able to detect the total particle mass concentrations. SMPS and OPC detect particles in the range 14 nm - 650 nm and 250 nm - 30 μm , respectively. Therefore, for comparison purpose and scaling the data of LAAP-ToF-MS the particle concentrations in the range between 150 nm and 250 nm measured by SMPS were used in this study. Then for the other size ranges, the particle concentrations measured by the OPC are used. Different types of diameters are measured by the instruments applied in this study, i.e. electrical-mobility diameter for SMPS, geometrical diameter for OPC and aerodynamic diameter for LAAP-ToF-MS. Therefore, the determination of the detection efficiency requires a homogeneity between the three types of diameters.

DeCarlo et al. (2004) reported that geometrical diameter (d_g) is equal to the electrical-mobility diameter (d_{me}) in case of spherical particles (shape factor =1). In this study, the particles are considered as spherical which implies an equivalence between geometrical and electrical-mobility diameter [30]. Thereby, to convert these two diameters to vacuum aerodynamic diameter (d_{va}) the Eq-1 is applied

$$d_{va} = d_g \times \left(\frac{\rho_p}{\rho_0} \right)^{0.5} \times f_{shape} \quad \text{Eq-1}$$

In this case, the factor shape is also considered equal to 1 and the particle density is assumed to be 1.5 $\text{g}\cdot\text{cm}^{-3}$ [1]. However, one should be aware that use of these values for density and shape factor could induce errors to the quantification. The average of shape factor and the average of density is different even for different periods of the day [31,32]. On the other hand, Liu et al. (1999) demonstrated that OPC underestimates the size of ambient particles because they have a refractive index (n) lower than the refractive index of the model particles (in this case polystyrene latex $n= 1.6$) used to calibrate the OPC [33]. The maximum underestimation of the particle size occurs for particles size comparable to the wavelength ($\lambda=655$ nm) used for the OPC.

1

2 Scaling of the LAAP-ToF-MS data

3 For data analysis, approximately 112 000 dual-ion single-particle mass spectra (Figure
4 1) have been collected for particles in the range between $d_{va}= 225 \text{ nm}$ ($d_{me}= 150\text{nm}$) and
5 $d_{va}= 1.5\mu\text{m}$ ($d_{me}= 1\mu\text{m}$).

6 **Insert Figure 1**

7 A representative spectrum of the total ionised particles is presented in the Figure 1. In
8 this figure, the standard deviation between the intensities of each m/z during the whole
9 duration of the field campaign is presented in function of m/z for positive and negative
10 ions. The choice of the standard deviation is based on the aim of this study, to present a
11 maximum of detected ions and to emphasize the ions that change during the
12 measurements. These ions correspond mostly to the typical compounds observed in the
13 urban aerosols: sulfate, nitrate, potassium, ammonium, organic compounds (OC),
14 elemental carbon (EC). Other ions correspond to metals such as iron and lead and ions
15 corresponding to chloride and sodium. The presence of such ions is reasonable
16 considering that the experimental campaign was held near the highway and the harbour.
17 The chloride can be derived from different chemical combinations such as NaCl (sea
18 salt) or ammonium chloride (industrial aerosol). The discussion about the presence of
19 these ions is further elaborated in section "Discussion". All the collected data are
20 normalized in function of particles size.

21 Figure 2 shows the repartition of the number of particles in different size ranges as detected by
22 OPC/SMPS and the number of particles ionised by the LAAP-ToF-MS during all the campaign,
23 for particle size ranges corresponding to the one detected by OPC/SMPS. The number of
24 ionised particles by LAAP-ToF-MS is considered as the number of exploitable spectra. The
25 detection efficiency ($D \%$) of LAAP-ToF-MS is defined as the ratio between the number of

particles sized and ionised by LAAP-ToF-MS and the number of particles measured by OPC/SMPS.

Insert figure 2

The detection efficiency of LAAP-ToF-MS varies in function of particle size ranges. For example, it is very low ca. 0.01 % for particle size ranges below 400 nm. Then, starting from 400 nm it increases stepwise reaching about 1% for the size range 580-650 nm, and it remains constant up to 1 μ m.

The relative standard deviation (RSD) for the detection efficiency calculated for each 5 min measurements widely varies depending on the number of detected particles. For example, the number of detected particles is very high for the particle range 500-580 nm, and consequently the RSD on the detection efficiency is about 50 %. However, the uncertainty of the detection efficiency is more than 200% for particles size ranges below 400 nm and over 1000 nm.

To reduce the uncertainty of the quantification, the detection efficiency was averaged every 60 min as described in the section below and the totality of the spectra is grouped into eight different size ranges expressed as d_{me} , as follows: 150-250 nm, 250-400 nm, 400-450 nm, 450-500 nm, 500-580 nm, 580-650 nm, 650-1000 nm. The corresponding RSD are listed as follows 103; 106; 57; 51; 50; 45 and 73 %, respectively. The first range (150-250 nm) corresponds to the size range as detected by the SMPS while the others correspond to the size range detected by the OPC.

Methodology

Table 1 lists the ion markers chosen to monitor main particle components by the LAAP-ToF-MS. As previously used by single particle mass spectrometer studies (i.e. ATOFMS) [19,24,346 37], the list includes sulfate, ammonium, nitrate, chloride, organic compounds (OC) and elemental carbon (EC).

Insert table 1

The first step in reducing uncertainties on detection efficiency consists of reducing the shot-to-shot effect due to the inhomogeneity in the laser beam by normalizing the intensities of the specific marker ions produced by the LAAP-ToF-MS (as listed in table 1) to the total ion current (TIC) of the corresponding dual ion spectrum (I_i). As a second step, the mean value of the intensity of TIC-normalised marker ions ($^1I_{i,t}$) is calculated for each hour and assigned to different size ranges (different n) ($^1I_{i,n,t}$).

Then the corresponding intensities of different ions ($^1I_{i,n,t}$) measured by LAAP-ToF-MS are divided by the corresponding size-specific detection efficiency (D) calculated for each hour:

$$^1I_{i,n,t}^D$$

For each particle component and in each particle size range, an average response factor over the whole campaign ($\bar{R}_{i,n}$) is determined as

$$\bar{R}_{i,n} = \frac{\sum_{t=1}^T \frac{^1I_{i,n,t}}{D_{i,n,t}}}{\sum_{t=1}^T \bar{C}_{i,n,t}} \quad \text{Eq-2}$$

where $\frac{^1I_{i,n,t}}{D_{i,n,t}}$ is the LAAP-ToF-MS averaged scaled ion intensity and $\bar{C}_{i,n,t}$ is the mean mass concentration derived from the HR-ToF-AMS data acquired in the P-tof mode [38].

The hourly LAAP-ToF-MS reconstructed mass concentrations were then calculated by dividing the TIC-normalised marker ion intensities scaled by the corresponding size-specific detection efficiency, $\frac{^1I_{i,n,t}}{D_{i,n,t}}$, by the mean size-specific response factor $\bar{R}_{i,n}$ and by summing over the whole size range:

$$\bar{C}_{i,n,t} = \frac{\sum_{n=1}^N \frac{^1I_{i,n,t}}{D_{i,n,t}}}{\bar{R}_{i,n}} \quad \text{Eq-3}$$

The sum of the mass concentrations calculated over the whole size ranges are defined by $M_{i,\Sigma n,t}$.

The $M_{i,\Sigma n,t}$ are compared to the hourly averaged mass concentrations measured by HR-ToF-AMS for the same chemical moieties listed in Table 1 ($C_{i,\Sigma n,t}$). During this campaign, ammonium concentrations were insufficient to work with the P-ToF mode and thus to express these concentrations in function of size range. On the other hand, the MAAP measurements provide a total concentration of EC for particles ranging between 1 nm and 1000 nm. For the analysis of these two compounds, ammonium and EC an additional step was used to calculate the response factor. The developed methodology is given in the supplementary information.

Results

Response factor

The obtained response factors ($RF_{i,n}$) from this study were compared to those obtained by Healy et al. (2013) [18] using an ATOFMS which operational principle (laser ablation) is similar to the one by LAAP-ToF-MS. However, the comparison was not an easy task because the characteristics of the ionisation laser are not the same (ATOFMS: $\lambda = 266$ nm, $E = 1.3$ mJ; LAAP-ToF-MS: $\lambda = 193$ nm, $E = 4$ mJ). For this reason we compared only the ranking. Another major difference is that the RF in this study were calculated for each size range while the RF in Healy et al. (2013) were estimated for a bulk particle ranging between 150 nm and 1067 nm. For this reason, the RF of the most abundant size range (500-6580 nm) has been chosen for comparison. The relative importance of the RF obtained by the two instruments (LAAP-ToF-MS and ATOFMS) were equivalent as shown in Figure 3 B. The RF of the compounds is presented in a decreasing order as follows: EC > sulfate ~ nitrate > organic > ammonium. As a result, in spite of the difference between the two wavelengths used by the two compared instruments (i.e. 193 nm for LAAP-ToF-MS and 266 nm for ATOFMS), both instruments give the same result, i.e., the highest signal is obtained for EC and the lowest for OC and ammonium. Although the amount of chloride ions generated per mass unit is in the same order of magnitude as those of nitrate and sulfate, the chloride was not included in this comparison because it was not measured by ATOFMS [18].

Insert figure 3

As a second step, the influence of particle size on the absolute RF was examined. As the RF is the ratio between the signal intensity and the mass concentration, a decrease in RF implies that for the same mass concentration the quantity of generated ions decreases. Figure 3 A shows the behavior of the response factor of each chemical compound for different particle sizes. It can be noticed that the RF decreases considerably when the size range increases. As a result, the quantity of generated ions decreases with the particle size but this depends on the considered compound. This result is surprising because the bigger particles should contain higher amount of compounds. Two phenomena could explain this result. First, the light emerging from the excimer laser is independent of the particle size so that, the quantity of energy available per mass unit decreases when the particle size increases [10,39]. The second reason is linked to the particle composition which depends on the size. The composition of the smaller particle facilitates the ionization mechanism [8,10,39]. This effect of the matrix will be discussed more in details below.

Evolution of the mass concentration

The results of the comparison between the hourly averaged mass concentrations obtained by the HR-ToF-AMS and MAAP with those derived from the LAAP-ToF-MS are presented in Figure 4 and Figure 5. The time series of the reconstructed mass concentrations from the three instruments are in good agreement.

Insert figure 4

Insert figure 5

The difference between the mass concentrations derived from the LAAP-ToF-MS and those given by the HR-ToF-AMS and the MAAP is expressed as an σ_{error} and is calculated according to Eq-4:

$$1 \quad \text{Error} = \frac{\text{LAAP-ToF-MS} - \text{HR-ToF-AMS/MAAP}}{\text{HR-ToF-AMS/MAAP}} \times 100 \quad \text{Eq-4}$$

2 An error equal to zero signifies an excellent agreement between LAAP-ToF-MS and HR-ToF-
3 AMS/MAAP. An overestimation by the LAAP-ToF-MS is translated by an error above zero,
4 while an underestimation by the LAAP-ToF-MS is interpreted by an error below zero.

5 To interpret statistically the good agreement between the two instruments, the concentrations
6 of each compound measured with LAAP-ToF-MS versus HR-ToF-AMS/MAAP are reported
7 on Figure 6. The degree of the agreement increases in the following order $\text{NH}_4^+ < \text{SO}_4^{2-} < \text{NO}_3^-$
8 $< \text{EC} < \text{OC} < \text{Cl}^-$ with r^2 values in the range of 0.4-0.95 and slopes extrapolated from the linear
9 regressions range between 0.6 and 1.2. This result is satisfying given: i) the uncertainties and
10 assumptions associated to the conversion of ion intensity to mass concentration of chemical
11 species i.e., the factor shape was assumed equal to 1 (spherical particles), the density was
12 considered constant ($1.5 \text{ g} \cdot \text{cm}^{-3}$) and the detection efficiency was based on the particle detection
13 instead on the detection of the given compound considered in this study, ii) the limitations
14 generally attributed to single laser mass spectrometers in association with the matrix effect on
15 desorption and ionisation processes.

16 **Insert figure 6**

17 Three parameters could explain different results obtained during the comparison between
18 LAAP-ToF-MS and HR-ToF-AMS/MAAP: 1) the matrix effect influencing the results using
19 LDI techniques, 2) the detection limit in size (LAAP-ToF-MS: 150 nm-1000 nm; MAAP: PM1)
20 and 3) the effect of relative humidity. In this study, the sampled air was dried using an aerosol
21 diffusion dryer system, which considerably reduces the effect of the relative humidity.

22 Yet, in comparison to the HR-ToF-AMS/MAAP, LAAP-ToF-MS underestimate the mass
23 concentrations of ammonium, EC, sulfate, OC and nitrate (Figure 6). Nevertheless, a statistical
24 test (correlation coefficient) carried out on the errors calculated for the various species, indicates

that moderate correlations exist between nitrate-sulfate (0.50) and nitrate δ OC (0.48), suggesting that these errors are linked mostly to the matrix effect.

Discussion

The matrix could have an impact on particles detection by mean of the scattering laser and the ionisation rate. Since the effect of the scattering laser is already included in the detection efficiencies values, the attention was focused on the matrix effect on the ionisation rate.

This matrix effect in the laser desorption studies is still misinterpreted and under discussion [8,40]. Thomson et al (1997) suggested that the energetic level necessary to obtain an ion production of a constituent strongly depends on other compounds present within the particle under analysis [39]. These authors demonstrated that the ion formation threshold for NH_4NO_3 with a 193 nm ionisation laser, presents the lowest value (4.2-5.6 MW cm⁻²) compared to the combination of other species such as ammonium sulfate (23-57 MW cm⁻²). On the other hand, Reinard et al (2008) have shown that numerous interactions exist in the plasma between the ions produced. According to these authors two cases have to be considered:

In plasma with low ion density, few collisions occur between ions, as a result, ions having low electronic affinity will survive and will be dominant. On the other hand, ions with high electronic affinity constitute the lower fraction of the total ion signal [35]. In plasma with high ion density many collisions occur and ions with low electronic affinity are deactivated; the high electronic affinity ions are dominant [35].

In this work, because field experiments were performed and because the analytical results represent hourly averaged values, it is difficult to evaluate the individual relative importance of each phenomenon (the effect of a dense plasma and the effect of the threshold ionisation energy that is, the minimum energy required to produce a detectable ion signal).

Therefore, the interactions between mass concentration of one species and the intensity of other species's signal, were evaluated. For this purpose, the correlations between concentration levels of each species (EC, OC, chloride, ammonium, nitrate and sulfate), and the individual errors of other species were assessed. Errors were calculated based on Eq-4.

These calculations lead to 36 possible combinations for the six considered compounds in this study including the effect of each compound on its own. On a first glance (Figure 7), it can be noticed, for each species, the absence of a systematic correlation between the mass concentration ($\mu\text{g m}^{-3}$) and its own error (%). This observation immediately discards the possibility of an error associated with the concentrations of the compounds in the sample. Among the 30 remaining combinations, only 8 exhibits statistically relevant correlation coefficients (>0.4) which are shown in Figure 7.

Factors influencing the error on the mass concentrations of OC

One of the limitations of the single aerosol mass spectrometers based on laser ablation is the variation of the ablation efficiency in function of particle size and chemical composition [41,42]. Kane and Johnston (2000) evaluated the ablation efficiency with respect to simple laboratory-generated oleic acid particles. However, when ablating polydisperse multicomponent aerosols, such as that found in a real atmosphere, the measurements can be influenced greatly by the presence of additional chemical species in the particle. Therefore, they also measured the ablation efficiency of mixed oleic acid and ammonium nitrate in a ratio of about 10:1 by mass. They demonstrated that for each size, ablation efficiencies of the mixed particles are greater than pure oleic acid by a factor of 3 and smaller than pure ammonium nitrate [41].

Even though the average ratio between OC and nitrates is about 3 in this study, it is clear that the presence of ammonium and nitrates has improved the ionization of the OC (figure 7).

1 Since nitrate and ammonium are simultaneously present in both studies, it is impossible to
2 identify which species (nitrate or ammonium) is responsible for the variation of the signal.

3 In this study a dependence of the OC signal (the same order as the one for nitrates) with respect
4 to the concentration of sulfate was observed. Kane and Johnston (2000) did not evaluated this
5 influence because they followed only positive ions which does not allow sulfate analysis.

6 **Insert figure 7**

7 Considering that atmospheric aerosols contain nitrate and sulfate in the form of ammonium
8 nitrate and ammonium sulfate, it can be envisaged that the variation of OC signal is associated
9 with the ammonium rather than nitrate or sulfate. To confirm this hypothesis, additional
10 measurements are necessary with other types of ammonium salts. The ammonium chloride
11 could be a good choice because chloride concentrations does not affect the OC signal.

12 Factors influencing the error on the mass concentrations of sulfate

13 The signal of sulfate is largely influenced by the nitrates, the OC, and EC (Figure 7). To the best
14 of authors knowledge, the influence of nitrate and EC on the sulfate, as observed in this study,
15 was not previously evaluated in laboratory experiments. Kane and Johnston (2001) have studied
16 the influence of OC on the signal of sulfate. These authors have shown that coatings of 1-
17 naphthyl acetate applied on an aerosol of ammonium sulfate, in a volume ratio 3:1, resulted in
18 a 20 % increase in the detection of sulfate. In the present study, the ratio OC/ sulfate that is
19 10:1, is expressed in mass concentration; thus cannot be directly compared to the study by Kane
20 and Johnston (2001). The error of the OC signal varies between -100 and + 200 % which results
21 in a total variation of 300 %. This dependence with respect to the sulfate exhibits the same trend
22 as the study by Kane and Johnston, (2001) with much bigger values in this study (300 %) against
23 20 % [42]. The reason for such a big discrepancy is not clear yet.

1 Factors influencing the error on the mass concentrations of nitrates

2 The EC is the only species affecting the error on nitrates signal. This correlation, to the best of
3 authors knowledge was not studied before. On the other hand, in the present study no influence
4 of OC was observed on the nitrates as was also the case in the study by Kane and Johnston
5 (2001) [41].

6 Factors influencing the error on the mass concentrations of EC

7 In a very recent study, Ahern et al. (2106) have demonstrated that the ionization of organics
8 leads to the formation of ion C^+ which interferes with one of the fragments used to follow the
9 signal of EC [43]. Figure 7 shows that the error on EC is mostly positive ranging between -80%
10 and 380% and that this error is well correlated with the concentration of the organics ($R^2 =$
11 0.42). Thus, it seems that the presence of organics leads to an overestimation of EC.

12 On the other hand the specific ions which were chosen for the quantification of EC, are
13 exclusively of type C_x^+ and C_x^- where x ranges between 1 and 4. The fragments C_x ($x>4$) were
14 not considered, neither positive nor negative, because the signal/noise was too low. Thus, these
15 C_x ($x>4$) ions could affect the quantitative measurements of EC. Figure 7 shows that OC affects
16 the error of EC which origin could be the modification in the ionization mode of EC, in presence
17 of organic species. Several studies [36,37,43,44] observed such influence of OC on EC, and
18 they even used the fragmentation mode of EC to identify their sources.

19 In the present study, the response factors of EC are calculated from the C_x^+ fragments, where x
20 ranges between 1 and 4. Nevertheless, the presence of organics favors the fragmentation of EC
21 in a sense that x in the produced ions of EC, ($C_x^{+/-}$), ranges between 1 and 3. Under such
22 conditions, all the ions emerged from the ionization of EC were considered, which is not the
23 case for the particles poor in organic material, where EC is less fragmented ($C_x^{+/-}$ with $1<x<8$).

24 MAAP measurements do not cause fragmentation of the atmospheric aerosols in function of
25 particles diameter. Thus, it gives the average mass concentrations for the ensemble of PM1 (all

the particles with diameter inferior than 1 μm). On the other hand, LAAP-ToF-MS detects only the particles with diameter > 150 nm. Consequently, the presence of strong concentration of particles with diameter inferior than 150 nm containing EC would induce an underestimation of EC by LAAP-ToF-MS. This phenomenon was clearly observed during the field campaign on 21st and 26th of January (red square on Figure 5). In this figure it can be seen the concomitance between the high concentrations of particles with diameter < 150 nm as measured by SMPS and the underestimation of the EC concentrations by LAAP-ToF-MS in comparison to the MAAP measurements.

The correlations emerged from the concurrent measurements
Figure 6 shows the mass concentrations of each of the considered species (sulfate, ammonium, nitrate, chloride, OC, and EC) measured by LAAP-ToF-MS in function of those obtained by concurrent reference measurements by MAAP for EC, and HR-ToF-AMS for sulfate, OC, nitrate, chloride, and ammonium. The slopes of the regression lines for all the species except chloride are in the range between 0.61 and 0.95. This implies that a bigger fraction of these species is detected by HR-ToF-AMS.

Contrary, the slope of the regression line 1.2, obtained for chloride signifies that LAAP-ToF-MS detects better this species than HR-ToF-AMS. This compound can be present in form of NaCl in atmospheric aerosols which are considered to emerge from marine sources [45,46]. In this form the chloride represent refractory species; hence, they cannot be detected by HR-ToF-AMS. The field campaign was held in a place situated only 1 km from the Mediterranean Sea which explains the presence of NaCl in the detected aerosols by LAAP-ToF-MS. During this campaign, 45% of the particles containing chloride were containing sodium as well. The relative standard deviation obtained on the average value of chloride was 43% which explain a high variability in the percentage of chloride in the form of NaCl. Therefore, this variability in

the percentage of chloride detected by the LAAP-ToF-MS is responsible of the overestimation of these ions since the HR-ToF-AMS showed less variability.

Conclusion

LAAP-ToF-MS as a single aerosol mass spectrometer based on laser desorption ionization technique is capable of measuring the time evolution of mass concentrations. This study revealed the importance of four points (described below) which need to be optimized in order to quantify the chemical composition of aerosol particles by LAAP-ToF-MS.

1: An absolute quantification of chemical species during a field campaign is a complex issue due to matrix effects, especially if the measurement site is affected by different sources of aerosols, as was the case in this study. However, a semi-quantification allowing to monitor the evolution of a chemical composition in function of time is possible. Therefore, detected particles should be classified in different classes depending on their composition, i.e. particles containing sulfate and nitrate are considered as one class while those containing only sulfate represents another class. By this, the matrix effect would be surmounted. Then the next point should be applied.

2: The determination of the mass concentration of a chemical composition should respect three important points. First of all, it is important to consider the sum of all the specific ions of each chemical composition. Second, the corresponding intensities of these ions have to be normalized to the detection efficiency. And at last but not least, one have to choose a suitable time interval leading to an overall chemical composition statistically sufficient to overcome the shot-to-shot effect of lasers variability. In this study a time interval including 15 particles was found to be suitable for quantification of mass concentrations.

3: An absolute quantification would be possible in the case of measurements at the source. In this case the chemical composition within the aerosol is stable (not influenced by various sources) and the effect of matrix would not significantly affect the quantification.

4: In the present study, the HR-ToF-AMS used as a reference instrument does not detect chloride present as NaCl, explaining the overestimation of chloride obtained from the LAAP-ToF-MS measurements. Therefore, the quantification of chloride using the LAAP-ToF-MS needs to be compared to an instrument capable of measuring all the possible states of chloride possibly present in an aerosol. In the future, development of a methodology based on laboratory studies and calculation of the response factors by type of particles, would resolve certain difficulties associated with the matrix effect.

Acknowledgements.

This work is a contribution to the LABEX SERENADE (no. ANR-11-LABX-0064) funded by the "Investissements d'Avenir", French Government program of the French National Research Agency (ANR) through the A*MIDEX project (No. ANR-11-IDEX-0001-02). The authors gratefully acknowledge the support of this work by French National Agency of Research within the ANR-10-EQPX-39-01.

References

- [1] R.M. Healy, J. Sciare, L. Poulain, K. Kamili, M. Merkel, T. Muller, a. Wiedensohler, S. Eckhardt, a. Stohl, R. Sarda-Estève, E. McGillicuddy, I.P. O'Connor, J.R. Sodeau, J.C. Wenger, Sources and mixing state of size-resolved elemental carbon particles in a European megacity: Paris, *Atmospheric Chemistry and Physics*. 12 (2012) 168161700. doi:10.5194/acp-12-1681-2012.
- [2] C.. Pope, Invited Commentary: Particulate Matter-Mortality Exposure-Response Relations and Threshold, *American Journal of Epidemiology*. 152 (2000) 4076412.
- [3] J. Lelieveld, J. Evans, M. Fnais, D. Giannadaki, A. Pozzer, The contribution of outdoor air pollution sources to premature mortality on a global scale, *Nature*. 525 (2015) 3676371.
- [4] IPCC, Intergovernmental Panel on Climate Change, *Climate Change fifth report*, 2013.
- [5] S. Gligorovski, J.. Van Elteren, I. Grgic, A multi-element mapping approach for size-segregated atmospheric particles using laser ablation ICP-MS combined with image analysis, *Science of the Total Environment*. 407 (2008) 5946602.
- [6] R. Gemayel, S. Hellebust, B. Temime-Roussel, N. Hayeck, J.T. Van Elteren, H. Wortham, S. Gligorovski, The performance and the characterization of laser ablation aerosol particle time-of-flight mass spectrometry (LAAP-ToF-MS), *Atmospheric Measurement Techniques*. 9 (2016) 194761959. doi:10.5194/amt-9-1947-2016.
- [7] R.J. Wenzel, K.A. Prather, Improvements in ion signal reproducibility obtained using a homogeneous laser beam for on-line laser desorption/ionization of single particles, *Rapid Communications in Mass Spectrometry*. 18 (2004) 152561533. doi:10.1002/rcm.1509.

- [8] D.S. Gross, M.E. Galli, P.J. Silva, K. a Prather, Relative sensitivity factors for alkali metal and ammonium cations in single particle aerosol time-of-flight mass spectra, *Analytical Chemistry*. 72 (2000) 4166422. <Go to ISI>://000084855300022.
- [9] D.-Y. Liu, Aerosol time-of-flight mass spectrometry during the Atlanta Supersite Experiment: 1. Measurements, *Journal of Geophysical Research*. 108 (2003) 8426. doi:10.1029/2001JD001562.
- [10] D.P. Fergenson, X.H. Song, Z. Ramadan, J.O. Allen, L.S. Hughes, G.R. Cass, P.K. Hopke, K.A. Prather, Quantification of ATOFMS data by multivariate methods, *Analytical Chemistry*. 73 (2001) 353563541. doi:10.1021/ac010022j.
- [11] D.A. Lake, M.P. Tolocka, M. V. Johnston, A.S. Wexler, Mass spectrometry of individual particles between 50 and 750 nm in diameter at the Baltimore supersite, *Environmental Science and Technology*. 37 (2003) 326863274. doi:10.1021/es026270u.
- [12] R.C. Moffet, L.G. Shields, J. Berntsen, R.B. Devlin, K.A. Prather, Characterization of an ambient coarse particle concentrator used for human exposure studies: Aerosol size distributions, chemical composition, and concentration enrichment, *Aerosol Science and Technology*. 38 (2004) 112361137. doi:10.1080/027868290890344.
- [13] D.A. Sodeman, S.M. Toner, K.A. Prather, Determination of single particle mass spectral signatures from light-duty vehicle emissions, *Environmental Science & Technology*. 39 (2005) 456964580. doi:10.1021/es0489947.
- [14] Y. Zhou, X.H.H. Huang, S.M. Griffith, M. Li, L. Li, Z. Zhou, C. Wu, J. Meng, C.K. Chan, P.K.K. Louie, J.Z. Yu, A field measurement based scaling approach for quantification of major ions, organic carbon, and elemental carbon using a single

particle aerosol mass spectrometer, *Atmospheric Environment*. 143 (2016) 3006312.
doi:10.1016/j.atmosenv.2016.08.054.

[15] P. V. Bhavé, J.O. Allen, B.D. Morrical, D.P. Fergenson, G.R. Cass, K.A. Prather, A
field-based approach for determining ATOFMS instrument sensitivities to ammonium
and nitrate, *Environmental Science and Technology*. 36 (2002) 486864879.
doi:10.1021/es015823i.

[16] E. Woods, G.D. Smith, Y. Dessiaterik, T. Baer, R.E. Miller, Quantitative detection of
aromatic compounds in single aerosol particle mass spectrometry, *Analytical
Chemistry*. 73 (2001) 231762322. doi:10.1021/ac001166l.

[17] D. Lee, K. Park, M.R. Zachariah, Determination of the Size Distribution of
Polydisperse Nanoparticles with Single-Particle Mass Spectrometry: The Role of Ion
Kinetic Energy, *Aerosol Science and Technology*. 39 (2005) 1626169.

[18] R.M. Healy, J. Sciare, L. Poulain, M. Crippa, A. Wiedensohler, A.S.H. Prévôt, U.
Baltensperger, R. Sarda-Estève, M.L. McGuire, C.H. Jeong, E. McGillicuddy, I.P.
O'Connor, J.R. Sodeau, G.J. Evans, J.C. Wenger, Quantitative determination of
carbonaceous particle mixing state in Paris using single-particle mass spectrometer and
aerosol mass spectrometer measurements, *Atmospheric Chemistry and Physics*. 13
(2013) 947969496. doi:10.5194/acp-13-9479-2013.

[19] C.H. Jeong, M.L. McGuire, K.J. Godri, J.G. Slowik, P.J.G. Rehbein, G.J. Evans,
Quantification of aerosol chemical composition using continuous single particle
measurements, *Atmospheric Chemistry and Physics*. 11 (2011) 702767044.
doi:10.5194/acp-11-7027-2011.

[20] D.C.. Beddows, R.. Donovan, R.. Harrison, M.. Heal, R. Kinnersley, M.. King, D..

Nicholson, M.-J. Schofield, K.. Thompson, Variations in the Chemical Composition of the Rural Background Atmospheric Aerosol Determined Real Time Using Time-of-Flight Mass Spectroscopy, *Journal of Environmental Monitoring*. 2 (2004) 1246133.

[21] S.R. Guazzotti, K.R. Coffee, K.A. Prather, Continuous measurements of size resolved particle chemistry during INDOEX-IFP 99, *Journal of Geophysical Research*. 106 (2001) 6076627.

[22] D.Y. Liu, R.J. Wenzel, K.A. Prather, Aerosol Time-of-Flight Mass Spectrometry during the Atlanta Supersite Experiment: 1. Measurements, *Journal of Geophysical Research Atmospheres*. 108 (2003).

[23] S.H. Pastor, J.O. Allen, L.S. Hughes, P. Bhawe, G.R. Cass, K.A. Prather, Ambient Single Particle Analysis in Riverside, California by Aerosol Time-of-Flight Mass Spectrometry during the SCOS97-NARSTO, *Atmospheric Environment*. 37 (2003) 2396258.

[24] X. Qin, P. V. Bhawe, K.A. Prather, Comparison of two methods for obtaining quantitative mass concentrations from aerosol time-of-flight mass spectrometry measurements, *Analytical Chemistry*. 78 (2006) 616966178. doi:10.1021/ac060395q.

[25] Y. Su, M.F. Sipin, H. Furutani, K.A. Prather, Development and Characterization of an Aerosol Time-of-Flight Mass Spectrometer with Increased Detection Efficiency, *Analytical Chemistry*. 76 (2004) 7126719. doi:10.1021/ac034797z.

[26] F. Gaie-Levrel, S. Perrier, E. Perraudin, C. Stoll, N. Grand, M. Schwell, Development and characterization of a single particle laser ablation mass spectrometer (SPLAM) for organic aerosol studies, *Atmospheric Measurement Techniques*. 5 (2012) 2256241. doi:10.5194/amt-5-225-2012.

- [27] J.T. Jayne, D.C. Leard, X.F. Zhang, P. Davidovits, K. a Smith, C.E. Kolb, D.R. Worsnop, Development of an aerosol mass spectrometer for size and composition analysis of submicron particles, *Aerosol Science and Technology*. 33 (2000) 49670. doi:10.1080/027868200410840.
- [28] P.F. Decarlo, J.R. Kimmel, A. Trimborn, M.J. Northway, J.T. Jayne, A.C. Aiken, M. Gonin, K. Fuhrer, T. Horvath, K.S. Docherty, D.R. Worsnop, J.L. Jimenez, Aerosol Mass Spectrometer, *Analytical Chemistry*. 78 (2006) 828168289. doi:8410.1029/2001JD001213.Analytical.
- [29] A. Petzold, M. Schönlinner, Multi-angle absorption photometry - A new method for the measurement of aerosol light absorption and atmospheric black carbon, *Journal of Aerosol Science*. 35 (2004) 4216441. doi:10.1016/j.jaerosci.2003.09.005.
- [30] P. DeCarlo, J. Slowik, K. Stainken, P. Davidovits, L.R. Williams, J.T. Jayne, C.E. Kolb, D. Worsnop, Y. Rudich, J. Jimenez, Particle Morphology and Density Characterization by Combined Mobility and Aerodynamic Diameter Measurements. Part 2: Application to Combustion-Generated Soot Aerosols as a Function of Fuel Equivalence Ratio, *Aerosol Science and Technology*. 38 (2004) 120661222. doi:10.1080/02786826.2004.10399462.
- [31] M. Pitz, J. Cyrys, E. Karg, A. Wiedensohler, H.E. Wichmann, J. Heinrich, Variability of apparent particle density of an urban aerosol, *Environmental Science and Technology*. 37 (2003) 433664342. doi:10.1021/es034322p.
- [32] R. McDonald, P. Biswas, *NIH Public Access*, 54 (2008) 106961078.
- [33] Y. Liu, P.H. Daum, *Corresponding Author, (1999) 1636.
- [34] A. Bharti, P.C. Ma, R. Salgia, Biomarker discovery in lung cancer--promises and

challenges of clinical proteomics., *Mass Spectrometry Reviews*. 26 (2007) 4516466.
doi:10.1002/mas.

[35] M.S. Reinard, M. V. Johnston, Ion Formation Mechanism in Laser Desorption
Ionization of Individual Nanoparticles, *Journal of the American Society for Mass
Spectrometry*. 19 (2008) 3896399. doi:10.1016/j.jasms.2007.11.017.

[36] M.T. Spencer, K. a. Prather, Using ATOFMS to Determine OC/EC Mass Fractions in
Particles, *Aerosol Science and Technology*. 40 (2006) 5856594.
doi:10.1080/02786820600729138.

[37] J.F. Cahill, K. Suski, J.H. Seinfeld, R.A. Zaveri, K.A. Prather, The mixing state of
carbonaceous aerosol particles in northern and southern California measured during
CARES and CalNex 2010, *Atmospheric Chemistry and Physics*. 12 (2012) 109896
11002. doi:10.5194/acp-12-10989-2012.

[38] P. DeCarlo, J. Kimmel, A. Trimborn, M. Northway, J. Jayne, A. Aiken, M. Gonin, K.
Fuhrer, T. Horvath, K. Docherty, D. Worsnop, J. Jimenez, Field-deployable, high-
resolution, time-of-flight aerosol mass spectrometer., *Analytical Chemistry*. 78 (2006)
828168289.

[39] D.S. Thomson, a M. Middlebrook, D.M. Murphy, Thresholds for laser-induced ion
formation from aerosols in a vacuum using ultraviolet and vacuum-ultraviolet laser
wavelengths, *Aerosol Science and Technology*. 26 (1997) 5446559.
doi:10.1080/02786829708965452.

[40] A.L. Van Wuijckhuijse, M.A. Stowers, W.A. Kleefsman, B.L.M. Van Baar, C.E.
Kientz, J.C.M. Marijnissen, Matrix-assisted laser desorption/ionisation aerosol time-of-
flight mass spectrometry for the analysis of bioaerosols: Development of a fast detector

for airborne biological pathogens, *Journal of Aerosol Science*. 36 (2005) 677-687.
doi:10.1016/j.jaerosci.2004.11.003.

[41] D.B. Kane, M. V. Johnston, Size and Composition Biases on the Detection of Individual Ultrafine Particles by Aerosol Mass Spectrometry, *Environmental Science & Technology*. 34 (2000) 4887-4893. doi:10.1021/es001323y.

[42] D.B. Kane, M. V Johnston, Enhancing the detection of sulfate particles for laser ablation aerosol mass spectrometry., *Analytical Chemistry*. 73 (2001) 5365-69. doi:10.1021/ac010469s.

[43] A.T. Ahern, R. Subramanian, G. Saliba, E.M. Lipsky, N.M. Donahue, R.C. Sullivan, Effect of secondary organic aerosol coating thickness on the real-time detection and characterization of biomass burning soot by two particle mass spectrometers, *Atmospheric Measurement Techniques Discussions*. 2 (2016) 1627. doi:10.5194/amt-2016-201.

[44] J.C. Corbin, B. Sierau, M. Gysel, M. Laborde, A. Keller, J. Kim, A. Petzold, T.B. Onasch, U. Lohmann, A.A. Mensah, Mass spectrometry of refractory black carbon particles from six sources: Carbon-cluster and oxygenated ions, *Atmospheric Chemistry and Physics*. 14 (2014) 2591-2603. doi:10.5194/acp-14-2591-2014.

[45] M. Millet, A. Sanusi, H. Wotham, Chemical composition of fogwater in an urban area: Strasbourg (France), *Environmental Pollution*. 3 (1996) 345-354.

[46] A. Sanusi, H. Wortham, M. Millet, P. Mirabel, Chemical composition of rain water and pollution origin in Eastern France., *Atmospheric Environment*. 30 (1996) 59-72.