Quantification of the gas phase methyl iodide using O_2^+ as the reagent ion in the PTR-ToF-MS technique

Badr R'Mili, Brice Temime-Roussel, Anne Monod, Henri Wortham, Rafal S. Strekowski*
Aix-Marseille Univ, CNRS, LCE, Marseille, France

Abstract

The charge-transfer-reaction between molecular oxygen ions (O_2^+) and methyl iodide (CH_3I) is studied to investigate if consistent environmental quantification of the gas phase CH_3I is possible without prior calibration. The neutral CH_3I molecule was chosen because this compound is of atmospheric chemistry and environmental importance in the field of nuclear power plant safety and nuclear energy. Molecular oxygen was used as a reagent ion source in a commercial Proton-Transfer-Reaction Time-of-Flight Mass Spectrometer (PTR-ToF-MS) to produce molecular oxygen ions (O_2^+). The use of O_2^+ ions allows for fast, sensitive and specific detection of gas phase CH_3I via an electron exchange reaction $O_2^+ + CH_3I \rightarrow CH_3I^+ + O_2$. The instrument response was linear in the 1-150 ppbv range and its sensitivity was humidity independent. The detection sensitivity of CH_3I normalized by the O_2^+ count rate of 10^6 cps was found to be $S = 22.6 \pm 0.3$ ncps/ppbv, independent of relative humidity. A typical O_2^+ primary ion signal was $(2.0 \pm 0.2) \times 10^6$ cps. The lowest measured CH_3I concentration was 0.23 ± 0.10 ppb. Error is $\pm \sigma$. The theoretical collision rate based on the dipole moment and molecular polarizability values is calculated using the Langevin collision rate (k_L) approximation, the average-dipole-orientation (ADO) theory and the capture rate coefficient (k_{CAP}) based on trajectory calculations. The experimental rate constant, k_{exp} , for the electron transfer reaction between O_2^+ ions and CH_3I is calculated to be $(1.72 \pm 0.22) \times 10^{-9} \text{ cm}^3\text{s}^{-1}$. Listed errors are $\pm \sigma$ and represent precision only. The experimentally determined value agrees very well with the theoretical collision rate values, $k_L = 1.24 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$, $k_{ADO} = 1.73 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$ and $k_{CAP} = 1.48 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$. The obtained results indicate that the PTR-MS technique is an excellent analytical method to quantify gas phase CH_3I .

Introduction

First observed by Lovelock and coworkers (1973) [1], methyl iodide is a natural photochemical source of iodine in the atmosphere [2]. As first suggested by Chameides and Davis (1980) [2], iodine and its oxides play an important role in the oxidizing capacity of the troposphere [3], aerosol formation and in the ozone depleting cycles in the troposphere [4,5] and stratosphere [6]. The atmospheric photochemistry of methyl iodide is also of interest in the nuclear industry safety field to better understand chemical processes responsible for the formation of different fission products if a major nuclear power plant accident type Three Mile Island (U.S.A.) [7], Chernobyl (Ukraine) [8] and Fukushima (Japan) [9] were to occur, again. It is known that iodine and methyl iodide are two of the more important fission products that are released from UO_2 fuel during a major nuclear power plant accident in pressurized water reactors due to their volatility and radiological health impact.[10] Furthermore, methyl iodide is a fission product that is difficult to retain by post-accident filtration systems [11] and is currently used in the nuclear industry field to test the organic iodine capture ability, ageing and performance of emergency charcoal filters.[12,13] Other important methyl iodide applications include agrarian control of insects, plant parasitic metabolites, soil-borne pathogens and weed seeds.[14]

To date, most gas phase environmental methyl iodide measurements are based on gas chromatographic (GC) separation equipped with a mass spectrometer (MS) or an electron capture detector (ECD). A more recent atmospheric gas phase methyl iodide measurement method is based on resonant fluorescence spectroscopy.[15] While GC-ECD (LD~10 ppt) and GC-MS (LD~100 ppt) methods are two very sensitive analytical techniques used to detect gas phase methyl iodide in various environment settings, they do not offer the time resolution and response needed to measure rapid flux or concentration changes in the gas phase. The newer resonant fluorescence technique used to detect gas phase methyl iodide in the atmosphere is fast, sensitive and selective but remains a research-grade instrument that requires advanced technical expertise and skills and may be difficult to apply in routine analyses. The chemical ionization mass spectrometric (CIMS)

technique has the potential for fast, sensitive, specific and real time methyl iodide measurements when rapid changes in the gas phase mixing ratios need to be known or monitored continuously.

The proton-transfer-reaction time-of-flight mass spectrometry (PTR-ToF-MS) is a type of CIMS instrument. The PTR-ToF-MS combines the concept of chemical ionization [16] with the flow-drift-tube technique.[17] This technique is often used for sensitive detection of volatile organic compounds (VOCs) in the atmosphere. The two more common CI reagents used in CIMS to detect atmospheric VOCs include hydronium ions (H_3O^+) and molecular oxygens ions (O_2^+). While H_3O^+ are good proton sources, these CI reagent ions can be applied only to gas-phase compounds with proton affinities (PA) higher than that of water, $\text{PA}(\text{H}_2\text{O}) = 691.0 \text{ kJ.mol}^{-1}$. [18,19] Since, the proton affinity of CH_3I , $\text{PA}(\text{CH}_3\text{I}) = 691.7 \text{ kJ.mol}^{-1}$ [18,19], is only slightly higher than that of water, H_3O^+ ions may not be the ideal reagent to use to detect gas phase methyl iodide. In theory, H_3O^+ ions may be used to ionize CH_3I , whilst this reaction is only slightly exothermic ($\Delta\text{PA} \sim 0.7 \text{ kJ.mol}^{-1}$). A recent study showed that H_3O^+ ions may still be successfully used to quantify methyl iodide in the atmosphere.[20] On the other hand, O_2^+ ions are often used as an alternative chemical ionization reagent in PTR-ToF-MS. Contrary to H_3O^+ , which is a soft ionizer, the O_2^+ ion is observed to be a more aggressive chemical ionization reagent. That is, the O_2^+ reagent ion ionization results in a mass spectrum not dissimilar to electron impact albeit less fragmented. That is, the more energetic O_2^+ ions produce characteristic mass spectral "fingerprint" that is less fragmented compared to electron impact and the mother ion is clearly identifiable and can be easily quantified. Contrary to the proton transfer reaction between the H_3O^+ ion and CH_3I molecule, the electron transfer reaction between O_2^+ ion ($\text{IE} = 12.0697 \pm 0.0002 \text{ eV}$) [19] and methyl iodide molecule ($\text{IE} = 9.54 \pm 0.02 \text{ eV}$) [19] is highly exothermic. As a result, the use of molecular oxygen as a CI reagent the PTR-ToF-MS instruments to detect CH_3I is proposed. It is shown that the proposed PTR-ToF-MS technique is an ideal tool for online analyses of relatively fast changing concentrations of methyl iodide

Experimental

The charge transfer reaction between the O_2^+ reagent ion and CH_3I was studied using a commercial PTR-ToF 8000 mass analyzer (Ionicon Analytik GmbH, Innsbruck, Austria). The experimental details that are particularly relevant to this work are given below.

Generation of gas phase CH_3I . The gas phase CH_3I concentration was generated using the gas saturation method, one of the oldest and most versatile ways of studying heterogeneous equilibria involving low vapor pressure compounds, first developed by Regnault in 1845.[21] The gas saturation method used in this work is similar to the one described originally by Markham (1940).[22] Briefly, nitrogen carrier gas was allowed to flow through the volume containing the CH_3I sample that itself was mixed with glass beads and supported on a fritted glass surface. The saturator volume itself was immersed in a temperature-controlled fluid and kept at a constant temperature using a temperature-controlled circulating bath with an accuracy of $\pm 0.1 \text{ K}$. The temperature inside the saturator volume was measured using a Type-J thermocouple (Omega) with an accuracy of $\pm 0.1 \text{ K}$. The carrier gas was allowed to enter the saturator volume, equilibrate with the sample and was then allowed to exit through a capillary exit and allow to flow through a glass tube. The geometry of the exiting glass tube was such that the diameter of the glass tube increased with increasing length. This was done to avoid any sample condensation as the sample and the carrier gas were allowed to leave the saturator system. Concentration of CH_3I at the exit of the saturation system was calculated from the given vapor pressure, mass flow rates, pressure within the saturator and the total pressure. Concentration of CH_3I at the exit of the saturation system was calculated using the Dalton's law of partial pressures. Assuming that the gas within the saturation system is a mixture of dry nitrogen and CH_3I , the volume mixing ratio of the latter can be defined as the following:

$$[\text{CH}_3\text{I}] = \frac{p_{\text{CH}_3\text{I}}}{p_s}$$

where $P_{\text{CH}_3\text{I}}$ is the vapor pressure of CH_3I and P_s , the total pressure within the saturation system, was $P_s = 957 \text{ Torr}$. The Antoine type equation used to calculate the vapor pressure of CH_3I is the following [23]:

$$\log_{10}(p) = 4.1554 - \frac{1177.78}{T(\text{K}) - 32.058}$$

where the pressure p is in units of bar and the temperature T is in Kelvin and in the range from 218K to 315.6.[23] Under normal operating conditions, the saturator was kept at $T = 274 \text{ K}$. At this temperature, the

vapor pressure of CH₃I within the saturator was calculated to be 146.37±3.40 Torr. Under typical experimental conditions, the gas phase CH₃I mixing ratio within the saturation system was about 15%. This vapor pressure was then further diluted using dry nitrogen carrier gas and a system of mass flow controllers to obtain the desired mixing ratio in the range 1-150 ppbv. Further, the carrier gas was allowed to pass through a bubbler filled with deionizer water at room temperature to evaluate the impact of relative humidity (%RH) on the signal sensitivity. The relative humidity was varied between 0.3±0.2 and 76±0.5%.

PTR-ToF-MS instrument. A commercial PTR-ToF 8000 mass analyzer (Ionicon Analytik GmbH, Innsbruck, Austria) was used to study the O₂⁺ reagent ion ionization process with CH₃I. The reaction chamber pressure (pdrift) was 2.10 mbar, drift tube voltage was 544 V and the drift tube temperature was held constant at T=333.15K. The corresponding *E/N* ratio was 127 Td. (1Td=10⁻¹⁷ Vcm²) where *E* is the electric field strength applied to the drift tube and *N* is the buffer gas density.

Materials. The nitrogen carrier gases used in this study was generated using the N2LCMS 1 Nitrogen Generator (Claind S.r.l., Italy). Iodomethane was purchased at Acros Organics (Belgium) and the stated minimum purity was 99%. To limit any photo-catalytic or thermal decomposition during storage, CH₃I original container bottles were stored under dark conditions at T=6°C. Oxygen gas was purchased at Linde Gas (Linde France S.A.) and had the stated minimum purity of 99.9995%. Deionized water with a resistivity greater than 18MΩ was prepared by allowing tap water to pass first through a reverse osmosis demineralization filter (ATS Groupe Osmose) and then through a commercial deionizer (Milli-pore, Milli-Q).

Results

Chemical ionization scheme for CH₃I

The electron transfer reaction between the O₂⁺ ion (IE=12.0697±0.0002) [19] and methyl iodide molecule (IE=9.54±0.02) [19] is highly exothermic. As a result, molecular oxygen appears to be an ideal reagent source to be used in the PTR-ToF-MS technique to quantify gas phase methyl iodide. Based on other works [24–26] where molecular oxygen was used as a chemical ionization reagent to detect gas phase compounds of atmospheric interest, the electron exchange reaction between the O₂⁺ ion and CH₃I may proceed as follows:

Although the O₂⁺ + CH₃I electron exchange reaction is highly exothermic, the reaction may result in the formation of other dissociation products in addition to the mother ion as observed by Španěl and coworkers [25,26] in their selected ion flow tube mass spectrometry (SIFT-MS) study of aldehydes and ketones [26]. The formation of partially dissociative products is discussed below.

Experimental determination of *k*_(O₂⁺ + CH₃I)

Based on the original work of Lindinger and coworkers (1998) [27] the gas phase concentration of CH₃I may be calculated using the following equation (1):

$$[\text{CH}_3\text{I}]_{\text{cm}^{-3}} = \frac{1}{k \times t} \times \frac{I(\text{CH}_3\text{I}^+)}{I(\text{O}_2^+)} \times \frac{\text{Tr}_{\text{O}_2^+}}{\text{Tr}_{\text{CH}_3\text{I}^+}} \quad (1)$$

In the equation above, [CH₃I⁺] is the methyl iodide concentration in units of cm⁻³, I(CH₃I⁺) and I(O₂⁺) are the observed ion count rates, *k* is the charge transfer rate constant between the molecular oxygen ion (cm³.s⁻¹), O₂⁺, and the methyl iodide molecule, and *t* (μs) is the residence or reaction time of the selected ion within the drift tube.[28] As shown in the equation (1) above, to correct mass dependence losses of the selected ions that are allowed to travel through the transfer lens system between the drift tube and the ToF mass analyzer, the detected ion count rates are normalized by their corresponding transmission factors, that is, Tr_{O₂⁺} and Tr_{CH₃I⁺}.

The equation (1) above may be rewritten in the following manner:

$$[\text{CH}_3\text{I}]_{\text{ppbv}} = \frac{10^9}{N} \times \frac{1}{k \times t} \times \frac{I(\text{CH}_3\text{I}^+)}{I(\text{O}_2^+)} \times \frac{\text{Tr}_{\text{O}_2^+}}{\text{Tr}_{\text{CH}_3\text{I}^+}} \quad (2)$$

142 In the equation (2) above, $[\text{CH}_3\text{I}]_{\text{ppbv}}$ is the methyl iodide mixing ratio in units of ppbv, N is the number of
 143 molecules per cm^3 within the drift-tube volume. The reaction time, t , may be calculated using the drift tube
 144 length ($L = 9.2 \text{ cm}$) and the O_2^+ ion drift velocity (v_d).

$$145 \quad t = \frac{L}{v_d} \quad (3)$$

146 In the equation (3), the drift tube velocity (v_d) may be determined from the reduced mobility of the O_2^+ ion (μ_0
 147 $= 2.5 \text{ cm}^2\text{V}^{-1}\text{s}^{-1}$) at given E/N ratio (equation (4)).

$$148 \quad v_d = \mu \cdot E = N_0 \cdot \mu_0 \cdot \left(\frac{E}{N}\right) \quad (4)$$

149 In equation (4), $N_0 = 2.68 \times 10^{19} \text{ cm}^{-3}$ and is the number of molecules per cm^3 under normal conditions of
 150 temperature and pressure.

151 If the methyl iodide mixing ratio is known, the experimental rate constant, k_{exp} , for the charge transfer reaction
 152 between the O_2^+ ion and CH_3I molecule may be calculated using the following equation (5):

$$153 \quad k_{\text{exp}} = \frac{10^9}{t \times N} \times \frac{I(\text{CH}_3\text{I}^+)}{[\text{CH}_3\text{I}]_{\text{ppbv}} \times I(\text{O}_2^+)} \times \frac{\text{Tr}_{\text{O}_2^+}}{\text{Tr}_{\text{CH}_3\text{I}^+}} \quad (5)$$

154 A typical calibration curve is shown in Figure 1. The MS signal is observed to be linear for the given methyl
 155 iodide mixing ratio.

156

157 **Figure 1. Calibration curve of the PTR-ToF-MS system used to quantify the gas phase CH_3I molecules using O_2^+ as reagent ions.**
 158 **The CH_3I signal is normalized to 1×10^6 cps of O_2^+ ions. Solid line is obtained from linear least-squares analysis and gives the**
 159 **slope that is signal sensitivity of 22.6 ± 0.3 ncps/ppbv. The lowest measured CH_3I concentration was 0.23 ± 0.10 ppb. Error is $\pm \sigma$.**
 160 **Error in the detection limit is $\pm \sigma$. Coefficient of determination (r^2) is calculated to be 0.993. The horizontal bars ($\pm \sigma$) represent**
 161 **experimental uncertainties in the generation of the gas phase methyl iodide using the saturation method followed by**
 162 **subsequent dilutions. The experimentally determined vertical error bars are $\pm 2\%$ of the measured counts.**

163 As shown in Figure 1, the measured methyl iodide PTR-ToF MS signal in counts per second (cps) was
 164 normalized using primary ion signal due to O_2^+ ions. The methyl iodide signal was normalized to aid in
 165 intercomparing of different PTR-MS techniques and to remove measured signal fluctuations in the amount of
 166 primary reagent ions and drift tube temperature and pressure.

167 As shown in Figure 1, methyl iodide signal sensitivity is obtained from linear least-squares analysis and gives
 168 $S = 22.6 \pm 0.3$ ncps/ppbv (normalized to 1×10^6 cps of O_2^+). If the solid line is forced through zero, linear
 169 least-squares analysis results in $S = 23.6 \pm 0.3$ ncps/ppbv (normalized to 1×10^6 cps of O_2^+). The lowest
 170 measured CH_3I concentration was 0.23 ± 0.10 ppb. Errors are $\pm \sigma$ and represent precision only. The count rate
 171 of O_2^+ , calculated from the count rate at m/z 34 ($^{16}\text{O}^{18}\text{O}^+$) multiplied by 250, was typically $(2.0 \pm 0.2) \times 10^6$
 172 cps. Under dry conditions, $\text{RH} = 0\%$, the hydronium ion (H_3O^+) signal is very weak (1 ± 0.2 ppm to that of
 173 O_2^+). The $\text{Tr}_{\text{O}_2^+}$ and $\text{Tr}_{\text{CH}_3\text{I}^+}$ values were measured to be 0.36 and 0.94, respectively.

174 The k_{exp} value may also be directly determined using the normalized detection sensitivity, S , as shown in the
 175 equation (6).

$$176 \quad k_{\text{exp}} = \frac{10^3}{t \times N} \times S \times \frac{\text{Tr}_{\text{O}_2^+}}{\text{Tr}_{\text{CH}_3\text{I}^+}} \quad (6)$$

177 The resulting experimental rate constant for the charge transfer reaction between the O_2^+ ion and the CH_3I
 178 neutral molecule is calculated to be $(1.72 \pm 0.22) \times 10^{-9} \text{ cm}^3\text{s}^{-1}$.

179 Theoretical determination of $k_{(\text{O}_2^+ + \text{CH}_3\text{I})}$

180 Given a large exothermicity of the electron exchange reaction between the molecular oxygen ion (O_2^+) and the
 181 neutral methyl iodide molecule (CH_3I), ion-neutral collision theories may be used to calculate the theoretical
 182 collision rate constants. Current theoretical approximations of the collision rate constants between a reagent
 183 ion and a polar neutral molecule that take place within the drift tube of a PTR-MS instrument include the
 184 Langevin collision rate k_L [29] approximation, the average-dipole-orientation (ADO) theory [30,31] and the
 185 capture rate coefficient k_{CAP} [32–34] trajectory calculation.

186 The Langevin rate constant k_L gives a first good approximation for the collision rate between a reagent ion and
 187 a non-polar molecule in the high temperature limit.[28] The following equation is the Langevin collision rate
 188 approximation [29] used to calculate k_L for the electron exchange reaction between the O_2^+ reagent ion and the
 189 polar neutral CH_3I molecule:

$$190 \quad k_L = q \sqrt{\frac{\pi\alpha}{\mu_m \epsilon_0}} \quad (7)$$

191 where α is the polarizability of the neutral molecule, q is the fundamental charge, ϵ_0 is the permittivity of the
 192 free space, μ_m is the reduced mass of the colliding partners (here, O_2^+ and CH_3I).

193 A better approximation of the rate constant between the O_2^+ reagent ion and the polar CH_3I neutral molecule
 194 is obtained using the average-dipole-orientation (ADO) theory.[30,31] That is, it has been observed that the
 195 ADO theoretical calculations of the ion-neutral molecule reaction rates are typically within 10 to 20% of the
 196 experimentally determined values.[28,30,35] The following equation was used to calculate k_{ADO} [28]:

$$197 \quad k_{\text{ADO}} = q \sqrt{\frac{\pi\alpha}{\mu_m \epsilon_0}} + C \frac{q\mu_D}{\epsilon_0} \sqrt{\frac{1}{2\pi\mu_m k_B T}} \quad (8)$$

198 where μ_D is the dipole moment, T is the temperature within the drift tube and k_B is the Boltzmann constant.
 199 Also, C is the dipole locking constant with values that range between 0 and 1 and is a function of the $\mu_D/\sqrt{\alpha}$
 200 ratio and the temperature.[36]

201 The theory most widely used to estimate the ion-molecule rate constants is based on the extensive trajectory
 202 calculations carried out by Su, Chesnavich and Bowers.[32] In their work on ion-polar molecule collision rate
 203 constants, these investigators [32,37,38] showed that the capture rate constants k_{CAP} can be parametrized using
 204 the polarizability α , permanent dipole moment μ_D , and the center-of-mass kinetic energy KE_{cm} in the following
 205 manner [39]:

$$206 \quad k_{\text{CAP}}(T, \text{KE}_{\text{cm}}) = k_L K_{\text{CAP}}(\tau, \epsilon) \quad (9a)$$

$$208 \quad \tau = \frac{\mu_D}{\sqrt{\alpha T}} \quad (9b)$$

$$209 \quad \epsilon = \frac{\mu_D}{\sqrt{\alpha \cdot \text{KE}_{\text{cm}}}} \quad (9c)$$

$$210 \quad \text{KE}_{\text{cm}} = \frac{3}{2} \cdot k_B \cdot T \cdot \left(\frac{M_R}{M_R + M_{\text{ion}}} \right) \cdot \frac{1}{2} \cdot (M_{\text{ion}} + M_{\text{buffer}}) \cdot v_d^2 \quad (9d)$$

$$211 \quad K_{\text{CAP}}(\tau, \epsilon) = 1 + c_1 \cdot \tau^{0.4} \cdot \epsilon^2 \cdot S + c_2 \cdot (1 - S) \cdot \sin[c_3 \{c_4 + \ln(\tau)\}] \cdot \tau^{0.6} \cdot \sqrt{\epsilon - 0.5} \quad (9e)$$

$$212 \quad c_1 = 0.727143, c_2 = 3.71823, c_3 = 0.586920, c_4 = 4.97894 \quad (9f)$$

$$S = e^{[-2(\epsilon-1.5)]} \text{ if } \epsilon > 1.5 \text{ or } S = 1 \text{ if } \epsilon \leq 1.5 \quad (9g)$$

where M_{buffer} is the mass of the buffer gas (here, nitrogen + oxygen), M_{ion} is the mass of the reagent ion (here, O_2^+), M_{R} is the mass of the neutral molecule (here, CH_3I) and $v_d=930 \text{ ms}^{-1}$ and is the drift velocity of the ion. The structure of methyl iodide was taken from the NIST database [19]. In this work, the dipole moment $\mu_D=1.62$ Debye and polarizability $\alpha=7.325 \text{ \AA}^3$. [19] The theoretical collision rate constants for the charge transfer reaction between the O_2^+ ion and the CH_3I molecule are calculated to be $k_L = 1.24 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$, $k_{\text{ADO}} = 1.73 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$ and $k_{\text{CAP}} = 1.48 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$.

Discussion

As stated above, the electron transfer reaction between O_2^+ ions and methyl iodide molecule is highly exothermic. While the $\text{O}_2^+ + \text{CH}_3\text{I}$ reaction proceeds via a charge transfer, the reaction may result in the formation of other dissociation products in addition to the mother ion. [25,26] Such formation of partially dissociative products was observed by Španěl and coworkers [25,26] in their selected ion flow tube mass spectrometry (SIFT-MS) reaction studies of O_2^+ ions with a series of aldehydes and ketones [26]. However, in a similar study of halogenated compounds, Španěl and coworkers observed that the reaction between the O_2^+ ion and a selected short chain chloroalkanes proceeds mainly through a charge transfer, producing a single ionic product.[40] Relative abundancies of major ion signals observed when methyl iodide mix was analyzed by the PTR-ToF-MS instrument operated with O_2 gas as the reagent ion source are shown in Figure 2 and Figure 3. The ions shown in Figure 2 and Figure 3 were not present in the mass spectra when nitrogen buffer gas was allowed to flow to the PTR-MS and no methyl iodide was present in the system. Relative abundancies of major ion signals observed when methyl iodide was absent in the gas flow are shown in Figure 4 and Figure 5. It should be noted that the relative abundancy profiles of major ions shown in Figure 4 and Figure 5 are independent of the presence of CH_3I in the system. That is, same mass spectrum profiles and intensities were observed whether CH_3I was present in the gas flow or not.

236

237 **Figure 2.** Plot of the signal ion abundance as a function of relative humidity. Experimental conditions: Drift T=25°C; E/N = 126 Td

Figure 3. Plot of the signal ion abundance as a function of relative humidity. Experimental conditions: Drift T=25°C; E/N = 139 Td

Figure 4. Plot of the signal ion abundance as a function of relative humidity. Experimental conditions: Drift T=25°C; E/N = 126 Td

Figure 5. Plot of the signal ion abundance as a function of relative humidity. Experimental conditions: Drift T=25°C; E/N = 139 Td

As shown in Figure 2 to 5, the relative abundances of H_3O^+ (m/Q 19.01784), NO^+ (m/Q 29.99744), NO_2^+ (m/Q 45.99235) and $\text{H}_3\text{O}^+(\text{H}_2\text{O})$ (m/Q 37.02841) were calculated relative to the molecular oxygen ion (O_2^+) signal intensity only. On the other hand, the CH_5O^+ (m/Q 33.03349), I^+ (m/Q 126.90392), HI^+ (m/Q 127.91175), IO^+ (m/Q 142.89884), CH_4I^+ (m/Q 142.93522) and HIO^+ (m/Q 143.90666) ions were calculated relative to the sum of the methyl iodide mother ion (CH_3I^+) (m/Q 141.92740) signal intensity and all the other fragment ions, that is, CH_3O^+ , CH_5O^+ , I^+ , HI^+ , IO^+ , CH_4I^+ and HIO^+ . The H_3O^+ (m/Q 19.01784) and O_2^+ (m/Q 31.98928) ions were not counted directly. That is, to limit saturation of the electron multiplier detector of the ToF-MS system, the count rate of O_2^+ was calculated from the count rate at m/Q 33.99353 ($^{16}\text{O}^{18}\text{O}^+$) multiplied by 250. In a similar manner, the count rate of H_3O^+ was calculated from the count rate at m/Q 21.02209 ($\text{H}_3^{18}\text{O}^+$) multiplied by 500. Further, the count rate of NO^+ was calculated from the count rates at m/Q 30.99447 ($^{15}\text{N}^{16}\text{O}^+$) and m/Q 31.00166 ($^{14}\text{N}^{17}\text{O}^+$) multiplied by 250. All signal intensities were corrected for transmission.

Similar to the work of Španěl and coworkers [40] on the reaction between the O_2^+ ion and selected short chain chloroalkanes, we observe in majority a single ionic product that is CH_3I^+ . As shown in Figures 2 and 3, there is no clear dependence of the CH_3I^+ molecular ion signal on relative humidity. However, as shown in Figures 2 to 5, the electron transfer reaction between O_2^+ ions and the CH_3I neutral molecule results in the formation of other dissociation/association and ionization products in addition to the CH_3I^+ mother ion. Other dissociation and ionization products include I^+ , IO^+ , HOI^+ , CH_4I^+ , CH_3O^+ , CH_5O^+ , H_3O^+ , $\text{H}_3\text{O}^+\cdot\text{H}_2\text{O}$ complex, NO^+ and NO_2^+ ions. However, only traces (<2%) of CH_4I^+ , HIO^+ , I^+ , HI^+ , IO^+ and CH_3O^+ ions with no clear dependence on relative humidity are observed. The observed NO^+ (m/Q 29.99744) and NO_2^+ (m/Q 45.99235) ion signals are not believed to be a result of methyl iodide ionization reaction with O_2^+ reagent ions. In a similar manner, the observed NO^+ and NO_2^+ ion signal counts are not believed to be a result of the electron exchange between the O_2^+ reagent ions and background NO_x ($\text{NO}_x = \text{NO} + \text{NO}_2$) impurities that may be present in the nitrogen carrier gas. For example, a 1% of NO^+ impurity within the carrier gas would correspond to a signal count rate of 10000 cps or 500 ppbv of NO (assuming 20 cps/ppb), an unrealistic value even in extremely polluted environments. On the other hand, it is known that NO^+ and NO_2^+ ions are produced within the ionization source of the ToF mass analyzer. That is, the nitrogen carrier gas and molecular oxygen are ionized within the ionization source as a result of a “backflush” to produce atomic N^+ and O^+ ions. These ions, in turn, recombine to produce NO^+ and NO_2^+ ions. The NO^+ (m/Q 29.99744) and NO_2^+ (m/Q 45.99235) ion signal abundances do not show any humidity or E/N value dependence. It is possible that the observed H_3O^+ (m/Q 19.01784) and $\text{H}_3\text{O}^+(\text{H}_2\text{O})$ (m/Q 37.02841) complex ion signals are a result of ionization of H_2O by O_2^+ reagent ions within the drift tube. However, in a similar situation listed above for the observed NO^+ and NO_2^+ ion

signals, the H_3O^+ and $\text{H}_3\text{O}^+(\text{H}_2\text{O})$ ion signals are most likely a result of humidity “backflushing” into the ionization source of the ToF mass analyzer to form the observed H_3O^+ (m/Q 19.01784) and $\text{H}_3\text{O}^+(\text{H}_2\text{O})$ (m/Q 37.02841) ion signals. As seen in Figures 4 and 5, the H_3O^+ (m/Q 19) and $\text{H}_3\text{O}^+(\text{H}_2\text{O})$ (m/Q 37.02841) complex ion signals increase in abundance as the relative humidity is increased while other minor ion signals, namely CH_4I^+ , HIO^+ , I^+ , HI^+ , IO^+ , do not show any humidity dependence. However, as shown in Figures 2 and 3, a slight dependence on humidity is observed of CH_5O^+ and CH_3O^+ ions. That is, it is observed that while CH_5O^+ ion signal count increases with increasing humidity, the CH_3O^+ ion signal count decreases slightly with increasing humidity. It should be noted that methanol is known to produce CH_3O^+ ions when the PTR-MS is operated in the O_2^+ ionization mode. In turn, CH_3O^+ ions may react with H_3O^+ reagent ions to produce CH_5O^+ ions. However, at this point we cannot explain how (and if) methanol is produced in our system. As a result, at this point we cannot explain the effect of the observed behavior of the CH_5O^+ and CH_3O^+ ion signal abundance on humidity.

290

291 Conclusion

292 A commercial Ionicon Analytic PTR-MS was used to study the O_2^+ reagent ion ionization process to quantify
 293 gas phase methyl iodide. The electron exchange reaction between the O_2^+ reagent ion and the CH_3I molecule
 294 produces a specific ion at m/Q 142 (CH_3I^+) that allows for fast and sensitive detection and quantification of
 295 gas phase methyl iodide. The instrument response for CH_3I was shown to be linear in the 1 - 150 ppbV range.
 296 The lowest measured CH_3I concentration was 0.23 ± 0.10 ppb. Error is $\pm \sigma$. However, this value does not
 297 represent the detection limit for gas phase CH_3I . The theoretical collision rate constants are calculated based
 298 on the dipole moment $\mu(\text{CH}_3\text{I}) = 1.620$ D [19] and the molecular polarizability $\alpha(\text{CH}_3\text{I}) = 7.325 \text{ \AA}^3$ [19] and
 299 give $k_L = 1.24 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$, $k_{\text{ADO}} = 1.73 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$ and $k_{\text{CAP}} = 1.48 \times 10^{-9} \text{ cm}^3\text{s}^{-1}$. The experimental
 300 collision rate, k_{exp} for the electron exchange reaction $\text{O}_2^+ + \text{CH}_3\text{I}$ is measured to be $(1.72 \pm 0.22) \times$
 301 $10^{-9} \text{ cm}^3\text{s}^{-1}$. Error is $\pm \sigma$ and represents precision only. The two values agree very well with one another.
 302 A good agreement between the experimentally determined rate constant for the electron exchange reaction
 303 between the molecular oxygen ion and the neutral methyl iodide molecule and the theoretical values indicates
 304 that the PTR-MS technique is an excellent analytical method for fast and continuous quantification of gas
 305 phase CH_3I without prior calibration.

306

307 Acknowledgement

308 This research was supported by the Mitigation of Releases to the Environment (MiRE) project launched by
 309 the *Institut de Radioprotection et de Sûreté Nucléaire* (IRSN) and funded by the French National Research
 310 Agency (ANR) under the convention number 11-RSNR-0013. We gratefully acknowledge this support. The
 311 technological platform MASSALYA of the Aix- Marseille University is acknowledged for providing the
 312 PTR-ToF-MS instrument.

313

314 References

- 315 [1] J.E. Lovelock, R.J. Maggs, R.J. Wade, Halogenated Hydrocarbons in and over the Atlantic, *Nature*. 241
 316 (1973) 194–196. doi:10.1038/241194a0.
- 317 [2] W.L. Chameides, D.D. Davis, Iodine: Its possible role in tropospheric photochemistry, *J. Geophys. Res.*
 318 *Oceans*. 85 (1980) 7383–7398. doi:10.1029/JC085iC12p07383.
- 319 [3] W. Bloss, J. Lee, G. Johnson, R. Sommariva, D. Heard, A. Saiz-Lopez, J. Plane, G. McFiggans, H. Coe, M.
 320 Flynn, others, Impact of halogen monoxide chemistry upon boundary layer OH and HO2
 321 concentrations at a coastal site, *Geophys. Res. Lett.* 32 (2005) L06814. doi:10.1029/2004GL022084.
- 322 [4] D. Davis, J. Crawford, S. Liu, S. McKeen, A. Bandy, D. Thornton, Potential impact of iodine on
 323 tropospheric levels of ozone, *J. Geophys. Res.* 101 (1996) 2135–2147.
- 324 [5] G. McFiggans, J. Plane, B.J. Allan, L.J. Carpenter, H. Coe, C. O'Dowd, A modeling study of iodine
 325 chemistry in the marine boundary layer, *J. Geophys. Res. Atmospheres*. 105 (2000) 14371–14385.
- 326 [6] S. Solomon, R.R. Garcia, A. Ravishankara, On the role of iodine in ozone depletion, *J. Geophys. Res.*
 327 *Atmospheres*. 99 (1994) 20491–20499.

- [7] J. Cline, P. Roy, J. Hollcroft, J. Houbaugh Jr, T. McVey, C. Thomas Jr, C. Pelletier, P. Voilleque, Measurements of ^{129}I and radioactive particulate concentrations in the TMI-2 containment atmosphere during and after the venting, Science Applications, 1981.
- [8] H. Noguchi, M. Murata, Physicochemical speciation of airborne ^{131}I in Japan from Chernobyl, J. Environ. Radioact. 7 (1988) 65–74.
- [9] N. Kinoshita, K. Sueki, K. Sasa, J. Kitagawa, S. Ikarashi, T. Nishimura, Y.-S. Wong, Y. Satou, K. Handa, T. Takahashi, others, Assessment of individual radionuclide distributions from the Fukushima nuclear accident covering central-east Japan, Proc. Natl. Acad. Sci. 108 (2011) 19526–19529.
- [10] J. Hala, J.D. Navratil, Radioactivity, Ionizing Radiation, and Nuclear Energy, 1st ed., Konvoj, Brno, 2003.
- [11] J.M. Ball, G.A. Glowa, D. Boulianne, J.R. Mitchell, Behaviour of Iodine Project: Final Report on Organic Iodide Studies, Nucl. Energy Agency At. Energy Can. Ltd. Chalk River Ont. Can. (2011). <http://scholar.google.com/scholar?cluster=1001493427184488016&hl=en&oi=scholar> (accessed March 20, 2017).
- [12] D. Nacapricha, C.G. Taylor, Quality control of nuclear charcoals: Particle size effect and trapping mechanism, Carbon. 34 (1996) 155–163.
- [13] J.C. Wren, C.J. Moore, M.T. Rasmussen, K.R. Weaver, Methyl iodide trapping efficiency of aged charcoal samples from Bruce-A emergency filtered air discharge systems, Nucl. Technol. 125 (1999) 28–39.
- [14] M. Waggoner, H. Ohr, C. Adams, J. Sims, D. Gonzalez, Methyl iodide: an alternative to methyl bromide for insectary fumigation, J. Appl. Entomol. 124 (2000) 113–117.
- [15] C.S. Bale, T. Ingham, R. Commane, D.E. Heard, W.J. Bloss, Novel measurements of atmospheric iodine species by resonance fluorescence, J. Atmospheric Chem. 60 (2008) 51–70.
- [16] F. Field, N. Munson, Chemical ionization mass spectrometry. I. General information, J Amer Chem Soc. 88 (1966) 2621.
- [17] M. McFarland, D. Albritton, F. Fehsenfeld, E. Ferguson, A. Schmeltekopf, Flow-drift technique for ion mobility and ion-molecule reaction rate constant measurements. II. Positive ion reactions of N^+ , O^+ , and H_2^+ with O_2 and O^+ with N_2 from thermal to [inverted lazy s] 2 eV, J. Chem. Phys. 59 (1973) 6620–6628.
- [18] E.P. Hunter, S.G. Lias, Evaluated gas phase basicities and proton affinities of molecules: an update, J. Phys. Chem. Ref. Data. 27 (1998) 413–656.
- [19] WebBook de Chimie NIST, (n.d.). <http://webbook.nist.gov/chemistry/>.
- [20] M. Mekić, B. Temime-Roussel, A. Monod, R.S. Strekowski, Quantification of gas phase methyl iodide using H_3O^+ as the reagent ion in the PTR-MS technique, Int. J. Mass Spectrom. 424 (2018) 10–15. doi:10.1016/j.ijms.2017.11.004.
- [21] M.V. Regnault, Études sur l'hygrométrie, Ann. Chim. Phys. Séries 3, Tome 15 (1845) 129–136.
- [22] A.E. Markham, A Compact Gas Saturator, Ind. Eng. Chem. Anal. Ed. 12 (1940) 112–112. doi:10.1021/ac50142a020.
- [23] D.R. Stull, Vapor Pressure of Pure Substances. Organic and Inorganic Compounds, Ind. Eng. Chem. 39 (1947) 517–540. doi:10.1021/ie50448a022.
- [24] M. Norman, A. Hansel, A. Wisthaler, O_2^+ as reagent ion in the PTR-MS instrument: Detection of gas-phase ammonia, Int. J. Mass Spectrom. 265 (2007) 382–387. doi:10.1016/j.ijms.2007.06.010.
- [25] D. Smith, P. Španěl, Selected ion flow tube mass spectrometry (SIFT-MS) for on-line trace gas analysis, Mass Spectrom. Rev. 24 (2005) 661–700. doi:10.1002/mas.20033.
- [26] P. Španěl, Y. Ji, D. Smith, SIFT studies of the reactions of H_3O^+ , NO^+ and O_2^+ with a series of aldehydes and ketones, Int. J. Mass Spectrom. Ion Process. 165–166 (1997) 25–37. doi:https://doi.org/10.1016/S0168-1176(97)00166-3.
- [27] W. Lindinger, A. Hansel, A. Jordan, On-line monitoring of volatile organic compounds at pptv levels by means of proton-transfer-reaction mass spectrometry (PTR-MS) medical applications, food control and environmental research, Int. J. Mass Spectrom. Ion Process. 173 (1998) 191–241.
- [28] L. Cappellin, M. Probst, J. Limtrakul, F. Biasioli, E. Schuhfried, C. Soukoulis, T.D. Märk, F. Gasperi, Proton transfer reaction rate coefficients between H_3O^+ and some sulphur compounds, Int. J. Mass Spectrom. 295 (2010) 43–48. doi:https://doi.org/10.1016/j.ijms.2010.06.023.
- [29] P. Langevin, A fundamental formula of kinetic theory, Ann Chim Phys. 5 (1905) 245–288.

- 381 [30] T. Su, M.T. Bowers, Ion-Polar molecule collisions: the effect of ion size on ion-polar molecule rate
382 constants; the parameterization of the average-dipole-orientation theory, *Int. J. Mass Spectrom. Ion*
383 *Phys.* 12 (1973) 347–356. doi:10.1016/0020-7381(73)80104-4.
- 384 [31] T. Su, M.T. Bowers, Theory of ion-polar molecule collisions. Comparison with experimental charge
385 transfer reactions of rare gas ions to geometric isomers of difluorobenzene and dichloroethylene, *J.*
386 *Chem. Phys.* 58 (1973) 3027–3037. doi:10.1063/1.1679615.
- 387 [32] T. Su, W.J. Chesnavich, Parametrization of the ion–polar molecule collision rate constant by trajectory
388 calculations, *J. Chem. Phys.* 76 (1982) 5183–5185. doi:10.1063/1.442828.
- 389 [33] T. Su, Trajectory calculations of ion–polar molecule capture rate constants at low temperatures, *J.*
390 *Chem. Phys.* 88 (1988) 4102–4103. doi:10.1063/1.453817.
- 391 [34] T. Su, Erratum: Trajectory calculations of ion-polar molecule capture rate constants at low
392 temperatures [*J. Chem. Phys.* 88, 4102 (1988)], *J. Chem. Phys.* 89 (1988) 5355–5355.
393 doi:10.1063/1.455750.
- 394 [35] G.I. Mackay, L.D. Betowski, J.D. Payzant, H.I. Schiff, D.K. Bohme, Rate constants at 297.degree.K for
395 proton-transfer reactions with hydrocyanic acid and acetonitrile. Comparisons with classical theories
396 and exothermicity, *J. Phys. Chem.* 80 (1976) 2919–2922. doi:10.1021/j100567a019.
- 397 [36] T. Su, M.T. Bowers, Parameterization of the average dipole orientation theory: temperature
398 dependence, *Int. J. Mass Spectrom. Ion Phys.* 17 (1975) 211–212. doi:10.1016/0020-7381(75)85046-7.
- 399 [37] T. Su, Parametrization of kinetic energy dependences of ion–polar molecule collision rate constants
400 by trajectory calculations, *J. Chem. Phys.* 100 (1994) 4703–4703. doi:10.1063/1.466255.
- 401 [38] W.J. Chesnavich, T. Su, M.T. Bowers, Collisions in a noncentral field: A variational and trajectory
402 investigation of ion–dipole capture, *J. Chem. Phys.* 72 (1980) 2641–2655. doi:10.1063/1.439409.
- 403 [39] K. Sekimoto, S.-M. Li, B. Yuan, A. Koss, M. Coggon, C. Warneke, J. de Gouw, Calculation of the
404 sensitivity of proton-transfer-reaction mass spectrometry (PTR-MS) for organic trace gases using
405 molecular properties, *Int. J. Mass Spectrom.* 421 (2017) 71–94. doi:10.1016/j.ijms.2017.04.006.
- 406 [40] P. Španěl, D. Smith, Selected ion flow tube studies of the reactions of H₃O⁺, NO⁺, and O₂⁺ with some
407 chloroalkanes and chloroalkenes, *Int. J. Mass Spectrom.* 184 (1999) 175–181. doi:10.1016/S1387-
408 3806(98)14296-3.
- 409