Manuscript Details

Manuscript number	COELEC_2018_39
Title	Electrochemical detection of nitric oxide and S-nitrosothiols in biological systems: past, present & future
Short title	Electrochemical detection of nitric oxide and S-nitrosothiols
Article type	Review article

Abstract

Electrochemical detection of nitric oxide using different electrode materials and strategies exploded after the discovery of nitric oxide as important biological messenger. S-nitrosothiols (RSNOs), which result from interaction of NO with peptides and proteins, were shown to be important pools of NO that interfere in different physiological and pathological conditions. This lead to development of several decomposition methods to detect RSNOs electrochemically. This mini-review summarizes the beginning and the current investigations in electrochemical methods to detect NO and RSNOs. Indeed, it describes the latest trends to detect NO and RSNO using microfluidic technologies coupled to electrochemistry and discuss the future of NO and RSNOs detection.

Keywords	S-nitrosothiols; nitric oxide; electrochemical detection; ultramicroelectrodes; microfluidics; microchip capillary electrophoresis; nanoparticles; nanomaterials
Corresponding Author	Fethi Bedioui
Corresponding Author's Institution	CNRS
Order of Authors	Fethi Bedioui, Abdulghani Ismail, Sophie Griveau
Suggested reviewers	Frederic Lemaître, Tebello Nyokong, Stéphane ARBAULT

Submission Files Included in this PDF

File Name [File Type]

Highlights.docx [Highlights]

Bedioui_et_al_final.docx [Manuscript File]

To view all the submission files, including those not included in the PDF, click on the manuscript title on your EVISE Homepage, then click 'Download zip file'.

Electrochemical detection of nitric oxide and S-nitrosothiols in biological systems: past, present & future

highlights:

- Electrochemical detection of nitric oxide using ultramicroelectrodes and the methods envisaged to enhance its selectivity and sensitivity
- RSNO decomposition methods in order to indirectly quantify RSNOs by electrochemical methods in biological fluids are summarized
- Detection of NO and of RSNO inside microfluidic devices is the new trend

Electrochemical detection of nitric oxide and S-nitrosothiols in biological systems: past, present & future

Fethi Bedioui^{1*}, Abdulghani Ismail^{1,2} & Sophie Griveau¹

¹ Chimie ParisTech, PSL Research University, Unité de Technologies Chimiques et Biologiques pour la Santé, CNRS 8258, INSERM 1022, 75005 Paris (France)

² Univ. Grenoble Alpes, CNRS, CEA, INAC-SyMMES, 38000 Grenoble, France

*fethi.bedioui@chimieparistech.psl.eu

Research highlights:

- Electrochemical detection of nitric oxide using ultramicroelectrodes and the methods envisaged to enhance its selectivity and sensitivity
- RSNO decomposition methods in order to indirectly quantify RSNOs by electrochemical methods in biological fluids are summarized
- Detection of NO and of RSNO inside microfluidic devices is the new trend

Keywords :

S-nitrosothiols, nitric oxide, electrochemical detection, ultramicroelectrodes, microfluidics, microchip capillary electrophoresis, nanoparticles, nanomaterials.

Abstract :

Electrochemical detection of nitric oxide using different electrode materials and strategies exploded after the discovery of nitric oxide as important biological messenger. S-nitrosothiols (RSNOs), which result from interaction of NO with peptides and proteins, were shown to be important pools of NO that interfere in different physiological and pathological conditions. This lead to development of several decomposition methods to detect RSNOs electrochemically. This mini-review summarizes the beginning and the current investigations in electrochemical methods to detect NO and RSNOs. Indeed, it describes the latest trends to detect NO and RSNO using microfluidic technologies coupled to electrochemistry and discuss the future of NO and RSNOs detection.

Electrochemical detection of NO

Nitric oxide, NO was first characterized in 1987 as endothelial-derived relaxing factor and it was reported that it is produced by endothelial cells in blood vessels and diffuses to the adjacent smooth muscles to cause vasodilatation** [1]. Since this initial report that resulted in *R. F. Furchgott*, *L. J. Ignarro* and *F. Murad* receiving the Nobel Prize in Physiology and Medicine in 1998, there has been an explosion of research activity showing that NO release occurs not only from endothelial cells but also from neuronal [2], tumoral [3] and immune system cells [4] etc. The huge and intense research activities in these fields have resulted in more than 268 000 papers being published in the literature during the last 20 years. NO is a free radical that reacts very fast with oxygen, peroxides, O₂-radicals (superoxide O₂^{•-}), metallic ions and metalloproteins [5]. This explains its fleeting existence and extremely low concentrations in biological systems. Its total free concentration in physiological conditions has been established recently in a range to be 0.1 μM down to 5 nM, which is orders of magnitude lower than once though a decade ago [6].

Several strategies have been proposed for NO detection. The only ones that allow direct, real time, label free and *in vivo* detection of NO are those based on the electrochemical detection of NO using ultramicroelectrodes (UMEs), electrodes having one dimension less than 25 μm **[7-44]. Electrochemical NO-sensors based on UMEs offer: (i) good selectivity factors > 100;

(ii) good sensitivity (down to the nanomolar range); (iii) fast response (within the millisecond scale time); (iv) long-term stability (over 1-2 hours); (v) non-destructive technique in close proximity to the site of release (single cells, organelles). Several groups have worked on the detection of NO itself which can be oxidized or reduced depending on the potential utilized. It can undergo mild cathodic reduction by gaining one electron at potentials from -0.5 to -1.4 V *vs* Ag / AgCl. However, the detection of NO through its reduction suffers from severe limitations due to the presence of O₂ which is a major interfering because it is more easily reduced than NO. Direct electro-oxidation of NO is the most used approach for its detection. NO oxidizes at potentials higher than 0.8 V *vs* Ag / AgCl. This leads to selectivity problems because a lot of biologically relevant molecules such as nitrite, dopamine, and urea are oxidized at such high potentials.

The electrode materials usually employed are platinum** [7], platinum alloy (90% Pt, 10% Ir) [9], gold** [10], glassy carbon [18] and carbon fibers [12]. The electrode material and surface characteristics affect the potential at which NO is oxidized, the selectivity, the sensitivity, the signal stability and the quality of the ensuing analytical measurements [16,**19,30,38]. For example, the platinization of the platinum electrode or carbon nanoelectrode surface gives faster electron transfer and lowers NO reduction [31,32] or oxidation [33] potentials, respectively. Recently, modification of glassy-carbon electrodes with graphene-gold nanocomposite [34] or elaborating a 3D nanoporous gold microelectrode [35] improved the performances of the sensors. Govindhan et al. [36] summarized the recent advances in platinum-based nanomaterials for amelioration of sensitivity and selectivity of NO detection by changing the electronic exchange surface properties. To overcome the problems of selectivity two main approaches have been developed through the use of membranes selective to NO and / or the use of electrocatalytic moieties that decrease the potential of NO oxidation.

The first “NO sensor” in brain tissue was Shibuki’s electrode** [7] (**Figure 1a**). The author succeeded in detecting NO by modifying of the Clark’s electrode that was developed for O₂. The gas permeable membrane used was wax printed and sealed twice with chloroprene rubber. This approach improved the selectivity against nitrite but lacked reproducibility. Although this sensor allows the very first studies of *in vivo* NO, its response time (*ca.* 1 s) and the fluctuations of the background signals make it useless for most biomedical applications.

The second approach of elaborating NO-sensor is based on the immobilization of NO-permeable membranes directly on the metal without the use of internal solution (**Figure 1b**). Such an electrode is amenable to miniaturization and uses a multilayer membranes that permits selectivity against several possible biological interferences [*16,**19,30,38]. Indeed, several kinds of membranes have been deposited on electrodes. Anionic and cationic membranes act by electrostatic repulsion improving selectivity against interferences. Bedioui et al.** [10] used for the first time microelectrodes of gold coated by Nafion for detection of NO. The linearity of the calibration curve was between 10-100 μM . Other membranes were evaluated, including polycarbazole, polydimethylsiloxane, polystyrene (PS), fluorinated xerogel, polytetrafluoroethylene (PTFE) and o-phenylenediamine (o-PD) [*16,**19,30,31,38]. Multilayered polymers have also been used to ensure the impermeability to the largest amount possible of analytes except NO. A compromise should be found between selectivity and sensitivity since thicker membranes lead to lower sensitivity. Among the various methods of depositing membranes on the surface of the electrodes, electropolymerisation represents an elegant method since it permits controlling the thickness of film and covering small irregular spaces* [16]. The monomers usually used are eugenol, phenol, aniline, o-PD.

The third approach involves the use of an electrocatalyst (metallo-porphyrins or metallo-phthalocyanines) to improve the electron transfer inducing a negative shift of the oxidation potential by 0.15 V and an increase in sensitivity by *ca* 1.5-3 times [**19,**39-41] (**Figure 1c**). The electrocatalyst layer is deposited directly on the metal and other layers (size or charge exclusion) can be added to provide the sensor with more selectivity. A fourth rarely used approach consists in the use of a composite material made from catalyst and permselective membrane [38].

In all cases the performances of the prepared sensors depend on several parameters such as (i) the electrode material and the conditions of polymerization that affect the properties of an electrochemically prepared polymer modified electrode* [17], (ii) the potential necessary for NO detection which depends on the nature of the substrate, the electrocatalytic properties of the membrane and on the surface roughness [30] and (iii) the electrochemical techniques which are usually simple amperometry, pulsed chronamperometry, differential normal pulse amperometry and differential normal pulse voltamperometry* [16].

Figure 1: General types of electrochemical NO sensors (a) Clark's type electrode, (b) whole solid type electrode and (c) composite electrode. Adapted from [30]

Electrochemical detection of S-nitrosothiols

In order to be transported and stored in biological fluids, NO binds to the sulfhydryl groups of peptides and proteins forming S-nitrosothiols that play important roles in several physiological functions ^{**}[45-52] and physiopathological events [53-58]. They exist in biological media at concentrations that vary between tenth of nanomolar to less than ten micromolar ^{**}[59-61] and there is no gold standard method to determine their biological concentrations. They can be directly detected if the RS-NO bond remains intact or otherwise indirectly determined [62,63]. Indirect detection methods are based on a two-step protocol: decomposition of the RS-NO bond followed by the detection of the decomposition products (NO, nitrite or thionyl moiety) using electrochemical, spectrophotometric, or fluorescent methods, biotin switch methods or chemiluminescence assays [31,^{**}59,64-67] (**figure 2**). Electrochemical methods represent direct, real time, and label-free detection approach that can be used for *in vivo* applications [31,^{*}60]. RSNOs can be decomposed through different pathways [68] such as metal cation catalysis ^{*}[69], ascorbic acid [70], heat [69,71], infrared, ultraviolet [72-74] or visible light [46,72,75,76]. However, they mainly lead to partial and non-reproducible decomposition that can be detrimental for accurate detection. Furthermore the decomposition processes can be multiple: homolytic cleavage giving rise to the formation of unstable NO[•] and RS[•] that can lead to nitrite and other end-products, or heterolytic cleavage leading to RS⁻ and NO⁺ which rapidly forms nitrite (**figure 2**).

Figure 2: General scheme for S-nitrosothiols (RSNOs) indirect quantitation. Decomposition is done using mercuric (II) ion (Hg^{2+}), cuprous ions (Cu^+), heat (Δ) or light ($h\nu$). This decomposition is followed, depending on the nascent product, by adequate detection method (fluorescence, colorimetry, chemiluminescence, or electrochemistry)..

Electrochemistry can be implemented for RSNOs detection through the determination of NO released upon their decomposition. Only one study reported on the direct RSNOs detection thanks to their reduction at gold and glassy carbon electrodes in acidic medium at pH 4 [77]. Very little application of RSNOs electrochemical reduction (between -0.6 and -0.9 V vs Ag / AgCl) is found in literature, probably due to the possible interferences of O_2 and other molecules and to the low sensitivity of the method. Thus, the most used strategy consists in chemical decomposition of RSNOs into NO or nitrite and then the detection of the resulting molecules by electrochemistry. The decomposition is accomplished mostly by Cu^+ catalyst. Several methods have been used (**Table 1**) (i) Cu^+ can be added directly using CuCl, (ii) Cu^{2+} can be added with a reducing agent such as thiol or ascorbic acid to give Cu^+ [78] or (iii) Cu metal can be oxidized to produce Cu^+ . For example, Meyerhoff et al. [79-82] have used a strategy where the catalyst is immobilized in a polymeric membrane near to the NO sensor. The catalysts used were copper, organoselenium and organotelluride nanoparticles. Recently, a new

method was developed by Baldim et al. to analyze RSNO electrochemically after decomposition by gold nanoparticles *[83]. Light decomposition and electrochemical detection of NO have also been used to analyze RSNOs [76,**84].

Table 1: Different methods of decomposition by copper and other organometals to detect RSNOs.

Adapted from [*60,76,*78,*83]

RSNOs decomposition	RSNOs sensor configuration	E (V/ref)	LOD	RSNO tested	Sensor lifetime	Interfering compounds tested	Biological samples
Catalyst: Cu(NO ₃) ₂ + thiol (L-cysteine or glutathione)	ISO-NO WPI	0.8 V	≈ 50 nM	GSNO, SNAP, S-nitrosated bovine serum albumin	n.d.	Nitrosomorpholine (no response up to 1 nM), N-nitroso-N-methylurea (no response up to 1 mM), Isoamyl nitrite (no response up to 0.7 mM), Nitroglycerol (no response up to 4 mM), NO ₂ ⁻ (no response up to 50 μM)	–
Catalyst: CuCl	ISO-NO Mark II WPI: carbon fiber (=100 nm diameter) / Nafion / WPI membrane	0.865 V	≈ 10 nM	GSNO, SNAP, AlbsNO	n.d.	NO ₂ ⁻ (no response at 50 μM), AA (no response at 50 μM), L-Arg (no response up to 100 μM), DA (no response at 10 μM), NH _{3(g)} (no response at 1 μM), CO _{2(g)} (no response at 1 μM), CO _{2(g)} (no response at 1 μM)	–
Catalyst: CuCl	Pt disk (200 μM diameter) / poly-Cu(II)TAPc / Nafion	DPV	≈ 4 nM	GSNO	n.d.	NO ₂ ⁻ (no response at 10 μM), UA (no response at 100 μM), AA (no response at 100 μM), 5-hydroxyindole-3-acetic acid (no response at 10 μM), 3,4-dihydroxyphenylacetic acid (no response at 10 μM), DA (peak at 0.3 V above 1 μM), Epinephrine (above 1 μM, peak observed at 0.3 V), 5-hydroxytryptamine (above 1 μM, peak observed at 0.3 V)	Whole blood
Catalyst: Cu(II)-ligand complex or Cu(II)phosphate or Cu(0) particles as the copper source + ascorbate	Platinized Pt disk (250 μM diameter) / PTFE / Copper-based catalytic membrane	0.75 V	n.d.	SNAP, SNAC, CysNO, GSNO, AlbsNO	10 days	NO ₂ ⁻ (no response from 0.1 to 100 μM)	Whole blood
Catalyst: CuSO ₄ + GSH	UME Platinum (25 μm)	0.8 V	100 nM	GSNO	1 day	nitrite	–
Catalyst: organoselenium	Platinized Pt disk (250 μM diameter) / PTFE / organoselenium catalytic hydrogel	0.75 V	<0.1 μM	SNAC, SNAP, SPA, CysNO, GSNO, AlbsNO	10 days	n.d.	Whole blood
Catalyst: organotelluride	Platinized Pt disk (250 μM diameter) / PTFE / organotelluride based hydrogel	0.75 V	<0.1 μM	SNAP, GSNO, CysNO, AlbsNO	>1 month	n.d.	Whole blood
Catalyst: organoselenium	Platinized Pt disk (250 μM diameter) / PTFE / organoselenium	0.75 V	<20 μM	GSNO, CysNO, AlbsNO	10 days	NO ₂ ⁻ : sensitivity ratio S(NO ₂ ⁻)/S(NO) = 10 ⁻⁶ , AA: sensitivity ratio S(NO)/S(AA) = 10 ⁻⁶ , N-nitroso-1-proline sensitivity ratio: S(N-	Whole blood

						nitrosoproline)/S(NO)= 10^{-4} , $\text{NH}_3/\text{NH}_4^+$	
Catalyst: electrochemically oxidized Cu(0) in the presence of ascorbate	Micrometric ring-disc: central disc (50 μM diameter) = Cu(0), ring = NO sensor (gold / poly-eugenol / polyphenol)	0.70 V	n.d.	GSNO	n.d.	NO_2^- (sensitivity ratio: 9×10^{-4}), H_2O_2 (sensitivity ratio: 3.9×10^{-2}), AA (sensitivity ratio: 1.8×10^{-2})	GSNO in serum
Catalyst: gold nanoparticles	UME Platinum (25 μm)	0.8 V	100 nM	GSNO and total RSNOs in plasma	1 day	nitrite	Total RSNOs in human plasma
Catalyst: visible light	Xerogel-modified platinum disk (2 mm) electrode	0.8 V	CySNO (40 nM) GSNO (30 nM) AlbSNO (0.42 μM)	CySNO, GSNO, AlbSNO		nitrite	Low molecular weight RSNOs in Porcine plasma

Microfluidic devices with electrochemical detection for analysis of NO and RSNOs : current prospects and their future development

Microfluidic devices permit the use of significantly reduced sample volumes, improving amenability to clinical analysis [85,86]. Bedioui et al. [87] was the first to elaborate an amperometric fluidic microchip array for the detection of NO. Later on several researchers have detected NO in microfluidic devices using amperometry [88-90]. Most of these NO detections were done for NO produced from cells cultured inside the microfluidic device. Only Hunter et al [88] have detected NO in a spiked wound fluid and in whole blood. Gunasekara et al. [91] used microchip electrophoresis with amperometric detection for the study of the generation of NO by NO-donor molecules such as NONOate salts. They separated NO from nitrite and NONOate in less than 1 min. To the best of our knowledge, only one study analyzed total RSNOs in microfluidic devices with electrochemical detection [84]. No separation step was performed and the decomposition duration was quite long (100 s). The use of a microfluidic device permitted more complete sample irradiation and thus higher conversion of RSNOs to NO after a specific time. Detection methods, other than electrochemistry, were applied to detect RSNOs after their separation [92-94]. Introducing electrochemical detection instead of fluorescence detection after RSNOs separation will offer more simple, specific, and portable devices.

Concluding remarks

Evolution of NO detection in biological media using UME is still under investigation mainly in the variation of electrodes surface composition (nanoparticles and nanocomposites) and in

membrane's permselective and catalytic properties. In case of NO bound to peptides and proteins in form of RSNO analysis, advance has been made mainly in decomposition pathways such as copper catalysis and gold nanoparticles techniques. Microfluidics is the new trend in biomedical applications. Few studies were done about the detection of RSNO and NO in microfluidic devices despite its importance in lowering sample volume and improving sensitivity due to confined diffusion volumes. Elaborating a point of care devices as diagnostic tool to separate then detect electrochemically a mixture of RSNO's and NO is still a challenge to be resolved in the coming years

References

* Paper of special interest

** Paper of outstanding interest

1. Ignarro LJ, Buga GM, Wood KS, Byrns RE, Chaudhuri G: **Endothelium-derived relaxing factor produced and released from artery and vein is nitric oxide**. *Proceedings of the National Academy of Sciences* 1987, **84**:9265-9269.
** the first work reporting on the identification of NO as relaxing factor
2. Prast H, Philippu A: **Nitric oxide as modulator of neuronal function**. *Progress in neurobiology* 2001, **64**:51-68.
3. Lamas S: *Nitric oxide, cell signaling, and gene expression*: CRC Press; 2005.
4. Bogdan C: **The function of nitric oxide in the immune system**. In *Nitric Oxide*. Edited by: Springer; 2000:443-492.
5. Beckman JS, Chen J, Ischiropoulos H, Crow JP: **[23] oxidative chemistry of peroxynitrite**. In *Methods in enzymology*. Edited by: Elsevier; 1994:229-240. vol 233.]
6. Hall CN, Garthwaite J: **What is the real physiological NO concentration in vivo?** *Nitric oxide* 2009, **21**:92-103.
7. Shibuki K: **An electrochemical microprobe for detecting nitric oxide release in brain tissue**. *Neurosc. Res.* 1990, **9**:69-76.
** the first electrochemical NO sensor exemplifying the use of permselective membranes
8. Malinski T, Taha Z: **Nitric oxide release from a single cell measured in situ by a porphyrinic-based microsensor**. *Nature* 1992, **358**:676.

****** The first electrochemical NO-sensor combining the use of ion selective membrannes (i.e. Nafion) and electrocatalysis (based on Ni porprphyrin film)

9. Ichimori K, Ishida H, Fukahori M, Nakazawa H, Murakami E: **Practical Nitric-Oxide Measurement Employing a Nitric Oxide-Selective Electrode**. *Review of Scientific Instruments* 1994, **65**:2714-2718.

10. Bedioui F, Trevin S, Devynck J: **The Use of Gold Electrodes in the Electrochemical Detection of Nitric-Oxide in Aqueous-Solution**. *J. Electroanal. Chem.* 1994, **377**:295-298.

****** the first use of gold and Nafion to elaborated a whole solid electrochemical NO-sensor

11. Lantoine F, Trevin S, Bedioui F, Devynck J: **Selective and sensitive electrochemical measurement of nitric oxide in aqueous solution: discussion and new results**. *J. Electroanal. Chem.* 1995, **392**:85-89.

****** the first calibration cuve of an electrochemivcal NO-sensor within nanomolar range

12. Malinski T, Taha Z, Grunfeld S, Burewicz A, Tombouliau P, Kiechle F: **Measurements of nitric oxide in biological materials using a porphyrinic microsensor**. *Analytica Chimica Acta* 1993, **279**:135-140.

***** 13. Amatore C, Arbault S, Bruce D, de Oliveira P, Erard M, Vuillaume M: **Analysis of individual biochemical events based on artificial synapses using ultramicroelectrodes: cellular oxidative burst**. *Faraday Discussions* 2000, **116**:319-333.

14. Friedemann MN, Robinson SW, Gerhardt GA: **o-Phenylenediamine-modified carbon fiber electrodes for the detection of nitric oxide**. *Analytical chemistry* 1996, **68**:2621-2628.

****** first approach attempting the use of electropolymerized o-PD and Nafion and its thourough use for the detection of NO

15. Ciszewski A, Milczarek G: **Electrochemical detection of nitric oxide using polymer modified electrodes**. *Talanta* 2003, **61**:11-26.

***** 16. Bedioui F, Villeneuve N: **Electrochemical nitric oxide sensors for biological samples - Principle, selected examples and applications**. *Electroanalysis* 2003, **15**:5-18.

***** 17. Ciszewski A, Milczarek G: **Preparation and General Properties of Chemically Modified Electrodes Based on Electrosynthesized Thin Polymeric Films Derived from Eugenol**. *Electroanalysis* 2001, **13**:860-867.

18. Ciszewski A, Milczarek G: **A New Nafion-Free Bipolymeric Sensor for Selective and Sensitive Detection of Nitric Oxide**. *Electroanalysis* 1998, **10**:791-793.

19. Bedioui F, Griveau S: **Electrochemical Detection of Nitric Oxide: Assesement of Twenty Years of Strategies**. *Electroanalysis* 2012, **25**:587-600.

****** a fully comprehensive review that recapitulates 20 years of elaborating electrochemical NO sensors

- *20. Amatore C, Arbault S, Guille M, Lemaitre F: **Electrochemical monitoring of single cell secretion: Vesicular exocytosis and oxidative stress.** *Chemical Reviews* 2008, **108**:2585-2621.
21. Amatore C, Arbault S, Bouton C, Coffi K, Drapier JC, Ghandour H, Tong YH: **Monitoring in real time with a microelectrode the release of reactive oxygen and nitrogen species by a single macrophage stimulated by its membrane mechanical depolarization.** *Chembiochem* 2006, **7**:653-661.
22. Amatore C, Arbault S, Bouret Y, Cauli B, Guille M, Rancillac A, Rossier J: **Nitric oxide release during evoked neuronal activity in cerebellum slices: Detection with platinized carbon-fiber microelectrodes.** *Chemphyschem* 2006, **7**:181-187.
23. Fletcher BL, Fern JT, Rhodes K, McKnight TE, Fowlkes JD, Retterer ST, Keffer DJ, Simpson ML, Doktycz MJ: **Effects of ultramicroelectrode dimensions on the electropolymerization of polypyrrole.** *Journal of Applied Physics* 2009, **105**.
- *24. Santos RM, Lourenço CF, Piedade AP, Andrews R, Pomerleau F, Huettl P, Gerhardt GA, Laranjinha J, Barbosa RM: **A comparative study of carbon fiber-based microelectrodes for the measurement of nitric oxide in brain tissue.** *Biosensors and Bioelectronics* 2008, **24**:704-709.
- *25. Brown FO, Finnerty NJ, Lowry JP: **Nitric oxide monitoring in brain extracellular fluid: characterisation of Nafion®-modified Pt electrodes in vitro and in vivo.** *Analyst* 2009, **134**:2012-2020.
26. Pashai E, Darzi GN, Jahanshahi M, Yazdian F, Rahimnejad M: **An electrochemical nitric oxide biosensor based on immobilized cytochrome c on a chitosan-gold nanocomposite modified gold electrode.** *International journal of biological macromolecules* 2018, **108**:250-258.
27. Kim MY, Naveen MH, Gurudatt NG, Shim YB: **Detection of Nitric Oxide from Living Cells Using Polymeric Zinc Organic Framework-Derived Zinc Oxide Composite with Conducting Polymer.** *small* 2017, **13**.
28. Tang L, Li Y, Xie H, Shu Q, Yang F, Liu Y-I, Liang F, Wang H, Huang W, Zhang G-J: **A sensitive acupuncture needle microsensor for real-time monitoring of nitric oxide in acupoints of rats.** *Scientific Reports* 2017, **7**:6446.
29. Lee G-J, Lee YJ, Park H-K: **Real-Time Monitoring of Nitric Oxide Dynamics in the Myocardium: Biomedical Application of Nitric Oxide Sensor.** In *Nitric Oxide Synthase-Simple Enzyme-Complex Roles*. Edited by: InTech; 2017.
30. Privett BJ, Shin JH, Schoenfish MH: **Electrochemical nitric oxide sensors for physiological measurements.** *Chem. Soc. Rev.* 2010, **39**:1925-1935.
31. Hetrick EM, Schoenfish MH: **Analytical Chemistry of Nitric Oxide.** In *Ann. Rev. Anal. Chem.* Edited by: Annual Reviews; 2009:409-433. *Ann. Rev. Anal. Chem.*, vol 2.]
32. Lee Y, Oh BK, Meyerhoff ME: **Improved planar amperometric nitric oxide sensor based on platinized platinum anode. 1. Experimental results and theory when applied for**

- monitoring NO release from diazeniumdiolate-doped polymeric films. *Anal. Chem.* 2004, **76**:536-544.
33. Li Y, Hu K, Yu Y, Rotenberg SA, Amatore C, Mirkin MV: **Direct Electrochemical Measurements of Reactive Oxygen and Nitrogen Species in Nontransformed and Metastatic Human Breast Cells.** *Journal of the American Chemical Society* 2017, **139**:13055-13062.
 34. Bai RG, Muthoosamy K, Zhou M, Ashokkumar M, Huang NM, Manickam S: **Sonochemical and sustainable synthesis of graphene-gold (G-Au) nanocomposites for enzymeless and selective electrochemical detection of nitric oxide.** *Biosensors and Bioelectronics* 2017, **87**:622-629.
 35. Liu Z, Nemec-Bakk A, Khaper N, Chen A: **Sensitive electrochemical detection of nitric oxide release from cardiac and cancer cells via a hierarchical nanoporous gold microelectrode.** *Analytical chemistry* 2017, **89**:8036-8043.
 36. Govindhan M, Liu Z, Chen A: **Design and electrochemical study of platinum-based nanomaterials for sensitive detection of nitric oxide in biomedical applications.** *Nanomaterials* 2016, **6**:211.
 37. Lee Y, Kim J: **Simultaneous Electrochemical Detection of Nitric Oxide and Carbon Monoxide Generated from Mouse Kidney Organ Tissues.** *Anal. Chem.* 2007, **79**:7669-7675.
 - * 38. Bedioui F, Quinton D, Griveau S, Nyokong T: **Designing molecular materials and strategies for the electrochemical detection of nitric oxide, superoxide and peroxynitrite in biological systems.** *Phys. Chem. Chem. Phys.* 2010, **12**:9976-9988.
 39. Caro CA, Zagal JH, Bedioui F: **Electrocatalytic Activity of Substituted Metallophthalocyanines Adsorbed on Vitreous Carbon Electrode for Nitric Oxide Oxidation.** *J. Electrochem. Soc.* 2003, **150**:E95-E103.
 - ** a complete experimental study of the catalytic effects of a large family of metallophthalocyanines on the electrochemical oxidation of NO
 - * 40. Brown MD, Schoenfish MH: **Catalytic selectivity of metallophthalocyanines for electrochemical nitric oxide sensing.** *Electrochimica Acta* 2018.
 41. Xu H, Liao C, Liu Y, Ye B-C, Liu B: **Iron phthalocyanine decorated nitrogen-doped graphene biosensing platform for real-time detection of nitric oxide released from living cells.** *Analytical chemistry* 2018.
 42. Dumitrescu E, Wallace KN, Andreescu S: **Real time electrochemical investigation of the release, distribution and modulation of nitric oxide in the intestine of individual zebrafish embryos.** *Nitric Oxide* 2018, **74**:32-38.
 43. Matsuoka R, Kobayashi C, Nakagawa A, Aoyagi S, Aikawa T, Kondo T, Kasai S, Yuasa M: **A Reactive Oxygen/Nitrogen Species Sensor Fabricated from an Electrode Modified with a Polymerized Iron Porphyrin and a Polymer Electrolyte Membrane.** *Analytical Sciences* 2017, **33**:911-915.

44. Musameh MM, Dunn CJ, Uddin MH, Sutherland TD, Rapson TD: **Silk provides a new avenue for third generation biosensors: Sensitive, selective and stable electrochemical detection of nitric oxide.** *Biosensors and Bioelectronics* 2018, **103**:26-31.
 45. Mellion BT, Ignarro LJ, Myers CB, Ohlstein EH, Ballot BA, Hyman AL, Kadowitz PJ: **Inhibition of human platelet aggregation by S-nitrosothiols. Heme-dependent activation of soluble guanylate cyclase and stimulation of cyclic GMP accumulation.** *Mol. Pharmacol.* 1983, **23**:653-664.
 46. Stamler JS, Simon DI, Osborne JA, Mullins ME, Jaraki O, Michel T, Singel DJ, Loscalzo J: **S-nitrosylation of proteins with nitric oxide: synthesis and characterization of biologically active compounds.** *Proc. Natl. Acad. Sci. U.S.A.* 1992, **89**:444-448.
 47. Moynihan HA, Roberts SM: **Preparation of Some Novel S-Nitroso Compounds as Potential Slow-Release - Agents of Nitric-Oxide in-Vivo.** *Journal of the Chemical Society-Perkin Transactions 1* 1994:797-805.
 48. Langford EJ, Brown AS, Wainwright RJ, Debelder AJ, Thomas MR, Smith REA, Radomski MW, Martin JF, Moncada S: **Inhibition of Platelet Activity by S-Nitrosoglutathione During Coronary Angioplasty.** *Lancet* 1994, **344**:1458-1460.
 49. Radomski MW, Rees DD, Dutra A, Moncada S: **S-Nitroso-Glutathione Inhibits Platelet Activation In vitro and In vivo.** *British Journal of Pharmacology* 1992, **107**:745-749.
 50. Kowaluk EA, Fung HL: **Spontaneous Liberation of Nitric-Oxide Cannot Account for In vitro Vascular Relaxation by S-Nitrosothiols.** *Journal of Pharmacology and Experimental Therapeutics* 1990, **255**:1256-1264.
 51. Ignarro LJ, Lipton H, Edwards JC, Baricos WH, Hyman AL, Kadowitz PJ, Gruetter CA: **Mechanism of Vascular Smooth-Muscle Relaxation by Organic Nitrates, Nitrites, Nitroprusside and Nitric-Oxide - Evidence for the Involvement of S-Nitrosothiols as Active Intermediates.** *Journal of Pharmacology and Experimental Therapeutics* 1981, **218**:739-749.
- * * First evidence of the effect of S-nitrosothiols in the vasodilatation and in the mechanism of action of NO in human body
52. de Souza GFP, Yokoyama-Yasunaka JKU, Seabra AB, Miguel DC, de Oliveira MG, Uliana SRB: **Leishmanicidal activity of primary S-nitrosothiols against Leishmania major and Leishmania amazonensis: Implications for the treatment of cutaneous leishmaniasis.** *Nitric Oxide-Biology and Chemistry* 2006, **15**:209-216.
 53. Butler AR, Rhodes P: **Chemistry, analysis, and biological roles of S-nitrosothiols.** *Anal. Biochem.* 1997, **249**:1-9.
 54. Beeh KM, Beier J, Koppenhoefer N, Buhl R: **Increased glutathione disulfide and nitrosothiols in sputum supernatant of patients with stable COPD.** *Chest* 2004, **126**:1116-1122.
 55. Tyurin VA, Liu SX, Tyurina YY, Sussman NB, Hubel CA, Roberts JM, Taylor R, Kagan VE: **Elevated levels of S-nitrosoalbumin in preeclampsia plasma.** *Circ. Res.* 2001, **88**:1210-1215.

56. Milsom AB, Jones CJH, Goodfellow J, Frenneaux MP, Peters JR, James PE: **Abnormal metabolic fate of nitric oxide in Type I diabetes mellitus.** *Diabetologia* 2002, **45**:1515-1522.
57. Ito M: **Long-Term Depression.** *Annual Review of Neuroscience* 1989, **12**:85-102.
58. Dorheim MA, Tracey WR, Pollock JS, Grammas P: **Nitric-Oxide Synthase Activity Is Elevated in Brain Microvessels in Alzheimers-Disease.** *Biochem. Biophys. Res. Commun.* 1994, **205**:659-665.
59. Giustarini D, Milzani A, Dalle-Donne I, Rossi R: **Detection of S-nitrosothiols in biological fluids: A comparison among the most widely applied methodologies.** *J. Chromatogr. B Analyt. Technol. Biomed. Life Sci.* 2007, **851**:124-139.
- * * Review that summarizes the RSNO quantification in biological fluids by the most widely used analytical techniques citing different obtained values and methodological errors.
- * 60. Griveau S, Bedioui F: **Electroanalytical methodologies for the detection of S-nitrosothiols in biological fluids.** *Analyst* 2013, **138**:5173-5181.
61. Gow A, Doctor A, Mannick J, Gaston B: **S-nitrosothiol measurements in biological systems.** *J. Chromatogr. B* 2007, **851**:140-151.
- * * a critical review on the methodologies of RSNO detection in biological fluids
62. Ismail A, d'Orlyé F, Griveau S, Da Silva JAF, Bedioui F, Varenne A: **Capillary electrophoresis with mass spectrometric detection for separation of S-nitrosoglutathione and its decomposition products: a deeper insight into the decomposition pathways.** *Analytical and bioanalytical chemistry* 2015, **407**:6221-6226.
63. Ismail A, d'Orlye F, Griveau S, Bedioui F, Varenne A, da Silva JAF: **Capillary electrophoresis coupled to contactless conductivity detection for the analysis of S-nitrosothiols decomposition and reactivity.** *Electrophoresis* 2015, **36**:1982-1988.
64. Bryan NS, Grisham MB: **Methods to detect nitric oxide and its metabolites in biological samples.** *Free Radic. Biol. Med.* 2007, **43**:645-657.
65. Foster MW: **Methodologies for the characterization, identification and quantification of S-nitrosylated proteins.** *Biochimica Et Biophysica Acta-General Subjects* 2012, **1820**:675-683.
- * 66. Diers AR, Keszler A, Hogg N: **Detection of S-nitrosothiols.** *Biochim. Biophys. Acta-Gen. Subj.* 2014, **1840**:892-900.
67. Wang H, Xian M: **Chemical methods to detect S-nitrosation.** *Current Opinion in Chemical Biology* 2010, **15**:32-37.
68. Wo Y, Brisbois EJ, Bartlett RH, Meyerhoff ME: **Recent advances in thromboresistant and antimicrobial polymers for biomedical applications: just say yes to nitric oxide (NO).** *Biomaterials science* 2016, **4**:1161-1183.
- * 69. Williams DLH: **The chemistry of S-nitrosothiols.** *Acc. Chem. Res.* 1999, **32**:869-876.

70. Smith JN, Dasgupta TP: **Kinetics and mechanism of the decomposition of S-nitrosoglutathione by L-ascorbic acid and copper ions in aqueous solution to produce nitric oxide.** *Nitric Oxide-Biology and Chemistry* 2000, **4**:57-66.
 71. de Oliveira MG, Shishido SM, Seabra AB, Morgon NH: **Thermal Stability of Primary S-Nitrosothiols: Roles of Autocatalysis and Structural Effects on the Rate of Nitric Oxide Release.** *J. Phys. Chem. A* 2002, **106**:8963-8970.
 72. Veleeparampil MM, Aravind UK, Aravindakumar CT: **Decomposition of S-Nitrosothiols Induced by UV and Sunlight.** *Adv. Phys. Chem.* 2009, **2009**:5.
 73. Dejam A, Kleinbongard P, Rassaf T, Hamada S, Gharini P, Rodriguez J, Feelisch M, Kelm M: **Thiols enhance NO formation from nitrate photolysis.** *Free Radic. Biol. Med.* 2003, **35**:S141-S141.
 74. Ismail A, Araujo MO, Chagas CLS, Griveau S, D'Orlye F, Varenne A, Bedioui F, Coltro WKT: **Colorimetric analysis of the decomposition of S-nitrosothiols on paper-based microfluidic devices.** *Analyst* 2016, **141**:6314-6320.
- ** first application of microPAD system for the quantification of RSNO in biological samples**
75. Sexton DJ, Muruganandam A, McKenney DJ, Mutus B: **Visible-Light Photochemical Release of Nitric-Oxide from S-Nitrosoglutathione - Potential Photochemotherapeutic Applications.** *Photochem. Photobiol.* 1994, **59**:463-467.
 76. Riccio DA, Nutz ST, Schoenfisch MH: **Visible Photolysis and Amperometric Detection of S-Nitrosothiols.** *Anal. Chem.* 2012, **84**:851-856.
 77. Peng B, Meyerhoff ME: **Reexamination of the Direct Electrochemical Reduction of S-Nitrosothiols.** *Electroanalysis* 2013, **25**:914-921.
 - *78. Ismail A, Griveau S, d'Orlyé F, Varenne A, Bedioui F: Quantitation of Cu⁺-catalyzed decomposition of S-nitrosoglutathione using Saville and electrochemical detection: a pronounced effect of glutathione and copper concentrations.** *Electroanal.* 2015, **27**:2857-2863.
 79. Yang J, Welby JL, Meyerhoff ME: **Generic Nitric Oxide (NO) Generating Surface by Immobilizing Organoselenium Species via Layer-by-Layer Assembly.** *Langmuir* 2008, **24**:10265-10272.
 80. Hofler L, Meyerhoff ME: **Modeling the Effect of Oxygen on the Amperometric Response of Immobilized Organoselenium-Based S-Nitrosothiol Sensors.** *Anal. Chem.* 2011, **83**:619-624.
 81. Hwang S, Cha W, Meyerhoff ME: **Amperometric nitrosothiol sensor using immobilized organoditelluride species as selective catalytic layer.** *Electroanal.* 2008, **20**:270-279.
 82. Cha W, Lee Y, Oh BK, Meyerhoff ME: **Direct detection of S-nitrosothiols using planar amperometric nitric oxide sensor modified with polymeric films containing catalytic copper species.** *Anal. Chem.* 2005, **77**:3516-3524.

- * * important contribution showing the catalytic cleavage of RSNO within the NO sensor vicinity and the excellent analytical performances of the device
- * 83. Baldim V, Ismail A, Taladriz-Blanco P, Griveau S, de Oliveira MG, Bedioui F: **Amperometric quantification of S-nitrosoglutathione using gold nanoparticles: a step towards determination of S-nitrosothiols in plasma.** *Anal. Chem.* 2016: doi: 10.1021/acs.analchem.1025b04035.
- 84. Hunter RA, Schoenfisch MH: **S-Nitrosothiol Analysis via Photolysis and Amperometric Nitric Oxide Detection in a Microfluidic Device.** *Anal. Chem.* 2015, **87**:3171-3176.
- * * First paper that analyses total RSNOs in biological fluids in microfluidic device using light decomposition and electrochemical detection. They showed the importance of microfluidic devices in enhancing analytical performance of RSNO detection
- * 85. Whitesides GM: **The origins and the future of microfluidics.** *Nature* 2006, **442**:368-373.
- 86. Tetala KKR, Vijayalakshmi MA: **A review on recent developments for biomolecule separation at analytical scale using microfluidic devices.** *Analytica Chimica Acta* 2016, **906**:7-21.
- 87. Wartelle C, Rodrigues NP, Koudelka-Hep M, Bedioui F: **Amperometric fluidic microchip array sensing device for nitric oxide determination in solution.** *Materials Science & Engineering C-Biomimetic and Supramolecular Systems* 2006, **26**:534-537.
- * * First paper on amperometric detection of nitric oxide in microfluidic devices
- 88. Hunter RA, Privett BJ, Henley WH, Breed ER, Liang Z, Mittal R, Yoseph BP, McDunn JE, Burd EM, Coopersmith CM, et al.: **Microfluidic Amperometric Sensor for Analysis of Nitric Oxide in Whole Blood.** *Anal. Chem.* 2013, **85**:6066-6072.
- 89. Vogel PA, Halpin ST, Martin RS, Spence DM: **Microfluidic Transendothelial Electrical Resistance Measurement Device that Enables Blood Flow and Postgrowth Experiments.** *Anal. Chem.* 2011, **83**:4296-4301.
- 90. Hulvey M, Martin R: **A microchip-based endothelium mimic utilizing open reservoirs for cell immobilization and integrated carbon ink microelectrodes for detection.** *Anal. Bioanal. Chem.* 2009, **393**:599-605.
- * 91. Gunasekara DB, Hulvey MK, Lunte SM, da Silva JAF: **Microchip electrophoresis with amperometric detection for the study of the generation of nitric oxide by NONOate salts.** *Anal. Bioanal. Chem.* 2012, **403**:2377-2384.
- 92. Wang SY, Circu ML, Zhou H, Figeys D, Aw TY, Feng J: **Highly sensitive detection of S-nitrosylated proteins by capillary gel electrophoresis with laser induced fluorescence.** *J. Chromatogr. A* 2011, **1218**:6756-6762.
- 93. Wang SY, Njoroge SK, Battle K, Zhang C, Hollins BC, Soper SA, Feng J: **Two-dimensional nitrosylated protein fingerprinting by using poly (methyl methacrylate) microchips.** *Lab on a Chip* 2012, **12**:3362-3369.
- 94. Tu FQ, Zhang LY, Guo XF, Zhang ZX, Wang H, Zhang HS: **Dual labeling for simultaneous determination of nitric oxide, glutathione and cysteine in macrophage RAW264.7 cells**

by microchip electrophoresis with fluorescence detection. *J. Chromatogr. A* 2014, **1359**:309-316.