Simultaneous reconstruction of absorption, scattering and anisotropy factor distributions in quantitative photoacoustic tomography

Fatmir Asllanaj, Ahmad Addoum

► To cite this version:

Fatmir Asllanaj, Ahmad Addoum. Simultaneous reconstruction of absorption, scattering and anisotropy factor distributions in quantitative photoacoustic tomography. *Biomed.Phys.Eng.Express*, 2020, 6 (4), pp.045010. 10.1088/2057-1976/ab90a0 . hal-02870842

HAL Id: hal-02870842

<https://hal.science/hal-02870842>

Submitted on 22 Nov 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Simultaneous reconstruction of absorption, scattering and anisotropy factor distributions in quantitative photoacoustic tomography

Fatmir Asllanaj^{1,2*}, Ahmad Addoum³

¹ Université de Lorraine, LEMTA, UMR 7563, Vandœuvre-lès-Nancy, F-54500, France

² CNRS, LEMTA, UMR 7563, Vandœuvre-lès-Nancy, F-54500, France

³ Université de Lyon, IP2I, UMR 5822, Lyon, France

* Correspondence author - Phone : +33 3 72 74 42 73

E-mail: Fatmir.Asllanaj@univ-lorraine.fr

Abstract We present for the first time the simultaneous reconstruction of three optical parameters distributions of biological tissues namely, the absorption μ_a and scattering μ_s coefficients, as well as the anisotropy factor g of the Henyey-Greenstein phase function as a new optical contrast. The 2D images are obtained from the simulation experiments and multi-source quantitative photoacoustic tomography with the radiative transfer equation (RTE) as light transport model. The image reconstruction method is based on a gradient-based optimization scheme. The adjoint method applied to the RTE is used to efficiently compute the gradient of the objective function. The results show simultaneous reconstructions of the three optical properties even with noisy data. The crosstalk problem between the three parameters is highlighted. Superior quality images are obtained for μ_a compared to those of μ_s and g . Moreover, our algorithm allows reconstructing inserts-like heterogeneities with very good spatial resolution and qualitative accuracy.

Keywords quantitative photoacoustic tomography, radiative transfer equation, inverse problem, image reconstruction, adjoint method, biological tissues.

24 Nomenclature

25	A	absorbed energy density field, W mm^{-3}
26	M	measured data of the absorbed optical energy density, W mm^{-3}
27	N_s	number of collimated sources
28	g	anisotropy factor of the Henyey-Greenstein phase function
29	H	observation equation
30	J	objective function
31	$\mathbf{n}$	outward unit vector normal to the medium boundary
32	p	scattering phase function
33	$\mathbf{r}$	spatial position ($= x, y$), mm
34	S	source term in the RTE

35

36 *Greek symbols*

37	ψ	radiance, $\text{W mm}^{-2} \text{sr}^{-1}$
38	$\Delta\Omega$	control solid angle
39	$\mathcal{L}$	Lagrangian
40	μ_a	absorption coefficient, mm^{-1}
41	Ω	direction vector
42	μ_t	attenuation coefficient ($= \mu_a + \mu_s$), mm^{-1}
43	ψ	radiance, $\text{W mm}^{-2} \text{sr}^{-1}$
44	μ_s	scattering coefficient, mm^{-1}
45	ρ	directional reflection coefficient
46	$\mathcal{R}$	state equation
47	Θ	angle between two directions, rad
48	θ	parameter to be reconstructed
49	Υ	intensity of the collimated laser beam, W mm^{-2}

50	Φ	fluence rate, W mm^{-2}
51	ϕ	adjoint variable
52	Σ	interval $[0, 2\pi]$
53	$\mathcal{D}$	domain of $\mathbb{R}^2$
54	$\partial\mathcal{D}$	boundary of the medium $\mathcal{D}$
55		
56	<i>Subscripts</i>	
57	$\mathcal{D}$	spatial
58	Ω	angular
59	$\mathcal{D}\Omega$	spatial-angular
60	c	collimated
61	s	scattered
62	sp	specular
63	x, y, z	(Ox) -axis, (Oy) -axis, (Oz) -axis
64	<i>Superscripts</i>	
65	$*$	adjoint operator
66	$+$	incoming boundary
67	$-$	outgoing boundary

68

69 **1 Introduction**

70 Photoacoustic tomography (PAT) is an emerging technique for non-invasive imaging of biological tissues. It is
71 based on the photoacoustic effect, which refers to the generation of acoustic waves by the absorption of the optical
72 energy in the tissue illuminated by an ultrashort pulsed laser [1–13]. The hybrid modality of PAT combines the
73 high tissue contrast of optical imaging methods and the good spatial resolution of ultrasound imaging methods.

PAT allows imaging at depths and resolutions unprecedented for optical methods. The optical part of PAT provides information on the distribution of chromophores, which are light-absorbing molecules within the tissue. The chromophores of interest are, for example, haemoglobin, melanin, collagen, and various contrast agents. PAT has been successfully applied to the visualization of different structures in biological tissues, such as microvasculature of tumors, human blood vessels, the cerebral cortex in small animals or breast cancer. However, this information is only a qualitative image and it does not give quantitative information on the concentrations of the chromophores. Quantitative photoacoustic tomography (QPAT) is a technique aimed at estimating the absolute concentration of the chromophores from a reconstructed PAT image. It reconstructs the optical parameters of biological tissue from data describing the absorbed energy distribution inside the tissue (assumed to be known in this work). The development of improved image reconstruction algorithm in QPAT constitutes a challenging problem [2]. An accurate forward model is essential to meet the requirements of clinical applications and to obtain a good quality reconstruction. There have been extensive studies on the optical inverse problem of QPAT, although most were using the Diffusion Equation (DE) in the diffusive regime typically assuming that the light propagation throughout the tissue is near-isotropic [7]. However, this model has well-known limitations [14–17]. In addition, the anisotropy factor also strongly affects the light propagation in tissue in the transport regime. Indeed, the biological tissues are highly forward scattering media where g is typically between 0.8 and 1 [9, 14, 18]. Although the absorption map is usually of the major clinic interest, it is necessary to reconstruct the scattering maps (μ_s , g) as well in order to accurately reconstruct the absorption map when the scattering coefficient and the anisotropy factor are unknown. Some works have shown that g can be modified when the tissue is affected by a tumor because cells and cell nuclei change their size and shape. Therefore, the morphological modification of the tissue changes the scattering coefficient μ_s . Since the anisotropy factor describes the anisotropy (angular distribution) of light scattering, this modification will also lead to a variation of g values between healthy and tumor tissues. For instance, quantitative phase imaging showed, on a prostate tissue biopsy with malignancy, that the anisotropy factor g can be a marker of disease [19]. Van Hillegersberg et al. [20] pointed that the anisotropy factor of rat liver decreases from 0.952 to 0.946 in a tumor at 633 nm. Germer et al. [21] reported experimentally that g was different for normal human liver tissue ($g = 0.902$) and liver metastases ($g = 0.955$) at three different

wavelengths. Consequently, the anisotropy factor can provide an additional intrinsic contrast for optical imaging.

To overcome the limitations of the DE, the Radiative Transfer Equation (RTE) has been addressed as a rigorous model for light transport in biological tissues and has become a focus of investigations in QPAT [15–17, 22–35].

In the optical inverse problem of QPAT, it has been shown that the absorption coefficient μ_a can be reconstructed if one light source is used whereas the simultaneous reconstruction of (μ_a, μ_s) needs multiple optical illuminations, so-called multi-source QPAT [16, 25, 36, 37]. Multiple measurements are often needed as well to eliminate non-uniqueness of the reconstruction problem [7, 23].

To the best of our knowledge, the simultaneous reconstruction of (μ_a, μ_s, g) has not been presented so far while the anisotropy factor is an important optical parameter [38]. In practical applications, the anisotropy factor is usually not known while this factor should be known to better describe light propagation. Recovering (μ_s, g) is especially difficult due to the weak dependence of the absorbed optical energy density on scattering. To overcome this problem, the approach has been to assume the anisotropy factor as known and estimate simultaneously (μ_a, μ_s) in QPAT based on the RTE [16, 17, 26, 29, 31–33]. However, this approach can bias the estimated value of μ_s and also of μ_a . In this regard, we investigated the simultaneous reconstruction of (μ_a, μ_s, g) in multi-source QPAT based on the RTE in the transport regime. For the inversion, a gradient-based scheme using the Lm-BFGS was considered to update the spatial distribution of optical parameters. In such scheme, the major challenge is the computation of the objective function gradient which is the most expensive step. Evaluating the gradient through perturbation methods is daunting and prohibitively expensive with the RTE, especially in this case where the parameters are spatially dependent. To overcome this difficulty, the adjoint method applied to the RTE [17, 26–28, 31, 35, 38–40] was used to efficiently compute the objective function gradient with respect to the three optical parameters (μ_a, μ_s, g) regardless the number of unknowns. In this work, a two-dimensional geometry was considered.

The inverse problem in QPAT is challenging since it is ill-posed due to the unsatisfying of both conditions uniqueness and stability. The uniqueness is caused by the strong under-determination nature of the problem where the spatially unknown number to retrieve is significantly higher than the spatially absorbed energy density data. This implies that different spatial distributions of parameters can lead to identical absorbed energy density data. Moreover, the measured noise due to the experimental setup causes an instability of the solution where small noise level

is able to significantly amplify the estimation errors of optical coefficients. The crosstalk problem is an interesting example in QPAT which allows to highlight the robustness of the reconstruction algorithm versus the uniqueness condition. The improvement of the quality reconstruction is expected when increasing the source number where the amount of measured data becomes more important leading to reduce the under-determination character. Furthermore, the noise level is also assessed in our study in order to test the stability of our method.

The remainder of this paper is organized as follows. Section 2 presents the optical forward model in QPAT. Section 3 deals with the optical inverse problem of QPAT. A continuous Lagrangian formulation is used to rigorously deduce the adjoint RTE and an objective function gradient. The results obtained on 2D reconstructions are presented and discussed in section 4. Concluding remarks are finally offered in the final section.

2 Optical forward model in QPAT

2.1 Light transport model

We assumed that the convex domain $\mathcal{D}$ of the medium is illuminated by a collimated laser beam of direction $\boldsymbol{\Omega}_c$. Then, the illuminated wall of the medium is defined by:

$$\partial\mathcal{D}_c = \left\{ \mathbf{r} \in \partial\mathcal{D}, \boldsymbol{\Omega}_c \cdot \mathbf{n}(\mathbf{r}) < 0 \right\}, \quad (1)$$

where $\mathbf{n}$ is the outward unit vector normal to the medium boundary. Let Σ be the interval $[0, 2\pi]$. We also define the incoming and outgoing boundaries:

$$\Gamma^- = \left\{ (\mathbf{r}, \boldsymbol{\Omega}) \in \partial\mathcal{D} \times \Sigma, \boldsymbol{\Omega} \cdot \mathbf{n}(\mathbf{r}) < 0 \right\} \quad \text{and} \quad \Gamma^+ = \left\{ (\mathbf{r}, \boldsymbol{\Omega}) \in \partial\mathcal{D} \times \Sigma, \boldsymbol{\Omega} \cdot \mathbf{n}(\mathbf{r}) > 0 \right\}. \quad (2)$$

The light source $\Upsilon(\mathbf{r})$, given at any location point $\mathbf{r} \in \partial\mathcal{D}$, penetrates from the outgoing into the medium. Part of it propagates through the medium without being deviated, while the rest is scattered in all directions. It is thus convenient to split the radiance ψ into two components [41]. These are denoted $\psi_c(\mathbf{r}) = \psi(\mathbf{r}, \boldsymbol{\Omega})\delta(\boldsymbol{\Omega} - \boldsymbol{\Omega}_c)$ for $\mathbf{r} \in \mathcal{D}$ (δ is the Dirac-delta function and $\boldsymbol{\Omega}_c$ is the direction of the collimated laser beam) and $\psi_s(\mathbf{r}, \boldsymbol{\Omega})$ for

$(\mathbf{r}, \mathbf{\Omega}) \in \mathcal{D} \times \Sigma$. They are respectively the collimated and scattered components of radiance [42]. The $\psi_c(\mathbf{r})$ collimated radiance is governed by the Bouguer-Beer-Lambert equation with its boundary conditions [42]:

$$\left[\mathbf{\Omega}_c \cdot \nabla + \mu_t(\mathbf{r}) \right] \psi_c(\mathbf{r}) = 0 \text{ for } \mathbf{r} \in \mathcal{D}, \quad (3)$$

$$\psi_c(\mathbf{r}) - \Upsilon(\mathbf{r}) = 0 \text{ for } \mathbf{r} \in \partial\mathcal{D}_c \text{ and } \psi_c(\mathbf{r}) = 0 \text{ for } \mathbf{r} \in \partial\mathcal{D} \setminus \partial\mathcal{D}_c. \quad (4)$$

141 where μ_t is the sum of the absorption and scattering coefficients. The scattered radiance $\psi_s(\mathbf{r}, \mathbf{\Omega})$ at location
142 $\mathbf{r} \in \mathcal{D} \subset \mathbb{R}^2$ in direction $\mathbf{\Omega} \in \Sigma$ is solution of the steady state RTE:

$$\left[\mathbf{\Omega} \cdot \nabla + \mu_t(\mathbf{r}) \right] \psi_s(\mathbf{r}, \mathbf{\Omega}) - \mu_s(\mathbf{r}) \int_{\Omega'=2\pi} p(\mathbf{\Omega}' \cdot \mathbf{\Omega}) \psi_s(\mathbf{r}, \mathbf{\Omega}') d\Omega' - S_c(\mathbf{r}, \mathbf{\Omega}) = 0, \quad (5)$$

143 for $(\mathbf{r}, \mathbf{\Omega}) \in \mathcal{D} \times \Sigma$ where S_c is an additional radiation source term to the RTE due to the scattered part of the
144 collimated laser beam within the medium [42]:

$$S_c(\mathbf{r}, \mathbf{\Omega}) = \mu_s(\mathbf{r}) p(\mathbf{\Omega}_c \cdot \mathbf{\Omega}) \psi_c(\mathbf{r}). \quad (6)$$

145 The Henyey-Greenstein (H-G) phase function is the most widely-adopted scattering phase function of biomedical
146 optics [14] and this has been used here. This function depends only on the inner product between the incident
147 direction $\mathbf{\Omega}'$ that scattered $\mathbf{\Omega}$ and the anisotropy factor g . It is expressed for 2D media as:

$$p(\mathbf{\Omega}' \cdot \mathbf{\Omega}) = \frac{1}{2\pi} \frac{1 - g^2}{(1 + g^2 - 2g \mathbf{\Omega}' \cdot \mathbf{\Omega})}. \quad (7)$$

148 The tissue surfaces are assumed to be semi-transparent boundaries due to the refractive index mismatch between
149 air and tissue. Thus, the boundary conditions for the scattered radiance are [43]:

$$\psi_s(\mathbf{r}, \mathbf{\Omega}) - \frac{1}{\pi} \int_{\mathbf{\Omega}' \cdot \mathbf{n} > 0} \rho(\Theta') \psi_s(\mathbf{r}, \mathbf{\Omega}') \mathbf{\Omega}' \cdot \mathbf{n} d\Omega' = 0 \text{ with } \cos \Theta' = \mathbf{\Omega}' \cdot \mathbf{n} \text{ (scattered reflection),} \quad (8)$$

for $(\mathbf{r}, \mathbf{\Omega}) \in \Gamma^-$. The directional reflection coefficient ρ is given by Snell-Descartes laws assuming that the refractive index of the outside medium (air) is unity and that of tissue is equal to 1.4 [42, 43]. The specular reflection $\mathbf{\Omega}_{sp} = \mathbf{\Omega} - 2(\mathbf{\Omega} \cdot \mathbf{n}) \mathbf{n}$ is defined as the direction from which a laser beam must hit the surface. Then, after a specular reflection it travels in the direction of $\mathbf{\Omega}$.

Similarly as for the radiance, the fluence may be separated into its collimated (Φ_c) and scattered (Φ_s) components:

$$\Phi(\mathbf{r}) = \Phi_s(\mathbf{r}) + \Phi_c(\mathbf{r}) \text{ for } \mathbf{r} \in \mathcal{D} \text{ with } \Phi_c(\mathbf{r}) = \psi_c(\mathbf{r}) \text{ and } \Phi_s(\mathbf{r}) = \int_{\Omega=2\pi} \psi_s(\mathbf{r}, \mathbf{\Omega}) d\Omega. \quad (9)$$

The absorption of light in the tissue results in the absorbed energy density field:

$$A(\mathbf{r}) = \mu_a(\mathbf{r})\Phi(\mathbf{r}) = \mu_a(\mathbf{r})\psi_c(\mathbf{r}) + \mu_a(\mathbf{r})\Phi_s(\mathbf{r}). \quad (10)$$

The function $\Phi(\mathbf{r})$ depends on the distribution of absorption and scattering within $\mathcal{D}$, as well as the light source.

The optical forward problem in QPAT is to compute (10) when the optical properties of the biological tissue and the input light source are given.

3 Optical inverse problem of QPAT

3.1 The objective function and observation equation

The optical inverse problem of QPAT is to estimate the optical parameters of the tissue when the absorbed energy density H is given. In this work, we intended to reconstruct the absorption μ_a and scattering μ_s coefficients as well as the anisotropy factor g . The spatial distribution of the vector of parameters $\theta = (\mu_a, \mu_s, g)$ is reconstructed by applying a nonlinear optimization technique to an objective function J that is an explicit function of θ . The real-value objective function describes the discrepancy between the measured absorbed energy density, $M(\mathbf{r})$ and the predicted numerical data, $A(\mathbf{r})$ (given from Eq. (10)). The objective function to be minimized, writes

$$J(\theta) = \frac{1}{2} \sum_{s=1}^{N_s} \left\| \frac{A_s(\theta) - M_s}{M_s} \right\|_{\mathcal{D}}^2, \quad (11)$$

167 where $A_s(\theta)$ and M_s are the predictions and measurements obtained with the s^{th} collimated source, respectively
 168 while N_s is the number of collimated sources. In (11), the fraction uses point-wise division and the norm is
 169 associated to $L^2(\mathcal{D})$, the space of real valued square-integrable functions on $\mathcal{D}$. In order to avoid round-off error
 170 due to the low level of the readings of the forward model, the function J is normalized with respect to M_s . Note
 171 that M_s can be small far away of the illuminated wall of the medium. But, this didn't affect the stability of the
 172 algorithm for the simulations presented further.

It can be noticed that all the mathematical development presented further can be made for one fixed collimated source. Then, we can omit the index s for simplicity. The observation equation $A(\theta)$ is defined as:

$$A(\theta)(\mathbf{r}) = (H \psi_{c,\theta})(\mathbf{r}) + (\tilde{H} \psi_{s,\theta})(\mathbf{r}) \text{ for } \mathbf{r} \in \mathcal{D}, \quad (12)$$

$$\text{with } (H \psi_{c,\theta})(\mathbf{r}) = \mu_a(\mathbf{r})\psi_c(\mathbf{r}) \text{ and } (\tilde{H} \psi_{s,\theta})(\mathbf{r}) = \mu_a(\mathbf{r}) \int_{\Omega=2\pi} \psi_s(\mathbf{r}, \mathbf{\Omega}) d\Omega. \quad (13)$$

173 To define compactly the state equation, we denote by $\mathcal{R}_c(\cdot, \psi_c)$ and $\mathcal{R}_s(\cdot, \psi_c, \psi_s)$ the right-hand sides in equations
 174 (3) and (5). Then:

$$\mathcal{R}(\theta, \psi_c, \psi_s) = \left\{ \mathcal{R}_c(\theta, \psi_c), \mathcal{R}_s(\theta, \psi_c, \psi_s) \right\}. \quad (14)$$

175 The reconstruction algorithm consists of minimizing J when (14) is satisfied. The nonlinear optimization algorithm
 176 chosen in this work requires knowledge of the objective function gradient with respect to unknown parameters. To
 177 compute this gradient, the adjoint method (starting from the Lagrangian method) is introduced.

178 3.2 The Lagrangian and adjoint method

179 For one fixed collimated source, the Lagrangian is written in the L^2 space as [44,45]:

$$\mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s) = \frac{1}{2} \left\| \frac{(H \psi_c) + (\tilde{H} \psi_s) - M}{M} \right\|_{\mathcal{D}}^2 + \left\langle \phi_c \middle| \mathcal{R}_c \right\rangle_{\mathcal{D}} + \left\langle \phi_s \middle| \mathcal{R}_s \right\rangle_{\mathcal{D}\Omega}, \quad (15)$$

180 where the Lagrangian multipliers are: $\phi_c = \phi_c(\mathbf{r})$ (with $\mathbf{r} \in \mathcal{D}$) and $\phi_s = \phi_s(\mathbf{r}, \mathbf{\Omega})$ (with $(\mathbf{r}, \mathbf{\Omega}) \in \mathcal{D} \times \Sigma$). They
 181 are real functions that represent the adjoint variables associated to (ψ_c, ψ_s) . The two last terms in (15) are the inner

182 products associated to $L^2(\mathcal{D})$ and $L^2(\mathcal{D}\Omega)$. It can be noticed that (which is trivial in appearance) if (ψ_c, ψ_s) is the
 183 solution of the state equation (14) for the true θ parameter, then we have the identity:

$$\mathcal{L}(\theta, \psi_c(\theta), \psi_s(\theta), \phi_c, \phi_s) = J(\theta), \text{ for all } \phi_c, \phi_s. \quad (16)$$

184 By deriving this equation it yields:

$$J'(\theta) \delta\theta = \frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \theta} \delta\theta + \frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \psi_c} \frac{\partial \psi_c(\theta)}{\partial \theta} \delta\theta + \frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \psi_s} \frac{\partial \psi_s(\theta)}{\partial \theta} \delta\theta. \quad (17)$$

185 We denote the following independant quantities by:

$$\delta\psi_c = \frac{\partial \psi_c(\theta)}{\partial \theta} \delta\theta \text{ and } \delta\psi_s = \frac{\partial \psi_s(\theta)}{\partial \theta} \delta\theta. \quad (18)$$

186 Then, the adjoint variables are solutions to the following equation [44, 45]:

$$\frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \psi_c} \delta\psi_c + \frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \psi_s} \delta\psi_s = 0, \quad (19)$$

187 and Eq. (17) is reduced to:

$$J'(\theta) \delta\theta = \left\langle \nabla J(\theta) \middle| \delta\theta \right\rangle_{\mathcal{D}} = \frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \theta} \delta\theta. \quad (20)$$

Using (15) and (19) we deduce that:

$$\begin{aligned} & \left\langle \phi_c \middle| \frac{\partial \mathcal{R}_c}{\partial \psi_c} \delta\psi_c \right\rangle_{\mathcal{D}} + \left\langle \phi_s \middle| \frac{\partial \mathcal{R}_s}{\partial \psi_s} \delta\psi_s \right\rangle_{\mathcal{D}\Omega} + \left\langle \phi_s \middle| \frac{\partial \mathcal{R}_s}{\partial \psi_c} \delta\psi_c \right\rangle_{\mathcal{D}\Omega} \\ & + \left\langle \frac{(H \psi_c) + (\tilde{H} \psi_s) - M}{M} \middle| \frac{(H \delta\psi_c)}{M} \right\rangle_{\mathcal{D}} + \left\langle \frac{(H \psi_c) + (\tilde{H} \psi_s) - M}{M} \middle| \frac{(\tilde{H} \delta\psi_s)}{M} \right\rangle_{\mathcal{D}} = 0. \end{aligned} \quad (21)$$

188 Using (12), we change $(H \psi_c) + (\tilde{H} \psi_s)$ by A in the second line of (21). As Eq. (21) has to be satisfied for all
 189 sensitivity directions $\delta\psi_c$ and $\delta\psi_s$, then it leads to the following set of equations (for each sensitivity directions):

$$\begin{aligned} \left\langle \frac{A-M}{M} \middle| \frac{(\tilde{H} \delta\psi_s)}{M} \right\rangle_{\mathcal{D}} + \left\langle \phi_s \middle| \frac{\partial \mathcal{R}_s}{\partial \psi_s} \delta\psi_s \right\rangle_{\mathcal{D}\Omega} &= 0, \\ \left\langle \frac{A-M}{M} \middle| \frac{(H \delta\psi_c)}{M} \right\rangle_{\mathcal{D}} + \left\langle \phi_c \middle| \frac{\partial \mathcal{R}_c}{\partial \psi_c} \delta\psi_c \right\rangle_{\mathcal{D}} + \left\langle \phi_s \middle| \frac{\partial \mathcal{R}_s}{\partial \psi_c} \delta\psi_c \right\rangle_{\mathcal{D}\Omega} &= 0. \end{aligned} \quad (22)$$

We denote $\mathcal{A}^*$ the adjoint operator of $\mathcal{A}$. Using its definition and Appendix A, the equations of (22) lead to:

$$\begin{aligned} \left\langle \left(\frac{\partial \mathcal{R}_s}{\partial \psi_s} \right)^* \phi_s \middle| \delta\psi_s \right\rangle_{\mathcal{D}\Omega} + \left\langle \frac{H(A-M)}{M^2} \middle| \delta\psi_s \right\rangle_{\mathcal{D}} &= 0, \\ \left\langle \left(\frac{\partial \mathcal{R}_c}{\partial \psi_c} \right)^* \phi_c \middle| \delta\psi_c \right\rangle_{\mathcal{D}} + \left\langle \left(\frac{\partial \mathcal{R}_s}{\partial \psi_c} \right)^{\bar{*}} \phi_s \middle| \delta\psi_c \right\rangle_{\mathcal{D}} + \left\langle \frac{H(A-M)}{M^2} \middle| \delta\psi_c \right\rangle_{\mathcal{D}} &= 0, \end{aligned} \quad (23)$$

where $\mathcal{A}^{\bar{*}} = 2\pi \mathcal{A}^*$. As the equations of (23) have to be satisfied for all sensitivity directions $\delta\psi_c$ and $\delta\psi_s$, the adjoint variables must be solutions to the following set of equations:

$$\left(\frac{\partial \mathcal{R}_s}{\partial \psi_s} \right)^* \phi_s + \frac{H(A-M)}{M^2} = 0 \quad \text{and} \quad \left(\frac{\partial \mathcal{R}_c}{\partial \psi_c} \right)^* \phi_c + \left(\frac{\partial \mathcal{R}_s}{\partial \psi_c} \right)^{\bar{*}} \phi_s + \frac{H(A-M)}{M^2} = 0. \quad (24)$$

Replacing $\mathcal{R}_c$ and $\mathcal{R}_s$ defined by (3) and (5) in Eq. (24), we obtain the following adjoint equations model:

$$\begin{aligned} \left[\boldsymbol{\Omega} \cdot \boldsymbol{\nabla} + \mu_t(\mathbf{r}) \right] \phi_s(\mathbf{r}, -\boldsymbol{\Omega}) &= \mu_s(\mathbf{r}) \int_{\Omega'=2\pi} p(\boldsymbol{\Omega}' \cdot (-\boldsymbol{\Omega})) \phi_s(\mathbf{r}, \boldsymbol{\Omega}') d\Omega' - \mu_a(\mathbf{r}) \frac{(A(\mathbf{r}) - M(\mathbf{r}))}{M^2}, \\ \left[\boldsymbol{\Omega}_c \cdot \boldsymbol{\nabla} + \mu_t(\mathbf{r}) \right] \phi_c(\mathbf{r}) &= \mu_s(\mathbf{r}) \int_{\Omega'=2\pi} p(\boldsymbol{\Omega}' \cdot (-\boldsymbol{\Omega}_c)) \phi_s(\mathbf{r}, \boldsymbol{\Omega}') d\Omega' - \mu_a(\mathbf{r}) \frac{(A(\mathbf{r}) - M(\mathbf{r}))}{M^2}. \end{aligned} \quad (25)$$

The directions $\boldsymbol{\Omega}$ and $\boldsymbol{\Omega}_c$ were changed to $-\boldsymbol{\Omega}$ and $-\boldsymbol{\Omega}_c$ for convenience.

If the tissue surfaces are assumed to be semi-transparent with specular reflection, the boundary conditions of the first equation are given by [39]:

$$\phi_s(\mathbf{r}, \boldsymbol{\Omega}) = \rho(\Theta_{sp}) \phi_s(\mathbf{r}, -\boldsymbol{\Omega}_{sp}) \quad \text{for } (\mathbf{r}, \boldsymbol{\Omega}) \in \Gamma^+ \quad \text{with} \quad \cos \Theta_{sp} = \boldsymbol{\Omega}_{sp} \cdot \mathbf{n} \quad \text{and} \quad \cos \Theta = \boldsymbol{\Omega} \cdot \mathbf{n}. \quad (26)$$

The boundary conditions for the adjoint collimated radiance fulfill the condition: $\phi_c(\mathbf{r}) = 0$ for $\mathbf{r} \in \partial\mathcal{D}$.

196 It can be seen that the adjoint equations model takes a similar form to the forward model. The adjoint equations
 197 model can be solved in a similar manner to that used to solve the forward model. It also shows that the first equation
 198 of (25) is solved to obtain $\phi_s(\mathbf{r}, \boldsymbol{\Omega})$ which is bring into the second equation to obtain $\phi_c(\mathbf{r})$.

199 3.3 Gradient of the objective function

200 The differentiation of the Lagrangian (for one fixed collimated source) with respect to θ in direction $\delta\theta$ satisfies:

$$\frac{\partial \mathcal{L}(\theta, \psi_c, \psi_s, \phi_c, \phi_s)}{\partial \theta} \delta\theta = \frac{\partial J(\theta)}{\partial \theta} \delta\theta + \left\langle \phi_c \left| \frac{\partial \mathcal{R}_c}{\partial \theta} \delta\theta \right\rangle_{\mathcal{D}} + \left\langle \phi_s \left| \frac{\partial \mathcal{R}_s}{\partial \theta} \delta\theta \right\rangle_{\mathcal{D}\Omega}. \quad (27)$$

201 It should be noticed that the function J (see Eqs. (11,12)) depends explicitly on θ only if $\theta = \mu_a$. For the other
 202 optical coefficients, $\frac{\partial J(\theta)}{\partial \theta} = 0$. Using (10) and Eq. (20), Eq. (27) is reduced to:

$$\left\langle \nabla J(\theta) \left| \delta\theta \right\rangle_{\mathcal{D}} = \left\langle \phi_c \left| \frac{\partial \mathcal{R}_c}{\partial \theta} \delta\theta \right\rangle_{\mathcal{D}} + \left\langle \phi_s \left| \frac{\partial \mathcal{R}_s}{\partial \theta} \delta\theta \right\rangle_{\mathcal{D}\Omega} \left(+ \left\langle \frac{\Phi(A(\theta) - M)}{M^2} \left| \delta\theta \right\rangle_{\mathcal{D}} \text{ if } \theta = \mu_a \right) \right). \quad (28)$$

This is the expression that evaluates the gradient of the objective function. Applying respectively Eq. (28) to
 $\theta = \mu_a$, $\theta = \mu_s$ and $\theta = g$ we deduce the objective function gradient, with respect to these parameters:

$$\nabla J(\mu_a) = \frac{\Phi(A(\mu_a) - M)}{M^2} + \phi_c \psi_c + \left\langle \phi_s \left| \psi_s \right\rangle_{\Omega}, \quad (29)$$

$$\nabla J(\mu_s) = \phi_c \psi_c + \left\langle \phi_s \left| \psi_s \right\rangle_{\Omega} - \left\langle \phi_s \left| \int_{\Omega'=2\pi} p(\boldsymbol{\Omega}' \cdot \boldsymbol{\Omega}) \psi_s(\mathbf{r}, \boldsymbol{\Omega}') d\Omega' + p(\boldsymbol{\Omega}_c \cdot \boldsymbol{\Omega}) \psi_c(\mathbf{r}) \right\rangle_{\Omega}, \quad (30)$$

$$\nabla J(g) = - \left\langle \phi_s \left| \mu_s(\mathbf{r}) \left(\int_{\Omega'=2\pi} \frac{\partial p(\boldsymbol{\Omega}' \cdot \boldsymbol{\Omega})}{\partial g} \psi_s(\mathbf{r}, \boldsymbol{\Omega}') d\Omega' + \frac{\partial p(\boldsymbol{\Omega}_c \cdot \boldsymbol{\Omega})}{\partial g} \psi_c(\mathbf{r}) \right) \right\rangle_{\Omega}. \quad (31)$$

203 It should be noticed that if more than one collimated sources are considered, the objective function gradient is
 204 obtained by summing the objective function gradient for each collimated source.

205 3.4 Parameter and data scaling

206 In order to speed-up the iterative convergence to the local minimum, a scaling strategy of the optical parameters [16]
 207 in the RTE based forward model was carried out in this work. Choosing an a priori function for each optical

parameters, say μ_a^r, μ_s^r, g_r , the parameters are searched that fluctuate about unity. This scaling leads to recover:

$$\sigma_a(\mathbf{r}) = \frac{\mu_a(\mathbf{r})}{\mu_a^r(\mathbf{r})}, \sigma_s(\mathbf{r}) = \frac{\mu_s(\mathbf{r})}{\mu_s^r(\mathbf{r})}, q(\mathbf{r}) = \frac{g(\mathbf{r})}{g_r(\mathbf{r})}, \quad (32)$$

for which magnitude is of order one approximately for all these three new parameters. It results that the considered objective function becomes $J(\sigma_a, \sigma_s, q)$ (instead of $J(\mu_a, \mu_s, g)$). In the results presented further, the a priori functions for each optical parameters were chosen as those of the background (homogeneous) medium. The objective function gradients with respect to σ_a, σ_s and q are:

$$\begin{aligned} \nabla J(\sigma_a) &= \mu_a^r(\mathbf{r}) \nabla J(\mu_a); \quad \nabla J(\sigma_s) = \mu_s^r(\mathbf{r}) \nabla J(\mu_s); \\ \nabla J(q) &= - \left\langle \phi_s \middle| \mu_s(\mathbf{r}) \left(\int_{\Omega'=2\pi} \frac{\partial p(\boldsymbol{\Omega}' \cdot \boldsymbol{\Omega})}{\partial q} \overline{\psi_s(\mathbf{r}, \boldsymbol{\Omega}')} d\Omega' + \frac{\partial p(\boldsymbol{\Omega}_c \cdot \boldsymbol{\Omega})}{\partial q} \overline{\psi_c(\mathbf{r})} \right) \right\rangle_{A\Omega}. \end{aligned} \quad (33)$$

A second scaling, as in [46], was necessary for the simultaneous reconstruction of three optical parameters distributions. Indeed, the optical coefficients to be reconstructed from the absorbed energy density are different in nature, and their order of magnitude also differs. As a consequence, the objective function gradient parts associated with these optical coefficients also differ by roughly the same order of magnitude, which is very bad for the convergence in the optimization problem when using a gradient-based method. Then, the objective function gradients with respect to σ_a, σ_s and q was scaled as:

$$\nabla J^{scaled}(\sigma_a) = c_{\sigma_a} \nabla J(\sigma_a); \quad \nabla J^{scaled}(\sigma_s) = c_{\sigma_s} \nabla J(\sigma_s) \quad \text{and} \quad \nabla J^{scaled}(q) = c_q \nabla J(q). \quad (34)$$

where $c_{\sigma_a}, c_{\sigma_s}$ and c_q are empirical coefficients that are determined after the first inverse iteration such that the largest element of the scaled gradient vector $\nabla J^{scaled}(\sigma_a)$ equals 5% of the largest element of vector σ_a^0 :

$$c_{\sigma_a} = 0.05 \frac{\max(\sigma_a^0)}{\max(|\nabla J(\sigma_a^0)|)}. \quad (35)$$

217 The same holds for c_{σ_s} . The results show that the best quality reconstruction are obtained when the largest element
 218 of $\nabla J^{scaled}(q)$ equals 1.5% of the largest element of q^0 . The scaling factors are kept constant during the recon-
 219 struction from the first iteration.

220 In QPAT, where the dynamic range of the measured light intensities can be very large, scaling of the data may be
 221 needed in order to ensure numerical stability of the optimization problem. Furthermore, in this work, the data space
 222 was scaled similarly as in [16], where we used the logarithm of amplitude as the data.

223 3.5 Implementation of the reconstruction algorithm

224 A Modified Finite Volume Method (MFVM) of high accuracy [42] was used for solving the equations of the
 225 forward and adjoint models. This MFVM can be applied to arbitrarily shaped geometries, by using unstructured
 226 triangular grids. The methodology of the employed method is not repeated here, we refer the reader to [42] for
 227 comprehensive details. The objective function J was iteratively minimized using the quasi-Newton algorithm with
 228 Lm-BFGS (limited-memory Broyden–Fletcher–Goldfarb–Shanno) [47]. It iteratively updates an initial estimate of
 229 the parameters distribution along a search descent direction denoted d . Once the minimum is found, the final result
 230 is the unknown parameters distribution. The updating procedure is formulated as: $\theta^{k+1} = \theta^k + \alpha^k d (\nabla J(\theta^k))$ [47],
 231 where k is the current iteration of the inverse procedure and α^k represents the step size obtained by the Armijo line
 232 search in order to provide a sufficient minimization of the objective function.

233 4 Results and discussion

234 We consider a 2D numerical phantom with a homogeneous background containing different inserts. In the first test
 235 case, the reconstruction were performed on a relatively large object of size $20 \times 20 \text{ mm}^2$ while the other test cases
 236 use a phantom of $10 \times 10 \text{ mm}^2$ (in all the simulations the length unit is the minimeter). The optical properties of
 237 the background are fixed to $\mu_a = 0.05 \text{ mm}^{-1}$, $\mu_s = 5 \text{ mm}^{-1}$ and $g = 0.9$ expect for the final test case wherein μ_s
 238 and g are assigned the value 6 mm^{-1} and 0.8, respectively. The optical background values were used to start the
 239 optimization procedure by assuming a homogeneous medium. The geometry and positions of the inserts differ for
 240 each test in order to carry out different situations that highlight the main issues encountered in QPAT. The intensity

of the Laser beam has a spatial Gaussian distribution along the x - axis or y - axis ($s = x$ or y) such as:

$$\Upsilon(s) = \frac{1}{\sigma_s \sqrt{2\pi}} \exp\left(\frac{-(s - s_c)^2}{2\sigma_s^2}\right), \quad (36)$$

where $\sigma_s = 1$ mm is the standard deviation of the Laser beam. The position s_c corresponds to the source location at the center of the illuminated wall of the medium. The angular space was discretized with 32 control solid angles whereas the number of nodes of the spatial mesh is given for each test case presented further. The quality of the algorithm is assessed thanks to the relative estimation error ε between the retrieved $\hat{\beta}$ and the exact vector β^* :

$$\varepsilon = 100\% \frac{\|\hat{\beta} - \beta^*\|_2}{\|\beta^*\|_2} \quad \text{where } \|\cdot\|_2 \text{ denotes Euclidian norm.} \quad (37)$$

4.1 Anisotropy factor reconstruction

This case points out the ability of the QPAT to reconstruct, in the multiple scattering regime, the anisotropy factor as an endogenous optical property of tissues. The spatial domain is discretized into 15,857 mesh nodes. Figure 1a illustrates the reference medium including three circular inserts centered at (4 mm; -2 mm), (10 mm; 4 mm) and (16 mm; -4 mm) with a radius of 2 mm. In this case, the absorption coefficient $\mu_a = 0.05 \text{ mm}^{-1}$ is kept constant while μ_s and g are chosen so that their values lead to a constant value of $\mu'_s = 0.5 \text{ mm}^{-1}$ in the whole phantom. This configuration avoids attributing the spatial variation of g to $\mu'_s = \mu_s(1 - g)$. One Laser source was used to illuminate the west surface ($x = 0$ mm). Figure 1b shows the reconstructed image of the anisotropy factor. It can be seen that the retrieved image presents a good agreement with the reference object. The three inserts were spatially well fitted with their original positions. The estimated mean values inside the inclusions are correctly retrieved with respect to their exact values, even though the red-insert presents a slightly over-estimated values $\hat{g}_{max} = 0.96$ against $g^* = 0.95$. The circular shape is also well reconstructed. This result implies that the spatial variation of g cannot be caused by a variation of μ'_s itself. The anisotropy factor can hence be independently reconstructed and separated from μ_s with our inverse algorithm.

Figure 1: Reconstruction of the anisotropy factor: (a) the reference medium containing three inclusions at different spatial positions and (b) the reconstructed image of g .

4.2 Source number effect

After testing the capability of the proposed algorithm to reconstruct the anisotropy factor with one Laser source, we attempt in this case to assess the robustness of the QPAT to simultaneously reconstruct the three optical properties μ_a , μ_s and g . This task has not been reported in the previous optical imaging related works, for our best knowledge. The original medium contains two circular inserts centered at (2 mm; 2 mm) and (-2 mm; -2 mm) with a radius of 1 mm. The exact optical values of the top-right and bottom-left inclusions are assigned as $\mu_a = 0.06 \text{ mm}^{-1}$, $\mu_s = 6 \text{ mm}^{-1}$, $g = 0.85$ and $\mu_a = 0.04 \text{ mm}^{-1}$, $\mu_s = 4 \text{ mm}^{-1}$ and $g = 0.95$, respectively. The unstructured triangular mesh used is composed of 2,821 nodes. Two illumination configurations were carried out: in the first one, the west surface was illuminated with one Laser source while in the second, the phantom is sequentially illuminated on its fourth tissue surfaces. The calculations were carried out with an Intel Xeon Processor E5-2683v4, 2.1GHz, 32 cores. This last uses Hyper-Threading and Intel C compiler. The computational time for the reconstruction in the second configuration was 40 min. where 103 iterations were required. It can be noticed that, while keeping the same quality of reconstruction, the logarithmic scaling allowed to reduce by about a factor four the number of iterations and decreased the norms of the objective function gradients (with respect to each parameter to be recover) by a factor 10^4 . The obtained results for the first and second configurations are depicted in the left and right column of figure 2, respectively. Despite the critical inversion conditions concerning the large unknowns number of parameters ($3 \times 2,821$) with only one source, the algorithm was still able to reveal the heterogeneities

277 in the medium. In addition, the retrieved local values are close to their exact values even for the relatively deeper
 278 insert (top-right). The localization and the circular edge were achieved with a better quality reconstruction for
 279 the μ_a coefficient compared to those of the μ_s coefficient and anisotropy factor. This is due to the measured
 280 absorbed energy density used for fitting that is directly related to the μ_a coefficient which explains, therefore, the
 281 superior quality estimation for μ_a (see Eq.10). From the right column, the reconstructed images were significantly
 282 improved and the estimated values and the localization are in a good agreement with the real solution. The relative
 283 estimation errors for the top-right inclusion have been decreased from 8.9%, 10.45% and 3.65% to 0.17%, 2.37%
 284 and 0.75% for μ_a , μ_s and g respectively when illuminating the medium with four Laser sources. Furthermore, the
 285 circular shape is correctly reconstructed for all the optical parameters. This configuration has led to increasing the
 286 amount of measured data in the inverse procedure which allows thus to better reconstruct the optical properties
 287 simultaneously.

288 This result highlights the potential interest of using multiple sources which indeed corresponds to real experimental
 289 scenarios with a tomographic context. It is worth noting that the simultaneous reconstruction of the three optical pa-
 290 rameters is not possible with the standard optical tomography since its inverse problem is usually under-determined
 291 and the measured data are collected on the tissue-surface.

Figure 2: Simultaneous reconstruction of μ_a , μ_s and g with: (left column) only one Laser source illuminating the west surface and (right column), four Laser sources illuminating sequentially the phantom, (top row) absorption coefficient μ_a , (middle row) scattering coefficient μ_s and (bottom row) anisotropy factor g .

4.3 Effect of anisotropy factor kept as a fixed constant

We consider the second configuration as in the previous test case by assuming that the two inclusions represent heterogeneities only in μ_a and μ_s coefficients and the anisotropy factor g is fixed at a constant value in the whole medium. The value of g is most often in the range $[0.8 - 1]$. We then chose four different values of g to represent this interval (0.8, 0.85, 0.9 and 0.95). Figure 3 displays the reconstructed images of μ_a and μ_s for each fixed value of g .

297 The relative estimation errors of the two inclusions for μ_a and μ_s are shown in Table 1. Qualitatively, the inclusions
 298 in μ_a and μ_s are recovered with low contrast level for $g = 0.8$ and $g = 0.85$ (due to the high backscattering of
 299 light) while the reconstruction is relatively improved when the g values are increased to 0.9 and 0.95. This has
 300 an important realistic interest because the biological tissues are known to be highly forward scattering. Therefore,
 301 more pronounced local artifacts are appeared near to the detectors and in the background especially in the μ_s
 302 images. For the four values of g , the μ_a coefficient is correctly retrieved with respect to the exact values of the two
 303 inclusions (~ 0.06 for the top-right and ~ 0.04 for the bottom-left). However, the reconstructed μ_s values become
 304 under- and over-estimated with respect to the inclusions-original values as the g is higher since the two scattering
 305 parameters (μ_s and g) are significantly correlated.
 306 To assess the anisotropy factor effect, the obtained result in the previous test case (Fig. 2 right column) is compared
 307 with the present test case when g is fixed at 0.9 in order to be closer as possible to the simulation conditions. For
 308 both cases, the background medium in the μ_a and μ_s images is recovered with the same quality (clear and homo-
 309 geneous). The relative estimation errors for the inclusions of μ_a (0.11 % for top-right and 0.13 % for bottom-left)
 310 is approximately similar while that of μ_s (46.47 % for top-right and 45.64 % for bottom-left) has been increased.

ε (%)	g values			
	0.8	0.85	0.9	0.95
μ_a (top-right inclusion)	7.59	4.49	0.11	2.72
μ_a (bottom-left inclusion)	8.23	5.42	0.13	2.6
μ_s (top-right inclusion)	18.9	8.82	46.47	167.79
μ_s (bottom-left inclusion)	70.57	62.98	45.64	30.22

Table 1: The relative estimation errors ε of inclusions of μ_a and μ_s parameters for the four different values of g .

Figure 3: Simultaneous reconstruction of μ_a and μ_s when g is kept as a fixed constant in the reconstruction procedure: (left column) absorption coefficient and (right column), scattering coefficient, (first line) $g = 0.8$, (second line) $g = 0.85$, (third line) $g = 0.9$ and (last line) $g = 0.95$.

4.4 Noise level effect

In QPAT the measurements noise due to the experimental acquisition setup is unavoidable. To mimics real situations, the simultaneous reconstructions of μ_a , μ_s and g are performed using corrupted data at different noise levels of 1%, 3% and 6% added as a random Gaussian distribution on the exact predictions (absorbed energy density). The original phantom of section 4.2 illuminated by four Laser sources is used in this test case with the same spatial mesh. The used data (i.e. the absorbed energy density) with 6% of noise when the top wall is illuminated is depicted in Figure 4. The reconstructed images are shown in Figure 5.

The relative estimation errors of μ_a , μ_s and g parameters are given in table 2 for the three noise levels with also the noiseless case for comparison. They were computed over the whole reconstructed image domain. The obtained results show that our QPAT algorithm is able to localize the spatial positions of the inserts for the three parameters even with noisy data. As expected, it is seen that the image quality (characterized by its relative error) is worse as the noise level increases (see Tab. 2). Qualitatively, the artifacts and local perturbations become more pronounced and the circular shape of the inserts is degraded especially for μ_s and g images. It can be seen that the μ_a images strongly handle the noise levels better than the scattering parameters μ_s and g . This is again explained by the fact that the fitted data in the inverse problem of the QPAT are directly dependent on the absorption coefficient. Quantitatively, the estimated mean values for the top-right insert become slightly over-estimated for μ_a and μ_s and under-estimated for g with the noise level. Concerning the bottom-left insert, the retrieved mean values are slightly under-estimated for μ_a and μ_s and over-estimated for g when the noise increases.

Figure 4: Used data (i.e. the absorbed energy density) with 6% of noise when the top wall is illuminated.

Figure 5: Simultaneous reconstruction with noisy data at different level of: (left column) 1%, (middle column) 3% and (right column) 6% for the absorption coefficient μ_a (top row), the scattering coefficient μ_s (middle row) and the anisotropy factor g (bottom row).

	Noise levels			
ε (%)	0 %	1 %	3 %	6 %
μ_a	0.54	1.53	3.83	8.26
μ_s	7.63	10.50	13.03	14.24
g	1.79	2.63	3.33	3.60

Table 2: The relative estimation errors ε of μ_a , μ_s and g parameters of the reconstruction algorithm for the four different noise levels on the absorbed density energy data.

4.5 Crosstalk effect

The crosstalk problem is often encountered in optical imaging when reconstructing simultaneously the μ_a and μ_s coefficients. It is worth noting that the simultaneous reconstruction of μ_a , μ_s and g parameters has not been previously reported in the literature, for authors' best knowledge. In this case the crosstalk problem and the inter-parameter effects become a challenge in the recovered images. In this work, we highlight the interest of the QPAT to simultaneously reconstruct the three optical properties while the conventional optical imaging fails to perform this task. We present a test case that mimics a crosstalk problem between μ_a , μ_s and g wherein their original images are depicted in Figs. 6a-c, respectively. The same spatial mesh that was previously presented (2,821 nodes) was used. The phantom was illuminated by four Laser sources. The corresponding recovered images are shown in Figs. 6d-e. The reconstructed results show that the crosstalk effect is only presented in the μ_s and g images. The impact of the g -insert appears with a high contrast heterogeneity in the μ_s image ($\varepsilon_{\mu_s}^{\text{crosstalk}} = 12\%$) while the μ_s insert produces, in turn, a small contrast heterogeneity in the g image ($\varepsilon_g^{\text{crosstalk}} = 1.56\%$). Therefore, the crosstalk error induced in the μ_s coefficient is more pronounced than that obtained for the anisotropy factor. This is due to the high sensitivity of this factor on the light scattering. On the other hand, no crosstalk effect was found in the μ_a image which is reconstructed in a very good agreement with its original image. Moreover, the μ_a inclusion has no impact on the μ_s and g images. That can be explained by the fact the absorbed energy density is directly dependent on the μ_a coefficient, which makes the fitted data more sensitive for μ_a than the μ_s and g parameters. Within the minimization scheme, when the algorithm is seeking to simultaneously reconstruct the three parameters (μ_a , μ_s and g), the μ_a image is accurately and quickly obtained after only a few iterations of the convergence. Therefore, the μ_s and g coefficients are reconstructed as in the case that assuming the μ_a coefficient is fixed to its exact value. The superior quality reconstruction of μ_a , under these crucial situations, implies that the QPAT has an important interest for pre-clinical applications because the μ_a coefficient can lead to further physiological properties such as oxygen saturation, hemoglobin concentration, blood oxygenation, etc.

The crosstalk is a challenging problem in QPAT which has been reported in previous works related to optical imaging. The crosstalk is induced when the μ_a coefficient is only changed for a particular tissue due to physiological

354 variation while its μ_s coefficient remains unchanged. Its practical implications can mainly concern the blood tumors
 355 or other tissues which not containing fibrous. The crosstalk assessment for cancer diagnosis can allow highlighting
 356 the robustness of the imaging system to handle the false positive tumoral inclusions. In order to overcome its effect,
 357 more data are needed in the reconstruction algorithm to reduce the uniqueness character of the inverse problem.

Figure 6: Simultaneous reconstruction of μ_a , μ_s and g with a crosstalk problem: (left column) original images and (right column) reconstructed images, (top row) absorption coefficient μ_a , (middle row) scattering coefficient μ_s and (bottom row) anisotropy factor g .

4.6 Slabs inserts reconstruction

In this case, we assess the performance of the QPAT to reconstruct rectangular heterogeneities with different thicknesses and distance separations. To this end, two examples are studied. In the first example, four thin slabs of $200\ \mu\text{m}$ of thickness separated by $2\ \text{mm}$ are inserted in the phantom (see Fig. 7 top-row). In the second example, three slabs have $1\ \text{mm}$ of thickness with $50\ \mu\text{m}$ of separation (see Fig. 7 bottom-row). The spatial mesh has been increased to 33,025 nodes for both cases to suitably represent the thin inserts and the small separation. The phantoms were illuminated by four Laser sources. The reconstructed results of the first and second example are shown in the left and right column of Fig. 8, respectively. Also, the reconstructions were achieved without a crosstalk problem.

Figure 7: The original images containing the rectangular slabs inserts of the first example (top row) and the second example (bottom row). The right column shows the exact values of the anisotropy factor while the left column indicates the exact values for the μ_a and μ_s coefficients.

Figure 8 shows the robustness of our QPAT algorithm to accurately reconstruct the thin slabs heterogeneities (left column in Fig. 8) and also to precisely separate the small-inter-distance thick inserts (right column in Fig. 8). The

algorithm is able to retrieve the localization, the size, the thicknesses and also the local optical values μ_a , μ_s and g of the rectangular slabs inserts. Therefore, the reconstruct images were achieved with a good quantitative and qualitative accuracy. For both examples, these slabs are recovered with a high contrast level, contrary to optical imaging. The use of the local absorbed energy density, related to the initial acoustic pressure, has advantageously provided a potential improvement of the spatial resolution to the optical properties images.

Figure 8: The reconstructed images of: (left column) the first example and (right column) the second example, (top row) the absorption coefficient μ_a , (middle row) the scattering coefficient μ_s and (bottom row) the anisotropy factor g .

5 Conclusion

The optical inverse problem of the QPAT using the RTE as light transport model was presented. The adjoint method was applied to efficiently compute the gradient of the objective function where its expressions for μ_a , μ_s , and g coefficients were explicitly obtained. For the first time, the simultaneous reconstruction of μ_a , μ_s , and g was possible thanks to our QPAT algorithm that uses spatially absorbed energy density data. The results showed that treating the anisotropy parameter as a fixed constant leads to low errors in the reconstructed images of the absorbing coefficient and significant errors in the reconstructed images of the scattering coefficient. Also, they showed that the μ_a images are reconstructed with a better estimation quality than μ_s and g images even with noisy data or when only one Laser source was used. Furthermore, the μ_a images are insensitive to the crosstalk issue contrary to those of μ_s and g parameters. The obtained results highlight the interest of the QPAT algorithm to advantageously complete the conventional imaging modalities for cancer diagnosis. This work was a necessary preliminary study to show that a complete optical imaging of tissue is possible through the QPAT modality. The extension of our method to 3D geometries for real applications is a significant numerical challenge which is not straightforward. We plan to investigate this problem.

Appendix A. Calculations of adjoint operators

It should be noted that in the calculations given below, the quantities are not divided by M^2 as in (22) for the sake of simplicity. This does not change the result. From (12), we have:

$$\begin{aligned}
 \left\langle A - M \left| (\tilde{H} \delta\psi_s) \right. \right\rangle_{\mathcal{D}} &= \int_{\mathcal{D}} (A - M)(\mathbf{r}) \mu_a(\mathbf{r}) \int_{\Omega=2\pi} \delta\psi_s(\mathbf{r}, \mathbf{\Omega}) d\Omega dr \\
 &= \int_{\mathcal{D}} \int_{\Omega=2\pi} (A - M)(\mathbf{r}) \mu_a(\mathbf{r}) \delta\psi_s(\mathbf{r}, \mathbf{\Omega}) d\Omega dr = \left\langle A - M \left| (H \delta\psi_s) \right. \right\rangle_{\mathcal{D}\Omega} \\
 &= \int_{\mathcal{D}} \int_{\Omega=2\pi} \mu_a(\mathbf{r}) (A - M)(\mathbf{r}) \delta\psi_s(\mathbf{r}, \mathbf{\Omega}) d\Omega dr = \left\langle H^*(A - M) \left| \delta\psi_s \right. \right\rangle_{\mathcal{D}\Omega}.
 \end{aligned} \tag{38}$$

Also,

$$\begin{aligned} \left\langle A - M \middle| (H \delta\psi_c) \right\rangle_{\mathcal{D}} &= \int_{\mathcal{D}} (A - M)(\mathbf{r}) \mu_a(\mathbf{r}) \delta\psi_c(\mathbf{r}, \mathbf{\Omega}) dr \\ \int_{\mathcal{D}} \mu_a(\mathbf{r}) (A - M)(\mathbf{r}) \delta\psi_c(\mathbf{r}, \mathbf{\Omega}) dr &= \left\langle H^*(A - M) \middle| \delta\psi_c \right\rangle_{\mathcal{D}}. \end{aligned} \quad (39)$$

³⁸⁹ Thus, $H^* = H$.

Bibliography

- [1] Pinkert M A, Salkowski L R, Keely P J, Hall T J, Block W F and Eliceiri K W 2018 Review of quantitative multiscale imaging of breast cancer *J. Med. Imaging* **5**(1) 010901
- [2] Anastasio M A, Beard P C and Zhu Q 2017 Photoacoustic imaging and sensing *J. Biomed. Opt.* **22**(4) 041001
- [3] Wang L V and Yao J 2016 A practical guide to photoacoustic tomography in the life sciences *Nat. Meth.* **13** 627-638
- [4] Xia J and Wang L V 2014 Small-animal whole-body photoacoustic tomography: a review *Phys. Med. Biol.* **61**(5) 1380-1389
- [5] Anastasio M A and Beard P C 2012 Photoacoustic imaging and sensing *J. Biomed. Opt.* **17**(6) 061201
- [6] Wang L V and Hu S 2012 Photoacoustic tomography: in vivo imaging from organelles to organs *Science* **335**(6075) 1458-1462
- [7] Cox B, Laufer J G, Arridge S R and Beard P C 2012 Quantitative spectroscopic photoacoustic imaging: a review *J. Biomed. Opt.* **17**(6) 061202
- [8] Beard P 2011 Biomedical photoacoustic imaging *Interface Focus* **1**(4) 602-631
- [9] Ntziachristos V 2010 Going deeper than microscopy: the optical imaging frontier in biology *Nat. Meth.* **7**(8) 603
- [10] Li C and Wang L V 2009 Photoacoustic tomography and sensing in biomedicine *Phys. Med. Biol.* **54**(19) R59

- 407 [11] Wang L V 2009 Photoacoustic imaging and spectroscopy, CRC Press, Florida
- 408 [12] Xu M and Wang L V 2006 Photoacoustic imaging in biomedicine *Rev. Scient. Instrum.* **77**(4) 041101
- 409 [13] Ntziachristos V, Ripoll J, Wang L V and Weissleder R 2005 Looking and listening to light: the evolution of
410 whole-body photonic imaging *Nat. Biotechnol.* **23**(3) 313
- 411 [14] Ripoll J 2012 Principles of diffuse light propagation. Light propagation in tissues with applications in biology
412 and medicine (Singapore: World Scientific)
- 413 [15] Yao L, Sun Y and Jiang H 2010 Transport-based quantitative photoacoustic tomography: Simulations and
414 experiments *Phys. Med. Biol.* **55**(7) 1917-1934
- 415 [16] Tarvainen T, Cox B T, Kaipio J P and Arridge S R 2012 Reconstructing absorption and scattering distributions
416 in quantitative photoacoustic tomography *Inverse Probl.* **28**(8) 084009
- 417 [17] Saratoon T, Tarvainen T, Cox B T and Arridge S R 2013 A gradient-based method for quantitative photoa-
418 coustic tomography using the radiative transfer equation *Inverse Probl.* **29**(7) 075006
- 419 [18] Jacques S L 2013 Optical properties of biological tissues: a review *Phys. Med. Biol.* **58**(11) R37
- 420 [19] Wang Z, Tangella K, Balla A, Popescu G 2011 Tissue refractive index as marker of disease, *J Biomed. Opt.*
421 **16**(11) 116017
- 422 [20] Hillegersberg R V, Pickering J W, Maurice A, and Beek J F 1993 Optical properties of rat liver and tumor at
423 633 nm and 1064 nm: Photofrin enhances scattering *Lasers Surg. Med.* **13**(1) 31-39
- 424 [21] Germer G T, Andre R, Ritz J P, Christoph I, Dirk A, Gerhard M and Buhr H J 1998 Optical properties of
425 native and coagulated human liver tissue and liver metastases in the near infrared range *Lasers Surg. Med.*
426 **23**(4) 194-203
- 427 [22] Yao L, Sun Y and Jiang H 2009 Quantitative photoacoustic tomography based on the radiative transfer equa-
428 tion *Opt. Lett.* **34**(12) 1765-1767

- 429 [23] Bal G, Jollivet A and Jugnon V 2010 Inverse transport theory of photoacoustics *Inverse Probl.* **26** 025011
- 430 [24] Xi L, Li X, Yao L, Grobmyer S and Jiang H 2012 Design and evaluation of a hybrid photoacoustic tomography
431 and diffuse optical tomography system for breast cancer detection *Med. Phys.* **39**(5) 2584-2594
- 432 [25] Gao H, Osher S and Zhao H 2012 Quantitative photoacoustic Tomography *Lecture Notes Math.* **2035** 131-158
- 433 [26] Mamonov A V and Ren K 2013 Quantitative photoacoustic imaging in the radiative transport regime *Commun.*
434 *Math. Sci.* **12**(2) 201-234
- 435 [27] Haltmeier M, Neumann L and Rabanser S 2015 Single-stage reconstruction algorithm for quantitative pho-
436 toacoustic tomography *Inverse Probl.* **31**(6) 065005
- 437 [28] De Cezaro A, Travessini De Cezaro F and Seije Suarez J 2015 Regularization approaches for quantitative
438 Photoacoustic tomography using the radiative transfer equation *J. Math. Anal. Appl.* **429**(1) 415-438
- 439 [29] Pulkkinen A, Cox B T, Arridge S R, Kaipio J P and Tarvainen T 2015 Quantitative photoacoustic tomography
440 using illuminations from a single direction *J. Biomed. Opt.* **20**(3) 036015
- 441 [30] Li X, Heldermon C D, Yao L, Xi L and Jiang H 2015 High resolution functional photoacoustic tomography
442 of breast cancer *Med. Phys.* **42**(9) 5321-5328
- 443 [31] Hochuli R, Powell S, Arridge S R and Cox B 2016 Quantitative photoacoustic tomography using forward and
444 adjoint Monte Carlo models of radiance *J. Biomed. Opt.* **21**(12) 126004
- 445 [32] Tarvainen T, Pulkkinen A, Cox B T and Arridge S R 2017 Utilizing the radiative transfer equation in quan-
446 titative photoacoustic tomography Proc. of SPIE Vol. 10064, Photons Plus Ultrasound: Imaging and Sensing,
447 100643E-1
- 448 [33] Wang C and Zhou T 2017 On iterative algorithms for quantitative photoacoustic tomography in the radiative
449 transport regime *Inverse Probl.* **33**(11) 115006

- 450 [34] Okawa S, Hirasawa T, Kushibiki T and Ishihara M 2017 Effects of the approximations of light propagation
451 on quantitative photoacoustic tomography using two-dimensional photon diffusion equation and linearization
452 *Opt. Rev.* **24**(6) 705-726
- 453 [35] Haltmeier M, Neumann L, Nguyen L and Rabanser S 2018 Analysis of the linearized problem of quantitative
454 photoacoustic tomography *SIAM J. Appl. Math.* **78**(1) 457-478
- 455 [36] Bal G and Ren K 2011 Multi-source quantitative photoacoustic tomography in a diffusive regime *Inverse*
456 *Probl.* **27** 075003
- 457 [37] Shao P, Cox B, and Zemp R J 2011 Estimating optical absorption, scattering, and Grueneisen distributions
458 with multiple-illumination photoacoustic tomography *Appl. Opt.* **50**(19) 3145-3154
- 459 [38] Addoum A, Contassot-Vivier S and Asllanaj F 2019 Three-dimensional frequency-domain optical anisotropy
460 imaging of biological tissues with near-infrared light *Med. Phys.* **46**(9) 4057-4069
- 461 [39] Asllanaj F, Addoum A and Roche J R 2018 Fluorescence molecular imaging based on the adjoint radiative
462 transport equation *Inverse Probl.* **34** 075009
- 463 [40] Addoum A, Farges O and Asllanaj F 2018 Optical properties reconstruction using the adjoint method based
464 on the radiative transfer equation *J. Quant. Spectrosc. Radiat. Transfer* **204** 179-189
- 465 [41] Saratoon T, Tarvainen T, Arridge S R and Cox B T 2013 3D quantitative photoacoustic tomography using
466 the delta-Eddington approximation Proc. of SPIE Vol. 8581, Photons Plus Ultrasound: Imaging and Sensing,
467 85810V
- 468 [42] Asllanaj F, Contassot-Vivier S, Liemert A and Kienle A 2014 Radiative transfer equation for predicting light
469 propagation in biological media: comparison of a modified finite volume method, the Monte Carlo technique
470 and an exact analytical solution *J. Biomed. Opt.* **19** 015002
- 471 [43] Klose A D 2009 Light Scattering. Reviews 4, Book Chapter: Radiative transfer of luminescence light in
472 biological tissue (Berlin Heidelberg: Springer:293-345)

- 473 [44] Fiacco A V and McCormick G P 1968 Nonlinear Programming: Sequential Unconstrained Minimization
474 Techniques (John Wiley and Sons)
- 475 [45] Bonnans J F, Gilbert J C, Lemarechal C and Sagastizabal C A 2016 Numerical Optimization Theoretical and
476 Practical Aspects (Springer Sciences and Business Media)
- 477 [46] Klose A 2001 Optical tomography based on the radiative transfer equation. PhD thesis, Department of Physics
478 Freie Universitat, Berlin Germany
- 479 [47] Klose A D, Ntziachristos V and Hielscher A H 2005 The inverse source problem based on the equation of
480 radiative transfer in optical molecular imaging *J. Comp. Phys.* **202** 323-345