Line-scan Compressive Raman imaging with spatio-spectral encoding

Camille Scotté, Siddharth Sivankutty, Randy A Bartels, Hervé Rigneault

► To cite this version:

Camille Scotté, Siddharth Sivankutty, Randy A Bartels, Hervé Rigneault. Line-scan Compressive Raman imaging with spatio-spectral encoding. *Optics Letters*, 2020, 45 (19), pp.5567. 10.1364/OL.400151 . hal-02955440

HAL Id: hal-02955440

<https://hal.science/hal-02955440>

Submitted on 5 Oct 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Line-scan Compressive Raman imaging with spatio-spectral encoding

CAMILLE SCOTTÉ¹, SIDDHARTH SIVANKUTTY¹, RANDY A. BARTELS^{2,3,4}, AND HERVÉ RIGNEAULT^{1,5*}

¹Aix Marseille Univ, CNRS, Centrale Marseille, Institut Fresnel, Marseille, France

²W.M. Keck Laboratory for Raman Imaging of Cell-to-Cell Communications, Colorado State University, Fort Collins, Colorado 80523, USA

³Department of Electrical and Computer Engineering, Colorado State University, Fort Collins, Colorado 80523, USA

⁴School of Biomedical Engineering, Colorado State University, Fort Collins, Colorado 80523, USA

⁵Lightcore Technologies, Marseille, France

*Corresponding author: herve.rigneault@fresnel.fr

Compiled July 29, 2020

We report a line-scanning imaging modality of Compressive Raman technology with a single-pixel detector. The spatial information along the illumination line is encoded onto one axis of a digital micromirror device, while spectral coding masks are applied along the orthogonal direction. We demonstrate imaging and classification of three different chemical species. © 2020

Optical Society of America

<http://dx.doi.org/10.1364/ao.XX.XXXXXX>

Spontaneous Raman spectroscopy permits characterization of molecular species with high specificity in a label-free manner. Typically, the Raman inelastically scattered light is spectrally dispersed and collected onto an array detector - for several positions of the sample - resulting in a Raman hyperspectral image. This measurement of a complete vibrational Raman spectrum per spatial pixel, coupled with the weak Raman scattering cross-section and detector array noise, requires lengthy acquisition times and generates large data sets. In situations where hyperspectral measurements simply aim to map the spatial distribution of known molecules, such an acquisition process is highly inefficient. Instead of unmixing the spectral data in a post-processing step to detect molecular species and/or estimate their concentrations [1, 2], higher acquisition efficiencies can be achieved by encompassing compressive techniques in the acquisition process. In Compressive Raman Technology (CRT), the measurement is directly designed to estimate quantities of interest (e.g., molecular concentrations), rather than deducing them from complete hyperspectral measurements [3–9]. This is achieved by replacing the array detector by a single-pixel detector combined with a fast programmable optical filter, typically a digital micromirror device (DMD). Based on the *a priori* knowledge of the Raman spectra of the pure molecular species contained in the sample, these filters select accurately chosen spectral components and combine them into the detector (Fig. 2). Such CRT developments have led to spontaneous Raman imaging with pixel dwell times down to 30 μ s [3, 10], which is orders of magnitude faster than state-of-the-art Raman hyperspectral imaging [7].

To further improve CRT, we recently demonstrated a CRT line-

scan strategy based on spatial frequency-modulated illumination imaging [11], referred to as CRISPY [12]. As opposed to recent spatial encoding approaches [13–15], CRISPY does not use an spatially-resolved detector but combines both spatial and spectral information into a single-pixel detector. It **encodes the 1D spatial information into temporal frequencies, by modulating modulates** the illumination line with chirped cosines imprinted on a rotating disk [11, 12]. In the present Letter, we explore an alternative line-scan CRT strategy that does not require an external modulator, but instead **makes use of the bi-dimensionality of the DMD exploits - for the first time in CRT - the bi-dimensionality of the DMD** (Fig. 2): (i) The Raman spectral information is dispersed across the λ -axis, and the DMD encodes this information with dedicated spectral filters; (ii) The spatial information along the line-focus is encoded onto the DMD x -axis, with Hadamard matrices shifted to contain only positive values. We demonstrate line-scan CRT with proof-of-concept experiments for chemically specific classification of molecular species, and give evidence that this approach is beneficial over point-scanning CRT in a number of cases.

In line-scan CRT, both spatial (along the line-focus) and spectral information is encoded in the signal measured onto the single-pixel detector. The aim is to estimate the relative abundances (proportions) of a set of Q molecular species (with known Raman spectra) in each of the N resolved points along the illumination line (Fig. 1). We denote P as the number of DMD patterns along the x -axis ($p = 1 \dots P$), Q as the number of pure chemical species present in the sample ($q = 1 \dots Q$), M as the number of spectral filters ($m = 1 \dots M$) and L as the number of resolved energy bins along a Raman spectrum ($l = 1 \dots L$). The $N \times Q$ matrix $\mathbf{C}$ is the quantity to estimate, i.e., the spatial distribution of the pure chemical species proportions. Each element of $\mathbf{C}$, $c_q(x_n)$, specifies the proportions of the q^{th} pure chemical species contained in the resolved point x_n of the illumination line. The $P \times N$ matrix $\mathbf{A}$ contains the P patterns displayed along the DMD x -axis. The $L \times M$ matrix $\mathbf{F}$ contains the M spectral filters $\mathbf{f}_m$. The $Q \times L$ matrix $\mathbf{S}$ contains the known Raman spectra of the pure chemical species. The measurements matrix, $\mathbf{H}$, contains the number of counts measured when displaying binary masks onto the DMD. Each mask is formed by the Kronecker product of a row of $\mathbf{A}$ and

a column of $\mathbf{F}$ (Fig. 2). Each column of $\mathbf{H}$ is recorded by keeping the spectral filter fixed, and scanning through the rows of $\mathbf{A}$ to obtain a complete set of spatial projections. Assuming the generated signal is linear to the excitation intensity and ignoring constant terms, the measurement process can be expressed as:

$$\mathbf{H} = \mathbf{A}\mathbf{C}\mathbf{S}\mathbf{F} = \mathbf{A}\mathbf{C}\mathbf{G}^T \quad (1)$$

where $\mathbf{G}^T = \mathbf{S}\mathbf{F}$.

Fig. 1. Visual representation of Eq. 1, with $Q = M = 2$.

Although the spatial and spectral dimensions could be considered conjointly, we handle them distinctly in this Letter. In the spectral domain, the filters $\mathbf{f}_m$ are calculated from the pure Raman spectra $\mathbf{s}_q$ with the same optimisation procedure as in [6]. In the spatial domain, we choose the matrix $\mathbf{A}$ to be a modified Hadamard matrix with positive coefficients (S-matrix). The S-matrix of size N is obtained by removing the first row and column of a Hadamard matrix of size $N + 1$, and changing the ones to zeros and minus ones to ones [16]. It is square ($P = N$), invertible, and its binary nature complies with the DMD design and its use at fast frame rates. **For each line**, we perform the estimation in two steps. First, we estimate a spatial line image of Raman intensities, for each spectral filter. In other words, we seek to estimate the global matrix $\mathbf{C}\mathbf{G}^T$, which elements $\eta_m(x_n)$ represent the Raman intensity in each pixel x_n , for the spectral filter $\mathbf{f}_m$. Since our measurements are shot-noise limited [7] and that $\eta_m(x_n) \geq 0$, we perform this estimation using the classical EM algorithm (Richardson-Lucy), which seeks to maximize the Poisson likelihood under positivity constraints [17–19]. **By reducing the estimation variance on the object pixels where the positivity constraint applies** **For relatively sparse objects**, it gives better performances in terms of mean-square error as compared to least-square estimation (data not shown), **for relatively sparse objects** [20–22]. **We will report more precisely on this effect in a forthcoming publication.** The resulting estimate is denoted $\hat{\eta}_m(x_n)$. After a transpose operation, the estimation of the species proportions reduces to a 1D CRT problem [6]:

$$\hat{\boldsymbol{\eta}}^T(x_n) = \mathbf{G}\mathbf{c}^T(x_n) \quad (2)$$

with $\hat{\boldsymbol{\eta}}(x_n) = (\hat{\eta}_1(x_n), \dots, \hat{\eta}_M(x_n))$ and $\mathbf{c}(x_n) = (c_1(x_n), \dots, c_Q(x_n))$. If $\mathbf{G}^T\mathbf{G}$ is not singular, the proportions **along each line** are finally estimated via least-square estimation [6]:

$$\hat{\mathbf{c}}^T(x_n) = [\mathbf{G}^T\mathbf{G}]^{-1}\mathbf{G}^T\hat{\boldsymbol{\eta}}^T(x_n) \quad (3)$$

The 2D image is built after performing these two estimation steps for each line of the sample.

A simplified schematic of the experimental setup is depicted in Fig. 2. On the illumination side, a continuous wave laser operating at 532 nm (Verdi, Coherent Inc) is spectrally filtered and expanded. The beam is brought to a line focus onto the sample plane with a combination of cylindrical lens, plano-convex lens and microscope objective (Nikon 20x, 0.5 NA). On the detection side, the scattered light from the object is relayed onto

Fig. 2. (a) Schematic of the experimental setup. $L_1 - L_6$ - convex lenses with focal lengths 50 mm, 150 mm, 150 mm, 150 mm, 100 mm, 150 mm and 50 mm, respectively. L_7 - combination of 2 lenses that image the DMD into the PMT with $\times 3$ de-magnification. L_{cyl} - cylindrical lens with focal length 150 mm, D - dichroic mirror, S - confocal slit, G - amplitude grating, PMT - photomultiplier tube. (b) The spectral axis gives access to the Raman spectrum, and allows to project optimized spectral filters. (c) The spatial axis gives access to the spatial information along the line focus.

a confocal slit. A combination of dichroic mirror and notch filter ensures only the Raman signal is retained. Next, it is dispersed with a blazed grating (600 mm^{-1} , Thorlabs), and the spatially dispersed wavelength components are imaged onto the DMD (V-7001, Vialux -1024 $\times$ 768 mirrors). The DMD λ -axis, in conjunction with the grating, acts as a programmable spectral filter. Since the line focus is imaged onto the DMD (≈ 22.5 de-magnification), its x -axis offers control over the corresponding spatial dimension of the object [Inset Fig. 2]. When the DMD pixels are in the 'ON' state, the signal impinging on these pixels is deflected into a photon-counting PMT (H7421-40, Hamamatsu), while the rest is sent into a beam dump. A piezoelectric stage scanner (P517, Physik Instrumente GmbH) holding the sample is used to scan the y -axis, yielding 2D images.

In this configuration, the spatial resolution along x - is limited by the imaging system and the size of one DMD mirror. In the experiments, the DMD mirrors are binned 2-by-2 along x -, resulting in an equivalent spatial resolution of about $1.2 \mu\text{m}$. The line length on the illumination plane, together with cropping from diverse optical elements, limits the field-of-view along x - to about $220 \mu\text{m}$. On the DMD plane, the spatial extent of the line image only represents 1/3 of the entire DMD size, but the S-patterns are displayed over the full DMD x -range ($P = 511$). In the y - direction, the spatial resolution is about $1.5 \mu\text{m}$ (extent of the line focus on the sample). The spectral resolution is estimated to 40 cm^{-1} , limited by the grating and the imaging system. This allows to bin the DMD mirrors 8-by-8 along λ -. Spatial calibration was performed by scanning a bead of known size and matching its profile onto the DMD. The

effective FOV was estimated by recording the Raman signal from a spatially homogeneous sample (microscope glass slide) and scanning it along the DMD x -axis.

The first experiment consisted in acquiring Raman line images with no spectral selectivity ($f_{lm} = 1 \forall l, m$) of two mixtures of beads (Sigma Aldrich). 30 μm polystyrene beads (PS), 20 μm polymethylmethacrylate (PMMA) and 12 μm melamin resin (MR) were dispersed in two different manners Fig. 3(b): (A) in a sparse way, onto a CaF_2 coverslip; (B) in a dense way, onto a glass coverslip. The total laser power on the illuminated line was set to 3.6 mW ($1.1 \times 10^{-5} \text{ W}/\mu\text{m}^2$), so that the total count rate lies in the linearity regime of the detector ($1.5 \times 10^6 \text{ Hz}$). Fig. 3(a) shows the signal obtained upon the projection of each S-pattern for 10 ms, averaged over 50 measurements, and Fig. 3(b) the associated estimated Raman line images. The signal along the lines of sample B is higher due to the presence of more chemical species along x - [Fig. 3(a, c)]. No further post-processing was performed, except from the removal of some systematic periodic artefacts on the estimated images, arising from yet unidentified technical issues.

Fig. 3. (a) Signal after projection of S-patterns along each line x -, for sample A and B. (b) Estimated images obtained with the EM algorithm. (c-d) Cross-sections along the dotted lines of (a, b). Results are averaged over 50 measures, and cropped to show only the FOV. Scale bar = 30 μm . x -: Line-scan axis, y -: Piezo-scan axis. Colorbars units: counts.

Next, we performed CRT experiments to obtain Raman images with chemical selectivity. The reference spectra were measured with using the DMD x -axis as a virtual pinhole [12] [Fig. 2], and the spectral filters f_m optimised as in [6, 7] - see Fig. 4(a). The resulting spectra and filters for sample A ($Q = M = 4$) are shown in red. In CRT experiments, the total laser power along the line was chosen higher ($10.6 \text{ mW} - 3.3 \times 10^{-5} \text{ W}/\mu\text{m}^2$) than for pure Raman line imaging described above, because of the lower signal due to the spectral filtering. The 4 line-images, corresponding to each spectral filter, are estimated, followed by the proportion estimation in each pixel along the line via Eq. 3. The resulting proportion maps for sample A, thresholded to [0 1], are shown in Fig. 4(b). The composite RGB maps are obtained by combining the normalised proportion maps of the three beads types. The results are shown for integration times of 10 ms and 1 ms per projected S-pattern, with no averaging. As a comparison, the same experiment was performed with

raster-scanning the DMD pixels along x - [Fig 2(c-left)], with same irradiance and integration times. This is formally equivalent to scanning the sample plane with a point-focus. Visually, the line-scan measurements seem superior to the raster-scan measurements, but this disparity is more striking for the sparse sample A than for sample B. The difference in image quality is quantitatively confirmed by the SNR values on the central pixels of a PS bead (Fig. 4(e-g)), defined as $\text{SNR}_{PS} = \langle \hat{c}(x_{PS}) \rangle / \sigma_{PS}$, where $\langle \hat{c}(x_{PS}) \rangle$ and σ_{PS} are the mean and standard deviation of the estimated proportions, respectively. On the PS pixels, the SNR improvement over raster-scanning is about 3 for object A, and 1.5 for object B.

Fig. 4. (a) Reference spectra and associated spectral filters, for sample A. (b) Estimated proportion maps with line-scan. (c-f) Visualization of proportion maps of PS (red), PMMA (green) and MR (blue), for line-scanning and raster-scanning.

To better assess the efficiency of the presented line-scanning technique over point-scanning, we compare the estimation variances of the corresponding Raman line images, with no spectral selectivity (Fig. 5). For sample A, the variance on the object is lower for line-scanning than for raster-scanning. For example, on the PS bead along central line of the sample, the variance σ^2 is about 3 times lower, leading to a SNR improvement of $\sqrt{3}$ (SNR on pixel i : $SNR_i = \langle \hat{\eta}(x_i) \rangle / \sigma_i$). On the PMMA bead 1, the variance is about 6 times lower for line-scanning, while for PMMA bead 2, it is only about 2 times lower. For sample B, line-scanning leads to a relatively uniform variance along the line which is similar to raster-scanning. This leads to a limited SNR improvement. These results are consistent with the CRT results of Fig. 4. They show that, in terms of SNR, line-scanning seems mostly beneficial for the lines along which the sample is sparse for the sparse sample A, and brings a limited improvement to the denser sample B. This stems from the shot-noise limited nature of our measurements [7]. Here, the noise scales with the square-root of the number of photons arising from the illuminated pixels across the line [23]. In addition, the signal from pixel x_n may be affected by photon noise coming from the average Raman signal along the entire line. This results in an object-dependent SNR improvement, which is expected to increase for sparser objects [20–23]. This is in contrast to measurements where detectors exhibit signal-independent additive noise [16]: In this case, the same S-multiplexing strategy would lead to an SNR improvement of $\sqrt{N}/2$, i.e. of 10 times, for both objects. The precise mechanisms for the dependence of SNR improvement due to the sparsity of the sample are out of the scope of this paper and will be investigated in more details in a forthcoming publication. These aspects will be investigated in more details in a forthcoming publication.

From this short analysis, we surmise that our line-scan CRT approach would bring a SNR advantage over point-scanning CRT, and thus be faster, when samples are relatively sparse or when the experimental configuration exhibits signal-independent noise contributions. We emphasize that the spectral filtering induced by CRT renders the samples more sparse, and thus further improves the SNR, as long as the chemical species spectral overlap is limited. The SNR and speed gain will then depend on the spatial and spectral structure of the sample along the multiplexed line.

In this Letter, we have demonstrated a simple approach for line-scan CRT with S-multiplexing, with proof-of-concept experiments. While a line-scan strategy includes some drawbacks such as non-uniform illumination profile [24] or potential loss in spatial or spectral resolution, we surmise line-scan CRT to be beneficial over point-scanning CRT for SNR for relatively sparse objects in many circumstances. While the mentioned speeds do not compete with point-scanning CRT, they are expected to be faster with approaches such as compressive-sensing [25, 26]. In addition, the results could further be improved by filters design and optimisation taking into account the spatial and spectral domain conjointly.

Funding Centre National de la Recherche Scientifique (CNRS), Institut Carnot STAR (IRAC 2018), Aix-Marseille University A*Midex (noANR-11-IDEX-0001-02 and A-M-AAP-ID-17-13-170228-15.22-RIGNEAULT), ANR grants France Bio Imaging (ANR-10-INSB-04-01), France Life Imaging (ANR-11-INSB-0006) infrastructure networks, Plan cancer INSERM PC201508 and

Fig. 5. Variance images (50 measurements), for line-scanning and raster-scanning. (a-b): sample A. (c-d): sample B. (e-f) Cross-sections along the dotted lines of (a-d).

18CP128-00. C.S. has received funding from the European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie grant agreement No713750.

Acknowledgements The authors acknowledge fruitful discussion with P. Réfrégier and F. Galland.

Disclosures Hervé Rigneault has financial interest in Lightcore Technologies SA, Marseille.

REFERENCES

- N. Keshava and J. Mustard, IEEE Signal Process. Mag. **19**, 44 (2002).
- R. D. Palkki, Opt. Eng. **49**, 113601 (2010).
- D. S. Wilcox, G. T. Buzzard, B. J. Lucier, P. Wang, and D. Ben-Amotz, Anal. Chimica Acta **755**, 17 (2012).
- D. S. Wilcox, G. T. Buzzard, B. J. Lucier, O. G. Rehrauer, P. Wang, and D. Ben-Amotz, The Analyst **138**, 4982 (2013).
- B. J. L. Gregery T. Buzzard, ProcSPIE **8657**, 8657 (2013).
- P. Réfrégier, C. Scotté, H. B. de Aguiar, H. Rigneault, and F. Galland, J. Opt. Soc. Am. A **35**, 125 (2018).
- C. Scotté, H. B. De Aguiar, D. Marguet, E. M. Green, P. Bouzy, S. Vergnole, C. P. Winlove, N. Stone, and H. Rigneault, Anal. Chem. **90**, 7197 (2018).
- D. Cebeci, B. Mankani, and D. Ben-Amotz, J. Imaging **5**, 1 (2018).
- B. Sturm, F. Soldevila, E. Tajahuerce, S. Gigan, H. Rigneault, and H. B. De Aguiar, ACS Photonics **6**, 1409 (2019).
- O. G. Rehrauer, V. C. Dinh, B. R. Mankani, G. T. Buzzard, B. J. Lucier, and D. Ben-Amotz, Appl. Spectrosc. **1**, 69 (2018).
- G. Futia, P. Schlup, D. G. Winters, and R. a. Bartels, Opt. express **19**, 1626 (2011).
- C. Scotté, S. Sivankutty, P. Stockton, R. A. Bartels, and H. Rigneault, Opt. Lett. **44**, 1936 (2019).
- D. Galvis-Carreño, Y. Mejía-Melgarejo, and H. Arguello-Fuentes, DYNA (Colombia) **81**, 116 (2014).
- J. Thompson, J. N. Bixler, J. N. Bixler, M. O. V. Scully, G. D. Noojin, V. V. Yakovlev, and B. H. Hokr, Opt. Lett. **42**, 2169 (2017).
- H. Choi, P. T. C. So, H. Heemskerk, Z. Yaqoob, P. Matsudaira, V. R. Singh, and D. N. Wadduwage, Optica. **4**, 546 (2017).
- M. Harwit, N. J. A. Sloane, I. Sloane, and N. James, Hadamard Transform Optics (Academic Press, 1979).
- W. H. Richardson, J. Opt. Soc. Am. **62**, 55 (1972).
- L. B. Lucy, The Astron. J. **79**, 745 (1974).
- L. A. Shepp and Y. Vardi, IEEE Transactions on Med. Imaging **MI-1**, 113 (1982).

20. S. E. Bialkowski, *Appl. Spectrosc.* **52**, 591 (1998).
21. D. R. Fuhrmann, C. Preza, J. A. O'sullivan, D. L. Snyder, and W. H. Smith, *IEEE Transactions on signal processing* **52**, 950 (2004).
22. Z. T. Harmany, R. F. Marcia, and R. M. Willett, *IEEE Transactions on Image Process.* **21**, 1084 (2012).
23. V. Studer, J. Bobin, M. Chahid, H. S. Mousavi, E. Candes, and M. Dahan, *Proc. Natl. Acad. Sci.* **109**, E1679 (2012).
24. J. Hutchings, C. Kendall, B. Smith, N. Shepherd, H. Ban, and N. Stone, *J. Biophotonics* **2**, 91 (2009).
25. E. Candes and T. Tao, *IEEE Transactions on Inf. Theory* **52**, 5406 (2006).
26. P. Berto, C. Scotté, F. Galland, H. Rigneault, and H. B. de Aguiar, *Opt. Lett.* **42**, 1696 (2017).

FULL REFERENCES

1. N. Keshava and J. Mustard, "Spectral Unmixing," *IEEE Signal Process. Mag.* **19**, 44–57 (2002).
2. R. D. Palkki, "Chemical mixture estimation under a Poisson Raman spectroscopy model," *Opt. Eng.* **49**, 113601 (2010).
3. D. S. Wilcox, G. T. Buzzard, B. J. Lucier, P. Wang, and D. Ben-Amotz, "Photon level chemical classification using digital compressive detection," *Anal. Chimica Acta* **755**, 17–27 (2012).
4. D. S. Wilcox, G. T. Buzzard, B. J. Lucier, O. G. Rehrauer, P. Wang, and D. Ben-Amotz, "Digital compressive chemical quantitation and hyperspectral imaging," *The Analyst* **138**, 4982–90 (2013).
5. B. J. L. Gregory T. Buzzard, "Optimal filters for high-speed compressive detection in spectroscopy," *ProcSPIE* **8657**, 8657 – 8657 – 10 (2013).
6. P. Réfrégier, C. Scotté, H. B. de Aguiar, H. Rigneault, and F. Galland, "Precision of proportion estimation with binary compressed Raman spectrum," *J. Opt. Soc. Am. A* **35**, 125 (2018).
7. C. Scotté, H. B. De Aguiar, D. Marguet, E. M. Green, P. Bouzy, S. Vergnole, C. P. Winlove, N. Stone, and H. Rigneault, "Assessment of compressive Raman versus hyperspectral Raman for microcalcification chemical imaging," *Anal. Chem.* **90**, 7197–7203 (2018).
8. D. Cebeci, B. Mankani, and D. Ben-Amotz, "Recent Trends in Compressive Raman Spectroscopy Using DMD-Based Binary Detection," *J. Imaging* **5**, 1 (2018).
9. B. Sturm, F. Soldevila, E. Tajahuerce, S. Gigan, H. Rigneault, and H. B. De Aguiar, "High-Sensitivity High-Speed Compressive Spectrometer for Raman Imaging," *ACS Photonics* **6**, 1409–1415 (2019).
10. O. G. Rehrauer, V. C. Dinh, B. R. Mankani, G. T. Buzzard, B. J. Lucier, and D. Ben-Amotz, "Binary Complementary Filters for Compressive Raman Spectroscopy," *Appl. Spectrosc.* **1**, 69–78 (2018).
11. G. Futia, P. Schlup, D. G. Winters, and R. a. Bartels, "Spatially-chirped modulation imaging of absorption and fluorescent objects on single-element optical detector," *Opt. express* **19**, 1626–1640 (2011).
12. C. Scotté, S. Sivankutty, P. Stockton, R. A. Bartels, and H. Rigneault, "Compressive raman imaging with spatial frequency modulated illumination," *Opt. Lett.* **44**, 1936–1939 (2019).
13. D. Galvis-Carreño, Y. Mejía-Melgarejo, and H. Arguello-Fuentes, "Efficient reconstruction of Raman spectroscopy imaging based on compressive sensing," *DYNA (Colombia)* **81**, 116–124 (2014).
14. J. Thompson, J. N. Bixler, J. N. Bixler, M. O. V. Scully, G. D. Noojin, V. V. Yakovlev, and B. H. Hokr, "Single-shot chemical detection and identification with compressed hyperspectral Raman imaging," *Opt. Lett.* **42**, 2169 (2017).
15. H. Choi, P. T. C. So, H. Heemskerk, Z. Yaqoob, P. Matsudaira, V. R. Singh, and D. N. Wadduwage, "Near-common-path interferometer for imaging Fourier-transform spectroscopy in wide-field microscopy," *Optica* **4**, 546 (2017).
16. M. Harwit, N. J. A. Sloane, I. Sloane, and N. James, *Hadamard Transform Optics* (Academic Press, 1979).
17. W. H. Richardson, "Bayesian-Based Iterative Method of Image Restoration," *J. Opt. Soc. Am.* **62**, 55–59 (1972).
18. L. B. Lucy, "An iterative technique for the rectification of observed distributions," *The Astron. J.* **79**, 745–754 (1974).
19. L. A. Shepp and Y. Vardi, "Maximum Likelihood Reconstruction for Emission Tomography," *IEEE Transactions on Med. Imaging* **MI-1**, 113–122 (1982).
20. S. E. Bialkowski, "Overcoming the multiplex disadvantage by using maximum-likelihood inversion," *Appl. Spectrosc.* **52**, 591–598 (1998).
21. D. R. Fuhrmann, C. Preza, J. A. O'sullivan, D. L. Snyder, and W. H. Smith, "Spectrum Estimation from Quantum-Limited Interferograms," *IEEE Transactions on signal processing* **52**, 950–961 (2004).
22. Z. T. Harmany, R. F. Marcia, and R. M. Willett, "This is SPIRAL-TAP: Sparse poisson intensity reconstruction algorithms-theory and practice," *IEEE Transactions on Image Process.* **21**, 1084–1096 (2012).
23. V. Studer, J. Bobin, M. Chahid, H. S. Mousavi, E. Candes, and M. Dahan, "Compressive fluorescence microscopy for biological and hyperspectral imaging," *Proc. Natl. Acad. Sci.* **109**, E1679–E1687 (2012).
24. J. Hutchings, C. Kendall, B. Smith, N. Shepherd, H. Ban, and N. Stone, "The potential for histological screening using a combination of rapid Raman mapping and principal component analysis," *J. Biophotonics* **2**, 91–103 (2009).
25. E. Candes and T. Tao, "Near Optimal Signal Recovery from Random Projections," *IEEE Transactions on Inf. Theory* **52**, 5406 – 5425 (2006).
26. P. Berto, C. Scotté, F. Galland, H. Rigneault, and H. B. de Aguiar, "Programmable single-pixel-based broadband stimulated Raman scattering," *Opt. Lett.* **42**, 1696 (2017).