



Controlling upper-limb prostheses with body compensations

Mathilde Legrand, Nathanael Jarrasse, Charlotte Marchand, Florian Richer, Amélie Touillet, Noël Martinet, Jean Paysant, Guillaume Morel

► To cite this version:

Mathilde Legrand, Nathanael Jarrasse, Charlotte Marchand, Florian Richer, Amélie Touillet, et al.. Controlling upper-limb prostheses with body compensations. WeRob, Oct 2020, Valencia, Spain. hal-03038844

HAL Id: hal-03038844

<https://hal.science/hal-03038844>

Submitted on 3 Dec 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Controlling upper-limb prostheses with body compensations

Mathilde Legrand, Nathanaël Jarrassé, Charlotte Marchand, Florian Richer, Amélie Touillet, Noël Martinet, Jean Paysant and Guillaume Morel

Abstract—With their advanced mechatronics, myoelectric upper-limb prostheses now have many motion possibilities. Yet, the latter are not fully employed because of the inconvenient control and prostheses wearers often use their device as a rigid tool while achieving hand positioning and orientation with compensatory movements. In this paper, we propose to take advantage of this natural human behaviour to control prosthesis motions: the user is in charge of the end-effector while the device role is to correct the human posture when necessary. We here apply this concept to control a prosthetic wrist pronosupination. A Rolyan clothespin test performed by two transradial amputees shows that the proposed control is as efficient as myoelectric control while requiring no learning.

I. INTRODUCTION

With the latest advances of mechatronics, the motion possibilities of upper-limb (UL) prostheses have increased a lot [1], [2] but they are not fully exploited because the control of these devices remains highly challenging.

The most widespread approach directly connects electromyographic signals generated by the user to the movement of the prosthesis. It includes conventional myoelectric control (on/off or proportional), integrated into most of the commercially available devices, and pattern-recognition based myoelectric control. Performant enough when controlling one degree of freedom (DOF), these two schemes become limited with an increasing number of DOF [3]. The muscular fatigue and the mental burden induced by these control approaches often lead prostheses users to employ their device as a rigid tool and move their end-effector with compensatory movements (see [4] e.g.). Body compensations are indeed efficient to achieve many tasks, but they are to be avoided since they cause musculoskeletal disorders [5].

In this paper, we propose to take advantage of this natural reaction of prostheses users to control the device: the end-effector task (placing and orientating the hand) will be achieved by the human subject while prosthesis motions will correct the human posture. This will be performed by servoing prosthetic joints motions to the user's compensatory movements. Note that this scheme is valid to control intermediate joints like wrist, elbow and shoulder, but not for hand since no body compensation substitutes grasping functions.

This paradigm has already been explored and validated

This work was supported by the ANR-BYCEPS, ANR-18-CE19-0004.

M. Legrand, N. Jarrassé, C. Marchand, F. Richer and G. Morel are with the Institute of Intelligent Systems and Robotics, Sorbonne University, CNRS UMR7222, INSERM U1150, Paris. Emails: name.surname@sorbonne-universite.fr

A. Touillet, N. Martinet and J. Paysant are with the Institut Régional de Réhabilitation, IRR UGECAM Nord-Est, Nancy. Emails: name.surname@uecam.assurance-maladie.fr

for the control of wrist pronosupination with able-bodied subjects [6]. In the work presented here, we extend this study to transradial amputated people, on a different task.

II. MATERIALS AND METHODS

Control law

The control law implemented to servo the prosthetic wrist pronosupination to the user's compensatory motions is the same as the one presented in [6]; we briefly recall it below. To supplant wrist pronosupination, amputated subjects tend to use trunk and arm compensatory motions to change the hand orientation [4]. To avoid an artificial mapping between these motions and prosthesis' motions, we directly measured the forearm rotation around its longitudinal axis (elbow-hand axis) and servoed the wrist motion to it.

First, the rotation between θ_{fa} , the current posture of the lower-arm, and θ_0 , an objective posture for the subject to go back to, is computed:

$$\epsilon(t) = \theta_{fa}(t) - \theta_0 \quad (1)$$

θ_0 is set here as the initial forearm posture but can change if necessary, depending on the task. ϵ is then used as the input of the control law, which pilots the prosthetic wrist angular velocity, $\dot{\theta}$:

$$\dot{\theta}(t) = \begin{cases} 0 & \text{if } |\epsilon(t)| < \epsilon_0 \\ \lambda(\epsilon(t) - \text{sign}(\epsilon(t) - \epsilon_0)\epsilon_0) & \text{otherwise,} \end{cases} \quad (2)$$

where ϵ_0 is a deadzone threshold (here set to 5 deg) and λ is a scalar gain (set to 2 s^{-1}) that tunes the rate of correction. These values were chosen experimentally to ensure stability.

Experimental set-up




Fig. 1: Set-up of the refined Rolyan Clothespin test with transradial amputees.

Two transradial amputees, regular myoelectric-prosthesis users, participated to this experiment. They were asked to

perform the refined Rolyan Clothespin test [7], both with their usual myoelectric control (MYO) and with the proposed control scheme, later called Compensations Cancellation Control (CCC). There were no specific training session and the Rolyan was achieved 5 times with each control mode. For the latter, the participants' wrist rotator was replaced with one from the laboratory, including an encoder, and controlled with an external Raspberry Pi 3[©]. As CCC cannot be implemented for the hand, the usual myoelectric control of each participant was conserved for the grasping function. For this experiment, θ_{fa} was obtained with an Inertial Measurement Unit located on the prosthetic forearm (see Figure 1). The protocol of this experiment was approved by the Université Paris Descartes ethic committee CERES. All participants gave their written informed consent in accordance with the Declaration of Helsinki.

III. RESULTS




Fig. 2: Rolyan assessment (mean and standard deviation over the 5 trials for each mode and subject). (a) Task performance; (b) wrist motion (wrist pronosupination for downwards pins relocation, by one subject) and (c) body compensations.

Task performance

Figure 2(a) shows the time, averaged over trials, obtained with the two control modes, for both subjects. The times are clearly similar between the direction of the motion (upwards –horizontal-to-vertical– or downwards –vertical-to-horizontal–) and the control modes. We can also notice that the standard deviation is small, for both MYO and CCC.

Joints motions

The example of wrist angular trajectories of Figure 2(b) illustrates that CCC gives trajectories very close to the ones given by MYO. Considering shoulder and trunk, it has to be checked that, with CCC, the use of compensatory motions as input of the controller does not enhance them. The maximum Ranges of Motion (over trials and pins location) of shoulder abduction (Abd.) and trunk angles (flexion, rotation and

lateral bending – T.Flex., T.Rot., T.Lat.Bend.–) with CCC are thus compared to the ones with MYO. Yet, as prosthesis users may employ undesirable compensatory motions to achieve the task with MYO, it was also verified that compensations of these participants were not troublesome (angles values and holding time) [8]. In Figure 2(c), we see that compensatory joints motions are not higher with CCC than with MYO and stay in acceptable ranges.

IV. DISCUSSION

The different metrics assessing Rolyan test all show that CCC was quickly mastered and was as good as MYO, while it was totally unknown from the participants, who discovered it during the experiment: the time of the task is similar, with a small standard deviation, wrist angular trajectories are close and body compensations are not heightened. Both participants also appreciated the absence of co-contraction with CCC and reported that this control mode induces less muscle fatigue than MYO. These results confirm what has been obtained with able-bodied subjects, on a different task.

Yet, as seen Figure 2(c), the subjects called different shoulder and trunk strategies (Subject 2 moved them much more than Subject 1). We are thus intending to perform an extensive study with more amputated participants to have a larger overview of users' strategies. The definition of the objective posture also merits further attention, to confirm that a same objective can be used for multiple arm movements.

V. CONCLUSION

The new control scheme for UL prostheses presented in this paper proposes to use the device to control the user's posture, while letting him/her in charge of the end-effector positioning and orientation. Tested with two transradial amputees, it was as good as conventional myoelectric control without any specific learning, while potentially reducing the muscle fatigue. Following these promising results, the concept developed here could be extended to the control of more than one DOF. We think that it could also be adapted to other rehabilitation devices like UL exoskeletons.

REFERENCES

- [1] F. Cordella et al., "Literature Review on Needs of Upper Limb Prosthesis Users," *Front. Neurosci.*, vol. 3, 2016.
- [2] N. M. Bajaj, A. J. Spiers and A. M. Dollar, "State of the art in prosthetic wrists: Commercial and research devices," *IEEE ICORR*, 2015.
- [3] G. Li, *Advances in Applied Electromyography*. IntechOpen, 2011.
- [4] S. L. Carey, M. J. Highsmith, M. E. Maitland, and R. V. Dubey, "Compensatory movements of transradial prosthesis users during common tasks," *Clin. Biomech.*, vol. 23, no. 9, pp. 1128–1135, 2008.
- [5] K. Østlie, R. J. Franklin, O. H. Skjeldal, A. Skrandal, and P. Magnus, "Musculoskeletal pain and overuse syndromes in adult acquired major upper-limb amputees," *Arch. Phys. Med. Rehabil.*, vol. 92, no. 12, pp. 1967–1973, 2011.
- [6] M. Legrand, N. Jarrassé, F. Richer, and G. Morel, "A closed-loop and ergonomic control for prosthetic wrist rotation," in *IEEE ICRA*, 2020.
- [7] A. Hussaini, W. Hill, and P. Kyberd, "Clinical evaluation of the refined clothespin relocation test: A pilot study," *Prosthet. Orthot. Int.*, vol. 43, no. 5, pp. 485–491, 2019.
- [8] D. Kee and W. Karwowski, "Luba: An assessment technique for postural loading on the upper body based on joint motion discomfort and maximum holding time," *Applied Ergonomics*, vol. 32, pp. 357–366, 2001.