



**HAL**  
open science

## Development of Polythiourethane/ZnO-Based Anti-Fouling Materials and Evaluation of the Adhesion of *Staphylococcus aureus* and *Candida glabrata* Using Single-Cell Force Spectroscopy

Sophie Klemm, Martina Baum, Haoyi Qiu, Zibin Nan, Mafalda Cavalheiro, Miguel Cacho Teixeira, Claire Tendero, Anna Gapeeva, Rainer Adelung, Etienne Dague, et al.

► **To cite this version:**

Sophie Klemm, Martina Baum, Haoyi Qiu, Zibin Nan, Mafalda Cavalheiro, et al.. Development of Polythiourethane/ZnO-Based Anti-Fouling Materials and Evaluation of the Adhesion of *Staphylococcus aureus* and *Candida glabrata* Using Single-Cell Force Spectroscopy. *Nanomaterials*, 2021, 11 (2), pp.271. 10.3390/nano11020271 . hal-03136104

**HAL Id: hal-03136104**

**<https://hal.science/hal-03136104>**

Submitted on 9 Feb 2021

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



## Open Archive Toulouse Archive Ouverte (OATAO)

OATAO is an open access repository that collects the work of Toulouse researchers and makes it freely available over the web where possible

This is a Publisher's version published in: <http://oatao.univ-toulouse.fr/27376>

**Official URL:** <https://doi.org/10.3390/nano11020271>

### **To cite this version:**

Klemm, Sophie and Baum, Martina and Qiu, Haoyi and Nan, Zibin and Cavalheiro, Mafalda and Teixeira, Miguel Cacho and Tendero, Claire  and Gapeeva, Anna and Adelong, Rainer and Dague, Etienne and Castelain, Mickaël and Formosa-Dague, Cécile *Development of Polythiourethane/ZnO-Based Anti-Fouling Materials and Evaluation of the Adhesion of Staphylococcus aureus and Candida glabrata Using Single-Cell Force Spectroscopy*. (2021) *Nanomaterials*, 11 (2). 271. ISSN 2079-4991

Any correspondence concerning this service should be sent to the repository administrator: [tech-oatao@listes-diff.inp-toulouse.fr](mailto:tech-oatao@listes-diff.inp-toulouse.fr)



## Article

# Development of Polythiourethane/ZnO-Based Anti-Fouling Materials and Evaluation of the Adhesion of *Staphylococcus aureus* and *Candida glabrata* Using Single-Cell Force Spectroscopy

Sophie Klemm <sup>1,2</sup>, Martina Baum <sup>1,\*</sup>, Haoyi Qiu <sup>1</sup>, Zibin Nan <sup>3</sup>, Mafalda Cavalheiro <sup>4</sup>, Miguel Cacho Teixeira <sup>4</sup>, Claire Tendo <sup>5,6</sup>, Anna Gapeeva <sup>1</sup>, Rainer Adelung <sup>1</sup>, Etienne Dague <sup>2,6</sup>, Mickaël Castelain <sup>3,6</sup> and Cécile Formosa-Dague <sup>3,6,\*</sup>

- <sup>1</sup> Functional Nanomaterials, Institute for Materials Science, Kiel University, 24143 Kiel, Germany; sophie.klemm@tu-berlin.de (S.K.); haq@tf.uni-kiel.de (H.Q.); ang@tf.uni-kiel.de (A.G.); ra@tf.uni-kiel.de (R.A.)
- <sup>2</sup> LAAS-CNRS, Université de Toulouse, CNRS, 31400 Toulouse, France; edague@laas.fr
- <sup>3</sup> TBI, Université de Toulouse, INSA, INRAE, CNRS, 31400 Toulouse, France; nan@insa-toulouse.fr (Z.N.); castelai@insa-toulouse.fr (M.C.)
- <sup>4</sup> Institute for Bioengineering and Biosciences (iBB), Instituto Superior Técnico, University of Lisbon, 1049-001 Lisbon, Portugal; mafalda.cavalheiro@tecnico.ulisboa.pt (M.C.); mnpct@tecnico.ulisboa.pt (M.C.T.)
- <sup>5</sup> CIRIMAT, Université de Toulouse, CNRS, INPT, UPS, 31400 Toulouse, France; claire.tendo@inp-toulouse.fr
- <sup>6</sup> Fédération de Recherche Fermat, CNRS, 31000 Toulouse, France
- \* Correspondence: marb@tf.uni-kiel.de (M.B.); formosa@insa-toulouse.fr (C.F.-D.); Tel.: +49-(0)431-880-6180 (M.B.); +33-(0)5-61-55-96-86 (C.F.-D.)



**Citation:** Klemm, S.; Baum, M.; Qiu, H.; Nan, Z.; Cavalheiro, M.; Teixeira, M.C.; Tendo, C.; Gapeeva, A.; Adelung, R.; Dague, E.; et al. Development of Polythiourethane/ZnO-Based Anti-Fouling Materials and Evaluation of the Adhesion of *Staphylococcus aureus* and *Candida glabrata* Using Single-Cell Force Spectroscopy. *Nanomaterials* **2021**, *11*, 271. <https://doi.org/10.3390/nano11020271>

Academic Editor: Antonios Kelarakis  
Received: 14 December 2020  
Accepted: 13 January 2021  
Published: 21 January 2021

**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

**Abstract:** The attachment of bacteria and other microbes to natural and artificial surfaces leads to the development of biofilms, which can further cause nosocomial infections. Thus, an important field of research is the development of new materials capable of preventing the initial adhesion of pathogenic microorganisms. In this work, novel polymer/particle composite materials, based on a polythiourethane (PTU) matrix and either spherical (s-ZnO) or tetrapodal (t-ZnO) shaped ZnO fillers, were developed and characterized with respect to their mechanical, chemical and surface properties. To then evaluate their potential as anti-fouling surfaces, the adhesion of two different pathogenic microorganism species, *Staphylococcus aureus* and *Candida glabrata*, was studied using atomic force microscopy (AFM). Our results show that the adhesion of both *S. aureus* and *C. glabrata* to PTU and PTU/ZnO is decreased compared to a model surface polydimethylsiloxane (PDMS). It was furthermore found that the amount of both s-ZnO and t-ZnO filler had a direct influence on the adhesion of *S. aureus*, as increasing amounts of ZnO particles resulted in reduced adhesion of the cells. For both microorganisms, material composites with 5 wt.% of t-ZnO particles showed the greatest potential for anti-fouling with significantly decreased adhesion of cells. Altogether, both pathogens exhibit a reduced capacity to adhere to the newly developed nanomaterials used in this study, thus showing their potential for bio-medical applications.

**Keywords:** polythiourethane; tetrapodal shaped ZnO; PTU/ZnO composite; microbial adhesion; initial attachment; *S. aureus*; *C. glabrata*; single cell force spectroscopy; anti-fouling

## 1. Introduction

Hospital-acquired microbial infections, also called nosocomial infections, are nowadays a leading cause of death on all continents [1]. Such infections are often caused by pathogenic bacteria and yeast, which are able to form biofilms on medical devices such as catheters, pacemakers or prosthetic joints. Biofilms are structured microbial communities that show increased resistance to antibiotics [2,3]. Such medical device-associated infections represent a heavy burden for societies, as they are responsible for increasing the

length of hospital stay for patients and result in high mortality rates [4]. Biofilm formation on surfaces usually involves three steps. The first step consists of the reversible adhesion of the microbial cells to surfaces involving physico-chemical forces. It is then followed by a transition from reversible to irreversible microbial adhesion where physico-chemical bonds are strengthened [5] and chemical forces, such as hydrophobic or hydrophilic forces [6], can be involved. Finally, after these adhesion steps, cells undergo important changes (cell wall deformation, production of exopolysaccharides, gene expression etc.) that lead to the formation of a mature biofilm [5].

So far, diverse strategies have been explored by scientists to combat the formation of biofilms on abiotic surfaces [7]. Such strategies include the use of anti-biofilm treatment agents, such as specific enzymes targeting the exopolysaccharide (EPS) matrix of mature biofilms [8], anti-quorum sensing molecules that interfere with biofilm development [9], and anti-oxidants that inhibit biofilm formation in many bacterial species [10]. Other studies have focused on developing anti-biofilm strategies based on the inhibition of cell–cell interactions, which also play a crucial role in biofilm formation [11,12]. Further, biofilm formation can be prevented by focusing on the surface properties of the material itself. From this point of view, anti-adhesive surfaces that directly inhibit the first steps of biofilm formation, i.e., cell adhesion, have been developed over the years. Some of these strategies include surface coatings with silver nanoparticles [13] or with polymeric hydrophilic materials [14,15]. Hydrophilic coatings prevent microbial adhesion thanks to a tightly correlated water layer that builds up on the material surface and forms an energetic barrier against microbial cells. In general, Geoghegan et al. showed stronger adhesion to hydrophobic materials for several bacterial strains (*Aquabacterium commune*, *Escherichia coli*, *Rhodococcus* spp., *Pseudomonas* sp., *Acidovorax* sp.) [16]. However, several studies have shown that hydrophobic or superhydrophobic properties of textured surfaces were able to reduce bacterial adhesion [17,18], for example in the cases of *S. aureus*, *E. coli*, or *Staphylococcus epidermidis*. Another well investigated mechanism involves steric repulsion. Polyethylene glycol is considered a gold standard polymer able to reduce adhesion via hydration forces and steric hindrance [19,20]. However, its oxidation in biological environments needs to be considered as it can be harmful to the human body [21]. As the surface roughness of a material also interferes with the adhesion, by changing the availability of the contact surface area [15,22], surface patterning is another strategy to combat initial cell attachment [14,20–22]. Several studies have reported that the patterning of titanium (Ti) surfaces in the nanometer range influences the adhesion of *S. aureus* [23–25]. However, while Ercan et al. observed reduced adhesion with increasing roughness (from 60 to 80 nm) [23], Wang et al. showed that for TiO<sub>2</sub>, the increase in roughness from 70 nm to 120 nm led to an increase in cell adhesion by three-fold [26]. These contradicting results show that the interpretation of the effects of physical parameters on the adhesion process is difficult and should be accompanied by chemical analysis to provide a full understanding. Furthermore, these examples of strategies to produce anti-fouling surfaces show that, so far, no ideal strategy has been developed to combat biofilm formation of multiple species; the research towards new materials to overcome this challenge has to be actively pursued.

As stated previously, the physical and chemical properties of the materials surface are of great importance. Therefore, the anti-fouling materials we developed in this work are based on a polymer/ceramic system. We combined polythiourethane (PTU) as a matrix polymer with zinc oxide (ZnO) particles of varying morphology and amount. The variation of morphology and amount of ZnO particles allows us to alter the topographical and physico-chemical properties of PTU. PTU is a two-component, full-solid, non-toxic, transparent thermoset polymer system, which is also resistant to UV-irradiation [27] and biocorrosion [28–30]. PTU forms very smooth surfaces ( $R_a$ : 0.037  $\mu\text{m}$ ) and does not swell in an aqueous environment [31]. Furthermore, PTU is easy to fabricate by a simple, solvent-free polyaddition reaction [28], and was successfully tested as an anti-fouling coating for ship hulls [28,31–33]. ZnO particles and ZnO thin films are widely used for various industrial applications, e.g., in optoelectronics, piezo electronic devices or anti-viral prod-

ucts [34,35]. Furthermore, the anti-microbial activity of ZnO nanoparticles is widely known and has been strongly investigated [34,36]. In this study we used tetrapodal ZnO microparticles next to spherical particles. T-ZnO has the advantage of bearing all ZnO's intrinsic properties, e.g., high Young's modulus and strength as well as biocompatibility [34], but with the added benefits of an even further reduced toxicity [37], improved force transduction [34,38] and anti-fouling properties [28]. The addition of t-ZnO particles allows for the modification of the composite material in terms of strength, elasticity and surface roughness which allows a broader spectrum of potential applications in the human body.

In the present study, we characterized the material properties of the developed PTU/ZnO composites (including PTU without addition of ZnO particles) in terms of surface roughness, mechanical properties, surface charge and hydrophobicity; properties that have been demonstrated to influence microorganism adhesion [39]. Their anti-adhesive properties towards two medically relevant microbial species, *S. aureus* and *C. glabrata*, were then evaluated using atomic force microscopy (AFM) [40]. To achieve this, *S. aureus*, the Gram-positive bacterial species most frequently found in health-care associated infections [41], and *C. glabrata*, a yeast species with dramatic implications in immune-suppressed patients [42], were used in single-cell force spectroscopy [43] experiments to probe their direct interactions with the different hydrophilic PTU/ZnO composites. Polydimethylsiloxane (PDMS) was used in this study as a hydrophobic model surface as it is widely used for medical devices [44–46]. These experiments allowed the comparison of the bio-adhesive properties of each material depending on the type of microorganism, but have also enlightened the molecular and biophysical mechanisms involved in such interactions. Altogether, this work provides a step forward towards tailoring efficient anti-adhesive materials that can be used in future biomedical applications.

## 2. Materials and Methods

### 2.1. Polymer Composites Preparation

The investigated PTU/ZnO composites are based on a two-component PTU thermoset (Fluid-& Prozesstechnik GmbH, Waltershausen, Germany), which consists of pentaerythritol tetrakis (3-mercaptopropionate) (PETMP) and hexamethylene diisocyanate (HDI). Respectively, 1 and 5 wt.% ZnO particles were dispersed within PETMP component by the MiniMaster Laboratory Disperser (Netzsch-Feinmahltechnik GmbH, Selb, Germany) with a dispersion speed of 6000 rpm for 15 min. Subsequently, the second component, HDI, was added and stirred by hand for approx. 1 min. The ratio of PETMP to HDI was kept at 1 to 1.4 in weight for all PTU/ZnO composites [31]. Hölken et al. showed that the ratio of 1:1.4 PETMP:HDI results in the highest ultimate tensile strength for this polymer system [28]. After degassing the polymer blend by using a desiccator, the final mixture was cast into silicone molds ( $9 \times 30 \times 0.9$  mm) and cured at 85 °C for 24 h under atmospheric pressure. Detailed information on the polymerization reaction of PTU can be found in Hölken et al. 2016 [28]. To investigate the influence of the amount of ZnO particles and their geometry on bio-adhesive properties of the polymer composites, 1 and 5 wt.% of tetrapodal (t-ZnO) and spherical (s-ZnO) ZnO particles were added. T-ZnO was produced by flame transport synthesis at Kiel University, Kiel, Germany [47,48]. The four arms of the t-ZnO form angles of 109.5° to each other, have a length of ca. 15–20 µm and a diameter of around 3–5 µm. S-ZnO with a grain size of 1–10 µm was purchased from Sigma-Aldrich Chemie GmbH, Munich, Germany. SEM images of both types of particles can be seen in [28]; SEM images of the different PTU/ZnO composites produced in this study are presented in Figure S1 (Supplementary Materials).

As a hydrophobic reference material, a two-component silicone rubber (ADDV-25, R&G Faserverbundwerkstoffe GmbH, Waldenbuch, Germany) was used and the samples were cast in the same dimensions and under the same conditions as for the PTU/ZnO composites.

## 2.2. AFM Surface Imaging

The nanoscale surface roughness of the PTU/ZnO composites and PDMS was performed using a Nanowizard III AFM (Bruker, Billerica, MA, USA) with MLCT AUWH probes (Bruker, Billerica, MA, USA, nominal spring constant of 0.6 N/m). The spring constant of the cantilevers was determined prior to imaging using the thermal noise method and ranged from 0.447 to 0.540 N/m [49]. Three images ( $20 \times 20 \mu\text{m}^2$ ) of each PTU/ZnO composite and the PDMS specimens were recorded in contact mode in air with a constant applied force of 2 nN and a line rate of 3.0 Hz. Figure S2 (Supplementary Materials) show examples of AFM height images recorded for each type of materials (PDMS and PTU/ZnO composites). For roughness analysis, the arithmetic roughness ( $R_a$ ) was calculated using JPK Data Processing Software on flattened (order 3) high-resolution ( $512 \times 512$  lines) height images, following Equation (1).

$$R_a = \frac{1}{L} \int_0^L |z(x)| dx \quad (1)$$

where  $L$  is the evaluation length and  $z(x)$  the profile height function.

## 2.3. Confocal Laser Scanning Microscopy (CLSM) Imaging

To obtain the mean arithmetic roughness ( $R_a$ ), the 3D Laser Scanning Confocal Microscope VK-X (Keyence Corporation, Osaka, Japan) with a red semiconductor laser source (wavelength of 658 nm) was used. For roughness determination, a magnification  $100\times$  was used. For obtaining the roughness, a total of ten-line profile measurements were conducted in x- and y-directions for PTU/ZnO composites and PDMS. The line profile was conducted on an image with the size of  $1 \times 1.5 \text{ mm}^2$ .

## 2.4. Microorganism Strains and Culture Conditions

The bacterial strain *S. aureus* ATCC 19123 and the yeast strain *C. glabrata* KChr606 were used in this study. *S. aureus* cells were cultured in Mueller–Hinton (MH) broth (Difco, Kansas, USA) at  $37^\circ\text{C}$  under agitation (150 rpm) overnight. For the AFM experiments, bacterial cells were harvested by centrifugation ( $1000\times g$ , 3 min) and washed two times in PBS 1X (phosphate-buffered saline, pH 7.4). *C. glabrata* cells were cultured in yeast extract peptone dextrose (YPD) broth (Difco, Kansas, USA) at  $30^\circ\text{C}$  under agitation (250 rpm) overnight. For AFM experiments, the yeast cells were harvested by centrifugation (5000 rpm, 5 min) and washed two times in acetate buffer (1.48 g/L sodium acetate, 111 mg/L  $\text{CaCl}_2$  and 198 mg/L  $\text{MgCl}_2$ , 5.2 pH).

## 2.5. Single-Cell Force Spectroscopy Experiments

These experiments were performed using a Nanowizard III AFM (Bruker, Billerica, MA, USA) equipped with an inverted light microscope (Zeiss, Jena, Germany). NP-O10 probes (Bruker, Billerica, MA, USA) with a nominal spring constant of 0.06 N/m for experiments with *S. aureus* and of 0.35 N/m for experiments with *C. glabrata* were used to measure the interactions between single-cells and the different material surfaces. For this, given their respective sizes *S. aureus* ( $\sim 1 \mu\text{m}$  of diameter), cells were first immobilized on colloidal probes [43], while *C. glabrata* cells ( $\sim 4\text{--}5 \mu\text{m}$  in diameter) were immobilized directly on the tipless cantilevers. All cantilevers were cleaned using oxygen plasma before use. For all conditions, force maps of  $10 \times 10$  pixels were recorded on  $1 \times 1 \mu\text{m}^2$  areas on the surfaces. In the case of *S. aureus*, colloidal probes were obtained by attaching a single silica microsphere ( $5 \mu\text{m}$  diameter, Bangs Laboratories, Fishers, Indiana, IN, USA) with a thin layer of UV-curable glue (NOA 63, Norland Edmund Optics, Barrington, NY, USA) on the cantilevers using the AFM. Afterwards these were immersed for 1 h in tris buffer saline (TBS, pH 8.5) containing 4 mg/mL of dopamine hydrochloride (Sigma-Aldrich Chemie GmbH, Munich, Germany), rinsed in PBS 1X, and used directly for cell probe preparation. The spring constant of the colloidal probes, ranging from 0.0554 to 0.0932 N/m was determined prior to cell immobilization using the thermal noise method [49]. Post

preparation, the colloidal probe was brought into contact with an isolated bacterium and subsequently retracted with the attached bacterial cell. The attachment of the cell on the silica microsphere was confirmed using optical microscopy and by performing a force distance curve on the Petri dish. The maximum force applied equaled 250 pN and a constant approach and retraction velocity of 1  $\mu\text{m/s}$  was applied. For each condition, experiments were repeated for seven cells originating from at least three different cultures, and 400 force curves were recorded for each cell.

In the case of *C. glabrata* cells, the cantilevers were incubated in a solution of Concanavalin A (ConA, Sigma-Aldrich, 0.1 mg/mL) overnight and rinsed in acetate buffer. The spring constant of the ConA functionalized cantilevers, ranging from 0.213 to 0.361 N/m, was determined prior to cell attachment using the thermal noise method [49]. The ConA cantilever was then brought into contact with an isolated yeast cell and retracted with the attached yeast cell. In this case also the proper attachment of the cell on the cantilever was confirmed using an inverted light microscope (Zeiss, Jena, Germany). In contrast to experiments with *S. aureus*, tipless cantilever can be used without a sphere because of the larger size of the *C. glabrata* cells. The maximum force applied equaled 1000 pN and a constant approach and retraction velocity of 5  $\mu\text{m/s}$  was applied. For each condition, experiments were repeated for seven cells, originating from at least three different cultures, and 400 force curves were recorded for each cell.

For both types of cell, probes were finally used to measure cell–surface interaction forces in PBS 1X for *S. aureus*, and in acetate buffer for *C. glabrata*, both at room temperature. Data were analyzed using the JPK Data Processing software (Bruker, Billerica, MA, USA). Adhesion force and rupture distance histograms were obtained by calculating the maximum adhesion force and the rupture distance of the last peak for each obtained curve. The mean values and standard deviations were calculated and the significance was tested by performing a two sample t-test by using OriginLab (OriginLab Corporation, Northampton, MA, USA). The force distance curves were finally analyzed by using the JPK Data Processing software (JPK Instruments, Berlin, Germany). The adhesion force and the rupture distance were obtained and summarized in histograms with a bin interval of 50 pN and 20 nm for *S. aureus* and 100 pN and 50 nm for *C. glabrata*, respectively. The mean values and standard deviations were calculated using OriginLab (OriginLab Corporation, Northampton, MA, USA).

## 2.6. Tensile Testing

In the underlying study, a Zwick 1445 universal testing machine (Zwick GmbH & Co.KG, Ulm, Germany) was used to perform the tensile tests (ISO 527) on the PTU/ZnO composites with an initial load of 5 N and an applied constant strain rate of 1 mm/min. For testing PDMS, the initial load was set to 0.1 N and the strain rate to 20 mm/min. The samples were prepared by the mixture casting into a dog-bone shaped silicone mold. The effective dimensions of the samples were 20 mm in length, 5 mm in width and 1 mm in thickness. The Young's Modulus was calculated from the stress–strain relationship using Hooks Law below (2), as described in [31]. The mean values and standard deviations of ten samples for every PTU/ZnO composite and PDMS were calculated using OriginLab (OriginLab Corporation, Northampton, MA, USA).

$$E = \frac{\sigma}{\epsilon} \quad (2)$$

where  $E$  is the elastic modulus,  $\sigma$  is the stress and  $\epsilon$  the strain.

## 2.7. Scanning Electron Microscopy (SEM) Imaging and Energy Dispersive X-ray Spectroscopy (EDX)

To evaluate the presence of chemical elements like Zn, a SEM (Zeiss ULTRA PLUS microscope with GEMINI column, Carl Zeiss AG, Oberkochen, Germany) equipped with an InLens detector and an EDX-detector (Carl Zeiss AG, Oberkochen, Germany) was used

to investigate the PTU/ZnO composites. The acceleration voltage was set to 10 keV and the magnification to 300 $\times$ . The topside of the casted samples was investigated. The polymer material was sputter coated with gold by a BAL-TEC SCD 050 Sputter Coater (Bal-Tec AG, Pfäffikon, Switzerland) for 80 s with a current of 30 mA. The amount of C and Zn detected by EDX on one sample was used to calculate the Zn/C ratio presented in Table S1 in the Supplementary Materials section.

### 2.8. Photometric Determination of Potential Zn-Ion ( $Zn^{2+}$ ) Release

Potential  $Zn^{2+}$  release from the PTU/ZnO composites was investigated by a photometric method, which is based on the reaction of zincon (2-carboxy-2'-hydroxy-5'-sulfoformazylbenzene). This chemical complex binds  $Zn^{2+}$  and forms a blue complex at pH 8.5–9.5. Prior to the  $Zn^{2+}$  concentration measurements, PTU/ZnO composites were washed with 2-propanol to remove potential superficial contaminations. To simulate a short-time exposure of those materials to an aqueous environment, the composite materials were immersed in deionized water for one day; for long-time exposure for 27 days. To investigate the potential  $Zn^{2+}$  release from PTU/ZnO composites, the topside and backside of the cast specimens with 1 wt.% and 5 wt.%, s- and t-ZnO were tested individually.

The 0.1 wt.% zincon solution contained 13 mg zincon in 0.2 mL 1 M NaOH solution and 10 mL deionized water. 0.1 mL of the aqueous test solution (incubated with the PTU/ZnO composites) was mixed with 0.3 mL zincon solution. The pH was adjusted by borax buffer ( $Na_2B_4O_7 \times 10H_2O$  (pH 9)). Deionized water was added to obtain a total volume of 5 mL. The intensity of the blue color was measured photometrically, by an Epoch 2 Microplate Spectrophotometer (BioTek Instruments Inc., Winooski, USA) at a wavelength of 620 nm. A calibration curve was constructed to assess the concentration of  $Zn^{2+}$  in tested samples. The amount of  $Zn^{2+}$  detected is presented in Table S2 in the Supplementary Materials section.

### 2.9. Wetting Properties

The water contact angle measurements were determined with the Drop Shape Analyser 25B (A. KRÜSS Optronic GmbH, Hamburg, Germany) by the sessile drop method with a drop volume of 8  $\mu$ L. Three surfaces were investigated for each PTU/ZnO composition and for PDMS and three measurements were conducted on each surface (nine measurements per material variation). The contact angle gives information about the wettability of a surface. If the contact angle exceeds 90 $^\circ$  the surface is considered hydrophobic otherwise the surface is considered hydrophilic.

### 2.10. Surface Charge Evaluation

The Zeta potential  $\zeta$  was measured on the surface of flat samples with a ZetaCAD<sup>®</sup> (CAD Instruments, Les Essarts le roi, France) setup. The various samples were immersed for one hour in  $KNO_3$   $10^{-3}$  M before being inserted in the flow cell. A pressure ramp (from 20 to 235 mBar with a step of 20 mBar) was applied to the flow cell. At each pressure step the electrolyte flows in two directions (from tank 1 to 2 and inversely). For each step and each flow direction, the resulting streaming potential was measured. The streaming potential is linked to the zeta potential  $\zeta$  through the Helmholtz–Smoluchowski equation [50]. Electrolyte pH, temperature and conductivity were measured for each experiment.

The values of the zeta potential are robust as far as both repeatability and reproducibility are concerned: each sample was tested three times and each kind of surface was evaluated with three couples of samples. Therefore, the streaming potential values given for each material variation result from nine measurements. The global deviation includes both correlation and experimental deviation.

### 3. Results and Discussion

#### 3.1. PTU/ZnO Composites Have Anti-Adhesive Surface Properties

Given the importance it may have for cell adhesion, the first property of the PTU/ZnO composites that we chose to explore in this study was surface roughness. For this, we used two established techniques to characterize the surface materials, confocal laser-scanning microscopy (CLSM) and atomic force microscopy (AFM) [51], and we measured the average roughness  $R_a$  of the PTU/ZnO composites and the PDMS specimen. The data obtained are presented in Table 1.

**Table 1.** Surface roughness ( $R_a$ ) of polymer variations containing 0 wt.%, 1 wt.% and 5 wt.% of t-ZnO and s-ZnO particles measured by confocal laser-scanning microscopy (CLSM) and atomic force microscopy (AFM).

		$R_a$ Measured by CLSM (nm)		$R_a$ Measured by AFM (nm)	
PDMS		110 ± 10		13.60 ± 4.60	
PTU		110 ± 40		0.40 ± 0.08	
Type of particle		s-ZnO	t-ZnO	s-ZnO	t-ZnO
PTU/1 wt.%		100 ± 10	150 ± 10	0.15 ± 0.34	0.41 ± 0.07
PTU/ 5 wt.%		140 ± 30	210 ± 50	0.78 ± 0.20	0.72 ± 0.18

The roughness analysis using CLSM was performed on images of  $1 \times 1.5 \text{ mm}^2$ , while for AFM it was performed on images of  $20 \times 20 \text{ }\mu\text{m}^2$ . Thus, the area scanned using CLSM was 3750 times larger than the area scanned using AFM, explaining the differences observed in the values between the two techniques. The first interesting point to note, is that while CLSM measurements gave a similar value for the roughness of PTU and PDMS, measurements at the micro-scale with AFM show an important difference between PDMS and PTU, with a roughness value of  $13.6 \pm 4.6 \text{ nm}$  for PDMS and of  $0.401 \pm 0.075 \text{ nm}$  for PTU. This difference may, indeed, be due to the size of the areas scanned in each case, but also to the different detection methods involved in the two techniques (CLSM, reflection of visible light; AFM, physical deflection of a cantilever). However, both techniques show the same tendency: while the addition of 1 wt.% of spherical ZnO particles tends to smoothen the surface, although the difference with PTU in both cases is not significant, increasing the amount of particles from 1 to 5 wt.% increases the roughness of the material. In the situation where tetrapodal particles are added to PTU, the roughness is slightly increased when 1 wt.% of particles were added compared to PTU; this increase is more pronounced with the addition of 5 wt.% of particles. It is important to note that in all cases, no  $\text{Zn}^{2+}$  was found at the surface of the materials, neither by EDX analysis nor by  $\text{Zn}^{2+}$  release measurements (Supplementary Tables S1 and S2), which implies that the roughness differences observed are only influenced by the filling material, and not by the physical presence of ZnO particles at the surface. The literature on the influence of roughness on cell adhesion is conflicting. While some studies suggest that an  $R_a$  smaller than 200 nm does not affect adhesion [52], different authors showed differences in the adhesion of cells on nanorough surfaces (30 nm to 120 nm) [20,23]. However, as far as we know, the roughness in the sub-nanometer range has not been extensively investigated in the context of microbial adhesion. Thus, it is difficult to state at this stage whether the small roughness differences measured between the PTU variations will influence cell adhesion.

Another component considered in this study is the material stiffness. In material science, material stiffness is a basic parameter, most relevant in evaluating fundamental phenomena like adhesion, friction or crack propagation [51,53]. However, within the interdisciplinary field of bio-adhesion between abiotic surfaces and microorganisms, this issue has not yet been widely addressed, although it has been shown that the stiffness of a material can influence the adhesion of bacteria onto it [39]. In the case of the PTU/ZnO composites produced in this study, the stiffness of the material variations was measured using tensile tests; the results are presented in Table 2. They show that while the stiffness of PTU is five times higher than that of PDMS, there are no significant differences observed be-

tween the PTU variations, which is in accordance with previous measurements performed on other batches of PTU/ZnO materials produced in our team [28]. Only PTU with 5 wt.% of t-ZnO shows a significant difference to the other PTU variations; this might be due to the geometric arrangement of the tetrapods in the material. If we put these results in light of the study by Song and Ren, then the adhesion to the stiffer material, namely PTU and PTU/ZnO variations, should be less favored by the microorganisms tested in this study.

**Table 2.** Elastic modulus from tensile tests for PTU and PTU/ZnO variations.

Material Variation	Elastic Modulus (MPa)
PDMS	120 ± 9
PTU	624 ± 63
PTU 1 wt.% s-ZnO	641 ± 31
PTU 5 wt.% s-ZnO	630 ± 24
PTU 1 wt.% t-ZnO	611 ± 56
PTU 5 wt.% t-ZnO	689 ± 32

The hydrophobicity of a material is also a crucial factor to consider when evaluating the adhesion of microorganisms to artificial surfaces [54–56]. For instance, it has been already shown that microorganisms such as *S. aureus* or *C. glabrata*, used in this study, adhere more to hydrophobic surfaces, as do many other microbes [17,57–61]. Thus, we characterized the wetting properties of the PTU/ZnO variation compared to PDMS; the results are presented in Table 3. Figure S3 (Supplementary Materials) shows examples of the corresponding micrographs from which the measurements were performed. While PDMS is hydrophobic with a Water Contact Angle (WCA) of  $114.9 \pm 1.4^\circ$ , PTU and PTU/ZnO variations are hydrophilic (WCA comprised between  $78.6$ – $73.3^\circ$ ). No significant differences can be observed among the PTU/ZnO variations. As well as for the increased stiffness, the hydrophilic properties of PTU/ZnO composites should also lead to the reduced adhesion of microbial cells.

**Table 3.** Water contact angle measurements of polymer variations.

	PTU	PTU/1 wt.% t-ZnO	PTU/5 wt.% t-ZnO	PTU/1 wt.% s-ZnO	PTU/5 wt.% s-ZnO	PDMS
Water Contact Angle (°)	78.6 ± 2.4	77.9 ± 1.3	75.1 ± 1.7	76.0 ± 1.6	73.3 ± 2.3	114.9 ± 1.4

Finally, the last parameter we measured to characterize the PTU/ZnO composites was their surface charge. Indeed, during the first phase of cell adhesion to an abiotic surface, electrostatic interactions are at play; alongside to van der Waals and hydrophobic forces, they mediate non-specific cell adhesion [61]. For this reason, notably, the surface charge is one of the most studied cell wall properties of bacteria cells [15]. The net surface charge of microbial cells in physiological medium is negative [62]. To evaluate the strength of electrostatic interactions between *S. aureus* respectively *C. glabrata* and the PTU/ZnO composites the  $\zeta$ -potential of PDMS, PTU and PTU with 5 wt.% of spherical and tetrapodal ZnO particles was measured: the results are presented in Table 4. The results show that all the surfaces are negatively charged; no significant differences can be observed between PTU and PTU functionalized by 5 wt.% ZnO, whether of spherical or tetrapodal filler shape. An important point to note, is that the PDMS surface is six times less charged than PTU: thus, on this surface there will be less repulsion with cells than for PTU and PTU variations, which may also have an important influence on cell adhesion. Indeed, as the cell wall of microbial cells is negatively charged, the repulsion forces will increase with increasing the negative charge of the material surface.

**Table 4.**  $\zeta$ -Potential after 1 h immersion in  $\text{KNO}_3$   $10^{-3}$  M.

Material Variation	$\zeta$ Zeta Potential (mV)
PDMS	$-3.5 \pm 0.6$
PTU	$-23.0 \pm 3.0$
PTU/5 wt.% t-ZnO	$-23.5 \pm 0.9$
PTU/5 wt.% s-ZnO	$-22.0 \pm 1.0$

All these experiments have thus allowed us to characterize the properties of the PTU/ZnO composites investigated in this study, and to compare them with the ones of PDMS. Our analyses show that PTU/ZnO composites have a decreased roughness compared to PDMS, an increased stiffness, they have a lower WCA and are hydrophilic, and finally, they present a more negatively charged surface. Within the PTU variations, the only significant difference was found concerning the stiffness, where PTU with 5 wt.% of tetrapodal ZnO particles had an increased stiffness compared to the other PTU variations. Based on these findings, the potential of these materials to reduce microbial adhesion, atomic force microscopy was used to probe the interactions between microbial cells and the different material surfaces at the molecular scale.

### 3.2. Adhesion of *S. aureus* to PTU/ZnO Composites Is Significantly Decreased Compared to PDMS

To evaluate whether the PTU/ZnO polymer variations are efficient anti-fouling materials, single-cell force spectroscopy experiments were conducted first using *S. aureus* cells. In this type of experiment, a single living bacterial cell is attached to a colloidal cantilever and directly used, in force spectroscopy mode, to probe the interactions with the surface. This gives access to matrices of force–distance curves recorded over a micrometer-sized ( $1 \times 1 \mu\text{m}^2$ ) area of the surface, from which different information, such as the maximum adhesion peak and the rupture distance, can be extracted. The maximum adhesion peak quantifies the maximum force of the interaction, and the rupture distance gives information on the length of the unfolded molecules at the surface of the cell, and thus, on the nature of the interaction probed. Moreover, the percentage of adhesion is determined by the number of force curves presenting adhesive events in each case.

The histograms of the results obtained are presented in Figure 1. For each type of polymer composite and PDMS, the interactions with seven different cells, coming from at least three independent cultures are presented. The results obtained show that PDMS presents the highest average adhesion force of  $726 \pm 397$  pN, as well as the highest average rupture distance of  $212 \pm 116$  nm among the material variations, containing PTU/ZnO composites and PDMS (Figure 1a). Compared to PDMS, PTU shows a significantly lower number of adhesive events (89.5% instead of 100% for PDMS), a lower average adhesion force with  $237 \pm 289$  pN and a corresponding average rupture distance of  $191 \pm 146$  nm (Figure 1b). With the addition of ZnO particles into the PTU, the number of force curves presenting adhesive events and adhesion forces decrease, as well as the corresponding rupture distances. The minimum average adhesion force recorded, of  $86 \pm 120$  pN, is reached for PTU with 5 wt.% t-ZnO (only 40.4% of adhesive curves, Figure 1f). The average rupture distance for PTU with 5 wt.% t-ZnO is  $103 \pm 61$  nm, which is comparable to PTU containing 5 wt.% s-ZnO. These results are also summarized in Table 5. Statistical analysis (two-sample t-test) showed that the differences recorded in the adhesion forces are significantly different ( $p$ -value  $< 0.001$ ), with only one exception made by PTU with a filler content of 1 wt.% of t-ZnO, for which the adhesion force is not significantly different from PTU and the other PTU/ZnO composite variations. From these results it thus seems that PTU variations display anti-adhesive properties as the number of adhesive events and the adhesion force values are significantly decreased compared to PDMS. While this was known for classic polyurethane materials coatings [63], it is the first time this has been shown for PTU. Moreover, the increasing amount of both types of ZnO particles present in the PTU further decreases both the frequency of adhesion of *S. aureus*; while in the case

of s-ZnO particles, the adhesion force does not significantly decrease with the increasing amount of particles, in the case of t-ZnO, the adhesion force is reduced by a factor of 2.2. Therefore, it seems the tetrapodal shape of ZnO particles is more efficient at decreasing the adhesion of cells. The significant decrease in the adhesion forces indicate the reduced bond-strengthening on PTU/ZnO composites; thus, irreversible adhesion is weaker. This proves the suitability of PTU/ZnO composites as anti-adhesive surfaces in the case of *S. aureus*.

The rupture distances obtained for PTU/ZnO composites are rather long and indicate that the interactions between the cells and the surfaces do not rely only on physico-chemical forces, but also on specific interactions involving specific cell-surface polymers. Indeed, when looking at the retract force curves obtained, multiple peaks can be observed after the contact point, thus proving that molecules at the surface of cells interact with the material and are unfolded upon retraction. Such molecules could be surface proteins or strands of peptidoglycan present on the surface of Gram-positive bacteria such as *S. aureus*, interacting with the polymer surface and thus strengthening the initial physico-chemical bonds between cells and surfaces, to transition to an irreversible adhesion [5].


**Table 5.** Mean adhesion forces and rupture distances of *S. aureus* on a variation of six different material surfaces. The right column presents the percentage of curves presenting adhesive events.

Material Variation	Adhesion Force (pN)	Rupture Distance (nm)	% of Adhesion
PDMS	726 ± 397	212 ± 116	100%
PTU	237 ± 289	191 ± 146	89.5%
PTU/1 wt.% s-ZnO	177 ± 237	122 ± 77	88.5%
PTU/1 wt.% t-ZnO	191 ± 256	169 ± 113	87.2%
PTU/5 wt.% s-ZnO	165 ± 240	98 ± 72	77.7%
PTU/5 wt.% t-ZnO	86 ± 120	103 ± 61	40.4%


The reasons explaining the decrease in adhesion force can be directly correlated to the characteristics of the surfaces described earlier. The first reason, is that PTU variations, compared to PDMS, are hydrophilic (for details see Table 3). Indeed, several studies have shown that an increased hydrophobicity of the material was responsible for a high level of bacterial initial binding to the surface [56,64]. A study conducted using *S. epidermidis* even showed that an increase in the degree of hydrophobicity of materials was linearly correlated with the number of adherent bacterial cells on the surface [57]. The propensity of *S. aureus* cells to adhere more to a hydrophobic surface such as PDMS compared to the hydrophilic PTU/ZnO variations, may be due to the presence of hydrophobic components at its cell surface, such as proteins also known as adhesins, which are important factors promoting strong adhesion to hydrophobic surfaces [58,65]. Furthermore, Maikranz et al. recently showed that *S. aureus* binds with many weakly binding macromolecules to hydrophobic surfaces, while only a few selected but strong macromolecules bind to hydrophilic surfaces. It is also suggested that the hydrophobic molecules form rather fast bonds while the hydrophilic binding molecules have to overcome a potential barrier [66], which may explain the decreased adhesion of cells on hydrophilic surfaces such as the PTU/ZnO composites produced in this study. Another reason that could explain the differences between PDMS and the PTU/ZnO composites is the surface charge of the material. The cell surface of *S. aureus* is moderately negatively charged (−4 to 6 mV) [67,68], thus, slightly negatively charged surfaces could be colonized by *S. aureus* cells, as long as van der Waals forces or other forces overcome the repulsion [69]. Furthermore, a study by Gross et al. suggests that increasing the repulsive forces between *S. aureus* and the polymer surface might complicate biofilm formation [70]. Thus, the fact that PTU/ZnO composites have a more negatively charged surface is also an important factor that can notably explain the smaller percentages of adhesion forces on these surfaces compared to PDMS. Another factor that seems to influence cell adhesion in the case of *S. aureus* is the stiffness of the materials. Only a few studies have reported on the influence of material stiffness on cell adhesion, but

no consistent trend has been found on the subject [71]. For example, Lichter et al. found that *S. epidermidis* adhered better to stiff polyelectrolyte multilayered thin films [72], while Wang et al. found that the adhesion of *S. aureus* was reduced on polyacrylamide hydrogels with increasing stiffness [73]. In our case, we show that adhesion to stiffer materials is reduced, as the adhesion force recorded on PDMS is much higher than the one recorded on PTU variations. The stiffnesses of the PTU variations are not significantly different from each other except in the case of PTU containing 5 wt.% of t-ZnO, which is stiffer than the other PTU/ZnO composites and where the adhesion force recorded is significantly lower ( $86 \pm 120$  pN). While it is difficult to evaluate the independent effects of each parameter (charge, wetting property and stiffness) on the cell adhesion, in this case, we could suggest the hypothesis that at similar charge and WCA, an increased stiffness of the material might reduce the adhesion of cells onto it.

The same type of experiment was conducted with *C. glabrata*. Given the large size of these cells compared to *S. aureus* (five times larger), they were directly attached to tipless cantilevers and used to probe the interactions with the different material variations. The adhesion force and rupture histograms obtained are presented in Figure 2. As for *S. aureus*, in this case PDMS also shows the highest adhesion force of  $11,164 \pm 5387$  pN and the highest rupture distance of  $1187 \pm 424$  nm among the material variations (Figure 2a), including PDMS and PTU/ZnO composites. In comparison, pure PTU shows a significantly lower average adhesion force with  $4881 \pm 1744$  pN and a corresponding rupture distance of  $825 \pm 404$  nm (Figure 2b). With addition of 1 wt.% of t-ZnO and s-ZnO, the average adhesion force and rupture distance values only change slightly, whereas the addition of 5 wt.% of s-ZnO raises the adhesion force to  $4909 \pm 1733$  pN. These results are summarized in Table 6. Statistical tests showed that the adhesion forces are significantly different from each other ( $p$ -value  $< 0.01$ ), with an exception made by PTU with a filler content of 5 wt.% of s-ZnO, which is not significantly different from pure PTU. Thus, the results that are obtained in the case of *C. glabrata* show the same tendency as the ones obtained with *S. aureus*: the adhesion of the cells is strongly decreased on all PTU variations compared to PDMS. However, for *C. glabrata*, increasing the amount of spherical ZnO particles present in PTU does not seem to further decrease cell adhesion, but it does in the case of PTU filled with 5 wt.% of t-ZnO, where the lowest adhesion force is recorded. This was already the case for *S. aureus*. Moreover, it is interesting to note that compared to *S. aureus*, the amount of force curves presenting adhesive events for all investigated polymer surfaces was 100%, and that both adhesion forces and rupture distance values were higher for *C. glabrata*. This is probably due to the increased size of cells (approximately  $5 \mu\text{m}$  in diameter), which, thus, increases the surface contact between the cells and the material, thus explaining the higher forces recorded. In this case, the long rupture distances measured and the presence of multiple peaks located after the contact point on the retract force curves also show that specific bonds are at play in the interaction. Here, strengthening the initial physico-chemical bonds also resulted in a transition from reversible to irreversible attachment [5].



**Figure 1.** Single-cell force spectroscopy to probe the interactions between *S. aureus* and material surfaces. Adhesion force histogram and corresponding rupture distance histogram obtained for seven cells interacting with (a) PDMS, (b) PTU, (c) PTU containing 1 wt.% of *s*-ZnO, (d) PTU containing 5 wt.% of *s*-ZnO, (e) PTU containing 1 wt.% of *t*-ZnO and (f) PTU containing 5 wt.% of *s*-ZnO. Insets in rupture distance histograms show representative force curves obtained in each case. % adh. is the % of force curves presenting adhesive events. Data were recorded using a set-point of 0.25 nN. 3.3 PTU/ZnO composites are also efficient anti-adhesive surfaces against yeast cells.



**Figure 2.** Single-cell force spectroscopy to probe the interactions between *C. glabrata* and material surfaces. Adhesion force histogram and corresponding rupture distance histogram obtained for seven cells interacting with (a) PDMS surface, (b) PTU surface, (c) PTU containing 1 wt.% of s-ZnO, (d) PTU containing 5 wt.% of s-ZnO, (e) PTU containing 1 wt.% of t-ZnO and (f) PTU containing 5 wt.% of s-ZnO. Insets in rupture distance histograms show representative force curves obtained in each case. % adh. is the % of force curves presenting adhesive events. Data were recorded using a set-point of 0.25 nN.


**Table 6.** The mean value of the adhesion force and the rupture distance of *C. glabrata* on a variation of six different material surfaces. The right column presents the percentage of curves presenting adhesive events.

Material Variation	Adhesion Force (pN)	Rupture Distance (nm)	% of Adhesion
PDMS	11,164 ± 5387	1187 ± 424	100%
PTU	4881 ± 1744	825 ± 404	100%
PTU/1 wt.% s-ZnO	3942 ± 1082	795 ± 401	100%
PTU/1 wt.% t-ZnO	4033 ± 1330	1000 ± 423	100%
PTU/5 wt.% s-ZnO	4909 ± 1733	757 ± 386	100%
PTU/5 wt.% t-ZnO	2684 ± 1066	784 ± 466	100%

The reasons for these lower adhesion values on PTU and PTU/ZnO variations compared to PDMS are probably the same as discussed previously in the case of *S. aureus*, i.e., lower water contact angle, higher negatively charged surface and higher stiffness of the PTU/ZnO composites. In particular, in the case of yeast cells, a study by El-Kirat-Chatel et al. demonstrated that the hydrophobic surface chemistry of the substrates plays an important role in the adhesion of the *C. glabrata* [60]. It shows that the adhesion of *C. glabrata* is mostly promoted by adhesins binding to hydrophobic surfaces. Which means that *C. glabrata* adheres more strongly to hydrophobic surfaces. This is in line with our findings. Regarding the surface charge of the materials, *C. glabrata* was found to bear a surface  $\zeta$ -potential of  $-24$  mV [74]. Thus, in this case, the repulsion of *C. glabrata* may also be more pronounced on PTU/ZnO composite surfaces than on PDMS, which has a smaller negative surface charge. Thus, perhaps explaining in part the reduced adhesion forces, although in this case, the percentage of force curves presenting adhesive events is still 100% for PTU/ZnO variations. Finally, regarding the stiffness of the materials, as far as we are aware, no studies have reported its influence on yeast cell adhesion. For PTU/ZnO composites, the adhesion force recorded is the lowest in the case of PTU with a filler content of 5 wt.% of t-ZnO particles. While the surface charge and the water contact angle are similar for all PTU/ZnO composites, the stiffness is significantly different for PTU/ZnO with 5 wt.% of t-ZnO. Thus, also in this case, it seems that stiffness is an important parameter to control the adhesion of microbes.

#### 4. Conclusions

Here, we demonstrate the development and characterization of new anti-fouling materials based on PTU filled with ZnO particles. These materials demonstrate promising anti-adhesion properties. Using single-cell force spectroscopy, we were able to show that the adhesion of two medically-relevant microorganisms, *S. aureus* and *C. glabrata* towards PTU/ZnO variations is successfully reduced when compared to PDMS (Figure 3). PTU/ZnO based materials exhibit a hydrophilic surface and a negative electro-static potential compared to conventional PDMS, which impedes the adhesion of *S. aureus* and *C. glabrata*, explaining the low adhesion forces compared to PDMS. Furthermore, the adhesion force recorded was significantly reduced for both microorganisms in the case of PTU containing 5 wt.% of t-ZnO, which might be the result of the increased stiffness of this particular variation compared to the others. Thus, this study shows that PTU/ZnO composite materials are efficient at reducing the initial attachment of microbial cells, and therefore show great potential for biomedical applications.



**Figure 3.** Recapitulative histogram showing the compared adhesion forces of *S. aureus* (orange bars) and *C. glabrata* (blue bars) to PDMS, PTU and PTU/ZnO composites.

**Supplementary Materials:** The following are available online at <https://www.mdpi.com/2079-4991/11/2/271/s1>, Figure S1: SEM micrographs of the different PTU/ZnO composites used in this study. Scale-bar = 100  $\mu\text{m}$ , Figure S2: AFM height images of PDMS, PTU and PTU/ZnO composites. Scale-bar = 5  $\mu\text{m}$ , Figure S3: Water contact angles photographs of PDMS, PTU and PTU/ZnO composites. Scale-bar = 500  $\mu\text{m}$ , Table S1: Zn/C ratio on material surface for variations of ZnO particles in the PTU matrix measured by EDX, Table S2:  $\text{Zn}^{2+}$  concentration on material surface for variations of ZnO particles in the PTU matrix measured by Zinc ion release method with zincon.

**Author Contributions:** Conceptualization, M.B., M.C.T., C.T., A.G., R.A., E.D., M.C. (Mickaël Castelain) and C.F.-D.; data collection and analysis, S.K., H.Q., Z.N., M.C. (Mafalda Cavalheiro), C.T.; data interpretation, S.K., M.B., H.Q., Z.N., M.C. (Mafalda Cavalheiro), M.C.T., C.T., A.G., R.A., E.D., M.C. (Mickaël Castelain), C.F.-D.; writing—original draft preparation, S.K., M.B. and C.F.-D. All authors revised the manuscript. All authors have read and agreed to the published version of the manuscript.

**Funding:** This research was funded by the Federal Ministry for Economic Affairs and Energy within the ZIM-program (16KN021202), and the Cluster of Excellence 80 “The Future Ocean”. The “Future Ocean” was funded within the framework of the Excellence Initiative by the Deutsche Forschungsgemeinschaft (DFG) on behalf of the German federal and state governments. S.K. acknowledges financial support from Kiel University, from the PROMOS scholarship program (German Academic Exchange Service—DAAD) and from LAAS-CNRS, department Micro Nano Bio Technology (MNBT) for her Master internship. C.F.-D. and E.D. are researchers at Centre National de la Recherche Scientifique (CNRS). C.F.-D. acknowledges financial support from the Agence Nationale de la Recherche, JCJC project FLOTALG (ANR-18-CE42-000). M.C. acknowledges financial support from Fundação para a Ciência e a Tecnologia (PhD scholarship PD/BD/116946/2016). Z.N. acknowledges financial support from the Chinese Scholarship Council (PhD scholarship).

**Institutional Review Board Statement:** Not applicable.

**Informed Consent Statement:** Not applicable.

**Data Availability Statement:** The data presented in this study are available on request from the corresponding author.

**Acknowledgments:** This research was funded by the Federal Ministry for Economic Affairs and Energy within the ZIM-program (16KN021202), and the Cluster of Excellence 80 “The Future Ocean”. The “Future Ocean” was funded within the framework of the Excellence Initiative by the Deutsche Forschungsgemeinschaft (DFG) on behalf of the German federal and state governments. S.K. acknowledges financial support from Kiel University, from the PROMOS scholarship program (German Academic Exchange Service—DAAD) and from LAAS-CNRS, department Micro Nano Bio Technology (MNBT) for her Master internship. C.F.-D. and E.D. are researchers at Centre National de la Recherche Scientifique (CNRS). C.F.-D. acknowledges financial support from the Agence Nationale de la Recherche, JJC project FLOTALG (ANR-18-CE42-000). M.C. acknowledges financial support from Fundação para a Ciência e a Tecnologia (PhD scholarship PD/BD/116946/2016). Z.N. acknowledges financial support from the Chinese Scholarship Council (PhD scholarship).

**Conflicts of Interest:** The authors declare no conflict of interest.


## References

1. Khan, H.A.; Ahmad, A.; Mehboob, R. Nosocomial Infections and Their Control Strategies. *Asian Pac. J. Trop. Biomed.* **2015**, *5*, 509–514. [[CrossRef](#)]
2. Chen, M.; Yu, Q.; Sun, H. Novel Strategies for the Prevention and Treatment of Biofilm Related Infections. *Int. J. Mol. Sci.* **2013**, *14*, 18488–18501. [[CrossRef](#)] [[PubMed](#)]
3. Cavalheiro, M.; Teixeira, M.C. Candida Biofilms: Threats, Challenges, and Promising Strategies. *Front. Med.* **2018**, *5*, 28. [[CrossRef](#)]
4. Percival, S.L.; Suleman, L.; Vuotto, C.; Donelli, G. Healthcare-Associated Infections, Medical Devices and Biofilms: Risk, Tolerance and Control. *J. Med. Microbiol.* **2015**, *64*, 323–334. [[CrossRef](#)] [[PubMed](#)]
5. Carniello, V.; Peterson, B.W.; van der Mei, H.C.; Busscher, H.J. Physico-Chemistry from Initial Bacterial Adhesion to Surface-Programmed Biofilm Growth. *Adv. Colloid Interface Sci.* **2018**, *261*, 1–14. [[CrossRef](#)]
6. Terada, A.; Yuasa, A.; Kushimoto, T.; Tsuneda, S.; Katakai, A.; Tamada, M. Bacterial Adhesion to and Viability on Positively Charged Polymer Surfaces. *Microbiology* **2006**, *152*, 3575–3583. [[CrossRef](#)]
7. Roy, R.; Tiwari, M.; Donelli, G.; Tiwari, V. Strategies for Combating Bacterial Biofilms: A Focus on Anti-Biofilm Agents and Their Mechanisms of Action. *Virulence* **2018**, *9*, 522–554. [[CrossRef](#)]
8. Kaplan, J.B. Biofilm Matrix-Degrading Enzymes. *Methods Mol. Biol.* **2014**, *1147*, 203–213. [[CrossRef](#)]
9. Kalia, V.C. Quorum Sensing Inhibitors: An Overview. *Biotechnol. Adv.* **2013**, *31*, 224–245. [[CrossRef](#)]
10. Ong, K.S.; Mawang, C.I.; Daniel-Jambun, D.; Lim, Y.Y.; Lee, S.M. Current Anti-Biofilm Strategies and Potential of Antioxidants in Biofilm Control. *Expert Rev. Anti-Infect. Ther.* **2018**, *16*, 855–864. [[CrossRef](#)]
11. Formosa-Dague, C.; Speziale, P.; Foster, T.J.; Geoghegan, J.A.; Dufrene, Y.F. Zinc-Dependent Mechanical Properties of Staphylococcus Aureus Biofilm-Forming Surface Protein SasG. *Proc. Natl. Acad. Sci. USA* **2016**, *113*, 410–415. [[CrossRef](#)]
12. Feuillie, C.; Formosa-Dague, C.; Hays, L.M.C.; Vervaeck, O.; Derclaye, S.; Brennan, M.P.; Foster, T.J.; Geoghegan, J.A.; Dufrene, Y.F. Molecular Interactions and Inhibition of the Staphylococcal Biofilm-Forming Protein SdrC. *Proc. Natl. Acad. Sci. USA* **2017**, *114*, 3738–3743. [[CrossRef](#)] [[PubMed](#)]
13. Antonelli, M.; De Pascale, G.; Ranieri, V.M.; Pelaia, P.; Tufano, R.; Piazza, O.; Zangrillo, A.; Ferrario, A.; De Gaetano, A.; Guaglianone, E.; et al. Comparison of Triple-Lumen Central Venous Catheters Impregnated with Silver Nanoparticles (AgTive®) vs Conventional Catheters in Intensive Care Unit Patients. *J. Hosp. Infect.* **2012**, *82*, 101–107. [[CrossRef](#)] [[PubMed](#)]
14. Legeay, G.; Poncin-Epaillard, F.; Arciola, C.R. New Surfaces with Hydrophilic/Hydrophobic Characteristics in Relation to (No)Bioadhesion. *Int. J. Artif. Organs* **2018**. [[CrossRef](#)] [[PubMed](#)]
15. Desrousseaux, C.; Sautou, V.; Descamps, S.; Traoré, O. Modification of the Surfaces of Medical Devices to Prevent Microbial Adhesion and Biofilm Formation. *J. Hosp. Infect.* **2013**, *85*, 87–93. [[CrossRef](#)]
16. Geoghegan, M.; Andrews, J.S.; Biggs, C.A.; Eboigbodin, K.E.; Elliott, D.R.; Rolfe, S.; Scholes, J.; Ojeda, J.J.; Romero-González, M.E.; Edyvean, R.G.J.; et al. The Polymer Physics and Chemistry of Microbial Cell Attachment and Adhesion. *Faraday Discuss.* **2008**, *139*, 85–103. [[CrossRef](#)]
17. Xu, L.-C.; Siedlecki, C.A. Staphylococcus Epidermidis Adhesion on Hydrophobic and Hydrophilic Textured Biomaterial Surfaces. *Biomed. Mater.* **2014**, *9*, 035003. [[CrossRef](#)]
18. Hizal, F.; Rungraeng, N.; Lee, J.; Jun, S.; Busscher, H.J.; van der Mei, H.C.; Choi, C.-H. Nanoengineered Superhydrophobic Surfaces of Aluminum with Extremely Low Bacterial Adhesivity. *ACS Appl. Mater. Interfaces* **2017**, *9*, 12118–12129. [[CrossRef](#)]
19. Banerjee, I.; Pangule, R.C.; Kane, R.S. Antifouling Coatings: Recent Developments in the Design of Surfaces That Prevent Fouling by Proteins, Bacteria, and Marine Organisms. *Adv. Mater.* **2011**, *23*, 690–718. [[CrossRef](#)]
20. Vendra, V.; Wu, L.; Krishnan, S. Polymer Thin Films for Biomedical Applications. In *Nanostructured Thin Films and Surfaces*; Wiley: Hoboken, NJ, USA, 2011; Volume 5, pp. 1–51, ISBN 978-3-527-32155-1.
21. Herold, D.A.; Keil, K.; Bruns, D.E. Oxidation of Polyethylene Glycols by Alcohol Dehydrogenase. *Biochem. Pharmacol.* **1989**, *38*, 73–76. [[CrossRef](#)]
22. Berne, C.; Ellison, C.K.; Ducret, A.; Brun, Y.V. Bacterial Adhesion at the Single-Cell Level. *Nat. Rev. Microbiol.* **2018**, *16*, 616–627. [[CrossRef](#)] [[PubMed](#)]


23. Ercan, B.; Taylor, E.; Alpaslan, E.; Webster, T.J. Diameter of Titanium Nanotubes Influences Anti-Bacterial Efficacy. *Nanotechnology* **2011**, *22*, 295102. [[CrossRef](#)] [[PubMed](#)]
24. Puckett, S.D.; Taylor, E.; Raimondo, T.; Webster, T.J. The Relationship between the Nanostructure of Titanium Surfaces and Bacterial Attachment. *Biomaterials* **2010**, *31*, 706–713. [[CrossRef](#)] [[PubMed](#)]
25. Fadeeva, E.; Truong, V.K.; Stiesch, M.; Chichkov, B.N.; Crawford, R.J.; Wang, J.; Ivanova, E.P. Bacterial Retention on Superhydrophobic Titanium Surfaces Fabricated by Femtosecond Laser Ablation. *Langmuir* **2011**, *27*, 3012–3019. [[CrossRef](#)] [[PubMed](#)]
26. Wang, Y.; Wen, C.; Hodgson, P.; Li, Y. Biocompatibility of TiO<sub>2</sub> Nanotubes with Different Topographies. *J. Biomed. Mater. Res. Part A* **2014**, *102*, 743–751. [[CrossRef](#)]
27. Pavličević, J.; Špirková, M.; Bera, O.; Jovičić, M.; Pilić, B.; Baloš, S.; Budinski-Simendić, J. The Influence of ZnO Nanoparticles on Thermal and Mechanical Behavior of Polycarbonate-Based Polyurethane Composites. *Compos. Part B Eng.* **2014**, *60*, 673–679. [[CrossRef](#)]
28. Hölken, I.; Hoppe, M.; Mishra, Y.K.; Gorb, S.N.; Adelung, R.; Baum, M.J. Complex Shaped ZnO Nano- and Microstructure Based Polymer Composites: Mechanically Stable and Environmentally Friendly Coatings for Potential Antifouling Applications. *Phys. Chem. Chem. Phys.* **2016**, *18*, 7114–7123. [[CrossRef](#)]
29. Kultys, A.; Rogulska, M.; Pikus, S. The Synthesis and Characterization of New Thermoplastic Poly(Thiourethane-Urethane)s. *J. Polym. Sci. Part A Polym. Chem.* **2008**, *46*, 1770–1782. [[CrossRef](#)]
30. Rogulska, M.; Kultys, A.; Olszewska, E. New Thermoplastic Poly(Thiourethane-Urethane) Elastomers Based on Hexane-1,6-Diyl Diisocyanate (HDI). *J. Therm. Anal. Calorim.* **2013**. [[CrossRef](#)]
31. Hölken, I. Mechanically Stable and Environmentally Friendly Polymer/Particle Composites for the Application as Low-Fouling Coating in the Marine Sector. Ph.D. Thesis, Technischen Fakultät der Christian-Albrechts-Universität zu Kiel, Kiel, Germany, 2016.
32. Hölken, I.; Hoppe, M.; Adelung, R.; Baum, M. Functional Ecofriendly Coatings for Marine Applications. Proceedings on the 3rd International Conference on Nanotechnologies and Biomedical Engineering, Chisinau, Republic of Moldova, 23–26 September 2015. [[CrossRef](#)]
33. Qiu, H.; Hölken, I.; Gapeeva, A.; Filiz, V.; Adelung, R.; Baum, M. Development and Characterization of Mechanically Durable Silicone-Polythiourethane Composites Modified with Tetrapodal Shaped ZnO Particles for the Potential Application as Fouling-Release Coating in the Marine Sector. *Materials* **2018**, *11*, 2413. [[CrossRef](#)]
34. Mishra, Y.K.; Adelung, R. ZnO Tetrapod Materials for Functional Applications. *Mater. Today* **2018**, *21*, 631–651. [[CrossRef](#)]
35. Antoine, T.E.; Hadigal, S.R.; Yakoub, A.M.; Mishra, Y.K.; Bhattacharya, P.; Haddad, C.; Valyi-Nagy, T.; Adelung, R.; Prabhakar, B.S.; Shukla, D. Intravaginal Zinc Oxide Tetrapod Nanoparticles as Novel Immunoprotective Agents against Genital Herpes. *J. Immunol.* **2016**, *196*, 4566–4575. [[CrossRef](#)] [[PubMed](#)]
36. Sirelkhatim, A.; Mahmud, S.; Seeni, A.; Kaus, N.H.M.; Ann, L.C.; Bakhori, S.K.M.; Hasan, H.; Mohamad, D. Review on Zinc Oxide Nanoparticles: Antibacterial Activity and Toxicity Mechanism. *Nanomicro Lett.* **2015**, *7*, 219–242. [[CrossRef](#)] [[PubMed](#)]
37. Papavlassopoulos, H.; Mishra, Y.K.; Kaps, S.; Paulowicz, I.; Abdelaziz, R.; Elbahri, M.; Maser, E.; Adelung, R.; Röhl, C. Toxicity of Functional Nano-Micro Zinc Oxide Tetrapods: Impact of Cell Culture Conditions, Cellular Age and Material Properties. *PLoS ONE* **2014**, *9*, e84983. [[CrossRef](#)] [[PubMed](#)]
38. Niu, L.N.; Fang, M.; Jiao, K.; Tang, L.H.; Xiao, Y.H.; Shen, L.J.; Chen, J.H. Tetrapod-like Zinc Oxide Whisker Enhancement of Resin Composite. *J. Dent. Res.* **2010**, *89*, 746–750. [[CrossRef](#)]
39. Song, F.; Koo, H.; Ren, D. Effects of Material Properties on Bacterial Adhesion and Biofilm Formation. *J. Dent. Res.* **2015**, *94*, 1027–1034. [[CrossRef](#)] [[PubMed](#)]
40. Binnig, G.; Quate, C.F.; Gerber, C. Atomic Force Microscope. *Phys. Rev. Lett.* **1986**, *56*, 930–934. [[CrossRef](#)] [[PubMed](#)]
41. Otto, M. Staphylococcal Biofilms. In *Bacterial Biofilms*; Romeo, T., Ed.; Current Topics in Microbiology and Immunology; Springer: Berlin/Heidelberg, Germany, 2008; pp. 207–228, ISBN 978-3-540-75418-3.
42. Rodrigues, C.F.; Silva, S.; Henriques, M. Candida Glabrata: A Review of Its Features and Resistance. *Eur. J. Clin. Microbiol. Infect. Dis.* **2014**, *33*, 673–688. [[CrossRef](#)]
43. Beaussart, A.; El-Kirat-Chatel, S.; Sullan, R.M.A.; Alsteens, D.; Herman, P.; Derclaye, S.; Dufrêne, Y.F. Quantifying the Forces Guiding Microbial Cell Adhesion Using Single-Cell Force Spectroscopy. *Nat. Protoc.* **2014**, *9*, 1049–1055. [[CrossRef](#)]
44. Teo, A.J.T.; Mishra, A.; Park, I.; Kim, Y.-J.; Park, W.-T.; Yoon, Y.-J. Polymeric Biomaterials for Medical Implants and Devices. *ACS Biomater. Sci. Eng.* **2016**, *2*, 454–472. [[CrossRef](#)]
45. Mata, A.; Fleischman, A.J.; Roy, S. Characterization of Polydimethylsiloxane (PDMS) Properties for Biomedical Micro/Nanosystems. *Biomed. Microdevices* **2005**, *7*, 281–293. [[CrossRef](#)] [[PubMed](#)]
46. Lima, R.; Wada, S.; Tanaka, S.; Takeda, M.; Ishikawa, T.; Tsubota, K.; Imai, Y.; Yamaguchi, T. In Vitro Blood Flow in a Rectangular PDMS Microchannel: Experimental Observations Using a Confocal Micro-PIV System. *Biomed. Microdevices* **2008**, *10*, 153–167. [[CrossRef](#)] [[PubMed](#)]
47. Mishra, Y.K.; Kaps, S.; Schuchardt, A.; Paulowicz, I.; Jin, X.; Gedamu, D.; Freitag, S.; Claus, M.; Wille, S.; Kovalev, A.; et al. Fabrication of Macroscopically Flexible and Highly Porous 3D Semiconductor Networks from Interpenetrating Nanostructures by a Simple Flame Transport Approach. *Part. Part. Syst. Charact.* **2013**, *30*, 775–783. [[CrossRef](#)]

48. Mishra, Y.K.; Modi, G.; Cretu, V.; Postica, V.; Lupan, O.; Reimer, T.; Paulowicz, I.; Hrkac, V.; Benecke, W.; Kienle, L.; et al. Direct Growth of Freestanding ZnO Tetrapod Networks for Multifunctional Applications in Photocatalysis, UV Photodetection, and Gas Sensing. *ACS Appl. Mater. Interfaces* **2015**, *7*, 14303–14316. [[CrossRef](#)]
49. Hutter, J.L.; Bechhoefer, J. Calibration of Atomic-Force Microscope Tips. *Rev. Sci. Instrum.* **1993**, *64*, 1868–1873. [[CrossRef](#)]
50. Delgado, A.V.; González-Caballero, F.; Hunter, R.J.; Koopal, L.K.; Lyklema, J. Measurement and Interpretation of Electrokinetic Phenomena (IUPAC Technical Report). *Pure Appl. Chem.* **2005**, *77*, 1753–1805. [[CrossRef](#)]
51. Bhushan, B. *Introduction to Tribology*, 2nd ed.; Wiley: Chichester, West Sussex, UK, 2013; ISBN 978-1-119-94453-9.
52. Bollen, C.M.; Papaioanno, W.; Van Eldere, J.; Schepers, E.; Quirynen, M.; van Steenberghe, D. The Influence of Abutment Surface Roughness on Plaque Accumulation and Peri-Implant Mucositis. *Clin. Oral. Implant. Res.* **1996**, *7*, 201–211. [[CrossRef](#)] [[PubMed](#)]
53. Popov, V.L. *Contact Mechanics and Friction: Physical Principles and Applications*; Springer: Berlin/Heidelberg, Germany, 2010; ISBN 978-3-642-10803-7.
54. Busscher, H.J.; Norde, W.; van der Mei, H.C. Specific Molecular Recognition and Nonspecific Contributions to Bacterial Interaction Forces. *Appl. Environ. Microbiol.* **2008**, *74*, 2559–2564. [[CrossRef](#)]
55. Rosenberg, M.; Kjelleberg, S. Hydrophobic Interactions: Role in Bacterial Adhesion. In *Advances in Microbial Ecology*; Marshall, K.C., Ed.; Advances in Microbial Ecology; Springer: Boston, MA, USA, 1986; pp. 353–393, ISBN 978-1-4757-0611-6.
56. Boks, N.P.; Busscher, H.J.; van der Mei, H.C.; Norde, W. Bond-Strengthening in Staphylococcal Adhesion to Hydrophilic and Hydrophobic Surfaces Using Atomic Force Microscopy. *Langmuir* **2008**, *24*, 12990–12994. [[CrossRef](#)]
57. Oliveira, R.; Azeredo, J.; Teixeira, P.; Fonseca, A.P. *The Role of Hydrophobicity in Bacterial Adhesion*; Bionline: Toronto, ON, USA, 2001.
58. Doyle, R.J. Contribution of the Hydrophobic Effect to Microbial Infection. *Microbes Infect.* **2000**, *2*, 391–400. [[CrossRef](#)]
59. Rosenberg, M.; Gutnick, D.; Rosenberg, E. Adherence of Bacteria to Hydrocarbons: A Simple Method for Measuring Cell-Surface Hydrophobicity. *FEMS Microbiol. Lett.* **1980**, *9*, 29–33. [[CrossRef](#)]
60. El-Kirat-Chatel, S.; Beaussart, A.; Derclaye, S.; Alsteens, D.; Kucharíková, S.; Van Dijck, P.; Dufrêne, Y.F. Force Nanoscopy of Hydrophobic Interactions in the Fungal Pathogen *Candida Glabrata*. *ACS Nano* **2015**, *9*, 1648–1655. [[CrossRef](#)] [[PubMed](#)]
61. Bos, R.; van der Mei, H.C.; Busscher, H.J. Physico-Chemistry of Initial Microbial Adhesive Interactions—Its Mechanisms and Methods for Study. *FEMS Microbiol. Rev.* **1999**, *23*, 179–230. [[CrossRef](#)]
62. Olitzki, L. Electric Charge of Bacterial Antigens. *J. Immunol.* **1932**, *22*, 251–256.
63. Francolini, I.; Piozzi, A.; Donelli, G. Efficacy Evaluation of Antimicrobial Drug-Releasing Polymer Matrices. In *Microbial Biofilms: Methods and Protocols*; Donelli, G., Ed.; Methods in Molecular Biology; Springer: New York, NY, USA, 2014; pp. 215–225, ISBN 978-1-4939-0467-9.
64. Sousa, C.; Teixeira, P.; Oliveira, R. Influence of Surface Properties on the Adhesion of *Staphylococcus Epidermidis* to Acrylic and Silicone. Available online: <https://www.hindawi.com/journals/ijbm/2009/718017/abs/> (accessed on 3 October 2019).
65. Krasowska, A.; Sigler, K. How Microorganisms Use Hydrophobicity and What Does This Mean for Human Needs? *Front. Cell. Infect. Microbiol.* **2014**, *4*. [[CrossRef](#)]
66. Maikranz, E.; Spengler, C.; Thewes, N.; Thewes, A.; Nolle, F.; Jung, P.; Bischoff, M.; Santen, L.; Jacobs, K. Different Binding Mechanisms of *Staphylococcus Aureus* to Hydrophobic and Hydrophilic Surfaces. *Nanoscale* **2020**, *12*, 19267–19275. [[CrossRef](#)]
67. Chen, L.-C.; Kung, S.-K.; Chen, H.-H.; Lin, S.-B. Evaluation of Zeta Potential Difference as an Indicator for Antibacterial Strength of Low Molecular Weight Chitosan. *Carbohydr. Polym.* **2010**, *82*, 913–919. [[CrossRef](#)]
68. Sonohara, R.; Muramatsu, N.; Ohshima, H.; Kondo, T. Difference in Surface Properties between *Escherichia Coli* and *Staphylococcus Aureus* as Revealed by Electrophoretic Mobility Measurements. *Biophys. Chem.* **1995**, *55*, 273–277. [[CrossRef](#)]
69. Rijnaarts, H.H.M.; Norde, W.; Bouwer, E.J.; Lyklema, J.; Zehnder, A.J.B. Reversibility and Mechanism of Bacterial Adhesion. *Coll. Surf. B. Biointerfaces* **1995**, *5*–22. [[CrossRef](#)]
70. Gross, M.; Cramton, S.E.; Götz, F.; Peschel, A. Key Role of Teichoic Acid Net Charge in *Staphylococcus Aureus* Colonization of Artificial Surfaces. *Infect. Immun.* **2001**, *69*, 3423–3426. [[CrossRef](#)]
71. Siddiqui, S.; Chandrasekaran, A.; Lin, N.; Tufenkji, N.; Moraes, C. Microfluidic Shear Assay to Distinguish between Bacterial Adhesion and Attachment Strength on Stiffness-Tunable Silicone Substrates. *Langmuir* **2019**, *35*, 8840–8849. [[CrossRef](#)] [[PubMed](#)]
72. Lichter, J.A.; Thompson, M.T.; Delgadillo, M.; Nishikawa, T.; Rubner, M.F.; Van Vliet, K.J. Substrata Mechanical Stiffness Can Regulate Adhesion of Viable Bacteria. *Biomacromolecules* **2008**, *9*, 1571–1578. [[CrossRef](#)]
73. Wang, Y.; Guan, A.; Isayeva, I.; Vorvolakos, K.; Das, S.; Li, Z.; Phillips, K.S. Interactions of *Staphylococcus Aureus* with Ultrasoft Hydrogel Biomaterials. *Biomaterials* **2016**, *95*, 74–85. [[CrossRef](#)] [[PubMed](#)]
74. Nikawa, H.; Sadamori, S.; Hamada, T.; Satou, N.; Okuda, K. Non-Specific Adherence of *Candida* Species to Surface-Modified Glass. *J. Med. Vet. Mycol.* **1989**, *27*, 269–271. [[CrossRef](#)] [[PubMed](#)]


## Supplementary materials



**Figure S1.** SEM micrographs of the different PTU/ZnO composites used in this study. Scale-bar = 100 µm.



**Figure S2.** AFM height images of PDMS, PTU and PTU/ZnO composites. Scale-bar = 5 μm



**Figure S3.** Water contact angles photographs of PDMS, PTU and PTU/ZnO composites. Scale-bar = 500  $\mu\text{m}$ .

**Table S1.** Zn/C ratio on material surface for variations of ZnO particles in the PTU matrix measured by EDX. As Zinc (ZnO) is only contained in ZnO and Carbon (C) is only contained in PTU, the amount of ZnO present on the surface was determined by the Zn-to-C (Zn/C) ratio.

<b>Material Variation</b>	<b>Zn/C</b>
PTU/1 wt.% t-ZnO	ND
PTU/1 wt.% s-ZnO	0,008
PTU/5 wt.% t-ZnO	ND
PTU/5 wt.% s-ZnO	ND

**Table S2.** Zn<sup>2+</sup> concentration on material surface for variations of ZnO particles in the PTU matrix measured by Zinc ion release method with zincon.

<b>Material Variation</b>	<b>Zn<sup>2+</sup> (mg/mL)</b>
PTU/1 wt.% t-ZnO	ND
PTU/1 wt.% s-ZnO	ND
PTU/5 wt.% t-ZnO	ND
PTU/5 wt.% s-ZnO	ND