Microfoam formation in a capillary

Spiros Kotopoulis, Michiel Postema

► To cite this version:

Spiros Kotopoulis, Michiel Postema. Microfoam formation in a capillary. Ultrasonics, 2010, 50 (2), pp.260-268. 10.1016/j.ultras.2009.09.028 . hal-03193286

HAL Id: hal-03193286

<https://hal.science/hal-03193286>

Submitted on 12 Apr 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Distributed under a Creative Commons Attribution - NonCommercial - NoDerivatives| 4.0 International License

Microfoam formation in a capillary

Spiros Kotopoulos, Michiel Postema¹

*Emmy-Noether Group
Institute of Medical Engineering
Department of Electrical Engineering and Information Technology
Ruhr-Universität Bochum
D-44780 Bochum
Germany*

*Department of Engineering
The University of Hull
Kingston upon Hull
HU6 7RX
United Kingdom*

¹ Corresponding author. Address: Emmy-Noether Group, Institute of Medical Engineering, Department of Electrical Engineering and Information Technology, Ruhr-Universität Bochum, D-44780 Bochum, Germany.
Telephone: +49 234 3226504. E-mail: michiel.postema@rub.de

Abstract

The ultrasound-induced formation of bubble clusters may be of interest as a therapeutic means. If the clusters behave as one entity, *i.e.*, one mega-bubble, its ultrasonic manipulation towards a boundary is straightforward and quick. If the clusters can be forced to accumulate to a microfoam, entire vessels might be blocked on purpose using an ultrasound contrast agent and a sound source.

In this paper, we analyse how ultrasound contrast agent clusters are formed in a capillary and what happens to the clusters if sonication is continued, using continuous driving frequencies in the range 1–10 MHz. Furthermore, we show high-speed camera footage of microbubble clustering phenomena.

We observed the following stages of microfoam formation within a dense population of microbubbles before ultrasound arrival. After the sonication started, contrast microbubbles collided, forming small clusters, owing to secondary radiation forces. These clusters coalesced within the space of a quarter of the ultrasonic wavelength, owing to primary radiation forces. The resulting microfoams translated in the direction of the ultrasound field, hitting the capillary wall, also owing to primary radiation forces.

We have demonstrated that as soon as the bubble clusters are formed and as long as they are in the sound field, they behave as one entity. At our acoustic settings, it takes seconds to force the bubble clusters to positions approximately a quarter wavelength apart. It also just takes seconds to drive the clusters towards the capillary wall.

Subjecting an ultrasound contrast agent of given concentration to a continuous low-amplitude signal makes it cluster to a microfoam of known position and known size, allowing for sonic manipulation.

Key words: Capillary blocking, Embolisation, Microfoam, Radiation forces, Ultrasound contrast agent

PACS: 43.25.Yw; 47.55.df; 87.50.yt

1 Introduction

2 Ultrasound contrast agents are used in diagnostic imaging. They consist of
3 microscopically small bubbles containing slowly diffusing gas encapsulated
4 by biodegradable shells. When inserted in the blood stream, these bubbles
5 oscillate upon ultrasonic sonication, thereby creating detectable ultrasound
6 themselves. A brief overview of the most common ultrasound contrast agents
7 has been presented in [1]. It follows that albumin and lipids are currently
8 the most common bubble encapsulation materials. Because of the proven
9 feasibility to attach therapeutic compounds to albumin and lipids, therapeutic
10 applications of contrast agents have become of interest [2–5]. It is desirable that
11 the therapeutic load of any such contrast agent is released close to the vessel
12 wall. Therefore, pushing bubbles towards boundaries by means of primary
13 radiation forces has been studied [6]. Both primary and secondary radiation
14 forces resulting from oscillating bubbles, may cause the repulsion or mutual
15 attraction, and eventual collision and coalescence, of contrast agent bubbles.
16 This phenomenon has been less studied.

17 From the therapeutic point of view, the formation of bubble clusters may
18 be of interest. If the clusters behave as one entity, *i.e.*, one mega-bubble,
19 its ultrasonic manipulation towards a boundary is fairly straightforward and
20 quick. If the clusters can be forced to accumulate to a microfoam, entire vessels
21 might be blocked on purpose using an ultrasound contrast agent and a sound
22 source.

23 In this paper, we analyse how ultrasound contrast agent clusters are formed
24 and what happens to the clusters if sonication is continued. Furthermore, we
25 show high-speed camera footage of microbubble clustering phenomena and
26 discuss the therapeutic consequences of our findings.

27 2 Theory

28 A brief overview of theory on radiation forces and ultrasound contrast agent
 29 has been given in [7]. Bubble translation in the direction of the sound field is
 30 caused by a primary radiation force resulting from a pressure gradient across
 31 the bubble surface. The translation is maximal in contraction phase. The
 32 velocity v of a bubble in a steady fluid subjected to an ultrasound field can
 33 be calculated using [8]:

$$34 \quad F_r + F_d - \frac{d(mv)}{dt} \approx 0, \quad (1)$$

35 where F_r is the primary radiation force, F_d is the drag force, $m = \frac{2}{3}\pi\rho R_0^3$ is
 36 the added mass of the translating bubble, equivalent to half the mass of the
 37 displaced surrounding fluid, in which R_0 is the equilibrium bubble radius and
 38 ρ is the density of the surrounding fluid. Averaging over one acoustic cycle,
 39 the primary radiation force is given by [8,9]:

$$40 \quad F_r = \frac{p_a^2 R_0}{\rho c f} \frac{\delta \left(\frac{f_0}{f} \right)}{\left[\left(\frac{f_0}{f} \right)^2 - 1 \right]^2 + \left[\delta \left(\frac{f_0}{f} \right) \right]^2}, \quad (2)$$

41 where c is the speed of sound, p_a is the peak rarefactional acoustic pressure, δ
 42 is the dimensionless total damping coefficient [10], f is the driving frequency,
 43 and f_0 is the bubble resonance frequency [10]. The drag force is given by [9,11]:

$$44 \quad F_d = -\frac{\pi\eta}{4} C_d \text{Re } R_0 v(t), \quad (3)$$

45 where η is the shear (dynamic) viscosity of the fluid, $\text{Re} = \frac{2\rho R_0}{\eta} |v(t)|$ is the
 46 Reynolds number, and

$$47 \quad C_d = \frac{24}{\text{Re}} (1 + 0.15 \text{Re}^{0.687}) \quad (4)$$

48 is the drag coefficient of a contaminated system [12], such as a contrast agent.

Combining equations (1)–(3) and integrating over dt gives the following expression for the average velocity of a bubble:

$$v = \frac{4 p_a^2}{\rho c f \eta C_d \text{Re}} \frac{\delta \left(\frac{f_0}{f} \right)}{\left[\left(\frac{f_0}{f} \right)^2 - 1 \right]^2 + \left[\delta \left(\frac{f_0}{f} \right) \right]^2} \left[1 - e^{\left(-\frac{3 \eta C_d \text{Re}}{8 \rho R_0^2} t \right)} \right]. \quad (5)$$

Secondary radiation forces, resulting from oscillating bubbles under sonication, may cause the mutual attraction and subsequent coalescence of contrast microbubbles. Two bubbles that oscillate in phase approach each other, whereas two bubbles that oscillate out of phase recede from each other [13,14]. At low acoustic amplitudes, the phase angle difference ϕ between excursion of the oscillating bubble and the incident sound field is given by [13,14]:

$$\phi = \pi + \arctan \left(\frac{\delta \left(\frac{f}{f_0} \right)}{1 - \left(\frac{f}{f_0} \right)^2} \right). \quad (6)$$

The presence of an encapsulating shell increases the damping coefficient by a term δ_s [15]

$$\delta_s = \frac{S_f}{2\pi m f_0}, \quad (7)$$

and increases the squared resonance frequency f_0^2 by a term f_s^2 [15]

$$f_s^2 = \frac{\chi}{2\pi R_0^3 \rho}, \quad (8)$$

where S_f is the shell friction [15] and χ is the shell stiffness parameter [14,15]

$$\chi = \frac{E\epsilon}{1-v}, \quad (9)$$

in which E is Youngs modulus, ϵ is the shell thickness, and v is the Poisson ratio.

The mean approach velocity u of two identical bubbles is given by [8]:

$$u = \frac{dd}{dt} = -\frac{(2\pi f p_a)^2}{27 \eta} \rho \kappa^2 \frac{R_0^5}{d^2}, \quad (10)$$

where d is the distance between the centres of the two bubbles and κ is the compressibility of the bubble. Integrating from the initial distance between the bubbles d_0 to 0 yields the collision time

$$t_c = - \int_{d_0}^0 \frac{27\eta}{(2\pi f p_a)^2 \rho \kappa^2 R_0^5} d^2 dd = \frac{9\eta}{(2\pi f p_a)^2 \rho \kappa^2 R_0^5} d_0^3. \quad (11)$$

In a standing wave field, bubbles with resonance frequencies higher than the transmitted sound field aggregate at the pressure antinodes, whereas bubbles with resonance frequencies lower than the transmitted sound field aggregate at the pressure nodes [13]. Hence, the ultimate distance d_∞ between clusters must be a quarter of the wavelength, *i.e.*,

$$d_\infty = \frac{\lambda}{4} = \frac{c}{4f}. \quad (12)$$

Both processes of bubble clusters aggregating and the movement of clusters in the direction of the sound field can be described by a simplified version of (5):

$$v = \frac{dh}{dt} \approx \frac{p_a^2}{6 \rho c f \eta} \frac{\delta \left(\frac{f_c}{f} \right)}{\left[\left(\frac{f_c}{f} \right)^2 - 1 \right]^2 + \left[\delta \left(\frac{f_c}{f} \right) \right]^2}, \quad (13)$$

where h is the distance travelled by the cluster and f_c is the cluster resonance frequency, for which $f_c < f_0$ must hold, since the bubble cluster radius $R_c > R_0$. For the bubble cluster compressibility κ_c , $\kappa \leq \kappa_c < \kappa_f$ must hold, in which κ_f is the compressibility of a free (unencapsulated) gas bubble.

Bubble coalescence is the fusion of two or more bubbles. As adjacent bubbles collide or expand, the pressure in the film between them increases, resulting in a deformation (flattening) of the bubble surfaces. The continuing bubble expansion causes drainage of the interposed film. This thinning continues until a critical thickness around $0.1 \mu\text{m}$ is reached, at which the Van der Waals attractive forces result in film rupture and the coalescence of the bubbles [16]. Film drainage is generally much faster for free (unencapsulated) bubbles than for encapsulated bubbles, as a result of the flow pattern in the draining film [17].

The coalescence mechanism of lipid-encapsulated microbubbles was investigated, based on high-speed optical observations of sonicated lipid-encapsulated ultrasound contrast agent microbubbles [17]. It was

100 found that, when sonicated at high acoustic amplitudes, lipid-encapsulated
101 microbubbles expose free surfaces during the expansion phase, speeding up the
102 coalescence process dramatically. Hence, for the formation of bubble clouds or
103 microfoams, the use of low acoustic amplitudes is desirable.

104 3 Materials and Methods

105 A schematic overview of our experimental setup for simultaneous optical
106 observations during sonication is shown in Figure 1. A polycarbonate container
107 was built with internal dimensions: $24 \times 18 \times 15$ (cm)³. To give access to a
108 microscope objective lens, a hole with an 11-mm diameter had been drilled
109 in the base, covered with a 2-mm thick test slide (Jencons (Scientific) Ltd,
110 Leighton Buzzard, Bedfordshire, UK). The container was filled with 2.6 L tap
111 water. The container was placed on an x - y -table on top of a DM IRM inverted
112 microscope (Leica Microsystems Wetzlar GmbH, Wetzlar, Germany) with two
113 objective lenses: a 506075 C-Plan 10 $\times$ objective lens (Leica Microsystems
114 Wetzlar GmbH) with a 0.22 numerical aperture and a 506236 N-Plan 50 $\times$
115 objective lens (Leica Microsystems Wetzlar GmbH) with a 0.50 numerical
116 aperture. A Mille LuceTM Fiber Optic Illuminator Model M1000 (StockerYale,
117 Inc., Salem, NH, USA) was connected to an optic fibre with a 7-mm diameter
118 leading into the water of the container. It was placed in line with the objective
119 lens, as shown in Figure 2.

120 The charge couple device (CCD) of a FASTCAM MC1 high-speed camera
121 (Photron (Europe) Limited, West Wycombe, Bucks, United Kingdom) was
122 mounted to the microscope and connected to its processing unit, which was
123 capable of recording images at 10,000 frames per second. The camera was
124 controlled by a laptop computer.

125 3.1 Ultrasound

126 A laptop computer triggered a DATAMAN-530 arbitrary waveform generator
127 (Dataman Programmers Ltd, Maiden Newton, Dorset, UK), which was
128 connected to a 2100L 50-dB RF power amplifier (Electronics & Innovation
129 Ltd., Rochester, NY, USA). The power amplifier was connected to an
130 undamped broadband single element transducer containing a Pz37 Piezo
131 crystal (Ferroperm Piezoceramics A/S, Kvistgård, Denmark) with a centre
132 frequency of 2.2 MHz. The design of the transducer has been described
133 in [18]. Transmitted signals were typically continuous with frequencies in
134 the range 1-10 MHz. The peak-negative acoustic pressures were determined
135 using a PVdF needle hydrophone system with a 0.2-mm probe (Precision
136 Acoustics Ltd, Dorchester, Dorset, UK) connected to a TDS 420A oscilloscope
137 (Tektronix, Inc., Beaverton, OR, USA).

138 The ultrasound transducer was positioned in the container using a clamp
139 stand, at a focal distance of 38 mm from the region of interest to be studied.
140 The azimuth of the length axis of the transducer relative to the North of the

141 container was 37° and the elevation of the length axis of the transducer relative
142 to the base of the container was 17° , as shown in Figure 2.

143 3.2 *Ultrasound contrast agent*

144 DEFINITY® (Lantheus Medical Imaging, North Billerica, MA, USA) consists
145 of C_3F_8 gas microbubbles with mean diameters between 1.1 and $3.3\ \mu\text{m}$,
146 encapsulated by lipid/surfactant shells. Its resonance frequency had been
147 measured to be 2.7 MHz [19]. The 1.5-ml vials used in our experiments had
148 been stored at 9°C . Each vial was shaken for 45 s using a Vialmix® device
149 (Lantheus Medical Imaging). Before introducing the ultrasound contrast agent
150 in our setup, it was further diluted using a 0.9% saline solution.

151 The diluted ultrasound contrast agent was inserted using a syringe into
152 a microbore tube with a 0.51-mm inner diameter. The tube led to a
153 CUPROPHAN® RC55 cellulose capillary (Membrana GmbH, Wuppertal,
154 Germany) with a $200\text{-}\mu\text{m}$ inner diameter and an $8\text{-}\mu\text{m}$ wall thickness. The
155 middle of the capillary coincided with the optical focus of the objective lens
156 and with the acoustic focus of the ultrasound transducer, as shown in Figure 2.
157 The typical field of view using the $10\times$ objective lens was $500 \times 200\ (\mu\text{m})^2$,
158 whereas the diameter of the acoustic focus was greater than 5 mm. Hence,
159 the whole field of view could be considered in acoustic focus. The capillary
160 was positioned 2 mm from the base of the container. The flow speed of the
161 ultrasound contrast agent through the capillary was manually controlled.

162 In total, 48 experiments were performed. Bubble and cluster sizes were
163 measured and tracked using Image-Pro Plus (Media Cybernetics, Inc.,
164 Bethesda, MD, USA). Further analysis was done using MATLAB® (The
165 MathWorks, Inc., Natick, MA, USA).

At the high concentrations we used, clustering started instantaneously after the ultrasound generator was switched on. Figure 3 illustrates the speed of cluster formation of DEFINITY® ultrasound contrast agent that had been further diluted to 1:20. With distances between the microbubbles of only few micrometres, collision times from (11) should be within a second, indeed, as shown in Figure 4. Also, from (11) and Figure 4 it is explained why cluster formation must be faster at higher frequencies, if the other acoustic parameters and the concentration are not changed. Or, after a fixed duration, larger clusters must have formed using higher frequencies, since bubble can approach from larger d_0 at higher f . These deductions are confirmed by our experimental observations: In Figure 3, after 233ms two clusters had been formed of approximately $15\mu\text{m}$ each. These started to approach in the subsequent frames. Overall, newly-formed clusters collided to form larger clusters. This is illustrated by Figures 5 and 6. Each branch represents a cluster. The branches coming together represent the collision of clusters into larger clusters. The velocities of the clusters are on the order of tens of micrometres per second. Although increasing the acoustic pressure would increase the cluster velocities dramatically, as is evident from (13), they would also lead to microbubble disruption [14]. We did not observe phenomena associated with microbubble disruption.

The larger a cluster grows, the lower its resonance frequency becomes. Hence, the velocity of a cluster in the direction of the sound field, defined by (13), should decrease in time. If two identical clusters with resonance frequency f_0 merge, the resulting resonance frequency $f'_c \approx \left(2^{-\frac{1}{3}}\right) f_0 = 0.79 f_0$ [1]. Assuming that the compressibility and damping coefficient do not substantially change, a similar decrease in cluster velocity is expected. However, the decrease in slope magnitude of the main branch in Figure 5 is negligible. This might be explained if the resulting cluster is much stiffer than the original clusters, increasing the damping coefficient.

Also, 7 MHz must be further off the cluster resonance frequency than 2 MHz. Hence, the magnitudes of the slopes in Figure 6 are lower than those than in Figure 5. Secondary radiation forces of clusters onto each other do not explain the cluster colliding times observed. Even if the compressibility of the clusters would be equal to that of a single ultrasound contrast agent microbubble, under the acoustic conditions we used, the collision times from (11) would be just milliseconds. Hence, the bubble clusters cannot be regarded as identical monopoles in our setting. A close-up of two colliding clusters with $22\text{-}\mu\text{m}$ diameters forming a $25\text{-}\mu\text{m}$ cluster is shown in Figure 7. The total time spanning this process is slightly less than 1.8 s.

206 The clusters were initially formed in the middle of the capillary. These clusters
 207 were located at distances $d_0 < \frac{1}{4}\lambda$, as demonstrated in Figure 8. However,
 208 following further cluster coalescence during 17.55 seconds of sonication, the
 209 final distance between the larger clusters corresponded to $\frac{1}{4}\lambda = 54 \mu\text{m}$. These
 210 had been pushed towards the lower capillary wall, owing to primary radiation
 211 forces.

212 The cluster velocities towards the capillary wall were between $5 \mu\text{m s}^{-1}$ at
 213 7 MHz and 22 kPa peak-negative pressure and $15 \mu\text{m s}^{-1}$ at 2 MHz and 20 kPa
 214 peak-negative pressure sonication. These are of the same order as the left-hand
 215 side term in (13). The magnitudes of the slopes in Figure 6 did not change
 216 close to the capillary wall. Hence, in our experimental setup, we neglected any
 217 effect of the capillary wall on cluster translation. With cluster diameters less
 218 than $30 \mu\text{m}$, buoyancy effects may neglected on our timescales as well.

219 In summary, we observed the following stages of microfoam formation,
 220 illustrated in Figure 9. Our initial situation was a dense, random bubble
 221 distribution before ultrasound arrival. After the sonication started, contrast
 222 microbubbles collided, owing to secondary radiation forces. Subsequently,
 223 these clusters coalesced within the space of a quarter of the wavelength,
 224 owing to primary radiation forces. The resulting microfoams translated in
 225 the direction of the ultrasound field, owing to primary radiation forces.

226 Small deviations in microbubble sizes or shell properties lead to deviations
 227 in individual bubbles's resonance frequencies, as expressed in (8). These in
 228 turn cause oscillation phase differences, as expressed in (6), big enough to
 229 be optically observed [7]. Therefore, predicting and manipulating individual
 230 microbubbles is technically challenging. We have demonstrated that as soon
 231 as the bubble clusters were formed and as long as they were in the sound
 232 field, they behaved as one entity. At our acoustic settings, it took seconds to
 233 force the bubble clusters to positions approximately $\frac{1}{4}\lambda$ apart. It also just took
 234 seconds to drive the clusters towards a boundary.

235 We may assume that vessel blocking can only be successful if a microfoam is
 236 created with a diameter equal to or greater than the vessel diameter d_v . From
 237 this study it follows that in order to create such a foam, $\frac{1}{4}\lambda > d_v$, or, $f < \frac{c}{4d_v}$.

238 For therapeutic purposes, it would be of great interest to induce microjetting
 239 on entire clusters towards a vessel wall, presumably causing sonoporation or
 240 sonolysis. Although ultrasound-induced microjetting has been observed with
 241 ultrasound contrast agents, its occurrence in *in vivo* situations is hard to
 242 control [20,21]. Predictable sonic manipulation would be better feasible if the
 243 microbubbles would be forced to clusters of known size and position first.

244 5 Conclusions

245 We observed the following stages of microfoam formation within a densely
246 populated concentration of microbubbles. After the sonication started,
247 contrast microbubbles collided, forming small clusters, owing to secondary
248 radiation forces. These clusters coalesced within the space of a quarter of
249 the ultrasonic wavelength, owing to primary radiation forces. The resulting
250 microfoams translated in the direction of the ultrasound field, hitting the
251 capillary wall, also owing to primary radiation forces.

252 We have demonstrated that as soon as the bubble clusters were formed and
253 as long as they were in the sound field, they behaved as one entity. At our
254 acoustic settings, it took seconds to force the bubble clusters to positions
255 approximately a quarter wavelength apart. It also just took seconds to drive
256 the clusters towards the capillary wall.

257 Subjecting ultrasound contrast agent microbubbles to a continuous
258 low-amplitude signal makes them cluster to known positions and known
259 microfoam sizes, allowing for straightforward sonic manipulation.

260 6 Acknowledgements

261 The authors are grateful to Lantheus Medical Imaging, North Billerica, MA,
262 USA, for supplying the ultrasound contrast agent DEFINITY®. This work
263 has been supported by DFG Emmy-Noether Programme grant 38355133,
264 EPSRC grant EP/F037025/1 and the HERI Research Pump Priming Fund.

265 References

- 266 [1] M. Postema, G. Schmitz, Bubble dynamics involved in ultrasonic imaging,
267 Expert Rev. Mol. Diagn. 6 (3) (2006) 493–502.
- 268 [2] J. R. Lindner, S. Kaul, Delivery of drugs with ultrasound, Echocardiography
269 18 (4) (2001) 329–337.
- 270 [3] E. C. Unger, T. O. Matsunaga, T. McCreery, P. Schumann, R. Sweitzer,
271 R. Quigley, Therapeutic applications of microbubbles, Eur. J. Radiol. 42 (2002)
272 160–168.
- 273 [4] N. Kudo, T. Miyaoka, K. Okada, K. Yamamoto, K. Niwa, Study on mechanism
274 of cell damage caused by microbubbles exposed to ultrasound, Proc. IEEE
275 Ultrason. Symp. (2002) 1351–1354.
- 276 [5] M. Postema, O. H. Gilja, Ultrasound-directed drug delivery, Curr. Pharm.
277 Biotechnol. 8 (6) (2007) 355–361.
- 278 [6] P. A. Dayton, J. S. Allen, K. W. Ferrara, The magnitude of radiation force on
279 ultrasound contrast agents, J. Acoust. Soc. Am. 112 (5) (2002) 2183–2192.
- 280 [7] M. Postema, M. Mleczko, G. Schmitz, Mutual attraction of oscillating
281 microbubbles, in: T. M. Buzug, D. Holz, S. Weber, J. Bongartz, M. Kohl-Bareis,
282 U. Hartmann (Eds.) Advances in Medical Engineering, Springer, Berlin, 2007,
283 vol. 19 of *Methods of experimental physics*, pp. 75–80.
- 284 [8] P. A. Dayton, K. E. Morgan, A. L. Klibanov, G. Brandenburger, K. R.
285 Nightingale, K. W. Ferrara, A preliminary evaluation of the effects of primary
286 and secondary radiation forces on acoustic contrast agents, IEEE Trans.
287 Ultrason., Ferroelect., Freq. Contr. 44 (6) (1997) 1264–1277.
- 288 [9] P. Tortoli, V. Michelassi, M. Corsi, D. Righi, Y. Takeuchi, On the interaction
289 between ultrasound and contrast agents during Doppler investigations,
290 Ultrasound Med. Biol. 27 (9) (2001) 1265–1273.
- 291 [10] H. Medwin, Counting bubbles acoustically: a review, Ultrasonics 15 (1977) 7–13.
- 292 [11] J. Magnaudet, I. Eames, The motion of high-Reynolds-number bubbles in
293 inhomogeneous flows, Annu. Rev. Fluid Mech. 32 (2000) 659–708.
- 294 [12] P. Di Marco, W. Grassi, G. Memoli, Experimental study on rising velocity of
295 nitrogen bubbles in FC-72, Int. J. Therm. Sci. 42 (2003) 435–446.
- 296 [13] T. G. Leighton, The Acoustic Bubble, Academic Press Ltd, London, 1994.
- 297 [14] M. Postema, G. Schmitz, Ultrasonic bubble in medicine: influence of the shell,
298 Ultrason. Sonochem. 14 (4) (2007) 438–444.
- 299 [15] N. de Jong, R. Cornet, C. T. Lancée, Higher harmonics of vibrating gas-filled
300 microspheres. Part one: simulations, Ultrasonics 32 (6) (1994) 447–453.

- 301 [16] P. C. Duiveveld, Bouncing and coalescence of two bubbles in water, sine nomine,
302 Sine loco, 1994.
- 303 [17] M. Postema, P. Marmottant, C. T. Lancée, S. Hilgenfeldt, N. de Jong,
304 Ultrasound-induced microbubble coalescence, *Ultrasound Med. Biol.* 30 (10)
305 (2004) 1337–1344.
- 306 [18] S. Kotopoulis, A. Schommartz, M. Postema, Sonic cracking of blue-green algae,
307 *Appl. Acoust.* 70 (10) (2009) 1306–1312.
- 308 [19] E. Kimmel, B. Krasovitski, A. Hoogi, D. Razansky, D. Adam, Subharmonic
309 responce of encapsulated microbubbles: condition for existance and
310 amplification, *Ultrasound Med. Biol.* 33 (11) (2007) 1767–1776.
- 311 [20] M. Postema, A. van Wamel, F. J. ten Cate, N. de Jong, High-speed photography
312 during ultrasound illustrates potential therapeutic applications of microbubbles,
313 *Med. Phys.* 32 (12) (2005) 3707–3711.
- 314 [21] P. Prentice, A. Cuschieri, K. Dholakia, M. Prausnitz, P. Campbell, Membrane
315 disruption by optically controlled microbubble cavitation, *Nature Phys.* 1 (2005)
316 107–110.

317 List of Figures

318	1	Schematic overview of the experimental setup	17
319	2	Close-up of the sonication tank with coinciding sound, light	
320		beam, and objective focus (<i>top</i>) and definitions of the azimuth	
321		and elevation of the transducer relative to the North of the	
322		container (<i>bottom</i>).	18
323	3	Microfoam formation during continuous sonication at 2 MHz	
324		and 20 kPa peak-negative acoustic pressure. Each frame	
325		corresponds to a $120 \times 120 (\mu\text{m})^2$ area. Time $t = 0$ was defined	
326		by the start of the sonication.	19
327	4	Collision times of individual encapsulated microbubbles as	
328		a function of driving frequency at given distances d_0 , using	
329		$p_a = 20 \text{ kPa}$, $R_0 = 1.25 \mu\text{m}$, $\kappa = 5 \times 10^{-6} \text{ m}^2 \text{ N}^{-1}$, and	
330		$\rho = 998 \text{ kg m}^{-3}$.	20
331	5	Cluster positions as a function of time, during continuous	
332		sonication at 2 MHz and 20 kPa peak-negative acoustic	
333		pressure. Position in the capillary is defined from East ($0 \mu\text{m}$)	
334		to West ($500 \mu\text{m}$). Bold lines indicate merged clusters. The	
335		beginning (<i>left</i>) of a line indicates the formation of a cluster	
336		of diameter $> 6.8 \mu\text{m}$. The end (<i>right</i>) of a line indicates	
337		the disintegration or contraction of a cluster to a diameter	
338		$< 6.8 \mu\text{m}$.	21
339	6	Cluster positions as a function of time, during continuous	
340		sonication at 7 MHz and 22 kPa peak-negative acoustic	
341		pressure. Position in the capillary is defined from East ($0 \mu\text{m}$)	
342		to West ($500 \mu\text{m}$). Bold lines indicate merged clusters. The	
343		beginning (<i>left</i>) of a line indicates the formation of a cluster	
344		of diameter $> 6.8 \mu\text{m}$. The end (<i>right</i>) of a line indicates	
345		the disintegration or contraction of a cluster to a diameter	
346		$< 6.8 \mu\text{m}$.	22
347	7	Two clusters, with $22\text{-}\mu\text{m}$ diameters and an initial distance of	
348		$55 \mu\text{m}$, colliding and coalescing during continuous sonication	
349		at a 2-MHz driving frequency and a 20-kPa peak-negative	
350		pressure. The frame size corresponds to $81 \times 81 (\mu\text{m})^2$. Times	
351		were relative to the start of the sonication ($t=0$).	23

352	8	Clusters forming during sonication at 7 MHz and 22 kPa	
353		peak-negative pressure. The frame size corresponds to	
354		$560 \times 264 (\mu\text{m})^2$. Time $t = 0$ was defined by the start of the	
355		sonication.	24
356	9	Schematic representation of the four stages of microfoam	
357		formation in a capillary: (<i>left-right</i>) random bubble	
358		distribution before ultrasound arrival, bubbles colliding during	
359		sonication, cluster coalescing within the space of a quarter of	
360		the wavelength, microfoam translation.	25

Fig. 1. Schematic overview of the experimental setup

Fig. 2. Close-up of the sonication tank with coinciding sound, light beam, and objective focus (*top*) and definitions of the azimuth and elevation of the transducer relative to the North of the container (*bottom*).

Fig. 3. Microfoam formation during continuous sonication at 2 MHz and 20 kPa peak-negative acoustic pressure. Each frame corresponds to a $120 \times 120 (\mu\text{m})^2$ area. Time $t = 0$ was defined by the start of the sonication.

Fig. 4. Collision times of individual encapsulated microbubbles as a function of driving frequency at given distances d_0 , using $p_a = 20 \text{ kPa}$, $R_0 = 1.25 \mu\text{m}$, $\kappa = 5 \times 10^{-6} \text{ m}^2 \text{ N}^{-1}$, and $\rho = 998 \text{ kg m}^{-3}$.

Fig. 5. Cluster positions as a function of time, during continuous sonication at 2 MHz and 20 kPa peak-negative acoustic pressure. Position in the capillary is defined from East ($0 \mu\text{m}$) to West ($500 \mu\text{m}$). Bold lines indicate merged clusters. The beginning (*left*) of a line indicates the formation of a cluster of diameter $> 6.8 \mu\text{m}$. The end (*right*) of a line indicates the disintegration or contraction of a cluster to a diameter $< 6.8 \mu\text{m}$.

Fig. 6. Cluster positions as a function of time, during continuous sonication at 7 MHz and 22 kPa peak-negative acoustic pressure. Position in the capillary is defined from East ($0 \mu\text{m}$) to West ($500 \mu\text{m}$). Bold lines indicate merged clusters. The beginning (*left*) of a line indicates the formation of a cluster of diameter $> 6.8 \mu\text{m}$. The end (*right*) of a line indicates the disintegration or contraction of a cluster to a diameter $< 6.8 \mu\text{m}$.

Fig. 7. Two clusters, with $22\text{-}\mu\text{m}$ diameters and an initial distance of $55\text{ }\mu\text{m}$, colliding and coalescing during continuous sonication at a 2-MHz driving frequency and a 20-kPa peak-negative pressure. The frame size corresponds to $81 \times 81 (\mu\text{m})^2$. Times were relative to the start of the sonication ($t=0$).

Fig. 8. Clusters forming during sonication at 7 MHz and 22 kPa peak-negative pressure. The frame size corresponds to $560 \times 264 (\mu\text{m})^2$. Time $t = 0$ was defined by the start of the sonication.

Fig. 9. Schematic representation of the four stages of microfoam formation in a capillary: (*left-right*) random bubble distribution before ultrasound arrival, bubbles colliding during sonication, cluster coalescing within the space of a quarter of the wavelength, microfoam translation.