Projection-based dynamic tomography

Clément Jailin, Stéphane Roux, David Sarrut, Simon Rit

► To cite this version:

Clément Jailin, Stéphane Roux, David Sarrut, Simon Rit. Projection-based dynamic tomography. Physics in Medicine and Biology, 2021, 66, pp.215018. 10.1088/1361-6560/ac309e . hal-03391945

HAL Id: hal-03391945

<https://hal.science/hal-03391945>

Submitted on 21 Oct 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Projection-based dynamic tomography

Clément Jailin^{a,b}, Stéphane Roux^a, David Sarrut^c, Simon Rit^c

^a Université Paris-Saclay, ENS Paris-Saclay, CNRS, LMT - Laboratoire de Mécanique et Technologie, F-91190, Gif-sur-Yvette, France

^b GE Healthcare, 78530 Buc, France

^c Lyon University, CNRS, CREATIS UMR5220, Centre Léon Bérard, 69008 Lyon, France

E-mail: clement.jailin@ge.com

Abstract. This paper proposes a 4D dynamic tomography framework that allows a moving sample to be imaged in a tomograph as well as the associated space-time kinematics to be measured with nothing more than a single conventional x-ray scan. The method exploits the consistency of the projection/reconstruction operations through a multi-scale procedure. The procedure is composed of two main parts solved alternatively: a motion-compensated reconstruction algorithm and a projection-based measurement procedure that estimates the displacement field directly on each individual projection. The method is applied to two studies: a numerical simulation of breathing from chest computed tomography (4D-CT) and a clinical cone beam CT of a breathing patient acquired for image-guidance of radiotherapy. The reconstructed volume, initially blurred by the motion, is cleaned from motion artifacts. It results in an improved reconstruction quality showing sharper edges and finer details.

Keywords: Computed tomography, Motion compensation, Tomographic reconstruction, Breathing motion

Submitted to: *Phys. Med. Biol.*

1. Introduction

Being able to image faithfully and precisely a body or a specimen is of utmost importance for medical diagnoses or for non-destructive evaluation. Among many different 3D imaging techniques (Magnetic Resonance Imaging (MRI), Ultrasound imaging, ...), X-ray computed tomography (CT) is frequently used, because of, e.g., its spatial resolution, its rapidity, and its contrast between soft and hard biological tissues. The actual limitation is essentially set by the biological damage due to x-ray radiation.

Initially developed for medical imaging [1, 2], tomography is now widely used in a broad range of applications ranging from biology [3] to material science [4, 5, 6]. The reconstruction of the 3D volume from the sinogram (collection of all acquired projections) is an inverse problem [7]. The technique relies on the strict satisfaction

of conditions, in particular concerning the geometry of the set-up and the motion of the sample as a perfectly controlled rigid rotation with prescribed axis and angles.

Displacement in tomography can be either fought to avoid motion blur in the image or identified to feed models. If the motion happens during the imaging process, it leads to a poor quality reconstruction with blurry edges [8]. This displacement may come from various reasons: motion of the sample itself (*e.g.*, due to the experiment itself, heartbeat and breathing [9], creep, etc.), unexpected motion of the setup, uncalibrated geometry or related to quick phase and/or texture changes [10, 11]. Different solutions have been found to reduce the scanning motion: using smart sampling in time for periodic motions (*e.g.*, based on phase measurements [12, 13]), accelerating the scanning procedure (*e.g.*, faster setup, brighter sources [14], multiple sources [15, 16], developing ultrafast scanning procedure in the synchrotron [17, 18]), using, in medical imaging, sedative, beta-blocker to reduce the heartbeat rate (especially for pediatric use [19]), requesting the patient not to move, or to stop breathing, etc. On the contrary, if this motion is measured and controlled, it gives access to extremely rich quantitative data that are precious pieces of information for diagnosis [20], model identification and data assimilation (*e.g.*, in material science [5, 21, 22], bio-mechanics [23, 24, 25]). From volumes imaged without motion (from static or 4D acquisitions), spatio-temporal deformable registration has received considerable attention in the literature, in material science [26, 22] and in medical imaging, in cardiac image analysis [27, 28], but also for respiratory-correlated imaging of the thorax [29, 9, 30].

When the sample is imaged with motion, or with an uncalibrated tomograph, different methods have been developed in the literature to get a satisfactory reconstruction quality. They are often composed of two successive procedures: (1) evaluation of the motion (or calibration parameters), based on either additional sensors, or full field measurement methods on the acquired images and (2) correction of the reconstruction procedure using the estimated motion.

After having measured the displacement field, the reconstruction algorithm can be enriched with motion. Techniques using a motion-compensated Filtered Backprojection (FBP) algorithms have been developed [31, 32, 33] as well as iterative algorithms [34, 35, 36]. Each (filtered) back-projection is warped with the measured spatial displacement field (estimated at the corresponding acquisition time) so that the sum of all those warped back-projections reconstructs the volume in a static configuration.

Instead of developing techniques based on measurement between multiple volumes (*e.g.*, in [37, 38, 39, 40, 41]), the projection-based strategy is very well fitted to the displacement measurement as it provides a spatial displacement field for each frame [42]. Developed in the medical imaging field and called 2D/3D image registration [34, 43, 44, 45, 46, 47, 48], a similar approach has also been developed in material science and called Projection-based Digital Volume Correlation, P-DVC [49, 50, 51, 52]. In those approaches, an initially well reconstructed volume is warped in space and time such that its projections match the recorded deformed projections. However, knowing a reference static volume is not accessible in all medical acquisitions.

It is proposed to combine both the projection-based measurement and the motion-compensated algebraic reconstruction to achieve the desired dynamic tomography. An initially motion-blurred reconstructed volume is used as the reference volume and leads to an evaluation of the displacement field. The measured motion is then accounted for in a motion-compensated ART algorithm and these two steps are performed iteratively. In [53, 54, 55], an affine transformation model was used to calibrate the acquisition system or to reconstruct moving samples. In [56, 57], the coupled measurement/compensated-reconstruction framework was developed with 2D synthetic examples of moving phantom. Those synthetic test cases resulted in high quality reconstructions as well as the full space-time kinematics to be identified. In [34, 58], the simultaneous motion estimation and image reconstruction (SMEIR) is developed. After grouping the projections by phase, multiple 'approximately still' reconstructions of the volume (ART algorithm regularized with total variation) are used to measure and initialize the major part of the 4D motion. This motion is then precisely updated from projection-based approach (regularized with a 3D mesh and a bio-mechanical model in [30, 59, 60]). In the study we propose, the (quasi-)periodic aspect of the motion was not exploited (to potentially take into account all possible form of displacements) and no initial 'static' volumes could be reconstructed.

We propose in the present paper a dynamic tomography method based on the consistency of the projection. A fully coupled projection-based measurement method with a motion-compensated ART algorithm embedded in a multi-scale framework is developed. The technique is applied on the numerical simulation of breathing from static chest CT and the cone beam (CB-) CT of a breathing patient acquired for image guidance of radiotherapy at the Léon Bérard Cancer center (Lyon, France). The proposed method simultaneously provides, with nothing more than a standard acquisition, the accurate reconstruction of the lungs and their 4D regularized motion along the approximately 35 irregular respiratory cycles taking place during the acquisition. A quality criterion based on a projection residual error allows the volume improvement and the measured displacement field to be quantified.

2. Coupled reconstruction-measurement problem

2.1. Acquisition with motion

The collected intensity for each detector position $\mathbf{r} = [r, z]$ (where z is parallel to the specimen rotation axis, and r is perpendicular to it) and rotation angle, θ , is to be related to the relative beam intensity attenuation. The Beer-Lambert law relates the line integral of the material absorption along the X-ray path $\mathcal{L}(\mathbf{r}, t)$ from source to detector at position $\mathbf{r}$, to the recorded intensity $I(\mathbf{r}, t)$ at time t (and rotation angle $\theta(t)$). In the present work, the projection operator Π is computed using the ASTRA toolbox as defined in [61] and, more precisely, its GPU implementation,

$$p(\mathbf{r}, t) \equiv -\log(I(\mathbf{r}, t)/I_0(\mathbf{r})) = \int_{\mathcal{L}(\mathbf{r}, t)} g(\mathbf{x}, t) \, d\mathbf{x} \equiv \Pi_\theta[g(\mathbf{x}, t)] \quad (1)$$

with $g(\mathbf{x}, t)$ the linear attenuation coefficient, referred to as the volume in the following and $I_0(\mathbf{r}, t)$, the intensity recorded without sample along the beam. Possibly, typical tomography artifacts [62, 63] (*e.g.*, beam hardening, intensity variations, etc) can be corrected at this stage. The collection of N_θ projections $p(\cdot, t)$ for each angle $\theta(t)$ is called the sinogram.

With the acquisition noise η , the projection can now be written $p(\mathbf{r}, t) = \Pi_\theta[g(\mathbf{x}, t)] + \eta(\mathbf{x}, t)$. When motion $\mathbf{v}(\mathbf{x}, t)$ is considered during the experiment, the projection of the deformed volume $g(\mathbf{x}, t)$, defined for every voxel $\mathbf{x} = [x, y, z]$ is written

$$p(\mathbf{r}, t) = \Pi_\theta[g(\mathbf{x} + \mathbf{v}(\mathbf{x}, t))] + \eta(\mathbf{x}, t). \quad (2)$$

For given estimated reconstruction f and displacement field $\mathbf{u}$, consistency of the model [54] can be evaluated from the so-called *projection residuals*, $\rho(\mathbf{r}, t; f, \mathbf{u})$

$$\rho(\mathbf{r}, t; f, \mathbf{u}) = \Pi_\theta[f(\mathbf{x} + \mathbf{u}(\mathbf{x}, t))] - p(\mathbf{r}, t). \quad (3)$$

Reconstructed volume and motion estimate are both determined from the minimization of a functional $\Gamma(f, \mathbf{u})$ which is chosen as the quadratic norm of these residuals (that can be shown to be optimal for a white Gaussian noise affecting p). The mean square error minimization is hence written

$$\Gamma(f, \mathbf{u}) = \sum_{\mathbf{r}, t} (\rho(\mathbf{r}, t; f, \mathbf{u}))^2. \quad (4)$$

This functional also provides a validation metric of the procedure (*e.g.*, quality of the reconstruction and estimated displacement field). The general coupled problem is solved as a fixed point problem, considering two staggered steps iteratively, as sketched in Figure 1, such that

$$f = \text{Argmin}_{\bar{f}} \Gamma(\bar{f}, \mathbf{u}), \quad (5)$$

$$\mathbf{u} = \text{Argmin}_{\bar{\mathbf{u}}} \Gamma(f, \bar{\mathbf{u}}). \quad (6)$$

2.2. Motion-compensated reconstruction

The first part of the two step procedure consists in reconstructing the 3D structure of the sample from the acquired sinogram and the estimated displacement field (initialized at 0 unless additional sensors information or initial guess can be exploited). The use of an ART reconstruction algorithm is natural as it reduces to equation (5) in the absence of motion $\mathbf{u} = 0$.

The SART algorithm [64, 7] was chosen here. At each iteration, indexed by (n) , the volume is warped with the estimated displacement field and its projection is compared with the recorded one, and their difference defines the projection residual field (see equation (3)). The latter is normalized and back-projected $\Delta f^{(n)}(\mathbf{x}, t) = \mathcal{B}_\theta[\rho^{(n)}(\mathbf{r}, t)]$ with the backprojection operator $\mathcal{B}_\theta$. Finally, the correction term is unwarped to the frame of the undeformed state, so that it matches the reference configuration and it is added to the volume, written in a reference configuration, at $t = 0$ with $\mathbf{u}(\mathbf{x}, 0) = 0$: $f^{(n)}(\mathbf{x}, 0) = f^{(n-1)}(\mathbf{x}, 0) + \Delta f^{(n)}(\mathbf{x} - \mathbf{u}(\mathbf{x}, t), t)$. The volume warping with the $\mathbf{u}(\mathbf{x}, t)$

Figure 1. Scheme of the proposed procedure that requires only a standard scan acquisition of the moving scene. It consists of two staggered blocks: 1) motion measurement based on the current approximation of the reconstructed volume and 2) motion-corrected reconstruction based on the current determination of the motion. Initially, the motion may either be set to 0 or benefit from prior knowledge.

field is performed with tri-linear interpolation. A convergence criterion has to be chosen as in the ART procedure. However, the value of the cost functional Γ cannot be used directly in this case as the reconstruction is imperfect. A stopping criterion based on the incremental variation of Γ (or an escape condition limiting the maximum number of iterations) can be set. Because of the difficulty to choose a stopping criterion on the functional evolution and because minimizing the computation time was not a target, it was decided in this paper to limit the number of iterations. This limit N_{DynART} was set to be large enough to allow convergence of the reconstruction. The algorithm of the motion-compensated reconstruction is shown in the Supplementary materials.

While several other algorithms are frequently used to improve the quality of the reconstructed volumes (*e.g.*, , total variation), we did not use them to observe the benefit of our motion correction alone.

As additional metrics to evaluate the sharpness improvement of the reconstructed volumes, we computed (1) the L2 norm of the gradients: $1/N_x \cdot \sum_{\Omega} |\nabla f(\mathbf{x})|^2$ and (2) the 10%-90% rise response of the image edges [65] (distance between 10% and 90% of the edge intensity).

2.3. Space-time measurements

Constrained kinematics: The second part of the two-step procedure is the projection-based displacement field measurement following a P-DVC strategy. This procedure uses as the reference volume the previously reconstructed volume and minimizes, now

with respect to the displacement field, the same cost function Γ , *i.e.*, the RMS of the projection residuals. DVC algorithms are known to be ill-posed [21], consequently the displacement field has to be regularized. Among many possible choices to regularize the displacement field, our choice is to write the displacement field on a reduced kinematic basis composed of a limited number of degrees of freedom

$$\mathbf{u}(\mathbf{x}, t) = \sum_{ij} a_{ij} \phi_i(\mathbf{x}) \sigma_j(t) \quad (7)$$

with a_{ij} unknown amplitudes associated with $\phi_i(\mathbf{x})$ and $\sigma_j(t)$ respectively space and time functions that characterize the space-time evolution. The displacement space can be expressed as the collection of independent subsets (called local DVC [66]), or from a global approach [67], such as using a finite-element mesh [59] or b-spline interpolations, etc.

The time parameter t used in the σ_j function can be, for example, substituted with the signal of additional sensors measuring a force (as used in most mechanical experiments) or a phase signal (accessible in four-dimensional cardiac and thoracic imaging). Finally, both space and time variations can be coupled using a model-based kinematics (called “Integrated Digital Image Correlation” [68]) or using knowledge from prior similar experiments then reduced with singular value decomposition [69]. In [70], the authors show that the identification of a regularized breathing motion from different 3D dynamic volumes leads to better results allowing to work with partial data.

The choice of both space and time functions $\phi_i(\mathbf{x})$ and $\sigma_j(t)$ and their numbers are very important. Too many degrees of freedom leads to ill-conditioning and too few introduces model errors.

Relaxed kinematics: Because the imposed time evolutions (5) may not be perfect, it can be updated without constraints. The discrete time is written t_k where k labels the different instants where a radiograph is acquired. The updated discrete temporal evolution n is hence written as

$$\sigma_n(t_k) = b_k^n \quad (8)$$

with b_k^n time unknowns identified with the space unknowns from a fixed point procedure and initialized with the breathing signal. It is noteworthy that this displacement mode should converge to the one that maximizes the residual field correction (although it is highly influenced by the signal initialization). After having identified a first space-time couple, the breathing time signal is re-identified from the estimated breathing space motion, time-wise, at the next iteration. This allows slight amplitude corrections in the time evolution.

Resolution: The non-linear problem is solved using a Gauss-Newton’s algorithm, *i.e.*, successive linearizations around the current solution (detailed equations are presented in the Supplementary material), until a stopping criterion is reached. In the following application, a single linearization was performed before updating the volume. All components of the gradient of the cost function with respect to all degrees of freedom are computed using first order centered finite differences. Because of a small number

of degrees of freedom, each linearized problem can be easily inverted (this is to be contrasted with other convergence procedures based on multiple search directions on separated spaces [71]).

The residual field of this procedure ρ , as defined previously, informs on the quality of the procedure and chosen model. From this residual field, both model and procedure can be (un)validated and, if required, the model can be enriched with additional degrees of freedom.

Details of the P-DVC algorithm are given in the Supplementary information.

2.4. Multi-scale procedure

The key to be able to solve the previous two step framework, even with large displacements, is to apply a multi-scale pyramidal coarse-to-fine procedure. To improve numerical stability and local minima, a pyramidal approach starting with the coarse scale is introduced. After the displacement field has been estimated at a coarse level (tolerant to displacements of large magnitude but inaccurate), the finer details of the next pyramidal level are considered and the displacement field is corrected (with better precision, and lower tolerance to large magnitude). The P-DVC problem is convex only when the displacement is small as compared to the characteristic length of the projected texture (*e.g.*, the auto-correlation length). This multi-scale analysis provides an “adaptive convexity” in the neighborhood of the solution. To generate the coarse description, projections are convoluted with a Gaussian kernel whose characteristic length can be adapted to the displacement amplitude. Thereby, high frequency patterns of the projection and reconstruction are erased and the P-DVC functional is driven by the global shape motion. The projections can also be down-sampled, in space and time to accelerate the procedure as the resulting images carry about the same information as the filtered one. For each scale, the procedure is computed until stagnation of the residual fields and displacement variation.

The space kinematics is modeled as that of a finite element mesh on the pulmonary pleurae, meshed from the reconstructed volume intensity (obtained by thresholding and a closing morphological operation). Because the initial volume is corrupted by motion blur, the position of the envelop nodes is updated while the volume edges are more accurately reconstructed with the procedure iterations.

3. Experimental studies

3.1. Breathing simulations from multiple static chest reconstructions

The above proposed method is applied to a test case that has been constructed to be realistic and to provide a known ground truth. It is based on 10 real reconstructions of a patient’s chest with phase-triggered 4D CT-scans [9, 12] of the respiratory motion acquired at the CREATIS laboratory, Léon Bérard cancer center and available online[‡].

[‡] <https://github.com/open-vv/pop-model>, with data from [12]

These volumes were obtained from a time sampling based on the acquisition of a respiratory surrogate signal (Pneumo Chest pressure belt). Each one consists of $256 \times 256 \times 141$ voxels. From this set of tomographic images, a synthetic moving CT-scan is assembled as follows:

- A respiratory phase (rounded to the closest of the 10 available phases) as a function of acquisition time (or equivalently projection angle) is first constructed as shown in Fig. 2(a). The same breathing state is used for 8 successive projections in order to have a realistic number of breaths in the entire experiment. This choice has almost no effect on the following developments.
- The synthetic sinogram is constructed by projecting at each angle (chosen to be uniformly sampled at 1 degree interval, over one full rotation, resulting in 360 radiographs) the volume corresponding to the respiratory phase.

While those volumes were acquired using an helical CT scan, the simulated moving patient is generated with a beam covering most of the patient's chest (*i.e.*, CBCT conditions). For simplicity, faster computations and to avoid removing the top and bottom part of the volume due to the cone beam geometry, a parallel beam geometry is chosen here. The difference between parallel and cone projections only lies in the projection operator Π_θ but has no further consequence on the current procedure.

To illustrate the motion, Figure 2(b) shows a composite image, where one such projection (90°) at respiratory phase # 1 is built with green color, and, for the same angle, the projection of the respiratory phase # 7 is attributed the complementary magenta color. Hence, all regions not affected by the motion appear as gray shades, but moving parts appear either in magenta or green. This shows that the displacement is essentially vertical and localized at the diaphragm. The obtained sinogram is used as an input of our procedure.

Because the bottom part of the scan moves with the breathing, the mean intensity of the image, $\kappa(t)$ computed over a region that contains mostly the bottom part of the image, varies and is assumed, at first order, to be related to the phase of motion. This measured signal $\kappa(t)$ is hence used to regularize the time evolution. In addition, the absolute $|\kappa(t)|$ and squared $\kappa(t)^2$ values of this signal and two periodic functions of the rotation angle, sine and cosine, (that could be related in real cases to scan rotation issues) are added to the time function library.

For the present case study, the space kinematic model is embedded into a multi-scale framework: at first, a simple C8 element (*i.e.*, 8 noded-cube element with tri-linear interpolations) composed of 24 degrees of freedom (*i.e.*, a total of 96 dofs for a_{ij} for the 4 time steps) is considered. Then, when a coarse reconstruction has been obtained, a finer mesh composed of 579 nodes with T4 elements (*i.e.*, 4 noded-tetrahedra elements with tri-linear interpolation) is used resulting in 6,948 degrees of freedom.

The root mean square (RMS) of the projected residuals reaches a plateau after 4 iterations (an iteration being both the reconstruction with compensated motion and the projection-based measurement).

Figure 2. Synthetic test case generated from 10 real CT-scans. (a) Chosen sequence of the respiratory phase as a function of acquisition angle. The entire test is composed of 4.5 respiratory cycles. (b-c) Composite image formed by two projections for the same angle 90°, and two different respiratory phases, where each projection is given a complementary color (green and magenta). (The two states correspond to the red dots in figure (a).) (c) Composite image from a 3D slice generated from the same respiratory phases.

The volume was reconstructed with $N_{\text{DynART}} = 10$ at each scale.

3.2. Clinical CBCT with motion

The method is applied to a patient cone-beam CT (CBCT) scan acquired for 3D image-guided radiotherapy at the Léon Bérard cancer center. The CBCT scanner is integrated with the linear accelerator (Elekta Synergy; Elekta Oncology Systems Ltd., Crawley, West Sussex, UK). The flat panel recorded radiographs with 512×512 pixels of $0.8 \times 0.8 \text{ mm}^2$ at a frame rate of 5.5 fps over a full revolution. The distance "source to isocenter" is 1000 mm, the distance "source to detector" is 1536 mm. Image acquisition was performed at 120 kV, 25 mA and 1 mAs/projection.

The sinogram is composed of 676 projections equally spaced over a 360° revolution. The flat panel detector of size 41 cm×41 cm is offset with respect to the line defined by the source and the center of rotation and images only a part of the entire chest (a projection is displayed in figure 9 showing the missing area). A partial reconstruction

procedure is hence required. The different geometrical parameters required for the cone beam reconstruction are given by the tomograph (position of the source and detector for each radiograph).

During the acquisition, the phase of approximately 35 irregular breaths of the patient are extracted from the projections. Because of the corresponding large magnitude (10-20 mm at the diaphragm) of breathing motion, standard reconstruction procedures lead to a blurry reconstruction around the diaphragm area.

The sought space-time kinematics is expressed in a reduced basis composed of selected space and time functions. A scheme of the regularization is shown in figure 3.

Figure 3. Space-time regularization of the motion composed of (a) a 192-noded mesh with tetrahedral elements and (b) 4 simple time evolutions and one fully identified temporal function initialized with the breathing signal.

The unstructured mesh is composed of 192 nodes with 532 T4 elements. In addition, the lungs may slip on the rib cage [29, 72, 30] and hence a discontinuity of the tangential velocity may appear. Reciprocally, non-zero normal displacements of the mesh envelop are removed by projecting the full displacement on axes normal to the mesh.

The time kinematics could be modeled with 5 simple time functions, including (1-2) a sine and cosine of the rotation angle, (3-4) constant and linear evolution and (5) a breathing signal obtained from an independent intensity tracking measurement on projections. This latter time evolution is expected to be the one containing most of the breathing motion. As such, it was chosen to update this breathing signal (5) with the relaxation procedure. With a normalization for the amplitude (by convention), 675 degrees of freedom (dofs) are hence added in the identification procedure.

The 4D motion is finally composed of 192 nodes thus 576 spatial dofs, 4 time functions and 675 temporal dofs from the time relaxation generating a total of 2979 space-time dofs. The space mesh and time evolutions are shown in figure 3.

4. Results

The measurement part was performed on CPU: 20 cores @ 2.20 GHz while all projection operations are done on GPU: Nvidia Tesla K80 (with ASTRA toolbox). What is computationally demanding is volume warping at each projection/back-projection when performing the motion-compensated SART procedure and the generation of the projected sensitivity fields (*e.g.*, image variation with respect to a parameter change).

4.1. Results on the numerical simulation

Slices of the reconstructed volume before and after are shown in figure 4. After the displacement field correction, the diaphragm is visible while the reconstruction without motion correction was more fuzzy. The 10%-90% rise response of the diaphragm edge decreased from 11.2 voxels (22.4 mm) to 2.9 voxels (5.8 mm).

Figure 4. Sagittal and coronal slices of the reconstructed chest without (left column) and with (right column) motion correction (the vertical red lines indicates the slice positions). It can be observed (especially in the zoomed area) that the diaphragm is better reconstructed and sharper (as pointed out by the two white arrows). On the sagittal view, a horizontal plane that polluted the reconstruction disappeared (white arrow). The bottom plot shows 2 intensity curves (in arbitrary units) extracted from the left and right volume on the zoomed area (yellow lines). The dashed curve, corresponding to the non-corrected volume, is smooth. The solid curve is sharper thanks to the motion correction.

The mesh and estimated displacements are shown in figure 5(a) and 5(b) respectively. The vertical motion is large around the diaphragm and vanishes towards the top part of the chest. A maximum vertical displacement amplitude is estimated to be approximately 7 voxels or 14 mm. The validation of our procedure is based on the projection error

Figure 5. (a) Scale 1 mesh with 579 nodes is plotted highlighting the lungs (in red); (b) Color-coded vertical displacement field is shown for the acquisition at angle 48° . The displacement is expressed in voxels (Note that voxels are slightly anisotropic, of size $[1.95, 1.95, 2.0]$ mm along $[x, y, z]$).

(*i.e.*, the part of the initial data that is not explained by the reconstruction approach), and this is precisely captured in the projection residual (see figure 6 before and after correction).

Figure 6. Projection residual fields (difference between the acquired sinogram and the set of projections computed from the reconstructed volume at the corresponding deformation stage) before (left) and after (right) correction by the estimated displacement field. The oscillating positive and negative patterns are the signature of the 4 respiratory motion periods. After correction by the estimated displacement field, a large part of the error has been corrected.

The projection residual field consists of noise and artifacts affecting the acquisition but includes also all model errors (approximate kinematic model, inaccurate interpolations, etc). It can be seen that the residual values are very high in the beginning. On the

bottom part, large positive and negative values are visible resulting from the oscillating respiratory motion. At the end of the procedure, most of the oscillating motion has been erased meaning that the estimated displacement field was trustworthy. The RMS was 2.9 times smaller after correction.

Displacement field The reconstruction quality can be evaluated by comparison with respect to the reference 3D volume at the same respiratory phase. The estimated displacement field can be compared with “standard” DVC procedures (because we have access to the reference 3D volumes). Let us underline that both of these quality evaluations are generally not accessible, but they motivated the construction of the present test case.

(a) Standard DVC measurements

(b) Dynamic tomography measurements

Figure 7. Estimated vertical displacement field (in voxels) for angle 48° using (a) a standard DVC procedure and (b) the dynamic approach using the same color scale. The red parts correspond to the motion of the diaphragm .

The estimated 4D displacement field can be compared with full 3D DVC procedures using the 3D reference volumes. Using the Correli-3.0 software [73] that follows a

similar intensity based measurement in 2D, the displacement field was obtained with the same 579-node mesh but independent time states. The comparison of the vertical displacement field using the dynamic procedure and DVC is presented in figure 7 and shows very similar results. The mean and standard deviation of the difference of the vertical fields is respectively -0.21 and 0.92 voxels. This comparison has however to be considered as informative and not as a validation because (i) the cost functions of the two methods are different and (ii) some nodes have a low sensitivity because of the poor contrast and hence they are susceptible to noise, although this poor determination of the displacement has no perceptible consequence on the reconstruction quality precisely because of the low contrast of these zones.

Figure 8. Norm of the differences between the reference volumes and (a) the corrected volume using the dynamic reconstruction procedure and (b) using a standard static ART reconstruction.

Reconstructed volume The quality of the reconstructed volumes, for the 10 respiratory phases, can be evaluated from their comparison with the initial reference volumes used to generate this test-case. The norm of the difference is mostly controlled by a region containing the diaphragm (see figure 8). In order to quantify the error, these norms are normalized by the noise variance on the same domain. This noise is estimated from the residual obtained after a fine 3D registration between two arbitrary 3D references (two different respiratory phases). This value corresponds to the best residual that can be obtained considering an extremely fine full 3D kinematic correction, and hence the normalized residual norm for a perfect correction should approach unity.

It can be seen that the corrected dynamic procedure produces a residual error much lower than the standard static reconstruction. The oscillating signature of the red curve due to the periodic motion has been, for the most part, cancelled. The error level flattens to about 1.5 and is no more than 50% higher than a perfect correction (with a much finer mesh). This result is to be compared to the error obtained with no motion correction which can reach values higher than 3.

4.2. Application to clinical CBCT data

Figure 9 shows the re-projection of the volume at angle $\theta(t_k) = 15^\circ$ and the projected mesh envelop that defines the region of interest at this angle. Although the missing part of the detector now appears (as the full volume was reconstructed from all angles), the residual field of this area cannot be computed as the projections are not defined. However, thanks to the time regularization, the sensitivity of the motion on those masked areas is obtained from the other angles where the information becomes visible.

Figure 9. (a) Acquired projection $p(\mathbf{r}, t_k)$, with $\theta(t_k) = 15^\circ$ and (b) simulated projection from the volume reconstruction. The rotation axis is highlighted in red. The blue line is the envelop of the projected mesh. The projection is incomplete and involves an insensitive area. The top and bottom part of (b) are subjected to the projected volume due to missing data in areas outside the field-of-view.

The projected residual field and displacement field increment become stationary in 10 iterations at scale 3 and 4 iterations at scale 1. $N_{\text{DynART}} = 10$ and $N_{\text{DynART}} = 30$ were chosen as stopping criterion for the reconstruction respectively at scale 3 and 1. At each iteration, the nodes of the mesh are re-positioned onto the lung external boundary. The procedure takes approximately 6 hours to reach stationarity at scale 3 (projections reduced by a factor $1/3$ in each dimension) and 24 hours at scale 1.

The estimated 4D displacement field is shown figure 10. As expected, the vertical displacement is very large along the diaphragm (18 voxels at scale 1 corresponding to approximately 15 mm) and follows essentially the breathing motion. The normal displacement has been erased allowing sliding of the pleural cavity [59]. The refined breathing signal is in phase with the original one, estimated from projections and used for initialization, however, amplitudes are observed to be markedly different. The reason may be that the initialization was performed with a tracking of the diaphragm in the projections, thus submitted to perspective effect due to the cone beam while rotating.

Reconstruction improvement The correction naturally focuses on the large displacement areas, *i.e.*, the bottom of the chest around the diaphragm. It can be noted that the diaphragm is reconstructed much more accurately after the displacement correction.

Figure 10. (a) Vertical displacement field of the mesh envelop at the first maximal exhale position (projection 7), expressed in voxels (1 voxel $\leftrightarrow$ 800 μ m) and (b) measured (black) vs estimated (dashed blue) shape of breathing signal. The amplitude of this motion is on the corresponding spatial field.

Initially smooth and blurry, the edges get much sharper, as expected. Some details of the bronchus become visible. The comparison of the two reconstructions is shown figure 11 and highlights the quality improvement.

The plot in figure 11 shows the intensity profiles at the bottom of the lungs (visible in the slices above), before (dashed curve) and after (plain curve) which highlights the correction of the motion blur. Initially spread over 12 voxels (from $z = 152$ to $z = 164$ voxels), also measured by the ± 6 voxel displacements of the diaphragm figure 11, the diaphragm transition is reconstructed sharper. The 10%-90% rise response decreased from 13.2 voxels (10.6 mm) to 1.4 voxels (1.2 mm). Segmenting the lungs for example would be much easier and reliable after correction.

The metric based on the gradient norms gave $2.8 \cdot 10^{-7}$ and $3.4 \cdot 10^{-7}$ for respectively the standard and corrected image. As an order of magnitude, when filtered by an averaged squared kernel of size (3x3), the the L2 norm of the gradients from the corrected filtered image is $2.7 \cdot 10^{-7}$ which is the value for the standard unfiltered image. This improvement highlights the sharpness evolution of the reconstruction.

The true metric to validate our procedure is the projected residual field that corresponds to the reconstruction model error. This residual sinogram can be shown in 3D in figure 12 before and after the dynamic tomography procedure. The reconstruction inconsistency due to motion is clearly visible in the left image with the positive and negative patterns. Those error parts are mostly corrected when taking the kinematics into account and the RMS decreases by a factor of 2.31 in the region of interest described by the mesh volume (expressed in % of the gray level amplitude of the projections).

What remains in the left corner part of the final residual field is an error located at the intersection between the diaphragm and the bottom rib (the acute lung corner in the bottom left of the coronal slice in figure 11). The kinematics at this position is complex (slip and recovery) and was not perfectly captured. Moreover, this part of

Figure 11. (a) Comparison of two slices of the reconstruction, using (left) a standard ART procedure and (b) the proposed dynamic reconstruction method with the identified displacement field. It is to be noted that the diaphragm area is much sharper in the right corrected slices and bronchus (highly subjected to motion artefacts) are more visible. The red lines indicate the slice locations and the white segment corresponds to the place where intensity profiles are extracted. On the bottom plot, the intensity profiles (in arbitrary units) on the diaphragm area, (dashed) before and (plain) after the dynamic reconstruction procedure. The diaphragm interface is much sharper for the corrected procedure.

Figure 12. Sinograms of the projected residuals before (left) and after (right) the kinematics identification and correction. The plot is focused on the diaphragm part. The color amplitudes and transparencies are identical for the two plots. The high positive and negative patterns (signature of the 35 breaths) visible in the left image have been mostly erased.

the diaphragm is not always visible depending on the breathing and rotation angle. This residual field study gives indications on where and how to improve the kinematic model. The model could be further enriched as suggested by the interpretation of the final residual, adding space or time dofs.

5. Discussion and Conclusions

The present work follows the trend that has been outlined in the literature review of the introduction on motion corrected tomography. An iterative algorithm is proposed to determine both morphology and kinematics with no additional acquisition cost as compared to a classical scan. The procedure consists of two staggered parts within a pyramidal coarse-to-fine multi-scale framework [74]: (i) a motion compensated procedure extended from a SART algorithm and (ii) a displacement field measurement from the reconstructed volume and the initial sinogram.

The procedure is tested on (1) a numerical simulation of breathing from static chest CT and (2) a CBCT of a breathing patient during radiotherapy at different phases of the respiratory motion whose reconstruction leads to large motion blur artifacts. The multi-scale procedure, performed for the volume reconstruction, allows large motions to

be captured. Finally, a large part of the motion blur artifacts was removed from the reconstruction, leading to small root mean squared errors computed on the projection residual fields. The reconstructed volume and measured displacement field of the numerical application were compared to ground truth, thereby validating the procedure.

The originality of the present work lies in a very accurate description of the kinematics, using a finite element mesh that offers many degrees of freedom. This progress leads to two main difficulties:

- First, the mesh is to be suited to the anatomy, however, initially the chest morphology is unknown, so that not only the 3D image and its motion are to be determined simultaneously, but also the high discretization of the motion and suited boundary conditions. In our specific application, the thoracic cage is mostly static, and hence the velocity field at the boundary of the meshed lungs is constrained to be parallel to lung boundary, allowing for a possible slip. This calls for several iteration steps along which the morphology is progressively described with a better fidelity, the motion discretization is rendered more and more accurately, ending with a fine mesh. The simultaneous determination of these three items — morphology, mesh and motion — is an original problem.
- Second, with the above spatial description of the displacement field, involving many degrees of freedom, each of which being susceptible to vary in time all along the scan, makes the problem severely ill-posed. This difficulty is circumvented by a model reduction technique whereby “modes” are considered. Each mode is a separated form being the product of one spatial motion (supported by the mesh description) and a temporal part that describes the amplitude evolution of this mode. Several types of modes are introduced, some where the temporal shape function is prescribed (such as sine or cosine of the rotation angle, or a low order polynomial in time), and some more complicated ones, where the time evolution is that of breathing. The latter can be inferred initially by the apparent motion of the diaphragm, and — through iterations — space and time parts are progressively fine-tuned to match the observation. By the suited selection of these modes, the number of degrees of freedom is severely reduced, and makes the problem tractable. In some way, one may thus add a fourth category of unknowns, in addition to morphology, mesh and motion, that are the considered modes.

It is interesting to control the procedure by studying the projected residual fields that contains all features not explained by the projection model (*e.g.*, detector noise, non-linear absorption, unsuited kinematic model *e.g.*, due to the chosen discretization, etc.). This residual field allows for (un)validating the kinematic model. This feedback is essential when encountering an unexpected kinematics. Regularization should be considered as evolutionary and any feature that can be understood from the residuals should help to enrich the model up to the stage where the remaining residual is deemed acceptable. Additional degrees of freedom allowing to model and correct the heartbeat motion is an interesting perspective.

The proposed method is totally non-intrusive as it does not require more than a standard acquisition (*i.e.*, a single revolution, and a standard acquisition protocol). Therefore, this method can be applied to any type of scanner. Of course, all additional information (from sensors, from other modalities, from experts, etc.) would help the analysis and accelerate the optimization of the process, making the problem better posed. From an hardware side, using faster scans would allow reducing the motion blur that may happen in the projections and would improve the identify temporal motion model.

In the proposed approach, it is not required to initiate the computation with a subset of radiographs that could be considered as (approximately) still, (because they would originate from a similar respiratory cycle phase). This is *not* a small difference with respect to existing approaches such as SMEIR, because for instance, the mesh that is used for describing the motion, has to align perfectly the organs in order to allow for free slip at boundaries in [59, 30]. But when no decent 3D volume has been reconstructed yet, it is not possible to proceed with a fine adjustment of the mesh. The volume, mesh, and displacements are all three very intricate and need to be progressively improved better and better defined. As a result, the proposed method does not require a (quasi-)periodic motion nor extraneous information such as the measurement of the phase. Without this initialized motion, its measurement from truncated partial projections became impossible without the introduction of time regularization.

One current limitation for clinical practice is the substantial important computation time that includes the computation of all the sensitivity fields. In addition to hardware improvement (faster computational resources, GPU, parallel computation, etc.), the lightening of the computation operations (development of reduced breathing models, minimization of the projected gradients updates, etc.) will highly decrease the computation time. During the iterative process, improving the reconstruction quality and segmenting the moving organs with image processing (spatial smoothness, total variation) or Deep-Learning techniques would also allow to better mesh and model the breathing motion leading to faster convergences.

This procedure opens up new avenues in dynamic tomography enabling to image moving patients, even if the displacement is large and with no additional data acquisition compared to a standard CT-scan protocol. The clinical development proposed in [75] with an a-priori motion model could be extended without the need of the prior 4D-CT scan. Moreover, any additional sensors and detectors (*e.g.*, force or displacement measurements) could be used transparently in the above method to constrain further the analysis. Using dual source scanners is also a promising perspective, especially when the sources are orthogonal [15, 16] so that displacement fields could be completely estimated for each acquisition time (*i.e.*, providing a better conditioning of the problem). Finally, any additional regularization, either for the volume reconstruction (*e.g.*, assumptions on discrete phases [76, 77], total variation (TV) regularization [78], dictionary of commonly encountered tissue morphology, etc.) or for the kinematic model can be easily included within the proposed algorithm. Such developments are very relevant

for practical applications. Nevertheless, they were not included in the presented test case, in order to focus on the key ingredients for the success of the reconstruction. However, additional acquired or prior information is expected to be beneficial for faster convergence, more accurate results, or for further reducing the radiation dose (*e.g.*, acquiring fewer projections for an equivalent result).

Acknowledgements

The authors would like to thank Dr. Matthieu Vitse and the SciViz group, for having performed most of the 3D rendering in this article.

Ethical statement

This research study was conducted retrospectively using non identifiable human subject data. The POPI study has been approved by the local research ethics committee and data have been anonymously open to the community since 2007. Applicable law and standards of ethic have been respected in accordance with the principles embodied in the Declaration of Helsinki and in accordance with local statutory requirements.

References

- [1] W. A. Kalender. X-ray computed tomography. *Physics in Medicine & Biology*, 51(13):R29, 2006.
- [2] G. Wang, H. Yu, and B. De Man. An outlook on X-ray CT research and development. *Medical physics*, 35(3):1051–1064, 2008.
- [3] E. Fogelqvist, M. Kördel, V. Carannante, B. Önfelt, and H. M. Hertz. Laboratory cryo X-ray microscopy for 3D cell imaging. *Scientific reports*, 7(1):13433, 2017.
- [4] J. Baruchel, J.Y. Buffière, E. Maire, P. Merle, and G. Peix. *X-Ray Tomography in Material Sciences*. Hermès Science, Paris (France), 2000.
- [5] E. Maire and P. J. Withers. Quantitative X-ray tomography. *International materials reviews*, 59(1):1–43, 2014.
- [6] L. Salvo, M. Suéry, A. Marmottant, N. Limodin, and D. Bernard. 3D imaging in material science: Application of X-ray tomography. *Comptes Rendus Physique*, 11(9):641–649, 2010.
- [7] A.C. Kak and M. Slaney. *Principles of computerized tomographic imaging*. IEEE press, 1988.
- [8] P. Milanfar. A model of the effect of image motion in the radon transform domain. *IEEE Transactions on Image Processing*, 8(9):1276–1281, 1999.
- [9] J. Vandemeulebroucke, S. Rit, J. Kybic, P. Clarysse, and D. Sarrut. Spatiotemporal motion estimation for respiratory-correlated imaging of the lungs. *Medical physics*, 38(1):166–178, 2011.
- [10] P. Babin, G. Della Valle, H. Chiron, P. Cloetens, J. Hoszowska, P. Pernot, A.L. Réguerre, L. Salvo, and R. Dendievel. Fast X-ray tomography analysis of bubble growth and foam setting during breadmaking. *Journal of Cereal Science*, 43(3):393–397, 2006.
- [11] N. Limodin, L. Salvo, E. Boller, M. Suéry, M. Felberbaum, S. Gaillière, and K. Madi. In situ and real-time 3-D microtomography investigation of dendritic solidification in an Al–10 wt.% Cu alloy. *Acta Materialia*, 57(7):2300–2310, 2009.
- [12] J. Vandemeulebroucke, D. Sarrut, P. Clarysse, et al. The POPI-model, a point-validated pixel-based breathing thorax model. In *XVth international conference on the use of computers in radiation therapy (ICCR)*, volume 2, pages 195–199, 2007.

- [13] R. Mokso, D. A. Schwyn, S. M. Walker, M. Doube, M. Wicklein, T. Müller, M. Stampanoni, G. K. Taylor, and H. G. Krapp. Four-dimensional in vivo X-ray microscopy with projection-guided gating. *Scientific reports*, 5:8727, 2015.
- [14] M. Di Michiel, J. M. Merino, D. Fernandez-Carreiras, T. Buslaps, V. Honkimäki, P. Falus, T. Martins, and O. Svensson. Fast microtomography using high energy synchrotron radiation. *Review of Scientific Instruments*, 76(4):043702, 2005.
- [15] G. Wang, H. Yu, and Y. Ye. A scheme for multisource interior tomography. *Medical physics*, 36(8):3575–3581, 2009.
- [16] T. G Flohr, C. H. McCollough, H. Bruder, M. Petersilka, K. Gruber, C. Süß, M. Grasruck, K. Stierstorfer, B. Krauss, R. Raupach, et al. First performance evaluation of a dual-source CT (DSCT) system. *European radiology*, 16(2):256–268, 2006.
- [17] E. Maire, C. Le Boulrot, J. Adrien, A. Mortensen, and R. Mokso. 20 Hz X-ray tomography during an in situ tensile test. *International Journal of Fracture*, 200(1):3–12, 2016.
- [18] K.J. Dobson, S. B. Coban, S. A. McDonald, J.N. Walsh, R. C. Atwood, and P. J. Withers. 4-D imaging of sub-second dynamics in pore-scale processes using real-time synchrotron X-ray tomography. *Solid Earth*, 7(4):1059, 2016.
- [19] M. Rashed, P. Banka, A. Barthur, R. D. MacDougall, R. H. Rathod, A. J. Powell, and A. Prakash. Effects of dose reduction on diagnostic image quality of coronary computed tomography angiography in children using a third-generation dual-source computed tomography scanner. *The American Journal of Cardiology*, 2018.
- [20] Yue Pan, Gary E Christensen, Oguz C Durumeric, Sarah E Gerard, Surya P Bhatt, R Graham Barr, Eric A Hoffman, and Joseph M Reinhardt. Assessment of lung biomechanics in copd using image registration. In *2020 IEEE 17th International Symposium on Biomedical Imaging (ISBI)*, pages 1891–1895. IEEE, 2020.
- [21] M. Grédiac and F. Hild, editors. *Full-field measurements and identification in solid mechanics*. John Wiley & Sons, 2012.
- [22] A. Buljac, C. Jailin, A. Mendoza, J. Neggers, T. Taillandier-Thomas, A. Bouterf, B. Smaniotto, F. Hild, and S. Roux. Digital volume correlation: Review of progress and challenges. *Experimental Mechanics*, 2018.
- [23] P. Moireau, C. Bertoglio, N. Xiao, C. A. Figueroa, C.A. Taylor, D. Chapelle, and J.-F. Gerbeau. Sequential identification of boundary support parameters in a fluid-structure vascular model using patient image data. *Biomechanics and modeling in mechanobiology*, 12(3):475–496, 2013.
- [24] L. Han, H. Dong, J. R. McClelland, L. Han, D. J. Hawkes, and D. C. Barratt. A hybrid patient-specific biomechanical model based image registration method for the motion estimation of lungs. *Medical image analysis*, 39:87–100, 2017.
- [25] L.P. Argani, F. Torella, R.K. Fisher, R.G. McWilliams, M.L. Wall, and A.B. Movchan. Deformation and dynamic response of abdominal aortic aneurysm sealing. *Scientific reports*, 7(1):17712, 2017.
- [26] F. Hild, A. Fanget, J. Adrien, E. Maire, and S. Roux. Three-dimensional analysis of a tensile test on a propellant with digital volume correlation. *Archives of Mechanics*, 63(5-6):1–20, 2011.
- [27] J. Schaerer, C. Casta, J. Pousin, and P. Clarysse. A dynamic elastic model for segmentation and tracking of the heart in MR image sequences. *Medical Image Analysis*, 14(6):738–749, 2010.
- [28] J.M. Peyrat, H. Delingette, M. Sermesant, C. Xu, and N. Ayache. Registration of 4D cardiac CT sequences under trajectory constraints with multichannel diffeomorphic demons. *IEEE transactions on medical imaging*, 29(7):1351–1368, 2010.
- [29] J. Vandemeulebroucke, O. Bernard, S. Rit, J. Kybic, P. Clarysse, and D. Sarrut. Automated segmentation of a motion mask to preserve sliding motion in deformable registration of thoracic CT. *Medical physics*, 39(2):1006–1015, 2012.
- [30] Y Zhang, J N Tehrani, and J Wang. A biomechanical modeling guided CBCT estimation technique. *IEEE transactions on medical imaging*, 36(2):641–652, 2016.
- [31] M. Prummer, L. Wigstrom, J. Hornegger, J. Boese, G. Lauritsch, N. Strobel, and R. Fahrig.

- Cardiac C-arm CT: Efficient motion correction for 4d-fbp. In *Nuclear Science Symposium Conference Record, 2006. IEEE*, volume 4, pages 2620–2628. IEEE, 2006.
- [32] M. Prummer, J. Hornegger, G. Lauritsch, L. Wigstrom, E. Girard-Hughes, and R. Fahrig. Cardiac C-arm CT: a unified framework for motion estimation and dynamic CT. *IEEE Transactions on Medical Imaging*, 28(11):1836–1849, 2009.
- [33] S. Rit, D. Sarrut, and L. Desbat. Comparison of analytic and algebraic methods for motion-compensated cone-beam CT reconstruction of the thorax. *IEEE Transactions on Medical Imaging*, 28(10):1513–1525, 2009.
- [34] J. Wang and X. Gu. Simultaneous motion estimation and image reconstruction (SMEIR) for 4D cone-beam CT. *Med Phys*, 40(10):101912, Oct 2013.
- [35] J Dang, X Gu, T Pan, and J Wang. A pilot evaluation of a 4-dimensional cone-beam computed tomographic scheme based on simultaneous motion estimation and image reconstruction. *International Journal of Radiation Oncology, Biology, Physics*, 91(2):410–418, 2015.
- [36] M Guo, G Chee, D O’Connell, S Dhou, J Fu, K Singhrao, D Ionascu, D Ruan, P Lee, D A Low, et al. Reconstruction of a high-quality volumetric image and a respiratory motion model from patient CBCT projections. *Medical physics*, 2019.
- [37] A. Ruhlandt, M. Töpperwien, M. Krenkel, R. Mokso, and T. Salditt. Four dimensional material movies: High speed phase-contrast tomography by backprojection along dynamically curved paths. *Scientific reports*, 7(1):6487, 2017.
- [38] T. De Schryver, M. Dierick, M. Heyndrickx, J. Van Stappen, M. A. Boone, L. Van Hoorebeke, and M. N. Boone. Motion compensated micro-CT reconstruction for in-situ analysis of dynamic processes. *Scientific reports*, 8(1):7655, 2018.
- [39] G. Zang, R. Idoughi, R. Tao, G. Lubineau, P. Wonka, and W. Heidrich. Space-time tomography for continuously deforming objects. *ACM Transactions on Graphics*, 37(4):1–14, 2018.
- [40] Guangming Zang, Ramzi Idoughi, Ran Tao, Gilles Lubineau, Peter Wonka, and Wolfgang Heidrich. Warp-and-project tomography for rapidly deforming objects. *ACM Transactions on Graphics (Proc. SIGGRAPH)*, 2019.
- [41] Carlos Mira, Ernesto Moya-Albor, Boris Escalante-Ramírez, Jimena Olveres, Jorge Brieva, and Enrique Vallejo. 3d hermite transform optical flow estimation in left ventricle ct sequences. *Sensors*, 20(3):595, 2020.
- [42] F. Guillard, B. Marks, and I. Einav. Dynamic X-ray radiography reveals particle size and shape orientation fields during granular flow. *Scientific Reports*, 7(1):8155, 2017.
- [43] R. Zeng, J.A. Fessler, and J.M. Balter. Respiratory motion estimation from slowly rotating X-ray projections: Theory and simulation. *Medical physics*, 32(4):984–991, 2005.
- [44] R. Zeng, J.A. Fessler, and J.M. Balter. Estimating 3-D respiratory motion from orbiting views by tomographic image registration. *IEEE Transactions on Medical Imaging*, 26(2):153–163, 2007.
- [45] J Hinkle, M Szegedi, B Wang, B Salter, and S Joshi. 4D CT image reconstruction with diffeomorphic motion model. *Medical image analysis*, 16(6):1307–1316, 2012.
- [46] P. Markelj, D. Tomaževič, B. Likar, and F. Pernuš. A review of 3D/2D registration methods for image-guided interventions. *Medical image analysis*, 16(3):642–661, 2012.
- [47] V. Delmon, J. Vandemeulebroucke, R. Pinho, M. Vila Oliva, D. Sarrut, and S. Rit. In-room breathing motion estimation from limited projection views using a sliding deformation model. In *Journal of Physics: Conference Series*, volume 489, page 012026. IOP Publishing, 2014.
- [48] Y. Suzuki, G.SK. Fung, Z. Shen, Y. Otake, O. Lee, L. Ciuffo, H. Ashikaga, Y. Sato, and K. Taguchi. Projection-based motion estimation for cardiac functional analysis with high temporal resolution: a proof-of-concept study with digital phantom experiment. In *Medical Imaging 2017: Physics of Medical Imaging*, volume 10132, page 1013230. International Society for Optics and Photonics, 2017.
- [49] H. Leclerc, S. Roux, and F. Hild. Projection savings in CT-based digital volume correlation. *Experimental Mechanics*, 55(1):275–287, 2015.
- [50] T. Taillandier-Thomas, S. Roux, and F. Hild. A soft route toward 4D tomography. *Physical*

- Review Letters*, 117(2):025501, 2016.
- [51] C. Jailin, A. Bouterf, M. Poncelet, and S. Roux. In situ μ CT-scan Mechanical Tests: Fast 4D Mechanical Identification. *Experimental Mechanics*, 57(8):1327–1340, 2017.
 - [52] C. Jailin, M. Etxegarai, E. Tudisco, S.A. Hall, and S. Roux. Fast tracking of fluid invasion using time-resolved neutron tomography. *Transport in Porous Media*, pages 1–19, 2018.
 - [53] C. Jailin, A. Buljac, A. Bouterf, M. Poncelet, F. Hild, and S. Roux. Self-calibration for lab- μ CT using space-time regularized projection-based dvc and model reduction. *Measurement Science and Technology*, 29(2):024003, 2018.
 - [54] T. van Leeuwen, S. Maretzke, and K.J. Batenburg. Automatic alignment for three-dimensional tomographic reconstruction. *Inverse Problems*, 34(2):024004, 2018.
 - [55] Ningning Zhao, Daniel O’Connor, Adrian Basarab, Dan Ruan, and Ke Sheng. Motion compensated dynamic mri reconstruction with local affine optical flow estimation. *IEEE Transactions on Biomedical Engineering*, 66(11):3050–3059, 2019.
 - [56] Martin Burger, Hendrik Dirks, Lena Frerking, Andreas Hauptmann, Tapio Helin, and Samuli Siltanen. A variational reconstruction method for undersampled dynamic x-ray tomography based on physical motion models. *Inverse Problems*, 33(12):124008, 2017.
 - [57] C. Jailin and S. Roux. Dynamic tomographic reconstruction of deforming volumes. *Materials*, 11(8):1395, 2018.
 - [58] J Dang, F-F Yin, T You, C Dai, D Chen, and J Wang. Simultaneous 4D-CBCT reconstruction with sliding motion constraint. *Medical physics*, 43(10):5453–5463, 2016.
 - [59] Zichun Zhong, Xuejun Gu, Weihua Mao, and Jing Wang. 4d cone-beam ct reconstruction using multi-organ meshes for sliding motion modeling. *Physics in Medicine & Biology*, 61(3):996, 2016.
 - [60] Shiwei Zhou, Yujie Chi, Jing Wang, and Mingwu Jin. General simultaneous motion estimation and image reconstruction (g-smeir). *Biomedical Physics & Engineering Express*, 7(5):055011, 2021.
 - [61] W. Van Aarle, W.J. Palenstijn, J. De Beenhouwer, T. Altantzis, S. Bals, K.J. Batenburg, and J. Sijbers. The ASTRA Toolbox: A platform for advanced algorithm development in electron tomography. *Ultramicroscopy*, 157:35–47, 2015.
 - [62] F. E. Boas and D. Fleischmann. CT artifacts: causes and reduction techniques. *Imaging in Medicine*, 4(2):229–240, 2012.
 - [63] F.P. Vidal, J.M. Létang, G. Peix, and P. Cloetens. Investigation of artefact sources in synchrotron microtomography via virtual X-ray imaging. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 234(3):333–348, 2005.
 - [64] Anders H Andersen and Avinash C Kak. Simultaneous algebraic reconstruction technique (sart): a superior implementation of the art algorithm. *Ultrasonic imaging*, 6(1):81–94, 1984.
 - [65] Toru Higaki, Fuminari Tatsugami, Chikako Fujioka, Hiroaki Sakane, Yuko Nakamura, Yasutaka Baba, Makoto Iida, and Kazuo Awai. Visualization of simulated small vessels on computed tomography using a model-based iterative reconstruction technique. *Data in brief*, 13:437–443, 2017.
 - [66] T. S. Smith, B. K. Bay, and M. M. Rashid. Digital volume correlation including rotational degrees of freedom during minimization. *Experimental Mechanics*, 42(3):272–278, 2002.
 - [67] G. Besnard, F. Hild, and S. Roux. “finite-element” displacement fields analysis from digital images: application to portevin–le châtelier bands. *Experimental Mechanics*, 46(6):789–803, 2006.
 - [68] F. Mathieu, H. Leclerc, F. Hild, and S. Roux. Estimation of elastoplastic parameters via weighted FEMU and integrated-DIC. *Experimental Mechanics*, 55(1):105–119, 2015.
 - [69] A. Charbal, S. Roux, F. Hild, and L. Vincent. Spatiotemporal regularization for digital image correlation: Application to infrared camera frames. *International Journal for Numerical Methods in Engineering*, 114(12):1331–1349, 2018.
 - [70] J R McClelland, M Modat, S Arridge, H Grimes, D D’Souza, D Thomas, D O’Connell, D A Low, E Kaza, D J Collins, et al. A generalized framework unifying image registration and respiratory

- motion models and incorporating image reconstruction, for partial image data or full images. *Physics in Medicine & Biology*, 62(11):4273, 2017.
- [71] J.-C. Passieux, R. Bouclier, and J.-N. Périé. A space-time PGD-DIC algorithm. *Experimental Mechanics*, pages 1–12, 2018.
 - [72] Vivien Delmon. *Recalage déformable de projections de scanner X à faisceau conique*. PhD thesis, INSA de Lyon, 2013.
 - [73] H Leclerc, J Neggers, F Mathieu, F Hild, and S Roux. Correli 3.0, 2015. IDDN.FR.001.520008.000.S.P.2015.000.31500.
 - [74] Edward H Adelson, Charles H Anderson, James R Bergen, Peter J Burt, and Joan M Ogden. Pyramid methods in image processing. *RCA engineer*, 29(6):33–41, 1984.
 - [75] Simon Rit, Jochem WH Wolthaus, Marcel van Herk, and Jan-Jakob Sonke. On-the-fly motion-compensated cone-beam ct using an a priori model of the respiratory motion. *Medical physics*, 36(6Part1):2283–2296, 2009.
 - [76] K.J. Batenburg and J. Sijbers. DART: a practical reconstruction algorithm for discrete tomography. *IEEE Transactions on Image Processing*, 20(9):2542–2553, 2011.
 - [77] S. Roux, H. Leclerc, and F. Hild. Efficient binary tomographic reconstruction. *Journal of Mathematical Imaging and Vision*, 49(2):335–351, 2014.
 - [78] Emmanuel J Candès, Justin Romberg, and Terence Tao. Robust uncertainty principles: Exact signal reconstruction from highly incomplete frequency information. *IEEE Transactions on information theory*, 52(2):489–509, 2006.

Supplementary material

I. Projection-based measurements

The P-DVC measurement consists of solving, with a Newton-Raphson algorithm, the following equation

$$\mathbf{u} = \underset{\boldsymbol{\mu}}{\text{Argmin}} \Gamma(f, \boldsymbol{\mu}) \quad (9)$$

with $\boldsymbol{\mu} \in \mathcal{U} = \mathcal{P} \times \mathcal{S}$, $\mathcal{P}$ and $\mathcal{S}$ being respectively the space generated by the $\phi(\mathbf{x})$ space functions and time evolution $\sigma(t)$

$$\Gamma(f, \mathbf{u}) = \sum_{\mathbf{r}, t} (\rho(\mathbf{r}, t; f, \mathbf{u}))^2. \quad (10)$$

The projected error field,

$$\rho(\mathbf{r}, t; f, \mathbf{u}) = \Pi_{\theta}[f(\mathbf{x} + \mathbf{u}(\mathbf{x}, t))] - p(\mathbf{r}, t), \quad (11)$$

with $p(\mathbf{r}, t)$ the acquired projections, is linearized, with $\tilde{f}(\mathbf{x}, t) = f(\mathbf{x} + \mathbf{u}(\mathbf{x}, t))$

$$\rho_{\text{lin}}(\mathbf{r}, t; f, \mathbf{u} + \partial\mathbf{u}) = \Pi_{\theta}[\tilde{f}(\mathbf{x}, t)] + \Pi_{\theta}[\nabla \tilde{f}(\mathbf{x}, t) \partial\mathbf{u}(\mathbf{x}, t)] - p(\mathbf{r}, t) \quad (12)$$

Then, with $\mathbf{u}(\mathbf{x}, t) = \sum_{ij} a_{ij} \phi_i(\mathbf{x}) \sigma_j(t)$, the minimization with respect to $\mathbf{a}$ leads to the solution of the linear system

$$\mathbf{a} = \left[\sum_{\mathbf{r}, t} \mathbf{S} \otimes \mathbf{S} \right]^{-1} \sum_{\mathbf{r}, t} \rho \mathbf{S} \quad (13)$$

with $\mathbf{S}$ the projected sensitivity fields (signature in the detector space of the dof sensitivity fields)

$$S_{ij}(\mathbf{r}, t) = \sigma_j(t) \Pi_{\theta(t)}[\phi_i(\mathbf{x}) \nabla \tilde{f}(\mathbf{x}, t)] \quad (14)$$

II. Dynamic reconstruction algorithm

The global dynamic reconstruction algorithm consists of two main steps, detailed hereafter, which are run alternately, up to stationarity.

Algorithm 1 Proposed motion-compensated reconstruction

```

Initialization  $n \leftarrow 1$ 
Initialization  $f^{(1)} \leftarrow 0$ 
Choose a permutation,  $\pi$ , over  $N_t$  acquisition times
for  $i \leftarrow 1$  to  $N_{\text{DynART}}$  do
  for  $t \leftarrow 1$  to  $N_t$  do
     $\tau \leftarrow \pi(t)$ 
     $\tilde{f}^{(n)}(\mathbf{x}, \tau) = f^{(n)}(\mathbf{x} + \mathbf{u}(\mathbf{x}, \tau))$  ▷ warp
     $\rho(\mathbf{r}, \tau) \leftarrow p(\mathbf{r}, \theta(\tau)) - \Pi_{\theta(\tau)}[\tilde{f}^{(n)}(\mathbf{x}, \tau)]$ 
     $\Delta \tilde{f}^{(n+1)}(\mathbf{x}) = \mathcal{B}_{\theta(\tau)}[\rho(\mathbf{r}, \tau)]$ 
     $\widehat{\Delta f}^{(n+1)}(\mathbf{x}) = \Delta \tilde{f}^{(n+1)}(\mathbf{x} - \mathbf{u}(\mathbf{x}, \tau))$  ▷ unwarp
     $f^{(n+1)}(\mathbf{x}) = f^{(n)}(\mathbf{x}) + \widehat{\Delta f}^{(n+1)}(\mathbf{x})$ 
    Implement additional constraints on  $f^{(n+1)}$  (e.g., positivity)
     $n \leftarrow n + 1$ 
  end for
end for

```

Algorithm 2 Projection-based kinematic measurement

$\mathbf{u}(\mathbf{x}, t) \leftarrow \mathbf{u}_0(\mathbf{x}, t)$ ▷ Displacement initialization
for Measurement iteration $\leftarrow 1$ to $N_{\text{iteration}}$ **do**
 $p \leftarrow \check{p}_\lambda$ ▷ Gaussian filtering
 $f_\lambda(\mathbf{x}) \leftarrow \text{Motion-compensated reconstruction}(p, \mathbf{u})$ ▷ Algo. 1
 Compute $\rho_l(\mathbf{r}, t)$ ▷ Projection residual
 for $i \leftarrow 1$ to N_s **do**
 for $j \leftarrow 1$ to N_t **do**
 $\tilde{f}(\mathbf{x}, t) \leftarrow f_\lambda(\mathbf{x} + \mathbf{u}(\mathbf{x}, t))$ ▷ Volume advection
 $\chi_i(\mathbf{x}, t) \leftarrow \Pi_{\theta(t)}[\phi_i(\mathbf{x}) \nabla \tilde{f}(\mathbf{x}, t)]$
 $S_{ij}(\mathbf{r}, t) \leftarrow \chi_i(\mathbf{x}, t) \sigma_j(t)$ ▷ Projected sensitivities
 end for
 end for
 $\tilde{\mathbf{a}} \leftarrow \left[\sum_{\mathbf{r}, t} \mathbf{S} \otimes \mathbf{S} \right]^{-1} \sum_{\mathbf{r}, t} \rho \mathbf{S}$ ▷ Motion identification
 $\mathbf{a} \leftarrow \tilde{\mathbf{a}}$ ▷ Erasing normal displacements
 for $i \leftarrow 1$ to N_s **do**
 $\bar{S}(\mathbf{x}) \leftarrow \Pi_{\theta(t)}[a_{i1} \phi_i(\mathbf{x}) \nabla \tilde{f}(\mathbf{x}, t)]$ ▷ Sensitivity of the first mode
 end for
 $\tilde{\mathbf{b}} \leftarrow \left[\sum_{\mathbf{r}} \bar{\mathbf{S}} \otimes \bar{\mathbf{S}} \right]^{-1} \sum_{\mathbf{r}} \rho \bar{\mathbf{S}}$ ▷ Time identification
end for