

Evaluation of excitation schemes for indirect detection of 14N via solid-state HMQC NMR experiments

Andrew G.M. Rankin, Julien Trébosc, Piotr Paluch, Olivier Lafon, Jean-Paul

Amoureux

► To cite this version:

Andrew G.M. Rankin, Julien Trébosc, Piotr Paluch, Olivier Lafon, Jean-Paul Amoureux. Evaluation of excitation schemes for indirect detection of 14N via solid-state HMQC NMR experiments. Journal of Magnetic Resonance, 2019, 303, pp.28 - 41. 10.1016/j.jmr.2019.04.004 . hal-03484340

HAL Id: hal-03484340 https://hal.science/hal-03484340

Submitted on 20 Dec 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Evaluation of excitation schemes for indirect detection of ¹⁴N via solid-state **HMOC NMR experiments**

Andrew G.M. Rankin,^{1*} Julien Trébosc,^{1,2} Piotr Paluch,^{1,3} Olivier Lafon,^{1,4} Jean-Paul Amoureux^{1,5*}

¹ Univ. Lille, CNRS, Centrale Lille, ENSCL, Univ. Artois, UMR 8181 – UCCS – Unit of Catalysis and Chemistry of Solids, F-59000 Lille, France

² Univ. Lille, CNRS-FR2638, Fédération Chevreul, F-59000 Lille, France

³ Centre of Molecular and Macromolecular Studies, Polish Academy of Sciences, Sienkiewicza 112, PL-90363 Lodz, Poland.

⁴ Institut Universitaire de France, 1 rue Descartes, F-75231 Paris Cedex 05, France.

⁵ Bruker Biospin, 34 rue de l'industrie, F-67166 Wissembourg, France.

* Corresponding authors: jean-paul.amoureux@univ-lille.fr andrew.rankin@univ-lille.fr

Keywords: solid-state NMR; *D*-HMQC; *J*-HMQC; ¹⁴N; selective excitation; SLP; DANTE; XiX.

Abstract. It has previously been shown that ¹⁴N NMR spectra can be reliably obtained through indirect detection via HMQC experiments. This method exploits the transfer of coherence between single- (SQ) or double-quantum (DQ) ¹⁴N coherences, and SQ coherences of a suitable spin-1/2 'spy' nucleus, e.g., ¹H. It must be noted that SQ-SQ methods require a carefully optimized setup to minimize the broadening related to the first-order quadrupole interaction (*i.e.*, an extremely well-adjusted magic angle and a highly stable spinning speed), whereas DQ-SQ ones do not. In this work, the efficiencies of four ¹⁴N excitation schemes (DANTE, XiX, Hard Pulse (HP), and Selective Long Pulse (SLP)) are compared using J-HMQC based numerical simulations and either SQ-SQ or DQ-SQ ¹H-{¹⁴N} *D*-HMQC experiments on L-histidine HCl and N-acetyl-L-valine at 18.8 T and 62.5 kHz MAS. The results demonstrate that both DANTE and SLP provide a more efficient ¹⁴N excitation profile than XiX and HP. Furthermore, it is shown that the SLP scheme: (i) is efficient over a large range of quadrupole interaction, (ii) is highly robust to offset and rf-pulse length and amplitude, and (iii) is very simple to set up. These factors make SLP ideally suited to widespread, non-specialist use in solid-state NMR analyses of nitrogen-containing materials.

Introduction 1.

Nitrogen atoms are ubiquitous in biomolecules, e.g., DNA and proteins, and organic, hybrid and inorganic materials, such as pharmaceuticals, polyamides, functionalized silica, metal-organic frameworks or nitrides. However, the NMR observation of the two nitrogen isotopes, ¹⁵N and ¹⁴N, remains challenging owing to their unfavorable NMR properties.

1.1.¹⁵N

It is a spin-1/2 nucleus, which often leads to high-resolution spectra, even for solids. However, owing to its very low natural abundance (NA = 0.36 %) and its low gyromagnetic ratio, $\gamma_{15N} \approx 0.1 \gamma_{1H}$, a major limitation of ¹⁵N NMR is often the lack of sensitivity. This issue can be circumvented by ¹⁵N isotopic enrichment, but this approach can be costly, timeconsuming and difficult to achieve. Alternative methods to enhance the NMR sensitivity and detect the ¹⁵N signals in natural abundance consist in the use of (i) indirect detection *via* protons at high Magic Angle Spinning (MAS) frequencies using through-space or throughbond ${}^{1}\text{H} \rightarrow {}^{15}\text{N}$ polarization transfers [1,2] or (ii) dynamic nuclear polarization enhanced ${}^{1}\text{H} \rightarrow$ ¹⁵N cross-polarization under MAS (CPMAS) experiments [2,3].

1.2.¹⁴N

Conversely, the ¹⁴N isotope has two major advantages: (i) a very high natural abundance (NA = 99.64 %), and (ii) the fact that the knowledge of the quadrupole interaction of this spin-1 nucleus provides additional information about the local environment of nitrogen atoms.

Direct detection.

The direct NMR detection of ¹⁴N nuclei is challenging for several reasons [4-6]. First, it has a spin-1 value and hence no central transition between energy levels $m_{\rm N} = \pm 1/2$. For a powdered sample, the signal of the two single-quantum transitions, ¹⁴N^{SQ}, between energy levels $m_{\rm N} = 0$ and ± 1 is broadened by the first-order quadrupole interaction (H_{Q1}) over a width of $3C_0/2$, where C_0 is the quadrupolar coupling constant. For this isotope, a value of $C_0 = 7$ MHz has been reported [7], which corresponds to a powder pattern width of 10.5 MHz. This large line broadening severely decreases the sensitivity since the total integrated intensity is spread over a broad spectral width. Second, the sensitivity of directly-detected spectra being proportional to $\gamma^{3/2}$ that of ¹⁴N isotope is further decreased with respect to ¹⁵N by its low gyromagnetic ratio with $\gamma_{14N} \approx 0.072\gamma_{1H}$. Third, the pulse dead-times, which are inversely proportional to γ , are thus long for ¹⁴N nuclei, whereas the broad linewidths result in short Free-Induction Decays (FID). This combination also decreases the sensitivity of ¹⁴N direct detection. For ¹⁵N isotope, this issue can be circumvented by the use of a spin-echo sequence or the CPMG (Carr-Purcell Meiboom-Gill) variant, which improves the sensitivity by the acquisition of a train of echoes in every scan [8]. However, these methods are not easy to use for the ¹⁴N isotope, because π -pulses are difficult to implement for this spin-1 nucleus. Forth, an additional difficulty for the direct ¹⁴N detection is the need to excite a broad spectral width, which can exceed the excitation bandwidth of the radio-frequency (rf) pulses and the detection bandwidth of the probe. This limitation is exacerbated for ¹⁴N isotope because the rf-fields are proportional to γ , which is small for ¹⁴N. When the width of the ¹⁴N^{SQ} spectrum exceeds the detection bandwidth of the probe or the excitation bandwidth of the pulses, subspectra with different frequency offsets are acquired and then co-added to generate the full wide powder pattern [8]. This piecewise acquisition technique is called VOCS (Variable Offset Cumulative Spectrum). It has been shown that the excitation bandwidth of directly detected ¹⁴N experiments can be increased using adiabatic pulses, such as WURST (wideband uniform-rate smooth-truncation) under static conditions [9] or trains of short rotorsynchronized rf-pulses in the manner of DANTE (Delays Alternating with Nutation for Tailored Excitation) under MAS conditions [10-12]. The sensitivity of directly detected ¹⁴N experiments under static conditions can be further enhanced by an initial polarization transfer from protons using BRAIN-CP (broadband adiabatic inversion cross-polarization) [13].

Nevertheless, a major limitation of the direct ¹⁴N^{SQ} detection under static conditions is the lack of resolution since this signal is broadened by H_{Q1} and the powder patterns of distinct nitrogen sites extensively overlap. Hence, this approach is only suitable for samples with one or two nitrogen sites. MAS can improve the resolution of ¹⁴N^{SQ} spectra, but these experiments require a very precise adjustment and a high stability of the MAS frequency and angle [14–17]. The resolution can be improved by the detection of the ¹⁴N double-quantum (¹⁴N^{DQ}) coherences between energy levels $m_N = \pm 1$ since these coherences are only broadened by the second-order quadrupole interaction (H_{Q2}). These coherences can be directly excited and acquired at twice the ¹⁴N Larmor frequency with the overtone sequence. This type of direct excitation has been reported first under static conditions [18,19] and more recently under MAS [4,20,21]. The detection of this usually "forbidden" DQ transition is allowed by a perturbation of the Zeeman spin states by the large quadrupole interaction. The sensitivity of overtone experiments can be enhanced by ¹H \rightarrow ¹⁴N CPMAS transfer and DNP [22,23].

However, the overtone ¹⁴N direct excitation and detection experiments are usually less sensitive than the direct detection of ¹⁴N^{SQ} transitions. Furthermore, the slow nutation of the overtone signal results in a small excitation bandwidth, which may prevent the detection of distinct nitrogen sites. WURST pulses can improve the overtone excitation bandwidth, but they decrease the sensitivity [24].

Indirect detection.

For protonated samples, the lack of resolution and sensitivity of direct ¹⁴N detection can be partially overcome by the indirect detection of ¹⁴N^{SQ} or ¹⁴N^{DQ} transitions via protons, denoted 1 H-{ 14 N^{SQ}} and 1 H-{ 14 N^{DQ}} respectively [5,25–27], using two-dimensional (2D) experiments at high MAS frequencies, such as HMQC (heteronuclear multiple-quantum correlation) [25,28], HSQC (heteronuclear single-quantum correlation) [29,30], double cross-polarization [31], or sequences using ¹⁴N rf-irradiation lasting hundreds of microseconds to generate ¹H- 14 N multiple-quantum coherences [32]. It has been shown that the HSQC sequence is less sensitive than the HMQC one [29,33]. Furthermore, for HMQC experiments, coherence transfers via ¹H-¹⁴N dipolar interactions, which are reintroduced by the application of heteronuclear dipolar recoupling schemes, such as $SR4_1^2$ [34], in dipolar-mediated HMQC schemes, are usually faster and more efficient than coherence transfers via a combination of Jcoupling and second-order quadrupolar-dipolar cross-terms, also known as residual dipolar splitting, during the free evolution periods of the conventional *J*-mediated HMQC (*J*-HMQC) scheme [28,35]. For ¹H-{¹⁴N^{SQ}} HMOC experiments, the broadening of ¹⁴N^{SQ} transitions due to the first-order quadrupole interaction is refocused by the rotor synchronization of the indirect evolution period, *i.e.*, $t_1 = nT_R$.

As for the ¹⁴N direct detection, a challenge for the indirect detection is the creation of ¹⁴N coherences because the magnitude of H_{Q1} exceeds that of the rf-field by several orders of magnitude. In ¹H-{¹⁴N} HMQC experiments, ¹⁴N SQ and DQ coherences have been created and reconverted by: short 'hard' pulses (HP) using high rf-power [36,37], sideband selective long pulses (SLP) [25,28,35,38,39], and DANTE trains [10,40]. Recently, single-sideband XiX (x-inverse x) pulse trains have also been used in the ${}^{1}H{-}{{}^{14}N^{DQ}}$ HMQC sequence [41]. These four excitation schemes (HP, SLP, DANTE and XiX) have all been applied at the ¹⁴N Larmor frequency. ¹H-{¹⁴N^{DQ}} *D*-HMQC 2D spectra have also been recorded using overtone irradiation at twice the ¹⁴N Larmor frequency, in the manner of rectangular, WURST and composite pulses [42–44]. Nevertheless, it has been reported that ${}^{1}H{-}{}^{14}N^{DQ}$ D-HMQC experiments on amino-acids using rectangular pulses or DANTE trains at the ¹⁴N Larmor frequency, or overtone pulses are less sensitive than the ¹H-{¹⁴N^{SQ}} *D*-HMQC ones using rectangular pulses or DANTE trains [35,39,45]. Furthermore, for the γ polymorph of glycine, we have shown that the ${}^{1}H-{}^{14}N^{SQ}$ D-HMQC experiment using DANTE is slightly more sensitive, but less robust to offset, than its SLP variant [39,40]. The performances of ¹H-{¹⁴N^{DQ}} *D*-HMQC using XiX excitation have not been compared to those of the other existing ${}^{1}\text{H}-\{{}^{14}\text{N}\}$ *D*-HMQC schemes.

Here, we compare the performances of ${}^{1}H{}{{}^{14}N}$ *D*-HMQC experiments using HP, SLP, DANTE and XiX schemes, all at the ${}^{14}N$ Larmor frequency. These sequences are first described using an average Hamiltonian theory. We then analyze the efficiency of these experiments and their robustness to offset, rf-field inhomogeneity and quadrupole interaction. The performances of ${}^{1}H{}^{14}N$ *D*-HMQC schemes using HP, SLP and DANTE schemes are also assessed by experiments on L-histidine.HCl and N-acetyl-L-valine, denoted His and NAV hereafter, respectively.

2. Pulse sequences and theory

Fig.1 shows the ¹H-{¹⁴N} *J*- and *D*-HMQC sequences using rectangular pulses (HP or SLP), the basic D_1^K DANTE, and the XiX schemes for the excitation and the reconversion of ¹⁴N coherences evolving during the t_1 period. In the *J*-HMQC sequence, the defocusing and refocusing delays, denoted τ_J , are free evolution periods. In the *D*-HMQC sequence, the SR4²₁ recoupling was applied during the defocusing and refocusing delays, τ_D , in order to reintroduce the ¹H-¹⁴N dipolar couplings. During SR4²₁, the ¹H rf-field is equal to twice the MAS frequency: $v_1 = 2v_R$. This recoupling reintroduces the space component |m| = 1 of the ¹H-¹⁴N dipolar couplings and ¹H chemical shift anisotropy (CSA), whereas it suppresses the ¹H-¹H dipolar couplings, the ¹H isotropic chemical shifts and the hetero-nuclear *J*-couplings with protons. As the SR4²₁ scheme is non- γ -encoded, the beginning of the two SR4²₁ periods in the *D*-HMQC sequence must be separated by an integer number of rotor periods.




Fig.1. ¹H-{¹⁴N} (a) *J*- and (b) *D*-HMQC sequences using for the excitation and reconversion of ¹⁴N coherences in place of the open rectangles in (b) on the ¹⁴N channel: (c) HP, (d) SLP, (e) D_1^K and (f) XiX schemes. The t_1 period is defined as the delay between the middles of the excitation and reconversion blocks.

For ${}^{1}H-{{}^{14}N}$ HMQC experiments, the ${}^{1}H$ *x*-magnetization created by the first pulse evolves during the defocusing period into [40,46]

$$S_{x}\left(\frac{1+2\cos(2\omega_{HN}\tau)}{3}\right) + S_{x}Q_{z}\left(\frac{\cos(2\omega_{HN}\tau)-1}{3}\right) + S_{y}I_{z}\sin(2\omega_{HN}\tau)$$
(1)

where $S_x Q_z$ is the *x*-magnetization of proton, leading to a triplet with intensities 1:-2:1 for the coupling with ¹⁴N nucleus, $S_y I_z$ is the *y*-magnetization of proton, leading to a triplet with intensities 1:0:-1. The angular frequency ω_{HN} is equal to πJ_{HN} for *J*-HMQC and to $\omega_{D,HN}$ for *D*-HMQC:

$$\omega_{D,HN} = \frac{1}{4} b_{\rm HN} \sin^2 \left(\beta_{PR}^{D,\rm HN}\right) \cos(2\varphi). \tag{2}$$

with

$$\varphi = \gamma_{PR}^{D,\text{HN}} + \alpha_{RL}^0 - \omega_R t^0 \tag{3}$$

In Eqs.2 and 3, $b_{\rm HN}$ denotes the dipolar coupling constant in rad.s⁻¹, the Euler angles $\{0, \beta_{PR}^{D,\rm HN}, \gamma_{PR}^{D,\rm HN}\}$ relate the inter-nuclear ¹H-¹⁴N direction to the MAS rotor-fixed frame, and t^0 refers to the starting time of the symmetry-based scheme. The subsequent evolution of the density matrix depends on the used excitation scheme.

2.1. HP

HP can achieve the uniform excitation of ¹⁴N powder patterns provided its length τ_p is sufficiently short [39,40]

$$\tau_{\rm p} < 2/(3\pi C_Q). \tag{4}$$

For $C_Q = 1$ MHz, Eq.4 indicates that the uniform excitation requires $\tau_p < 0.2 \ \mu$ s. The maximal ¹⁴N rf-field compatible with the specifications of a 1.3 mm double-resonance MAS probe is $v_1 \approx 100$ kHz, which results in a flip angle of $\theta < 2\pi v_1 \tau_p \approx 8^\circ$. This maximum value is much smaller than the optimal angles, $\theta \approx 54.4$ and 90°, yielding maximal transfer efficiency for ¹H-{¹⁴N^{SQ}} and ¹H-{¹⁴N^{DQ}} *D*-HMQC experiments, respectively [40]. These small tilt angles decrease the efficiency of ¹H-{¹⁴N} HMQC experiments using HP. When Eq.4 is satisfied, the evolution under H_{Q1} during τ_p is negligible and the Hamiltonian in the jolting-frame during HP is equal to [39,41]

$$\widetilde{H} = 2\pi \nu_1 I_x \tag{5}$$

In Eq.1, a HP corresponding to this Hamiltonian converts I_z into I_y and the quadrupolar order operator Q_z into K_y and D_x operators, which represent respectively the ¹⁴N^{SQ} coherences along y-axis antiphase with respect to H_{Q1} and the ¹⁴N^{DQ} coherences along the x-axis. After the first HP, only the signal resulting from either S_xK_y and S_yI_y or S_xD_x operator is detected, according to ¹H-{¹⁴N^{SQ}} or ¹H-{¹⁴N^{DQ}} HMQC experiment is used, respectively. The coherences evolve under isotropic shift during t_1 . For ¹H-{¹⁴N^{SQ}}, the excitation and reconversion HPs must be rotor-synchronized in order to refocus the evolution under H_{Q1} during t_1 . In that case, it has been shown [40] that the signal is proportional to

$$S(\tau) \propto \frac{2}{3} \sin^2(\theta) \langle \sin^2(2\omega_{HN}\tau) \rangle + \frac{2}{3} \sin^2(2\theta) \langle \sin^4(\omega_{HN}\tau) \rangle$$
(6)

where $\langle ... \rangle$ denotes the powder average, and for ¹H-{¹⁴N^{2Q}} it is proportional to

$$S(\tau) \propto \frac{2}{3} \sin^4(\theta) \langle \sin^4(\omega_{HN}\tau) \rangle$$
. (7)

For *J*-HMQC experiments, $\omega_{\text{HN}} = \pi J_{\text{HN}}$, and with HP the maximal transfer efficiency is obtained for $\theta = \pi/2$ and a delay of

$$\tau = 1/(4J_{\rm HN}) \text{ for } {}^{1}\text{H} - \{{}^{14}\text{N}^{\rm SQ}\} \text{ or } 1/(2J_{\rm HN}) \text{ for } {}^{1}\text{H} - \{{}^{14}\text{N}^{\rm DQ}\}$$
(8)

This is contrary to *J*-HMQC experiments with a pair of two spin-1/2 nuclei, where the optimum signal is always observed for $\tau = 1/(2J)$.

2.2. D₁^K

¹⁴N SQ and DQ coherences in ¹H-{¹⁴N} HMQC experiments can also be created using rotor-synchronized DANTE schemes, denoted D_M^K , which consist of a train of rectangular pulses of length, τ_p , rf-amplitude, v_1 , with identical phases. The length of a D_M^K train is equal to KT_R and M equally spaced pulses are applied during each of these K rotor periods. The D_M^K schemes with M = 1 and $M \ge 2$ are referred to as basic and interleaved DANTE sequences, respectively [10-12]. The DANTE excitation profile in the frequency domain consists of a comb of rf-spikelets, all separated from the carrier frequency by a multiple of Mv_R , with amplitude decreasing with the separation from the carrier frequency according to a sinc function. For a D_1^K scheme resonant with the center-band and with individual pulses satisfying Eq.4, the first-order average Hamiltonian can be written as [40,47]

$$\bar{H} = 2\pi\nu_1\tau_p I_x/T_R \tag{9}$$

This equation is only valid when the evolution under H_{Q2} is negligible during KT_R . As the magnitude of H_{Q2} is proportional to $C_Q^2/v_{0,14N}$, for large C_Q values this assumption requires high B_0 field and MAS frequencies, which limit the losses due to H_{Q2} during D_1^K [40]. When Eq.9 is valid and the excitation and reconversion of D_1^K schemes are rotor-synchronized, *i.e.*, separated by an integer multiple of rotor periods, the signals of ${}^{1}H{-}\{{}^{14}N^{SQ}\}$ and ${}^{1}H{-}\{{}^{14}N^{DQ}\}$ HMQC experiments using DANTE schemes are given by Eqs. 6 and 7, respectively, where $\theta = 2\pi Kv_1\tau_p$, and the optimum delays by Eq.8. In a first approximation, the D_1^K train can hence be viewed as a HP with an effective rf-field of Kv_1. If S_{FWHM} denotes the offset width of each excitation **S**pikelet, and E_{FWHM} that of the global Envelope excitation, then these values are equal to [40]

$$S_{FWHM} \approx 1.1/KT_R; \quad E_{FWHM} \approx 1.1/\tau_p$$
 (10)

In order to obtain an excitation profile that is robust with respect to the offset, S_{FWHM} must be large and hence K must be small. This implies that most of the time, except when the isotropic frequency range, $\Delta v_{14N,iso}$, is small, the largest possible rf-field should be used. It is common that several different ¹⁴N species exist in the same sample, all with different C_Q values and isotropic resonances, $v_{14N,iso}$. Therefore, as a starting point for the experimental optimization, the number of rotor periods should accommodate the isotropic frequency range, and the individual pulse length should excite the majority of the static linewidth for the maximum C_Q value

$$K \approx 1/(T_R * \Delta v_{14N,iso}) \text{ and } \tau_p \le 0.6/C_{Q,max}$$
(11)

However, these simple rules are mainly applicable in the case of weak $H_{\rm Q2}$ effects, as discussed later.

It has been shown (i) that ${}^{1}H{-}{{}^{14}N^{DQ}}$ HMQC experiments are less efficient than the ${}^{1}H{-}{{}^{14}N^{SQ}}$ ones, and (ii) that HMQC with D_{1}^{K} is more efficient than with D_{M}^{K} (M > 1) [39,40,45]. For these reasons, we only studied here ${}^{1}H{-}{{}^{14}N^{SQ}}$ HMQC sequences using D_{1}^{K} trains.

2.3. SLP

The HP scheme is based on the assumption of a simultaneous excitation of all crystallites, which is why the maximum possible rf-field strength is then used. Recently, another scheme has been proposed, based on the rotor-driven sequential excitation of the crystallites, which

occurs when the time-dependent quadrupole interaction becomes small as a result of sample rotation [39]. This principle leads to the use of selective long pulses (SLP). It must be noted that the name of SLP may appear to be misleading because, as we will see below, this pulse is robust towards offsets (Figs.6 and 12). This is contrary to the way SLP are used in liquids. However, in MAS NMR it has very recently been demonstrated that, contrary to HP and D_1^K pulses that excite numerous sidebands, SLP only excites the sideband that coincides with the carrier frequency (Figs.9d and 14d in [59]). As a result, SLP can be considered as selective when only observing the sideband combs. The relatively high efficiency of SLP comes from the fact that all crystallites in the rotor provide a signal, which always occurs when their interaction frequency matches that of the carrier irradiation due to sample rotation. This is contrary to HP and D_1^K excitations, where a large part of the sample gives a small signal due to rf-inhomogeneity. As a consequence, the SLP signal is always at the spinning sideband that is the closest to the carrier frequency, not at the isotropic chemical shift, as shown in Fig.9d in [59]. The selective excitation of a single spinning sideband of the ¹⁴N^{SQ} spectrum requires a sufficiently long pulse with low rf-amplitude: $\tau_p > T_R$ and $v_1 < v_R$. We employed here rectangular pulses satisfying these conditions. The evolution of the density operator between time points separated by several rotor periods can be described by an average Hamiltonian. For a SLP with phase ϕ_p and resonant with the center-band of the ${}^{14}N^{SQ}$ spectrum, its firstorder value is given by [41]

$$\overline{H}^{(1)} = 2\pi v_1 A_0 R_z(\phi_p) U_Q(\xi_0)^{-1} I_x U_Q(\xi_0) R_z(\phi_p)^{-1}$$
(12)

 A_0 and ξ_0 are the intensity and the phase of the center-band of the ¹⁴N^{SQ} spinning sideband manifold. A_0 does not depend on the γ_{PR} Euler angle specifying the orientation of the electric field gradient tensor with respect to the rotor-fixed frame, but depends on the two other Euler angles { α_{PR} , β_{PR} }. The scaling of the rf-field strength by A_0 stems from the fact that the instantaneous resonance frequency of the nucleus changes rapidly during the pulse and hence the rf-field rotates the magnetization only during a part of the pulse. R_z (ϕ_p) is the rotation operator of the ¹⁴N spin about the z-axis through angle ϕ_p and U_Q (ξ_0) is the quadrupolar propagator defined as:

$$U_Q(\xi_0) = \exp\left[-i\xi_0 \frac{Q_z}{3}\right] \tag{13}$$

When considering only the Hamiltonian given by Eq.12 during SLP, it can be shown that the ${}^{1}H-{}^{14}N^{SQ}$ HMQC signal is proportional to

$$S(\tau) \propto \frac{2}{3} \langle \sin^2(\theta) \sin^2(2\omega_{HN}\tau) \rangle + \frac{2}{3} \langle \sin^2(2\theta) \sin^4(\omega_{HN}\tau) \rangle$$
(14)

and that of $^1\text{H-}\{^{14}\text{N}^{\text{DQ}}\}$ to

$$S(\tau) \propto \frac{2}{3} \langle \sin^4(\theta) \sin^4(\omega_{HN}\tau) \rangle$$
 (15)

where $\theta = 2\pi A_0 v_1 \tau_p$ depends on the Euler angles { α_{PR} , β_{PR} } through A_0 .

Furthermore, for a SLP with a phase ϕ_p and resonant with the center-band of the ¹⁴N^{SQ} spectrum, the second-order average Hamiltonian is given by [41]

$$\overline{H}^{(2)} = \frac{\omega_1^2}{4\omega_R} d\left[Q_z + R_z(\phi_p)D_x R_z(\phi_p)^{-1}\right]$$
(16)

with

$$d = \sum_{p=-\infty, p\neq 0}^{+\infty} \frac{A_p^2 - A_{-p}^2 - 2A_p A_0 \cos(\xi_p - \xi_0) + 2A_{-p} A_0 \cos(\xi_{-p} - \xi_0)}{p}$$
(17)

where A_p and ξ_p denote the intensity and the phase of the ¹⁴N^{SQ} *p*th-order spinning sideband manifold. In Eq.16, $\overline{H}^{(2)}$ contains Q_z and D_x terms, which commute [46]. The latter can convert the I_z operator in Eq.1 into D_y one, which represents the ¹⁴N^{DQ} coherences along the y-axis [39] and hence, this term contributes to the signal of ¹H-{¹⁴N^{DQ}} HMQC. However, the double-quantum term of $\overline{H}^{(2)}$ in Eq.16 does not commute with $\overline{H}^{(1)}$, which can thus perturb the creation of ¹⁴N^{DQ} coherences in ¹H-{¹⁴N^{DQ}} HMQC experiment, by creating unwanted ¹⁴N^{SQ} coherences.

2.4. XiX

This issue can be circumvented by the use of the XiX scheme, which is constructed as a train of 2*N* SLP pulses of the same length T_R , with alternating phase *x* and *-x* to eliminate $\overline{H}^{(1)}$ [41]. For XiX scheme resonant with the center-band, $\overline{H}^{(2)}$ is given by Eq.16 with $\phi_p = 0$ and the signal of ¹H-{¹⁴N^{DQ}} HMQC is equal to

$$S(\tau) \propto \frac{2}{3} \langle \sin^2(\theta) \sin^2(2\omega_{HN}\tau) \rangle$$
 (18)

with $\theta = 2\pi N d(v_1/v_R)^2$ [41]. Hence, for ¹H-{¹⁴N^{DQ}} J-HMQC experiments using XiX schemes, the maximal transfer efficiency is achieved for

$$\tau = 1/(4J_{\rm HN}) \tag{19}$$

3. 1D numerical simulations

3.1. Simulation parameters

The spin dynamics during the aforementioned pulse sequences were simulated using SIMPSON software [48]. The powder average was calculated using 3864 { α_{MR} , β_{MR} , γ_{MR} } Euler angles that described the orientation of the molecule in the rotor frame. The 966 { α_{MR} , β_{MR} } pairs were selected according to the REPULSION algorithm [49], whereas the four γ_{MR} angles were regularly stepped from 0 to 360°. The simulations were performed for one isolated ¹H-¹⁴N spin pair. The ¹⁴N nucleus was subject to H_{Q1} and H_{Q2} with {C_Q (MHz), η_Q } = {1.18, 0.5} or {3.21, 0.32}, which result in powder patterns with a full width of 1.77 and 4.82 MHz, respectively. For a powder-averaged resonance, two effects are related to H_{Q2}, both proportional to C_Q²/v_{0,14N}: a crystallite dependent shift which leads to a second-order broadening, and an overall positive quadrupole induced shift (QIS) of the gravity center

$$v_{\text{QIS}} = (3 + \eta_0^2) C_Q^2 / (40 v_{0,14\text{N}})$$
(20)

At $B_0 = 18.8$ T, $v_{QIS} = 1957$ and 13827 Hz, for the two previous cases with $C_Q = 1.18$ and 3.21 MHz, respectively. The former parameters correspond to those of ¹⁴N nucleus of α -glycine [15,50] and are typical of those measured in amino acids. The latter correspond to those of N-acetyl-L-valine (NAV) [51]. Except in Fig.S4, no CSA was introduced because the magnitude of this interaction is often negligible with respect to the quadrupole one [52,53]. In order to decrease the computing time, the simulations were done with the *J*-HMQC sequence shown in Fig.1a, instead of the targeted *D*-HMQC one (Fig.1b). This, however, does not change the analysis of the efficiency of the ¹⁴N excitation and reconversion, which is the main topic of this article. Therefore, in the simulations performed at $B_0 = 18.8$ T with $v_R = 62.5$ kHz, ¹H and ¹⁴N nuclei were only scalar coupled with $J_{1H-14N} = 400$ Hz.

We simulated the 1D signals of ¹H-{¹⁴N^{SQ}} *J*-HMQC sequences using HP, SLP, D₁^K and XiX schemes as well as ¹H-{¹⁴N^{DQ}} *J*-HMQC sequences using SLP and XiX schemes. The rf-field for the $\pi/2$ and π -pulses on the ¹H channel was fixed to $v_{1,1H} = 100$ kHz and was always applied on-resonance. Except in Fig.6, the pulses on ¹⁴N channel were applied on resonance, *i.e.*, their carrier frequency was equal to the sum of the isotropic chemical shift and the QIS. The maximum rf-field on ¹⁴N channel did not exceed 200 kHz in order to be consistent with the rf-power specifications of common MAS probes. For ¹H-{¹⁴N^{SQ}} *J*-HMQC simulations, the delays were set to their shortest optimal value of $\tau_J = 1/(4J_{HN}) = 2.5$ ms (Eq.8). For ¹H-{¹⁴N^{DQ}} simulations, we employed $\tau_J = 1/(4J_{HN})$ for XiX (Eq.19) and $\tau_J = 1/(2J_{HN})$ for SLP

(Eq.8). For these 1D simulations, the t_1 delay was fixed to an integer multiple of T_R , in order to refocus the evolution under H_{Q1} . For HP and SLP, $t_1 = 2T_R$, which is the largest τ_p value used in the simulations. For the same reason, t_1 was fixed to KT_R and $2NT_R$ for D_1^K and XiX schemes, respectively. In all *J*-HMQC simulations, the transfer efficiency was normalized with respect to the ¹H signal observed with an ideal spin-echo experiment with Dirac pulses. The maximum efficiency of ¹H-{¹⁴N^{SQ}} *J*-HMQC using two ideal Dirac ¹⁴N $\pi/2$ -pulses for one ¹H-¹⁴N spin pair with $C_Q = 0$ is equal to 0.35. This ideal value will be used as a benchmark in the following discussion.

3.2. Optimal rf-field amplitude and duration

The simulation-based analysis started by calculating a series of two-dimensional maps showing the transfer efficiency *versus* the rf-field and the duration of excitation and reconversion schemes applied to ¹⁴N channel on-resonance. In the case of HP or SLP these maps are displayed as ($v_1 vs. \tau_p$), whereas for XiX, they are displayed as ($v_1 vs. N$). In the case of D₁^K, each (v_1, τ_p) map was calculated for one particular *K* value. The maps corresponding to the best efficiencies are shown either in the main text (Figs.2-5) or in the supplementary information (Figs.S1 and S2).




Fig.2. Simulated on-resonance efficiency of ¹H-{¹⁴N^{SQ}} *J*-HMQC sequence using HP *versus* τ_p and ν_1 for spin systems corresponding to (a) glycine and (b) NAV.

As seen in Fig.2a, the maximal efficiency for ${}^{1}\text{H}-\{{}^{14}\text{N}^{SQ}\}\ J$ -HMQC using HP is achieved for $\tau_p = 0.75 \ \mu\text{s}$ and $\nu_1 = 200 \ \text{kHz}$, which corresponds to $\theta = 54^{\circ}$. This angle is lower than 90°. Furthermore, $\tau_p = 0.75 \ \mu\text{s}$ does not satisfy Eq.4 for glycine and hence, HP does not produce a uniform excitation of the ${}^{14}\text{N}^{SQ}$ powder pattern. Therefore, the efficiency of the sequence only reaches 0.2, which is 40% smaller than that of ${}^{1}\text{H}-\{{}^{14}\text{N}^{SQ}\}\ J$ -HMQC sequence using Dirac pulses for a ${}^{14}\text{N}$ nucleus not subject to quadrupole interaction. As seen in Fig.2b, the efficiency is further decreased for larger C_Q .




Fig.3. Simulated on-resonance efficiency of (a,c) ${}^{1}H-\{{}^{14}N^{SQ}\}\$ and (b,d) ${}^{1}H-\{{}^{14}N^{DQ}\}\$ *J*-HMQC sequence using SLP *versus* τ_{p} and ν_{1} for spin systems corresponding to (a,b) glycine and (c,d) NAV. The τ_{p} and ν_{1} values corresponding to $\pi/2$ -pulses are plotted as solid white lines.

For ¹H-{¹⁴N^{SQ}} *J*-HMQC using SLP (Fig.**3a** and **c**), the maximal transfer efficiency is obtained for $\tau_p \approx 2T_R$ and it reaches 0.23 with $v_1 \approx 35$ kHz for glycine and 0.20 with $v_1 \approx 50$ kHz for NAV. Those efficiencies are *ca*. the same and 100% higher than those obtained for glycine and NAV, respectively, using HP with $v_1 = 200$ kHz. These results indicate that (i) SLP is more efficient than HP for the excitation and the reconversion of ¹⁴N^{SQ} coherences, notably for ¹⁴N nuclei subject to large quadrupole interaction, and (ii) SLP requires much lower rf-field than HP. For ¹H-{¹⁴N^{DQ}} *J*-HMQC using SLP, maximal transfer efficiencies are obtained under similar conditions than for ¹H-{¹⁴N^{SQ}} (Fig.**3b** and **d**). However, the excitation and the reconversion of ¹⁴N^{SQ} coherences are significantly less efficient than those of ¹⁴N^{SQ} ones.




Fig.4. Simulated on-resonance efficiency of ¹H-{¹⁴N^{SQ}} *J*-HMQC sequence using D_1^K versus τ_p and v_1 for spin systems corresponding to (a) glycine with K = 5 and (b) NAV with K = 6.

For ¹H-{¹⁴N^{SQ}} *J*-HMQC using D₁⁵ applied to glycine, a maximal efficiency of 0.32 is obtained for rf-fields down to $v_1 = 80$ kHz. For such rf-field, the optimal τ_p length is 0.6 µs, which corresponds to $\theta \approx 86^{\circ}$. This angle is close to the theoretical optimal angle of 90°, whereas the achieved efficiency reaches almost the theoretical one of 0.35. This high efficiency indicates small losses due to H_{Q2}. For NAV, ¹H-{¹⁴N^{SQ}} *J*-HMQC with D₁⁶ yields a maximal efficiency of 0.21 for $\tau_p \approx 0.3$ µs and $v_1 = 100$ kHz, *i.e.*, $\theta \approx 65^{\circ}$. The reduced efficiency and tilt angle stem from faster H_{Q2} dephasings for NAV than glycine since H_{Q2} is proportional to C_Q². As a simple rule, when H_{Q2} dephasings are small, T_Rv_{QIS} ≤ 0.05 (e.g.

0.032 for glycine), the optimal total pulse duration corresponds to $\theta \approx 90^{\circ}$, but it decreases when it is not the case (e.g. 0.22 for NAV).

The simulations showing the simulated efficiency of ¹H-{¹⁴N^{SQ}} *J*-HMQC with D₁^K scheme and $3 \le K \le 8$ for a spin system corresponding to glycine or NAV are shown in the Supplementary Information (Figs.S1 or S2, respectively). Practically, the D₁^K optimization is performed by fixing v₁ at its maximum value (80 kHz with our MAS probe), and by optimizing τ_p for various *K* values.




Fig.5. Simulated on-resonance efficiency of (a,c) ${}^{1}H-{}^{14}N^{SQ}$ and (b,d) ${}^{1}H-{}^{14}N^{DQ}$ *J*-HMQC sequence using XiX *versus N* and v₁ for spin systems corresponding to (a,b) glycine and (c,d) NAV.

Fig.**5b** shows that maximal efficiency for ¹H-{¹⁴N^{DQ}} *J*-HMQC with XiX is obtained for v₁ = 30 kHz and N = 4 for glycine. For NAV, the maximal efficiency is obtained for smaller N values because of losses due to H_{Q2} (Fig.**5d**). Furthermore, for both glycine and NAV, XiX results in 45% lower efficiency for ¹H-{¹⁴N^{DQ}} *J*-HMQC than with SLP. Fig.**5a** and **c** indicate that XiX can also be used to excite and reconvert the ¹⁴N^{SQ} coherences. Shorter N values are required for NAV in order to limit the losses produced by H_{Q2}. ¹H-{¹⁴N^{SQ}} *J*-HMQC using XiX are twice less efficient than those using SLP for glycine and NAV.

The optimum parameters obtained from the aforementioned maps, as well as their resulting normalized efficiencies, Eff, are summarized in Table 1. These parameters were used in subsequent simulations. No further discussion of the XiX results will be presented in the remainder of this article, because their simulated and experimental efficiency and robustness (not shown) were consistently found to be much smaller than those of the SLP and D_1^K schemes.

| Excitation | $	au_{ m J}$ | τ _p / μs | v 1 / kHz | Eff | | |
|------------|--------------|---------------------|------------------|------|--|--|
| Glycine | | | | | | |
| HP SQ | 1/4J | 0.75 | 200 | 0.24 | | |
| SLP SQ | 1/4J | $32(2T_R)$ | 30 | 0.23 | | |
| SLP DQ | 1/2J | $32 (2T_R)$ | 35 | 0.16 | | |
| XiX SQ | 1/4J | 64 (N = 2) | 75 | 0.13 | | |
| XiX DQ | 1/4J | 128 (N = 4) | 30 | 0.11 | | |
| $D_1^K SQ$ | 1/4J | 0.5 (K = 5) | 80 | 0.32 | | |
| NAV | | | | | | |
| HP SQ | 1/4J | 0.75 | 200 | 0.11 | | |
| SLP SQ | 1/4J | $32(2T_R)$ | 50 | 0.20 | | |

Table 1. Optimum on-resonance simulation parameters used for glycine or NAV.

| SLP DQ | 1/2J | 32 (2T _R) | 55 | 0.10 |
|------------|------|-----------------------|-----|------|
| XiX SQ | 1/4J | 32 (N = 1) | 140 | 0.13 |
| XiX DQ | 1/4J | 64 (N = 2) | 65 | 0.05 |
| D_1^K SQ | 1/4J | 0.3 (K = 6) | 100 | 0.21 |

3.3. Robustness

3.3.1. Offset

In addition to the on-resonance sensitivity, another important criterion is the robustness to offset, which is shown in Fig.6a for glycine and Fig.6b for NAV. The two most robust excitations are those related to SLP. Their widths are the same with SQ or DQ selection and they depend very little upon the C_Q value (S_{FWHM} \approx 40 kHz). The D^K₁ scheme is less robust (S_{FWHM} \approx 20 kHz), and this value is the same for glycine and NAV. These results indicate that the robustness to offset mainly depends on the length of the excitation and reconversion blocks (see S_{FWHM} in Eq.10). The shorter length of SLP with respect to D^K₁ explains the higher robustness to offset of the former.




Fig.6. Simulated efficiency of ¹H-{¹⁴N^{SQ}} and ¹H-{¹⁴N^{DQ}} *J*-HMQC sequences using SLP and ¹H-{¹⁴N^{SQ}} with D_1^K with respect to ¹⁴N offset for spin systems corresponding to (a) glycine and (b) NAV. The rf parameters (v_1 , τ_p) are indicated with + on Figs.2-4.

3.3.2. Rf-inhomogeneity

Previous simulations were performed with one particular rf-field. However, there exists an inherent distribution of rf-fields in the coil of NMR probes, mostly along the axis of the solenoid coil in MAS probes. Typically, this amplitude is maximal at the center of the coil and it drops by *ca*. 50 % near the edges [54–57].

For SLP, the simulations displayed in Fig.7 show that, independently of the C_Q value, (i) the optimal rf-field is equal to *ca*. 50-80 % of v_R , and (ii) the relative widths of these curves, *i.e.*, the FWHM divided by the optimum rf-field, which are the relevant values with respect to rf-inhomogeneity, are approximately the same with SQ or DQ selection.

With D_1^K trains, the rf-requirement is higher than with SLP, but the efficiency is larger with small C_Q values. Conversely for large C_Q values, the losses due to H_{Q2} decrease the efficiency.





Fig.7. Simulated on-resonance efficiency of ${}^{1}H-{{}^{14}N^{SQ}}$ and ${}^{1}H-{{}^{14}N^{DQ}}$ *J*-HMQC sequences using SLP and ${}^{1}H-{{}^{14}N^{SQ}}$ *J*-HMQC sequences using D₁^K with respect to respect to ${}^{14}N$ rf-field strength for spin systems corresponding to (a) glycine and (b) NAV. The rf-parameters (v₁, τ_p) are indicated with + on Figs.2-4, with D₁⁵ in (a) and D₁⁶ in (b).

3.3.3. Spinning speed and CSA

The robustness with respect to spinning speed and CSA (Figs.S3 and S4, respectively) was also tested. The simulations show that SLP depends little upon these two parameters, but that D_1^K is sensitive with respect to spinning speed fluctuations. However, modern MAS speed controllers can provide sufficiently stable MAS frequency to allow the use of D_1^K schemes.



3.3.4. Quadrupole interaction

Fig.8. Simulated on-resonance optimized efficiency of ¹H-{¹⁴N^{SQ}} *J*-HMQC sequences using SLP and D^K₁ (green and purple curves, respectively) for a series of quadrupole parameters corresponding to $\eta_Q = 0.5$ and $C_Q = 0.4$ MHz, incremented in steps of 125 kHz for ($\nu_{1,max}/kHz$, ν_R/kHz , B_0/T) = (a) (80, 62.5, 18.8); (b) (160, 62.5, 18.8); (c) (80, 125, 18.8) and (d) (80, 62.5, 9.4). For SLP, the τ_p range investigated was 1.5-2.5T_R. For D^K₁, the investigated range of the total pulse duration, $K\tau_p$, is 0-10 µs, with *K* ranging from 3 to 8, as also shown in Fig.S1 and S2.

The last aspect analyzed via simulations was the robustness with respect to the C_Q value. This is an important criterion as this parameter can vary between 0 and 7 MHz, depending on the local environment of ¹⁴N nucleus [5,7]. To that end, the optimum efficiency of the two most efficient methods, SLP and D_1^K with SQ selection, was calculated for a series of quadrupole parameters corresponding to $C_Q = 0.4$ MHz. These simulations were performed with the following sets of parameters: $(v_{1,max}/kHz, v_R/kHz, B_0/T) = (80, 62.5, 18.8);$ (160, 62.5, 18.8); (80, 125, 18.8) and (80, 62.5, 9.4), and the corresponding results are shown in Fig.8a-d, respectively. These results show that D_1^K on-resonance excitation is more efficient than SLP for small and moderate C_Q values and that it is the contrary for large C_Q values. It can be seen that the efficiency of D_1^K increases: (i) with v_1 at moderate to high C_Q value (compare Figs.8a) and b), and (ii) with v_R and B_0 at high C_Q value (compare Figs.8a, c and d). Indeed, a large rf field allows the use of shorter D_1^K trains, *i.e.*, smaller K values (compare Figs.9a and b), which limits the losses due to H_{Q2}, which are proportional to C_Q^2 . Similarly high MAS frequency decreases the length, KT_R , of D_1^K train, hence reducing losses due to H_{Q2}. Finally, as H_{Q2} is inversely proportional to B_0 , losses produced by this interaction are smaller at high field. By contrast, the on-resonance SLP excitation exhibits very little dependence upon these parameters, and hence the range of C_Q values where D_1^K is more efficient than SLP extends with increasing v_1 , v_R and B_0 values. However, it must be pointed out that the robustness to offset of SLP is much larger than that of D_1^K (Fig.6), and that this robustness further increases with faster spinning speeds (Fig.5 of [39]). Therefore, for small or moderate C_0 values, the choice in between these two schemes also depends on the range of isotropic resonance frequencies.




Fig.9. τ_p and *K* values, green and purple curves respectively, yielding optimal simulated efficiency for ¹H-{¹⁴N^{SQ}} *J*-HMQC with D₁^K for a series of quadrupole parameters corresponding to $\eta_Q = 0.5$ and $C_Q = 0.4$ MHz, incremented in steps of 125 kHz for ($\nu_{1,max}$ /kHz, ν_R /kHz, B₀/T) = (a) (80, 62.5, 18.8); (b) (160, 62.5, 18.8); (c) (80, 125, 18.8) and (d) (80, 62.5, 9.4).

Fig.9a-d plots the variations in optimal τ_p and *K* values of the D₁^K scheme with respect to C_Q, over the same range of parameters as for Fig.8. These results show that, in general, to obtain a maximum efficiency with D₁^K, the higher the C_Q value, the shorter the train that should be used. Such an arrangement is effective because it limits the signal decay due to H_{Q2} dephasing [40]. This leads to a decrease of *K* with large C_Q values in Fig.9a and d. It can be observed that, at lower B₀ field (Fig.9d), the lengths of τ_p at high C_Q are noticeably longer than the equivalent values at 18.8 T. Long pulses allow for reduction in the length of D₁^K train, *i.e.*, the *K* value, hence limiting the losses due to H_{Q2}. Regardless of which values are used, excitation of high C_Q values at low magnetic field is likely to be inefficient with D₁^K.

It should be noted that the data presented in Figs.8 and 9 were sorted computationally and, as such, represent theoretical, numerical maxima. These efficiencies often only differ after several decimals and so the choice between, for example, one K value and another, may vary from a practical perspective.

4. Experimental results for ¹H-{¹⁴N} *D*-HMQC

4.1. Experimental conditions

All experiments were performed at $B_0 = 18.8$ T on a Bruker standard bore superconducting magnet equipped with an AV-IV console. Two model samples were selected for this study: Lhistidine HCl and N-acetyl-L-valine, referred to His and NAV, and obtained from Sigma-Aldrich and Alfa Aesar, respectively, and used as-received. Their ${}^{1}H{-}{{}^{14}N^{SQ}}$ and ${}^{1}H{-}{{}^{14}N^{DQ}}$ spectra were acquired with the D-HMQC sequence depicted in Fig.1b, which was employed since it is more sensitive than the J-HMQC variant for solids. The samples were packed into 1.3 mm rotors and spun under MAS at $v_R = 62.5$ kHz, using a standard HX MAS probe. SR4²₁ recoupling with $v_1 = 2v_R = 125$ kHz was applied on ¹H channel during the τ_D delays to reintroduce the ¹⁴N-¹H dipolar couplings. The ¹H rf-field for the $\pi/2$ and π -pulses was set to $v_1 \approx 116$ kHz. The rf-field on the ¹⁴N channel delivered by this probe is limited to $v_1 \approx 80$ kHz. HP, D₁^K trains and SLP were employed to excite and reconvert the ¹⁴N SQ or DQ coherences. The position of ¹⁴N carrier frequency for His is indicated in Fig.10. For NAV, the ¹⁴N carrier frequency was resonant with the NH peak. In Fig. 12, the ¹⁴N carrier frequency was varied. For 1D experiment, the t_1 delay was fixed to its minimum rotor-synchronized value equal either to $2T_R$ for rectangular pulses or KT_R for D_1^K trains. All 2D spectra were acquired using t_1 period equal to an integer multiple of T_R . This rotor-synchronization allows for refocusing of the broadening by H_{Q1} for ¹H-{¹⁴N^{SQ}} D-HMQC spectra. The spectra were acquired with a relaxation delay, $\tau_{RD} = 1$ s. The other experimental parameters are given in Table 2 as well as in the figure captions. Chemical shift referencing was carried out using solid samples of NH₄Cl ($\delta_{14N} = 0$ ppm) and His ($\delta_{1H} = 8.2$ ppm for N^tH). His has three ¹⁴N species that possess the following NMR parameters: $\delta_{iso} = 4.0/146.6/133.4$ ppm, C₀ = 1.26/1.46/1.29 MHz and $\eta_0 = 0.36/0.30/0.97$, for NH₃, N^{π}H and N^{τ}H, respectively [58]. When taking into account their v_{QIS} values, these peaks appear at -17200, -8500 and -9500 Hz, respectively. NAV has only one species with NMR parameters: $\delta_{iso} = 127.7$ ppm, $C_0 =$ 3.21 MHz and $\eta_0 = 0.32$ [51]. Its ¹⁴N resonance appears at -2000 Hz on the spectra presented in this work.

4.2. Optimized parameters

In this part, we present the results observed for His and NAV that were recorded either with conventional HP (SQ, $v_1 = 80$ kHz), D_1^K (SQ, $v_1 = 40$ or 80 kHz) or SLP (SQ and DQ) with optimized v_1 values. In Table 2, it can be observed that the optimized SLP length, $\tau_p = 20-28$ $\mu s \approx 1.5T_R$, depends very little upon the sample and hence on the C_Q value, and that these values are slightly smaller than the simulated one, $\tau_p \approx 32 \ \mu s = 2T_R$. This small decrease is attributed to experimental losses, such as those due to ¹H-¹H dipolar couplings that were not included in the simulations. Concerning the D_1^K scheme, the pulse length with $v_1 = 80$ kHz is longer than in the simulations: 1 instead of 0.6 μs for His and 0.5 instead of 0.3 μs for NAV, whereas the optimal *K* values are smaller: 4 instead of 5 and 6 for His and NAV, respectively. These smaller *K* values stem from experimental losses, such as those due to ¹H-¹H dipolar couplings, which were not included in the simulations.

| Excitation | τ _D / μs | $	au_{ m p}$ / μs | v 1 / kHz | | | |
|--------------------------------|---------------------|------------------------|------------------|--|--|--|
| His | | | | | | |
| HP SQ | 140 | 1 | 80 | | | |
| SLP SQ | 140 | 24 | 40 | | | |
| SLP DQ | 160 | 20 | 53 | | | |
| D ₁ ^K SQ | 140 | 1 (K = 6) | 40 | | | |
| $D_1^{\overline{K}}$ SQ | 140 | 1 (K = 4) | 80 | | | |
| NAV | | | | | | |
| HP SQ | | 1 | 80 | | | |
| SLP SQ | 170 | 28 | 45 | | | |
| SLP DQ | 130 | 28 | 64 | | | |
| D ₁ ^K SQ | 140 | 2 $(K = 3)$ | 40 | | | |
| D ₁ ^K SQ | 140 | 0.5 (K = 4) | 80 | | | |

Table 2. Optimal experimental parameters used for His or NAV.

4.3. Efficiencies

¹H-{ ^{14}N } *D*-HMQC spectra were acquired with the aforementioned optimized parameters from Table 2.




Fig.10. ¹H-{¹⁴N^{SQ}} *D*-HMQC spectra of His acquired using (a) D_1^4 train and (b) SLP excitation, with $B_0 = 18.8$ T, $v_1 = 80$ (D_1^4) or 40 kHz (SLP), $v_R = 62.5$ kHz. The 2D spectra result from averaging 8 transients for each of the 250 t_1 increments, resulting in a total experiment time of 33 min. Red arrows indicate the positions of the carrier frequencies.

Fig.10a and b demonstrate that such 2D spectra can be recorded with a good efficiency for His with experimentally achievable rf-fields of $v_1 = 80$ or 40 kHz, with SQ selection and either D_1^K or SLP excitation, respectively.




Fig.11. F₁ slices from 2D ¹H-{¹⁴N} *D*-HMQC spectra for (a) NH₃, (b) N^{π}H, (c) N^{τ}H of His and (d) NH of NAV, with B₀ = 18.8 T, v_R = 62.5 kHz, and v₁ as indicated in (b). In all cases, SQ line-shapes have been scaled up by 2 with respect to DQ ones. Additionally, the DQ ppm scale has been divided by two to allow for comparison with SQ slices. The 2D spectra result from averaging 8 (SQ) or 16 (DQ) transients for each of 250 (SQ) or 124 (DQ) *t*₁ increments, resulting in a total experiment time of 33 min.

The F₁ slices of the 2D spectra for NH₃, N^{π}H, N^{τ}H of His and NAV, are shown in Figs.**11a**, **b**, **c** and **d**, respectively. It can be seen that the S/N ratio observed with NAV is smaller than with His. The ¹⁴N MAS linewidth of NAV is *ca*. 5-6 times larger than those of His. However, this larger indirect linewidth should not significantly affect the sensitivity because it allows for the use of fewer t₁ increments. Therefore, the main cause of the reduced HMQC efficiency in NAV is likely to be brought about by the increase in *C*_Q value (Fig.**8**), and possibly a reduction in the ¹H-¹⁴N dipolar coupling or ¹H coherence lifetime.

It is evident that, even with the highest available rf-field, HP excitation is inefficient, especially for large C_Q values, as it is the case for NAV. With respect to SQ selection, DQ always leads to a large decrease in efficiency, as previously shown *via* simulations. Moreover, no resolution enhancement is observed with DQ selection over the SQ one, indicating that in the latter case the spinning speed was stable and the magic angle was well-adjusted, and hence, the broadening of SQ coherences by H_{Q1} was properly refocused. D_1^K excitation can provide almost the same level of efficiency with $v_1 = 40$ or 80 kHz, because the increased value of $K\tau_p$ is able to compensate for the decrease in rf-field strength, a finding which is in agreement with previous studies [12,40]. As seen in Fig.11a-c, the relative efficiency of SLP and D_1^K for His depends on the observed site. However, D_1^K is not efficient for large C_Q values, as shown in Fig.11d for NAV, owing to large losses due to H_{Q2} .

17




Fig.12. Experimental (a,b) ${}^{1}H{}^{14}N^{SQ}$ and (c,d) ${}^{1}H{}^{14}N^{DQ}$ *D*-HMQC 1D signal *versus* ${}^{14}N$ offset with SLP of (a,c) N^tH of His and (b,d) NH of NAV, at $B_0 = 18.8$ T with $v_R = 62.5$ kHz using $v_1 = 40$ (a,c) or (b,d) 45 kHz.

Fig.12 shows the sensitivity to ¹⁴N offset observed with SLP and SQ or DQ selection, for N^TH of His (Figs.12a and c) and NAV (Figs.12b and d). This robustness is very large, especially with SQ selection (Figs.12a and b), with full-width at half maximum excitation bandwidth of $S_{FWHM} \approx 60$ and 40 kHz for His and NAV respectively, whereas it is smaller with DQ selection: $S_{FWHM} \approx 40$ and 30 kHz (Figs.12c and d). As previously noted, HMQC experiments using DANTE trains are less robust to offset than those using SLP, although DANTE benefits from a slight increase in overall on-resonance excitation efficiency [40]. It must be noted that the robustness to ¹⁴N^{SQ} offset of ¹H-{¹⁴N} *D*-HMQC 2D experiment with SLP excitation ($S_{FWHM}/v_R \approx 0.6-1.0$) is much larger than that of the corresponding method with double CP MAS transfer ($S_{FWHM}/v_R \approx 0.25$) (Fig.4d-AM in [31]). It has very recently been demonstrated that optimizing the carrier frequency is recommended when using SLP and this is even more the case for D₁^K (Figs.11 and 12 in [59]).

5. Conclusions

In this work, various previously proposed schemes for ¹⁴N excitation in ¹H-{¹⁴N} HMQC experiments have been analyzed and compared. This analysis was performed with selection of either single- (SQ) or double- (DQ) quantum levels of ¹⁴N nuclei. DQ selection leads to a smaller efficiency than SQ selection, without any resolution enhancement, and is thus not generally recommended with a modern speed controller and MAS probe.

Conventional hard pulses are not efficient and should be avoided. The same recommendation is given for the XiX (X inverse X) scheme, which has recently been proposed for DQ selection.

The most efficient methods are the sideband selective long pulse (SLP) or the D_1^K (DANTE) train of pulses. It has been shown here that SLP excitation is much more efficient than D_1^K trains for large C_Q values. As SLP is also more robust with respect to offset, rf-inhomogeneity and experimental setup, this scheme should therefore be preferred, and is particularly recommended for use by non-specialists, due to its simplicity. For weak or moderate C_Q values, the choice between the two schemes depends on the rf and magnetic field strengths, the spinning speed, and the ¹⁴N isotropic frequency range. When this range is very large, SLP, which is very robust with respect to offsets, should also be the sequence of choice. In the case of paramagnetic samples, the ¹⁴N isotropic shift range may even be larger than v_R leading to folded sidebands. In this case, two 2D experiments with two different spinning speeds should be acquired in order to discriminate the center-bands from the spinning sidebands [59].

It must be emphasized that ¹H-{¹⁴N} HMQC experiments with SQ selection, which require the cancellation of the first-order quadrupole dephasing in order to be well-resolved along the ¹⁴N dimension, benefit from several recent technical developments. These include new MAS speed controllers that provide very stable spinning speeds, and new MAS probes with automatic magic angle adjustment with a level of precision that meets the so-called 'STMAS requirements'.

It must also be noted that very recently, a new variant of the SR4²₁ recoupling sequence has been proposed for ¹H-{X} *D*-HMQC experiments at ultra-fast MAS. This variant replaces the basic π -pulses with 90₋₄₅90₄₅90₋₄₅ composite π -pulses to decrease the effects of ¹H-¹H interactions [60]. As a result, this recoupling quenches the spin diffusion and hence long-lived coherence lifetimes can be obtained, which facilitate the detection of long N-H distances with this non-dipolar truncated sequence.

Finally, it must be noted that the ¹H-{¹⁴N} *D*-HMQC experiment is complementary to the double CP scheme recently proposed [**31**]. Both methods experimentally require the same 'STMAS specifications' and ultra-fast sample spinning to minimize the losses and maximize the indirect spectral width. Compared with *D*-HMQC using SLP excitation, the double CP method exhibits a lower robustness to offset, but a better ¹⁴N resolution when no ¹H decoupling is used. However, it should be pointed out that the ¹⁴N resolution of *D*-HMQC can be enhanced with ¹H decoupling, and this allowed for the first detection of the cross-peak between aliphatic ¹H and ¹⁴N ammonium site in Histidine [**61**]. It must also be reminded that two parameters have to be optimized with the double CP scheme: the contact times and the rf-fields to match the Hartmann-Hahn conditions [**62,63**], which are narrow at ultra-fast MAS [**64,65**]. On the contrary, there is no such rf-matching with *D*-HMQC and, moreover, with SLP excitation the ¹⁴N rf-field is small and efficient over a broad range. To conclude, it ought also to be noted that the ¹H resolution can be enhanced by using weak continuous wave ¹⁴N irradiation during *t*₂ [**65**].

Acknowledgments

Institut Chevreul (FR 2638), Ministère de l'Enseignement Supérieur et de la Recherche, Région Hauts-de-France and the European Union (FEDER/ERDF) are acknowledged for supporting and partially funding this work. Financial support from the IR-RMN-THC FR 3050 CNRS for conducting the research is gratefully acknowledged. Authors also thank ANR-17-ERC2-0022 (EOS) and ANR-18-CE08-0015-01 (ThinGlass). This project has also received funding from the European Union's Horizon 2020 research and innovation program under grant agreement no. 731019 (EUSMI). OL acknowledges financial support from Institut Universitaire de France (IUF). The computational resources were partially provided by the Polish Infrastructure for Supporting Computational Science in the European Research Space (PL-GRID). PP is grateful for funding from the MOBILITY PLUS program (grant no. DN/MOB/172/V/2017).

Supplementary Information: Figs.S1-S4.

References

[1] S.M. Althaus, K. Mao, J.A. Stringer, T. Kobayashi, M. Pruski, Indirectly detected heteronuclear correlation solid-state NMR spectroscopy of naturally abundant ¹⁵N nuclei, *Solid State Nucl. Magn. Reson.* 57–58 (**2014**) 17–21. doi:10.1016/j.ssnmr.2013.11.001.

[2] S.L. Veinberg, K.E. Johnston, M.J. Jaroszewicz, B.M. Kispal, C.R. Mireault, T. Kobayashi, M. Pruski, R.W. Schurko, Natural abundance ¹⁴N and ¹⁵N solid-state NMR of pharmaceuticals and their polymorphs, *Phys. Chem. Chem. Phys.* 18 (**2016**) 17713–17730. doi:10.1039/C6CP02855A.

[3] A.J. Rossini, A. Zagdoun, M. Lelli, J. Canivet, S. Aguado, O. Ouari, P. Tordo, M. Rosay, W.E. Maas, C. Copéret, D. Farrusseng, L. Emsley, A. Lesage, Dynamic Nuclear Polarization Enhanced Solid-State NMR Spectroscopy of Functionalized Metal-Organic Frameworks, *Angew. Chem. Int. Ed Engl.* 51 (**2012**) 123–127. doi:10.1002/anie.201106030.

[4] L.A. O'Dell, Direct detection of nitrogen-14 in solid-state NMR spectroscopy, *Prog. Nucl. Magn. Reson. Spectrosc.* 59 (**2011**) 295–318. doi:10.1016/j.pnmrs.2011.04.001.

[5] E. Dib, T. Mineva, B. Alonso, Chapter Three - Recent Advances in ¹⁴N Solid-State NMR, in: G.A. Webb (Ed.), *Annu. Rep. NMR Spectrosc.*, Academic Press, **2016**: 175–235. doi:10.1016/bs.arnmr.2015.08.002.

[6] L.A. O'Dell, ¹⁴N Overtone Magic Angle Spinning NMR, in: G.A. Webb (Ed.), *Annu. Rep. NMR Spectrosc.*, Academic Press, **2015**: 211–236. doi:10.1016/bs.arnmr.2015.04.003.

[7] J.P. Yesinowski, E.A. Hill, Wide-Line ¹⁴N NMR in Solids, in: *Solid-State NMR Spectrosc. Inorg. Mater.*, American Chemical Society, **1999**: 358–376. doi:10.1021/bk-1999-0717.ch013.

[8] E.A. Hill, J.P. Yesinowski, Wide-Line ¹⁴N NMR in Solids and Reorientation-Induced Redistribution of Isochromats, *J. Am. Chem. Soc.* 118 (**1996**) 6798–6799. doi:10.1021/ja960757j.

[9] L.A. O'Dell, R.W. Schurko, Fast and simple acquisition of solid-state ¹⁴N NMR

spectra with signal enhancement via population transfer., *J. Am. Chem. Soc.* 131 (**2009**) 6658–6659. doi:10.1021/ja901278q.

[10] V. Vitzthum, M.A. Caporini, S. Ulzega, G. Bodenhausen, Broadband excitation and indirect detection of nitrogen-14 in rotating solids using Delays Alternating with Nutation (DANTE)., *J. Magn. Reson.* 212 (**2011**) 234–239. doi:10.1016/j.jmr.2011.06.013.

[11] V. Vitzthum, M.A Caporini, S. Ulzega, J. Trébosc, O. Lafon, J.-P. Amoureux, G. Bodenhausen, Uniform broadband excitation of crystallites in rotating solids using interleaved sequences of delays alternating with nutation., *J. Magn. Reson.* 223 (**2012**) 228–236. doi:10.1016/j.jmr.2012.05.024.

[12] X. Lu, J. Trébosc, O. Lafon, D. Carnevale, S. Ulzega, G. Bodenhausen, J.-P. Amoureux, Broadband excitation in solid-state NMR using interleaved DANTE pulse trains with N pulses per rotor period., *J. Magn. Reson.* 236 (**2013**) 105–116. doi:10.1016/j.jmr.2013.09.003.

[13] K.J. Harris, S.L. Veinberg, C.R. Mireault, A. Lupulescu, L. Frydman, R.W. Schurko, Rapid Acquisition of ¹⁴N Solid-State NMR Spectra with Broadband Cross Polarization, *Chem. – Eur. J.* 19 (**2013**) 16469–16475. doi:10.1002/chem.201301862.

[14] H.J. Jakobsen, H. Bildsøe, J. Skibsted, T. Giavani, ¹⁴N MAS NMR Spectroscopy: The Nitrate Ion, *J. Am. Chem. Soc.* 123 (**2001**) 5098–5099. doi:10.1021/ja0100118.

[15] T. Giavani, H. Bildsøe, J. Skibsted, H.J. Jakobsen, A solid-state ¹⁴N magic-angle spinning NMR study of some amino acids, *J. Magn. Reson.* 166 (**2004**) 262–272. doi:10.1016/j.jmr.2003.10.023.

[16] A.R. Hove, H. Bildsøe, J. Skibsted, M. Brorson, H.J. Jakobsen, Probing Crystal Structures and Transformation Reactions of Ammonium Molybdates by ¹⁴N MAS NMR Spectroscopy, *Inorg. Chem.* 45 (**2006**) 10873–10881. doi:10.1021/ic061197k.

[17] H.J. Jakobsen, A.R. Hove, H. Bildsøe, J. Skibsted, M. Brorson, Long-term stability of rotor-controlled MAS frequencies to 0.1 Hz proved by ¹⁴N MAS NMR experiments and simulations., *J. Magn. Reson.* 185 (**2007**) 159–63. doi:10.1016/j.jmr.2006.12.008.

[18] M. Bloom, M.A. LeGros, Direct detection of two-quantum coherence, *Can. J. Phys.* 64 (**1986**) 1522–1528. doi:10.1139/p86-271.

[19] R. Tycko, S.J. Opella, Overtone NMR spectroscopy, J. Chem. Phys. 86 (**1987**) 1761– 1774. doi:10.1063/1.452176.

[20] L.A. O'Dell, A. Brinkmann, ¹⁴N overtone NMR spectra under magic angle spinning: Experiments and numerically exact simulations, *J. Chem. Phys.* 138 (**2013**) 64201. doi:10.1063/1.4775592.

[21] Z. Gan, I. Hung, Y. Nishiyama, J.-P. Amoureux, O. Lafon, H. Nagashima, J. Trébosc,
B. Hu, 14N overtone nuclear magnetic resonance of rotating solids, *J. Chem. Phys.* 149 (2018) 64201. doi:10.1063/1.5044653.

[22] A.J. Rossini, L. Emsley, L.A. O'Dell, Dynamic nuclear polarisation enhanced ¹⁴N overtone MAS NMR spectroscopy, *Phys. Chem. Chem. Phys.* 16 (**2014**) 12890–12899. doi:10.1039/C4CP00590B.

[23] I. Haies, J.A. Jarvis, H. Bentley, I. Kuprov, P.T.F. Williamson, M. Carravetta, ¹⁴N Overtone NMR under MAS: signal enhancement using symmetry-based sequences and novel simulation strategies, *Phys Chem Chem Phys.* 17 (**2015**) 6577–6587. doi:10.1039/C4CP03994G.

[24] L.A. O'Dell, C.I. Ratcliffe, ¹⁴N magic angle spinning overtone NMR spectra, *Chem. Phys. Lett.* 514 (**2011**) 168–173. doi:10.1016/j.cplett.2011.08.030.

[25] S. Cavadini, S. Antonijevic, A. Lupulescu, G. Bodenhausen, Indirect detection of nitrogen-14 in solids via protons by nuclear magnetic resonance spectroscopy., *J. Magn. Reson.* 182 (**2006**) 168–72. doi:10.1016/j.jmr.2006.06.003.

[26] S. Cavadini, Indirect detection of nitrogen-14 in solid-state NMR spectroscopy., *Prog. Nucl. Magn. Reson. Spectrosc.* 56 (**2010**) 46–77. doi:10.1016/j.pnmrs.2009.08.001.

[27] S.P. Brown, Nitrogen–Proton Correlation Experiments of Organic Solids at Natural Isotopic Abundance, *eMagRes*. 3 (**2014**) 243–254. doi:10.1002/9780470034590. emrstm1323.

[28] Z. Gan, J.-P. Amoureux, J. Trébosc, Proton-detected ¹⁴N MAS NMR using homonuclear decoupled rotary resonance, *Chem. Phys. Lett.* 435 (**2007**) 163–169. doi:10.1016/j.cplett.2006.12.066.

[29] S. Cavadini, A. Abraham, G. Bodenhausen, Coherence transfer between spy nuclei and nitrogen-14 in solids., *J. Magn. Reson.* 190 (**2008**) 160–164. doi:10.1016/j.jmr.2007.10.008.

[30] S. Antonijevic, N. Halpern-Manners, Probing amide bond nitrogens in solids using ¹⁴N NMR spectroscopy., *Solid State Nucl. Magn. Reson.* 33 (**2008**) 82–87. doi:10.1016/j.ssnmr.2008.04.006.

[31] D. Carnevale, X. Ji, G. Bodenhausen, Double cross polarization for the indirect detection of nitrogen-14 nuclei in magic angle spinning NMR spectroscopy, *J. Chem. Phys.* 147 (**2017**) 184201. doi:10.1063/1.5000689.

[32] J.A. Jarvis, M. Concistre, I.M. Haies, R.W. Bounds, I. Kuprov, M. Carravetta, P.T.F. Williamson, Quantitative analysis of ¹⁴N quadrupolar coupling using ¹H detected ¹⁴N solid-state NMR, *Phys. Chem. Chem. Phys.* 21 (**2019**) 5941–5949. doi:10.1039/C8CP06276E.

[33] O. Lafon, Q. Wang, B. Hu, F. Vasconcelos, J. Trébosc, S. Cristol, F. Deng, J.-P. Amoureux, Indirect Detection via Spin-1/2 Nuclei in Solid State NMR Spectroscopy: Application to the Observation of Proximities between Protons and Quadrupolar Nuclei, *J. Phys. Chem. A.* 113 (**2009**) 12864–12878. doi:10.1021/jp906099k.

[34] A. Brinkmann, A.P.M. Kentgens, Proton-Selective ¹⁷O-¹H Distance Measurements in Fast Magic-Angle-Spinning Solid-State NMR Spectroscopy for the Determination of Hydrogen Bond Lengths, *J. Am. Chem. Soc.* 128 (**2006**) 14758–14759. doi:10.1021/ja065415k

[35] S. Cavadini, A. Abraham, G. Bodenhausen, Proton-detected nitrogen-14 NMR by recoupling of heteronuclear dipolar interactions using symmetry-based sequences, *Chem. Phys. Lett.* 445 (**2007**) 1–5. doi:10.1016/j.cplett.2007.07.060.

[36] Z. Gan, Measuring amide nitrogen quadrupolar coupling by high-resolution ${}^{14}N/{}^{13}C$

NMR correlation under magic-angle spinning., J. Am. Chem. Soc. 128 (2006) 6040–6041. doi:10.1021/ja0578597.

[37] Z. Gan, ¹³C/¹⁴N heteronuclear multiple-quantum correlation with rotary resonance and REDOR dipolar recoupling., *J. Magn. Reson.* 184 (**2007**) 39–43. doi:10.1016/j.jmr.2006.09.016.

[38] S. Cavadini, A. Lupulescu, S. Antonijevic, G. Bodenhausen, Nitrogen-14 NMR Spectroscopy Using Residual Dipolar Splittings in Solids, *J. Am. Chem. Soc.* 128 (2006) 7706–7707. doi:10.1021/ja0618898.

[39] M. Shen, J. Trébosc, O. Lafon, Z. Gan, F. Pourpoint, B. Hu, Q. Chen, J.-P. Amoureux, Solid-state NMR indirect detection of nuclei experiencing large anisotropic interactions using spinning sideband-selective pulses, *Solid State Nucl. Magn. Reson.* 72 (2015) 104–117. doi:10.1016/j.ssnmr.2015.09.003.

[40] Y. Li, J. Trébosc, B. Hu, M. Shen, J.-P. Amoureux, O. Lafon, Indirect detection of broad spectra in solid-state NMR using interleaved DANTE trains, *J. Magn. Reson.* 294 (2018) 101–114. doi:10.1016/j.jmr.2018.07.005.

[41] A.J. Pell, K.J. Sanders, S. Wegner, G. Pintacuda, C.P. Grey, Low-power broadband solid-state MAS NMR of ¹⁴N, *J. Chem. Phys.* 146 (**2017**) 194202. doi:10.1063/1.4983220.

[42] Y. Nishiyama, M. Malon, Z. Gan, Y. Endo, T. Nemoto, Proton – nitrogen-14 overtone two-dimensional correlation NMR spectroscopy of solid-sample at very fast magic angle sample spinning, *J. Magn. Reson.* 230 (**2013**) 160–164. doi:10.1016/j.jmr.2013.02.015.

[43] L.A. O'Dell, R. He, J. Pandohee, Identifying H–N proximities in solid-state NMR using ¹⁴N overtone irradiation under fast MAS, *CrystEngComm.* 15 (**2013**) 8657–8667. doi:10.1039/c3ce40967h.

[44] M. Shen, Q. Chen, J.-P. Amoureux, B. Hu, Broad-band excitation in indirectly detected ¹⁴N overtone spectroscopy with composite pulses, *Solid State Nucl. Magn. Reson.* 78 (**2016**) 5–8. doi:10.1016/j.ssnmr.2016.05.001.

[45] M. Shen, J. Trébosc, L.A. O'Dell, O. Lafon, F. Pourpoint, B. Hu, Q. Chen, J.-P. Amoureux, Comparison of various NMR methods for the indirect detection of nitrogen-14 nuclei via protons in solids, *J. Magn. Reson.* 258 (2015) 86–95. doi:10.1016/j.jmr.2015.06.008.

[46] D. Merlet, M. Sarfati, B. Ancian, J. Courtieu, P. Lesot, Description of natural abundance deuterium 2D-NMR experiments in weakly ordered liquid-crystalline solvents using a tailored cartesian spin-operator formalism, *Phys. Chem. Chem. Phys.* 2 (2000) 2283–2290. doi:10.1039/B000834F.

[47] A.J. Pell, G. Pintacuda, Broadband solid-state MAS NMR of paramagnetic systems, *Prog. Nucl. Magn. Reson. Spectrosc.* 84–85 (**2015**) 33–72. doi:10.1016/j.pnmrs.2014.12.002.

[48] M. Bak, J.T. Rasmussen, N.C. Nielsen, SIMPSON: a general simulation program for solid-state NMR spectroscopy., *J. Magn. Reson.* 147 (**2000**) 296–330. doi:10.1006/jmre.2000.2179.

[49] M. Bak, N.C. Nielsen, REPULSION, A Novel Approach to Efficient Powder

Averaging in Solid-State NMR, J. Magn. Reson. 125 (1997) 132–139. doi: 10.1006/jmre.1996.1087.

[50] S.L. Veinberg, Z.W. Friedl, A.W. Lindquist, B. Kispal, K.J. Harris, L.A. O'Dell, R.W. Schurko, ¹⁴N Solid-State NMR Spectroscopy of Amino Acids, *ChemPhysChem.* 17 (**2015**) 4011–4027. doi:10.1002/cphc.201600873.

[51] R.E. Stark, R.A. Haberkorn, R.G. Griffin, ¹⁴N NMR determination of NH bond lengths in solids, *J. Chem. Phys.* 68 (**1978**) 1996–1997. doi:10.1063/1.435879.

[52] R.A. Santos, P. Tang, W.J. Chien, S. Kwan, G.S. Harbison, NMR resolution of a crystallographic controversy: nitrogen-14 single-crystal studies of silver, barium, and lead nitrates, *J. Phys. Chem.* 94 (**1990**) 2717–2721. doi:10.1021/j100369a088.

[53] T. Giavani, H. Bildsøe, J. Skibsted, H.J. Jakobsen, Determination of nitrogen chemical shift anisotropy from the second-order cross-term in ¹⁴N MAS NMR spectroscopy, *Chem. Phys. Lett.* 377 (**2003**) 426–432. doi:10.1016/S0009-2614(03)01140-0.

[54] R. Gupta, G. Hou, T. Polenova, A.J. Vega, RF inhomogeneity and how it controls CPMAS, *Solid State Nucl. Magn. Reson.* 72 (**2015**) 17–26. doi:10.1016/j.ssnmr.2015.09.005.

[55] H. Nagashima, J. Trébosc, O. Lafon, F. Pourpoint, P. Paluch, M.J. Potrzebowski, J.-P. Amoureux, Imaging the spatial distribution of radiofrequency field, sample and temperature in MAS NMR rotor, *Solid State Nucl. Magn. Reson.* 87 (**2017**) 137–142. doi:10.1016/j.ssnmr.2017.08.001.

[56] B. Dillmann, L. Dubois, E. Paleczny, J. Trébosc, J.-P. Amoureux, F. Pourpoint, O. Lafon, Local measure of the electromagnetic field in magnetic resonance coils: How do simulations help to disentangle the contributions of the electric and magnetic fields?, *Solid State Nucl. Magn. Reson.* 82–83 (**2017**) 1–9. doi:10.1016/j.ssnmr.2016.12.005.

[57] Z. Tošner, A. Purea, J.O. Struppe, S. Wegner, F. Engelke, S.J. Glaser, B. Reif, Radiofrequency fields in MAS solid state NMR probes, *J. Magn. Reson.* 284 (**2017**) 20–32. doi:10.1016/j.jmr.2017.09.002.

[58] R. Siegel, J. Trébosc, J.-P. Amoureux, Z. Gan, 3D ¹H-¹³C-¹⁴N correlation solid-state NMR spectrum., *J. Magn. Reson.* 193 (**2008**) 321–5. doi:10.1016/j.jmr.2008.05.013.

[59] P. Paluch, A.G.M. Rankin, J. Trébosc, O. Lafon, J.-P. Amoureux, Analysis of HMQC experiments applied to a spin ¹/₂ nucleus subject to very large CSA, *Solid State Nucl. Magn. Reson.* 100 (**2019**) 11–25. doi:10.1016/j.ssnmr.2019.03.001.

[60] F.A. Perras, T.W. Goh, L.-L. Wang, W. Huang, M. Pruski, Enhanced ¹H-X *D*-HMQC performance through improved ¹H homonuclear decoupling, *Solid State Nucl. Magn. Reson.* 98 (**2019**) 12–18. doi:10.1016/j.ssnmr.2019.01.001.


[61] M. Shen, J. Trébosc, O. Lafon, F. Pourpoint, B. Hu, Q. Chen, J.-P. Amoureux, Improving the resolution in proton-detected through-space heteronuclear multiple quantum correlation NMR spectroscopy., *J. Magn. Reson.* 245 (**2014**) 38–49. doi:10.1016/j.jmr.2014.05.006.

[62] E. Stejskal, J. Schaefer, J.S. Waugh, Magic-Angle Spinning and Polarization Transfer in Proton-Enhanced NMR, 28 (**1977**) 105–112. doi: 10.1016/0022-2364(77)90260-8.

[63] B.H. Meier, Cross polarization under fast magic angle spinning: thermodynamical considerations, *Chem. Phys. Lett.* 188 (**1992**) 201–207. doi:10.1016/0009-2614(92)90009-C.

[64] P. Paluch, J. Trébosc, J.-P. Amoureux, M.J. Potrzebowski, ¹H-³¹P CPVC NMR method under Very Fast Magic Angle Spinning for analysis of dipolar interactions and dynamics processes in the crystalline phosphonium tetrafluoroborate salts, *Solid State Nucl. Magn.* Reson. 87 (**2017**) 96–103. doi:10.1016/j.ssnmr.2017.05.004.

[65] M.K. Pandey, Y. Nishiyama, Determination of NH proton chemical shift anisotropy with ¹⁴N–¹H heteronuclear decoupling using ultrafast magic angle spinning solid-state NMR, *J. Magn. Reson.* 261 (**2015**) 133–140. doi:10.1016/j.jmr.2015.10.015.



 F_1 slices at 18.8 T and v_R = 62.5 kHz from ¹H-{¹⁴N} *D*-HMQC 2D spectra of NAV, with either Single-Quanta and v_1 = 45 (SLP), 40 (D_1^k), 80 (D_1^k), 80 kHz (HP), or Double-Quanta and v_1 = 64 kHz (SLP).